

101.531-48 – including *Taq* pol., IFU-01
 101.531-48u – without *Taq* pol., IFU-02

Visit www.olerup.com for
 “Instructions for Use” (IFU)

Lot No.: **3H1**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP[®] HLA-B*27 – unit dose

Product number:	101.531-48 – including <i>Taq</i> polymerase 101.531-48u – without <i>Taq</i> polymerase
Lot number:	3H1
Expiry date:	2021-07-01
Number of tests:	48
Number of wells per test:	2

Well specifications:

Well No.	Production No.
1	2019-011-01
2	2019-011-02

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2019-02-06

Approved by: *R. Becker-Selme*

Production Quality Control

101.531-48 – including *Taq* pol., IFU-01101.531-48u – without *Taq* pol., IFU-02Visit www.olerup.com for
“Instructions for Use” (IFU)Lot No.: **3H1**

Lot-specific information

Declaration of Conformity**Product name:** Olerup SSP[®] HLA-B*27 - unit dose**Product number:** 101.531-48/48u**Lot number:** 3H1**Intended use:** HLA-B*27 low resolution histocompatibility testing**Manufacturer:** Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, 1 Trinity Park, Bickenhill Lane, Birmingham B37 7ES, United Kingdom.
(Notified Body number: 0088.)

Stockholm, Sweden


Emil Jonsson
Head of QA