

CERTIFICATE OF ANALYSIS

Olerup SSP[®] DNA Size Marker

Product number: 103.202-100
Lot number: 76M
Volume: 2 x 500 µl
Concentration: 20 ng/µl
Expiry date: 2013-August-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 2011-September-02

Approved by:



Quality Control, Supervisor

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100
Lot number: 76M

Intended use: DNA Size Marker

Manufacturer: *Olerup* SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Stockholm, Sweden
2011-September-02



Ann-Cathrin Jareman
Head of QA and Regulatory Affairs