

## CERTIFICATE OF ANALYSIS

### Olerup SSP® DNA Size Marker

**Product number:** 103.202-100/-500  
**Lot number:** 46X  
**Volume:** Product No. 103.202-100 – 2 x 500 µl  
Product No. 103.202-500 – 10 x 500 µl  
**Concentration:** 20 ng/µl  
**Expiry date:** 2016-September-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

**Results:** Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

**Date of approval:** 2014-October-06

**Approved by:**

*Karin Mattsson*

**Production Quality Control**

## Declaration of Conformity

**Product name:** DNA Size Marker  
**Product number:** 103.202-100/500  
**Lot number:** 46X

**Intended use:** DNA Size Marker

**Manufacturer:** Olerup SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden  
2014-October-06



Daniel Malica  
Head of QA and Regulatory Affairs