

## CERTIFICATE OF ANALYSIS

### Olerup SSP® DNA Size Marker

**Product number:** 103.202-100/-500  
**Lot number:** 1E1  
**Volume:** Product No. 103.202-100 – 2 x 500 µl  
Product No. 103.202-500 – 10 x 500 µl  
**Concentration:** 20 ng/µl  
**Expiry date:** 2018-06-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

**Results:** Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

**Date of approval:** 2016-06-23

**Approved by:**



**Production Quality Control**

## **Declaration of Conformity**

**Product name:** DNA Size Marker  
**Product number:** 103.202-100/500  
**Lot number:** 1E1

**Intended use:** DNA Size Marker

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden  
2016-Jun-23



Daniel Malica  
Head of QA and Regulatory Affairs