

## Olerup SSP® DNA Size Marker

**Product number:** 103.202-100/500  
**Lot number:** OH3  
**Volume:** Product No. 103.202-100 – 2 x 500 µl  
 Product No. 103.202-500 – 10 x 500 µl  
**Concentration:** 20 ng/µl  
**Format:** Ready-to-load, 10 µl per gel lane  
**Expiry date:** 2020-10-01  
**Storage:** 2-8°C

**DESCRIPTION** 103.202-100: 2 vials, 500 µl each.  
 103.202-500: 10 vials, 500 µl each.  
 DNA ladder, 7 bands: 50, 100, 200, 300, 400, 500 and 1 000 bp,  
 20 ng/µl  
 in 10 mM Tris-HCl (pH 9.0), 50 mM KCl, 5% (v/v) glycerol  
 (99.5%), cresol red 100 ng/µl.

The DNA ladder contains 7 fragments of equal intensity:

<u>Fragment</u>	<u>Number of base pairs</u>
1	1000
2	500
3	400
4	300
5	200
6	100
7	50

**PRODUCT USE** The DNA Size marker can be resolved well in standard LE agarose gels of 1-2%, in NuSieve 3:1 or Metaphor agarose gels up to 4%.  
 The DNA Size Marker is supplied in a **ready-to-load format**. Typically load **10 µl per gel lane** of the ready-to-load DNA Size marker, equals 200 ng of DNA Size Marker.

**STORAGE** Store at 2-8°C.

**SHIPPING** The DNA Size Marker is shipped at ambient temperature.

**STABILITY** The DNA Size Marker is stable for 24 months when stored at 2-8°C.

**CERTIFICATE OF ANALYSIS****Olerup SSP® DNA Size Marker**

**Product number:** 103.202-100/-500  
**Lot number:** 0H3  
**Volume:** Product No. 103.202-100 – 2 x 500 µl  
Product No. 103.202-500 – 10 x 500 µl  
**Concentration:** 20 ng/µl  
**Expiry date:** 2020-10-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

**Results:** Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

**Date of approval:** 2018-10-16

**Approved by:**



**Production Quality Control**

## Declaration of Conformity

**Product name:** DNA Size Marker  
**Product number:** 103.202-100/500  
**Lot number:** 0H3

**Intended use:** DNA Size Marker

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden

  
Emil Jonsson  
Head of QA  
2018/10/19

**ADDRESSES:****Manufacturer:****Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.**Tel:** +46-8-717 88 27**Fax:** +46-8-717 88 18**E-mail:** [olerup-se@caredx.com](mailto:olerup-se@caredx.com)**Web page:** <http://www.olerup.com>**Distributed by:****Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.**Tel:** +43-1-710 15 00**Fax:** +43-1-710 15 00 10**E-mail:** [olerup-at@caredx.com](mailto:olerup-at@caredx.com)**Web page:** <http://www.olerup.com>**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382**Tel:** 1-877-OLERUP1**Fax:** 610-344-7989**E-mail:** [olerup-us@caredx.com](mailto:olerup-us@caredx.com)**Web page:** <http://www.olerup.com>For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.