

Olerup SSP® DNA Size Marker

Product number: 103.202-100/500
Lot number: 5G3
Volume: Product No. 103.202-100 – 2 x 500 µl
 Product No. 103.202-500 – 10 x 500 µl
Concentration: 20 ng/µl
Format: Ready-to-load, 10 µl per gel lane
Expiry date: 2020-05-01
Storage: 2-8°C

DESCRIPTION 103.202-100: 2 vials, 500 µl each.
 103.202-500: 10 vials, 500 µl each.
 DNA ladder, 7 bands: 50, 100, 200, 300, 400, 500 and 1 000 bp,
 20 ng/µl
 in 10 mM Tris-HCl (pH 9.0), 50 mM KCl, 5% (v/v) glycerol
 (99.5%), cresol red 100 ng/µl.

The DNA ladder contains 7 fragments of equal intensity:

<u>Fragment</u>	<u>Number of base pairs</u>
1	1000
2	500
3	400
4	300
5	200
6	100
7	50

PRODUCT USE The DNA Size marker can be resolved well in standard LE agarose gels of 1-2%, in NuSieve 3:1 or Metaphor agarose gels up to 4%.
 The DNA Size Marker is supplied in a **ready-to-load format**. Typically load **10 µl per gel lane** of the ready-to-load DNA Size marker, equals 200 ng of DNA Size Marker.

STORAGE Store at 2-8°C.

SHIPPING The DNA Size Marker is shipped at ambient temperature.

STABILITY The DNA Size Marker is stable for 24 months when stored at 2-8°C.

CERTIFICATE OF ANALYSIS**Olerup SSP® DNA Size Marker**

Product number: 103.202-100/-500
Lot number: 5G3
Volume: Product No. 103.202-100 – 2 x 500 µl
Product No. 103.202-500 – 10 x 500 µl
Concentration: 20 ng/µl
Expiry date: 2020-05-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 2018-06-19

Approved by: *Rebecca Salme*

Production Quality Control

Lot No.: **5G3**

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100/500
Lot number: 5G3

Intended use: DNA Size Marker

Manufacturer: *Olerup* SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden



Emil Jonsson
Head of QA

Lot No.: **5G3****ADDRESSES:****Manufacturer:****Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.**Tel:** +46-8-717 88 27**Fax:** +46-8-717 88 18**E-mail:** olerup-se@caredx.com**Web page:** <http://www.olerup.com>**Distributed by:****Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.**Tel:** +43-1-710 15 00**Fax:** +43-1-710 15 00 10**E-mail:** olerup-at@caredx.com**Web page:** <http://www.olerup.com>**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382**Tel:** 1-877-OLERUP1**Fax:** 610-344-7989**E-mail:** olerup-us@caredx.com**Web page:** <http://www.olerup.com>For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.