

DNA Size Marker 103.202-100/500 Certificates

Page 1 of 2 Visit <u>www.labproducts.caredx.com</u> for

"Instructions for Use" (IFU)

Lot No.: 4L5

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® DNA Size Marker

Product number:

103.202-100/-500

Lot number:

4L5

Volume:

Product No. 103.202-100 – 2 x 500 μ l

Product No. 103.202-500 – 10 x 500 μ l

Concentration:

20 ng/μl

Expiry date:

2022-06-01

10 μ l DNA Size Marker was size separated on a 2% agarose gel stained by ethidium bromide.

Results:

Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were

visible on a UV transilluminator.

Date of approval:

2020-08-12

En Ens

Approved by:

Production Quality Control



DNA Size Marker 103.202-100/500

Certificates

Page 2 of 2
Visit <u>www.labproducts.caredx.com</u> for
"Instructions for Use" (IFU)

Lot No.: 4L5

Lot-specific information

Declaration of Conformity

Product name: Product number:

DNA Size Marker 103.202-100/500

Lot number:

4L5

Intended use:

DNA Size Marker

Manufacturer:

CareDx AB Franzéngatan 5

SE-112 51 Stockholm, Sweden **Phone:** +46-8-508 939 00 **Fax:** +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden

2020-08.3

Date:

Quality Assurance

CE