

DNA Size Marker for short gel runs 103.203-100/500

Certificates

Page 1 of 2 Visit www.caredx.com for "Instructions for Use" (IFU)

Lot No.: 1V9

Lot-specific information

## CERTIFICATE OF ANALYSIS

Olerup SSP® DNA Size Marker for short gel runs

**Product number:** 

103.203-100/500

Lot number:

**1V9** 

Volume:

Product No. 103.203-100 – 2 x 500 μl

Product No. 103.203-500 – 10 x 500 μl

**Concentration:** 

20 ng/µl

**Expiry date:** 

2026-12-01

10 μl DNA Size Marker was size separated on a 2% agarose gel stained by ethidium bromide.

Results:

Four fragments of 50, 200, 500 and 1000 bp were visible on a UV

transilluminator. Double intensity of 200 bp fragment.

Date of approval: 2024-19-13

Approved by:

**Production Quality Control** 



DNA Size Marker for short gel runs 103.203-100/500

Certificates

Page 2 of 2
Visit <u>www.caredx.com</u> for "Instructions for Use" (IFU)

Lot No.: 1V9

Lot-specific information

**Declaration of Conformity** 

Product name:

DNA Size Marker 103.203-100/500

Product number: Lot number:

1V9

Intended use:

DNA Size Marker for short gel runs

Manufacturer:

CareDx AB

Franzéngatan 5

SE-112 51 Stockholm, Sweden **Phone:** +46-8-508 939 00

Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden

Date: 2025-01-21

**Quality Assurance**