

Lot No.: **0R7**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP[®] DNA Size Marker for short gel runs

Product number: 103.203-100/500
Lot number: 0R7
Volume: Product No. 103.203-100 – 2 x 500 µl
Product No. 103.203-500 – 10 x 500 µl
Concentration: 20 ng/µl
Expiry date: 2024-05-01

10 µl DNA Size Marker was size separated on a 2% agarose gel stained by ethidium bromide.

Results: Four fragments of 50, 200, 500 and 1000 bp were visible on a UV transilluminator. Double intensity of 200 bp fragment.

Date of approval: 2022-07-06

Approved by:



Production Quality Control



Lot No.: **0R7**

Lot-specific information

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.203-100/500
Lot number: 0R7

Intended use: DNA Size Marker for short gel runs

Manufacturer: *CareDx* AB
Franzégatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzégatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *CareDx* AB.

Stockholm, Sweden

Date: 2022-07-06

Quality Assurance

