

103.203-100/500

Lot No.: 88N

CERTIFICATE OF ANALYSIS**Olerup SSP[®] DNA Size Marker for short gel runs**

Product number: 103.203-100/500
Lot number: 88N
Volume: Product No. 103.203-100 – 2 x 500 µl
Product No. 103.203-500 – 10 x 500 µl
Concentration: 20 ng/µl
Expiry date: 2014-May-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Four fragments of 50, 200, 500 and 1000 bp were visible on a UV transilluminator. Double intensity of 200 bp fragment.

Date of approval: 2012-May-11

Approved by:



Production Quality Control

103.203-100/500

Lot No.: 88N

Declaration of Conformity

Product name: DNA Size Marker for short gel runs
Product number: 103.203-100/500
Lot number: 88N

Intended use: DNA Size Marker for short gel runs

Manufacturer: *Olerup* SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden
2012-May-11


Ann-Cathrin Jareman
Head of QA and Regulatory Affairs