

Lot No.: **5L8**

Lot-specific information

**CERTIFICATE OF ANALYSIS****Olerup SSP<sup>®</sup> DPB1**

Product number:

101.311-24/04 – including *Taq* pol.101.311-24u/04u – without *Taq* pol.

Lot number:

5L8

Expiry date:

2024-10-01

Number of tests:

24 tests – Product No. 101.311-24/24u

4 tests – Product No. 101.311-04/04u

Number of wells per test:

95+1

**Well specifications:**

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2016-683-01	33	2016-683-33	65	2017-863-65
2	2016-683-02	34	2016-683-34	66	2018-982-66
3	2020-172-03	35	2016-683-35	67	2020-172-67
4	2016-683-04	36	2016-683-36	68	2018-982-68
5	2016-683-05	37	2018-982-37	69	2018-982-69
6	2016-683-06	38	2018-982-38	70	2018-982-70
7	2016-683-07	39	2016-683-39	71	2016-683-71
8	2016-683-08	40	2017-863-40	72	2018-982-72
9	2016-683-09	41	2016-683-41	73	2017-776-73
10	2017-863-10	42	2016-683-42	74	2017-863-74
11	2016-683-11	43	2016-683-43	75	2016-683-75
12	2017-863-12	44	2018-982-44	76	2018-982-76
13	2016-683-13	45	2016-683-45	77	2018-982-77
14	2017-776-14	46	2016-683-46	78	2020-245-78
15	2018-982-15	47	2016-683-47	79	2018-982-79
16	2016-683-16	48	2016-683-48	80	2018-982-80
17	2016-683-17	49	2017-863-49	81	2016-683-81
18	2016-683-18	50	2016-683-50	82	2016-683-82
19	2016-683-19	51	2017-776-51	83	2016-683-83
20	2016-683-20	52	2017-776-52	84	2017-863-84
21	2016-683-21	53	2018-982-53	85	2017-776-85
22	2016-683-22	54	2018-982-54	86	2018-982-86
23	2016-683-23	55	2017-776-55	87	2018-982-87
24	2016-683-24	56	2017-863-56	88	2016-683-88
25	2016-683-25	57	2016-683-57	89	2017-776-89
26	2018-982-26	58	2020-172-58	90	2020-245-90
27	2020-172-27	59	2017-863-59	91	2018-982-91
28	2016-683-28	60	2018-982-60	92	2018-982-92
29	2018-982-29	61	2016-683-61	93	2018-982-93
30	2016-683-30	62	2016-683-62	94	2017-863-94
31	2016-683-31	63	2020-245-63	95	2016-683-95
32	2016-683-32	64	2017-863-64		



101.311-24/04 – including *Taq* pol., IFU-01

Visit <https://labproducts.caredx.com> for

101.311-24/04u – without *Taq* pol., IFU-02

“Instructions for Use” (IFU)

Lot No.: **5L8**

Lot-specific information

The negative control primer pairs, **Production No. 2020-205-01**, can detect contamination with PCR products diluted  $10^{-7}$ .

**Results of Quality Control:** No false positive or false negative amplifications obtained.

**Date of approval:** 2020-11-16

**Approved by:**



**Production Quality Control**



Lot No.: **5L8**

Lot-specific information

### Declaration of Conformity

**Product name:** Olerup SSP<sup>®</sup> DPB1  
**Product number:** 101.311-24/04, -24u/04u  
**Lot number:** 5L8

**Intended use:** DPB1 high resolution histocompatibility testing

**Manufacturer:** CareDx AB  
Franzégatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-508 939 00  
**Fax:** +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at CareDx AB, Franzégatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden

Date:

2020-11-19

Quality Assurance

