

101.201-48/12 – including *Taq* pol., IFU-01
101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.labproducts.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **9K6**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® DQ low resolution

Product number: 101.201-48/12 – including *Taq* pol.
101.201-48u/12u – without *Taq* pol.
Lot number: 9K6
Expiry date: 2024-02-01
Number of tests: 48 tests – Product No. 101.201-48/48u
12 tests – Product No. 101.201-12/12u
Number of wells per test: 15+1

Well specifications:

Well No.	Production No.	Well No.	Production No.
1	2019-140-01	9	2017-883-09
2	2019-140-02	10	2019-140-10
3	2020-159-03	11	2019-140-11
4	2019-140-04	12	2017-883-12
5	2020-159-05	13	2017-883-13
6	2020-159-06	14	2017-883-14
7	2019-140-07	15	2017-883-15
8	2019-140-08		

The negative control primer pairs, **Production No. 2019-109-01**, can detect contamination with PCR products diluted 10^{-7} .

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2020-03-05

Approved by:



Production Quality Control



101.201-48/12 – including *Taq* pol., IFU-01
101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.labproducts.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **9K6**

Lot-specific information

Declaration of Conformity

Product name: *Olerup* SSP[®] DQ low resolution
Product number: 101.201-48/48u, -12/12u
Lot number: 9K6

Intended use: DQB1 low resolution histocompatibility testing

Manufacturer: *CareDx* AB
Franzégatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, *CareDx* AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzégatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *CareDx* AB.

Stockholm, Sweden

Date: 2020-03-06

Quality Assurance

