

101.704-48/12 – including *Taq* polymerase
101.704-48u/12u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **2V4**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® DQ-DR SSP Combi Tray

Product number: 101.704-48/12 – including *Taq* pol.
101.704-48u/12u – without *Taq* pol.
Lot number: 2V4
Expiry date: 2029-03-01
Number of tests: 48 tests – Product No. 101.704-48/48u
12 tests – Product No. 101.704-12/12u
Number of wells per test: 46 + 1

Well specifications:

Well No.	Production No.	Well No.	Production No.
1	2025-630-01	9	2025-630-09
2	2025-630-02	10	2025-630-10
3	2025-630-03	11	2025-630-11
4	2025-630-04	12	2025-630-12
5	2025-630-05	13	2025-630-13
6	2025-630-06	14	2025-630-14
7	2025-630-07	15	2025-630-15
8	2025-630-08		

Well No.	Production No.	Well No.	Production No.
16	2024-617-01	32	2025-629-17
17	2024-617-02	33	2024-604-18
18	2024-617-03	34	2024-617-19
19	2024-617-04	35	2024-617-20
20	2024-617-05	36	2024-617-21
21	2024-617-06	37	2024-617-22
22	2024-617-07	38	2024-617-23
23	2024-604-08	39	2024-617-24
24	2024-617-09	40	2024-617-25
25	2024-617-10	41	2024-617-26
26	2024-617-11	42	2024-617-27
27	2024-617-12	43	2024-617-28
28	2024-617-13	44	2024-617-29
29	2024-617-14	45	2024-617-30
30	2024-617-15	46	2024-617-31
31	2024-617-16		

The negative control primer pairs, **Production No. 2025-632-01**, can detect contamination with PCR products diluted 10^{-7} .



0197

For In Vitro Diagnostic Use
MA100 v07 CoA_DoC IVD Annex II List B
Date: March 2025, Rev. No: 00

101.704-48/12 – including *Taq* polymerase
101.704-48u/12u – without *Taq* polymerase

Visit www.caredx.com for
"Instructions for Use" (IFU)

Lot No.: **2V4**

Lot-specific information

Results of Quality Control:

No false positive or false negative amplifications
obtained.

Date of approval: 2025-04-09

Approved by:



Production Quality Control

CE

0197

For *In Vitro* Diagnostic Use
MA100 v07 CoA_DoC IVD Annex II List B
Date: March 2025, Rev. No: 00

101.704-48/12 – including *Taq* polymerase
 101.704-48u/12u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **2V4**

Lot-specific information

Declaration of Conformity

Product name: Olerup SSP® DQ-DR SSP Combi Tray
Product number: 101.704-48/48u, -12/12u
Lot number: 2V4

Intended use: DQB1 and DRB1 low resolution histocompatibility testing

Manufacturer: CareDx AB
 Franzéngatan 5
 SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at CareDx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Notified Body: TÜV Rheinland LGA products, Tillystrasse 2, D-90431 Nürnberg, Germany. (Notified Body number: 0197.)

Stockholm, Sweden
 Date:

2025-04-15

Quality Assurance




0197

For *In Vitro* Diagnostic Use
 MA100 v07 CoA_DoC IVD Annex II List B
 Date: March 2025, Rev. No: 00

