

101.212-24/04 – including *Taq* pol., IFU-01
 101.212-24u/04u – without *Taq* pol., IFU-02

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **0R4**

Lot-specific information

CERTIFICATE OF ANALYSIS

OLERUP SSP® DQB1*06

Product number: 101.212-24/04 – including *Taq* pol.
 101.212-24u/04u – without *Taq* pol.

Lot number: 0R4

Expiry date: 2026-06-01

Number of tests: 24 tests – Product No. 101.212-24/24u
 4 tests – Product No. 101.212-04/04u

Number of wells per test: 63+1

Well specifications:

| Well No. | Production No. | Well No. | Production No. | Well No. | Production No. |
|----------|----------------|----------|----------------|----------|----------------|
| 1 | 2021-277-01 | 25 | 2021-277-25 | 49 | 2021-277-49 |
| 2 | 2021-277-02 | 26 | 2021-277-26 | 50 | 2021-277-50 |
| 3 | 2021-277-03 | 27 | 2021-277-27 | 51 | 2021-277-51 |
| 4 | 2021-277-04 | 28 | 2021-277-28 | 52 | 2021-277-52 |
| 5 | 2021-277-05 | 29 | 2021-277-29 | 53 | 2021-277-53 |
| 6 | 2021-277-06 | 30 | 2021-277-30 | 54 | 2021-277-54 |
| 7 | 2021-277-07 | 31 | 2021-277-31 | 55 | 2021-277-55 |
| 8 | 2021-277-08 | 32 | 2021-277-32 | 56 | 2022-407-56 |
| 9 | 2021-277-09 | 33 | 2021-277-33 | 57 | 2021-277-57 |
| 10 | 2021-277-10 | 34 | 2022-407-34 | 58 | 2021-277-58 |
| 11 | 2021-277-11 | 35 | 2021-277-35 | 59 | 2021-277-59 |
| 12 | 2021-277-12 | 36 | 2021-277-36 | 60 | 2021-277-60 |
| 13 | 2021-277-13 | 37 | 2021-277-37 | 61 | 2021-277-61 |
| 14 | 2021-277-14 | 38 | 2021-277-38 | 62 | 2022-407-62 |
| 15 | 2021-277-15 | 39 | 2021-277-39 | 63 | 2021-277-63 |
| 16 | 2021-277-16 | 40 | 2021-277-40 | | |
| 17 | 2021-277-17 | 41 | 2021-277-41 | | |
| 18 | 2021-277-18 | 42 | 2021-277-42 | | |
| 19 | 2021-277-19 | 43 | 2022-407-43 | | |
| 20 | 2021-277-20 | 44 | 2021-277-44 | | |
| 21 | 2021-277-21 | 45 | 2021-277-45 | | |
| 22 | 2021-277-22 | 46 | 2021-277-46 | | |
| 23 | 2021-277-23 | 47 | 2021-277-47 | | |
| 24 | 2022-407-24 | 48 | 2021-277-48 | | |

The negative control primer pairs, **Production No. 2021-336-01**, can detect contamination with PCR products diluted 10⁻⁷.



101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **0R4**

Lot-specific information

Results of Quality Control: No false positive or false negative amplifications
obtained.

Date of approval: 20220701

Approved by: 

Production Quality Control



101.212-24/04 – including *Taq* pol., IFU-01
 101.212-24u/04u – without *Taq* pol., IFU-02

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **OR4**

Lot-specific information

Declaration of Conformity

Product name: Olerup SSP® DQB1*06
Product number: 101.212-24/24u, -04/04u
Lot number: OR4

Intended use: DQB1*06 high resolution histocompatibility testing

Manufacturer: CareDx AB
 Franzégatan 5
 SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at CareDx AB, Franzégatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden

Date:

2022-07-06

Quality Assurance


