

Lot No.: **80M**

Lot-specific Information

www.olerup-ssp.com

## CERTIFICATE OF ANALYSIS

### **Olerup SSP® HLA Wipe Test – Negative Control**

Product number: 102.101-01u – without Taq polymerase  
Lot number: 80M  
Expiry date: 2014-March-01  
Number of tests: 96  
Number of wells per test: 1-2

#### Well specification:

Well No.	Production No.
1	2010-760-01

The negative control primer solution has been tested in a dilution series of the corresponding PCR products, 1 to 10<sup>3</sup> down to 1 to 10<sup>9</sup>.

The Positive Control DNA has been tested with the HLA Wipe Test kit and gives rise to PCR amplicons.

**Results:** The negative control primer pairs can detect contamination with the corresponding PCR products diluted 1 to 10<sup>7</sup>.

**Date of approval:** 2011-October-17

**Approved by:**



**Quality Control, Supervisor**

Lot No.: **80M**

Lot-specific Information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## Declaration of Conformity

**Product name:** *Olerup SSP*<sup>®</sup> HLA Wipe Test – Negative Control  
**Product number:** 102.101-01u  
**Lot number:** 80M

**Intended use:** Detection of contamination with HLA amplicons.

**Manufacturer:** *Olerup SSP AB*  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup SSP AB*, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup SSP AB*, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden  
2011-October-17



Ann-Cathrin Jareman  
Head of QA and Regulatory Affairs