

**102.101-01 – including *Taq* polymerase, IFU-01**  
**102.101-01u - without *Taq* polymerase, IFU-02**

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
**“Instructions for Use” (IFU)**

**Lot No.: 36V**

**Lot-specific information**

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## **CERTIFICATE OF ANALYSIS**

### **Olerup SSP® HLA Wipe Test – Negative Control**

**Product number:** 102.101-01 – including *Taq* polymerase  
**Product number:** 102.101-01u – without *Taq* polymerase  
**Lot number:** 36V  
**Expiry date:** 2016-August-01  
**Number of tests:** 96  
**Number of wells per test:** 1-2

**Well specification:**

<b>Well No.</b>	<b>Production No.</b>
1	2013-307-01

The negative control primer solution has been tested in a dilution series of the corresponding PCR products, 1 to 10<sup>3</sup> down to 1 to 10<sup>9</sup>.

The Positive Control DNA has been tested with the HLA Wipe Test kit and gives rise to PCR amplicons.

The negative control primer pairs can detect contamination with the corresponding PCR products diluted 1 to 10<sup>7</sup>.

**Results of Quality Control:** No false positive or false negative amplifications obtained.

**Date of approval:** 2014-March-04

**Approved by:**



**Production Quality Control**

102.101-01 – including *Taq* polymerase, IFU-01  
102.101-01u - without *Taq* polymerase, IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **36V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## Declaration of Conformity

**Product name:** Olerup SSP® HLA Wipe Test – Negative Control  
**Product number:** 102.101-01/01u  
**Lot number:** 36V

**Intended use:** Detection of contamination with HLA amplicons.

**Manufacturer:** Olerup SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden  
2014-March-04



Ann-Cathrin Jareman  
Head of QA and Regulatory Affairs