OLERUP SSP

HLA Wipe Test – Negative Control

Certificates

Page 1 of 2

102.101-01 – including *Taq* **polymerase**, IFU-01 **102.101-01u** – without *Taq* polymerase, IFU-02

Visit https://labproducts.caredx.com for "Instructions for Use" (IFU)

Lot No.: 5H5

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA Wipe Test – Negative Control

Product number:

102.101-01 - including Tag polymerase

Product number:

102.101-01u – without *Taq* polymerase

Lot number:

5H5

Expiry date:

2021-09-01

Number of tests:

96

Number of wells per test:

1-2

Well specification:

Well No. 1 Production No.

2019-044-01

The negative control primer solution has been tested in a dilution series of the corresponding PCR products, 1 to 10³ down to 1 to 10⁹. The Positive Control DNA has been tested with the HLA Wipe Test kit and gives rise to PCR amplicons.

The negative control primer pairs can detect contamination with the corresponding PCR products diluted 1 to 108, equivalent to 0,1 ng starting material.

Results of Quality Control: No false positive or false negative amplifications

obtained.

Date of approval: 2019 - 05 - 22

Approved by:

Rebecka Salve

Production Quality Control

OLERUP SSP

HLA Wipe Test – Negative Control

Certificates

Page 2 of 2

102.101-01 – including *Taq* polymerase, IFU-01 **102.101-01u** – without *Taq* polymerase, IFU-02

Visit https://labproducts.caredx.com
for "Instructions for Use" (IFU)

Lot No.: 5H5

Lot-specific information

Declaration of Conformity

Product name:

Olerup SSP® HLA Wipe Test - Negative Control

Product number:

102.101-01/01u

Lot number:

5H5

Intended use:

Detection of contamination with HLA amplicons.

Manufacturer:

Olerup SSP AB Franzéngatan 5

SE-112 51 Stockholm, Sweden **Phone:** +46-8-508 939 00

Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: Olerup SSP AB.

Stockholm, Sweden

Date:

Terese Wasell

Head of Regulatory Affairs

CE

For In Vitro Diagnostic Use MA099 v01 CoA_DoC General IVDs Date: May 2019, Rev. No: 01