

Lot No.: 70K

Lot-specific Information

www.olerup-ssp.com

## CERTIFICATE OF ANALYSIS

### Olerup SSP® HLA-A\*68 SSP

**Product number:** 101.418-12u/04u – without Taq pol.  
**Lot number:** 70K  
**Expiry date:** 2013-July-01  
**Number of tests:** 12 tests – Product No. 101.418-12  
4 tests – Product No. 101.418-04  
**Number of wells per test:** 32

#### Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2009-627-01	13	2010-795-13	25	2010-795-25
2	2007-326-02	14	2010-795-14	26	2010-795-26
3	2007-326-03	15	2007-326-15	27	2009-627-27
4	2007-326-04	16	2009-627-16	28	2009-627-28
5	2007-326-05	17	2007-326-17	29	2010-795-29
6	2010-795-06	18	2010-795-18	30	2008-462-30
7	2007-326-07	19	2007-326-19	31	2010-795-31
8	2009-627-08	20	2007-326-20	32	2010-795-32
9	2009-627-09	21	2007-326-21		
10	2007-326-10	22	2007-326-22		
11	2007-326-11	23	2007-326-23		
12	2009-627-12	24	2007-326-24		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 5, 7 to 9, 11 to 13, 15, 16, 19 to 21, 25 to 28 and 32 were available. The specificities of the primers in primer solutions 5, 7 to 9, 11, 13, 15, 16, 20, 21, 25 to 27 and 32 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer. The specificities of the 3'-primers in primer solutions 12, 19 and 28 were tested by separately adding one additional 5'-primer, it was not possible to test the 5'-primers. In primer solution 5, 7, 8, 11, 17, 25 and 27 one of the 5'-primers was not possible to test. In primer solutions 8, 9, 14 to 16, 18, 21, 26, 27, 29, 30 and 32 one or two of the 3'-primers were not possible to test. Additional primers in primer solutions 4, 6, 14 and 29 to 31 were tested tested by separately adding either one additional 5'-primer, or one additional 3'-primer.

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2011-January-31

**Approved by:**



**Quality Control, Supervisor**

Lot No.: **70K**

Lot-specific Information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## Declaration of Conformity

**Product name:** *Olerup* SSP® HLA-A\*68  
**Product number:** 101.418-12u/04u  
**Lot number:** 70K

**Intended use:** HLA-A\*68 high resolution histocompatibility testing

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden  
2012-October-05



Ann-Cathrin Jareman  
Head of QA and Regulatory Affairs