

101.418-12/04 – including *Taq* pol., IFU-01 Rev. No. 03
 101.418-12u/04– without *Taq* pol., IFU-02 Rev. No. 03

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **36N**

Lot-specific information
CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-A*68 SSP

Product number: 101.418-12/04 – including *Taq* pol.
 101.418-12u/04u – without *Taq* pol.
Lot number: 36N
Expiry date: 2014-September-01
Number of tests: 12 tests – Product No. 101.418-12/12u
 4 tests – Product No. 101.418-04/04u
Number of wells per test: 42

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2012-975-01	16	2012-975-16	31	2012-975-31
2	2012-975-02	17	2012-975-17	32	2012-975-32
3	2012-975-03	18	2012-975-18	33	2012-975-33
4	2012-975-04	19	2012-975-19	34	2012-975-34
5	2012-975-05	20	2012-975-20	35	2012-975-35
6	2012-975-06	21	2012-975-21	36	2012-975-36
7	2012-975-07	22	2012-975-22	37	2012-975-37
8	2012-975-08	23	2012-975-23	38	2012-975-38
9	2012-975-09	24	2012-975-24	39	2012-975-39
10	2012-975-10	25	2012-975-25	40	2012-975-40
11	2012-975-11	26	2012-975-26	41	2012-975-41
12	2012-975-12	27	2012-975-27	42	2012-975-42
13	2012-975-13	28	2012-975-28		
14	2012-975-14	29	2012-975-29		
15	2012-975-15	30	2012-975-30		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 5, 7 to 9, 11 to 13, 15, 16, 19 to 21, 25 to 28, 32 to 35 and 37 to 42 were available. The specificities of the primers in primer solutions 5, 7 to 9, 11, 13, 15, 16, 20, 21, 25 to 27, 37 and 41 were tested by separately adding additional 5'-primers, respectively additional 3'-primers. The specificities of the 3'-primers in primer solutions 12, 19, 28, 34, 39 and 42 were tested by separately adding one additional 5'-primer, it was not possible to test the 5'-primers. In primer solutions 32, 33, 35, 38 and 40 it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solution 5, 7, 8, 11, 17, 25 and 27 one of the 5'-primers was not possible to test. In primer solutions 8, 9, 14 to 16, 18, 21, 26, 27, 29 and 30 one or two of the 3'-primers were not possible to test. Additional primers in primer solutions 4, 6, 14, 29 and 30 were tested tested by separately adding either one additional 5'-primer, or one additional 3'-primer.

101.418-12/04 – including *Taq* pol., IFU-01 Rev. No. 03
101.418-12u/04– without *Taq* pol., IFU-02 Rev. No. 03

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **36N**

Lot-specific information

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2012-April-12

Approved by:

Karin Mattsson

Production Quality Control

101.418-12/04 – including *Taq* pol., IFU-01 Rev. No. 03
101.418-12u/04– without *Taq* pol., IFU-02 Rev. No. 03

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **36N**

Lot-specific information

Declaration of Conformity

Product name: Olerup SSP® HLA-A*68
Product number: 101.418-12/12u, 04/04u
Lot number: 36N

Intended use: HLA-A*68 high resolution histocompatibility testing

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden
2012-April-12



Ann-Cathrin Jareman
Head of QA and Regulatory Affairs