

101.545-06 – including *Taq* polymerase

101.545-06u – without *Taq* polymerase

Lot No.: **2S0**

Lot-specific Information

Visit www.caredx.com for
 “Instructions for Use” (IFU)

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-B*47

Product number:

101.545-06 – including *Taq* polymerase

101.545-06u – without *Taq* polymerase

Lot number:

2S0

Expiry date:

2027-06-01

Number of tests:

6

Number of wells per test:

10+1

Well specifications:

Well No.	Production No.	Well No.	Production No.
1	2018-946-01	9	2018-946-09
2	2018-946-02	10	2021-290-10
3	2018-946-03		
4	2018-946-04		
5	2018-946-05		
6	2018-946-06		
7	2018-946-07		
8	2018-946-08		

The negative control primer pairs, **Production No. 2022-404-01**, can detect contamination with PCR products diluted 10^{-7} .

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2023-08-28

Approved by:



Production Quality Control



0197

For *In Vitro* Diagnostic Use

MA100 v06 CoA_DoC IVD Annex II List B

Date: June 2023, Rev. No: 00

101.545-06 – including *Taq* polymerase101.545-06u – without *Taq* polymeraseVisit www.caredx.com for
“Instructions for Use” (IFU)Lot No.: **2S0**

Lot-specific Information

Declaration of Conformity**Product name:** Olerup SSP® HLA-B*47**Product number:** 101.545-06**Lot number:** 2S0**Intended use:** HLA-B*47 high resolution histocompatibility testing**Manufacturer:** CareDx AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at CareDx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Notified Body: TÜV Rheinland LGA products, Tillystrasse 2, D-90431 Nürnberg, Germany. (Notified Body number: 0197.)

Stockholm, Sweden

Date:

2023-08-28

Quality Assurance

**CE****0197**For *In Vitro* Diagnostic Use
MA100 v06 CoA_DoC IVD Annex II List B
Date: June 2023, Rev. No: 00