

101.571-06 – including *Taq* polymerase, IFU-01  
 101.571-06u – without *Taq* polymerase, IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **85R**

Lot-specific information

## CERTIFICATE OF ANALYSIS

### Olerup SSP® HLA-B\*56 SSP

Product number: 101.571-06 – including *Taq* polymerase  
 101.571-06u – without *Taq* polymerase  
 Lot number: 85R  
 Expiry date: 2015-July-01  
 Number of tests: 6  
 Number of wells per test: 24

#### Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2010-750-01	9	2011-893-09	17	2013-132-17
2	2010-750-02	10	2010-750-10	18	2011-893-18
3	2010-750-03	11	2010-750-11	19	2010-750-19
4	2010-750-04	12	2013-132-12	20	2010-750-20
5	2010-750-05	13	2011-893-13	21	2010-750-21
6	2011-893-06	14	2013-132-14	22	2013-132-22
7	2013-132-07	15	2010-750-15	23	2010-750-23
8	2010-750-08	16	2013-132-16	24	2010-750-24

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 8, 19, 21 and 23 were available. The specificities of the primers in primer solutions 8 and 23 were tested by separately adding one additional 5'-primer, respectively one or two additional 3'-primers. In primer solutions 19 and 21, it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solutions 1, 2, 3 and 15 one or two 3'-primers were not possible to test, and in primer solutions 13, 17 and 23, one 5'-primer was not possible to test. Additional 5'-primers and 3'-primers in primer solutions 6, 15 to 17 and 22 were tested by separately adding one 5'-primer and/or 3'-primer.

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2013-January-28

**Approved by:**

Production Quality Control

101.571-06 – including *Taq* polymerase, IFU-01  
101.571-06u – without *Taq* polymerase, IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
"Instructions for Use" (IFU)

Lot No.: **85R**

Lot-specific information

## Declaration of Conformity

**Product name:** Olerup SSP® HLA-B\*56  
**Product number:** 101.571-06/06u  
**Lot number:** 85R

**Intended use:** HLA-B\*56 high resolution histocompatibility testing

**Manufacturer:** Olerup SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlesbrough Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom.  
(Notified Body number: 0088.)

Stockholm, Sweden  
2013-January-28

Ann-Cathrin Jareman  
Head of QA and Regulatory Affairs