

101.601-24/12 – including *Taq* pol., IFU-01
 101.601-24u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
 “Instructions for Use” (IFU)

Lot No.: **3F9**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-C low resolution

Product number: 101.601-24/12 – including *Taq* pol.
 101.601-24u/12u – without *Taq* pol.
Lot number: 3F9
Expiry date: 2020-02-01
Number of tests: 24 tests – Product No. 101.601-24/24u
 12 tests – Product No. 101.601-12/12u
Number of wells per test: 31 + 1

Well specifications:

Well No.	Production No.	Well No.	Production No.
1	2017-817-01	17	2015-531-17
2	2015-531-02	18	2015-531-18
3	2015-531-03	19	2016-730-19
4	2016-730-04	20	2016-693-20
5	2015-531-05	21	2015-531-21
6	2015-531-06	22	2015-531-22
7	2016-730-07	23	2017-817-23
8	2016-760-08	24	2017-817-24
9	2015-531-09	25	2017-817-25
10	2015-531-10	26	2017-817-26
11	2016-730-11	27	2017-817-27
12	2015-531-12	28	2017-817-28
13	2015-531-13	29	2017-817-29
14	2016-730-14	30	2017-817-30
15	2016-672-15	31	2017-817-31
16	2015-531-16		

The negative control primer pairs, **Production No. 2016-746-01**, can detect contamination with PCR products diluted 10^{-7} .

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2017-09-20

Approved by: 

Production Quality Control

101.601-24/12 – including *Taq* pol., IFU-01
 101.601-24u/12u – without *Taq* pol., IFU-02

Visit www.olerup.com for
 “Instructions for Use” (IFU)

Lot No.: **3F9**

Lot-specific information

Declaration of Conformity

Product name: Olerup SSP® HLA-C low resolution
Product number: 101.601-24/24u, -12/12u
Lot number: 3F9


Intended use: HLA-C low resolution histocompatibility testing

Manufacturer: Olerup SSP AB
 Franzengatan 5
 SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden


 Emil Jonsson 2017-09-20
 Head of QA