**HLA-C** low resolution

Certificates

Page 1 of 2

**101.601-24/12 – including** *Taq* **pol.,** IFU-01 **101.601-24u/12u – without** *Taq* **pol.**, IFU-02 Visit www.labproducts.caredx.com for "Instructions for Use" (IFU)

Lot No.: 9L9

Lot-specific information

## CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-C low resolution

**Product number:** 

101.601-24/12 – including *Taq* pol.

101.601-24u/12u – without *Taq* pol.

Lot number:

**9L9** 

**Expiry date:** 

2025-02-01

Number of tests:

24 tests - Product No. 101.601-24/24u

12 tests - Product No. 101.601-12/12u

Number of wells per test:

31 + 1

Well specifications:

Well specifications.		
Production No.	Well No.	Production No.
2021-288-01	17	2021-288-17
2021-288-02	18	2021-288-18
2021-288-03	19	2021-288-19
2021-288-04	20	2021-288-20
2021-288-05	21	2021-288-21
2021-288-06	22	2021-288-22
2021-288-07	23	2021-288-23
2021-288-08	24	2017-817-24
2021-288-09	25	2021-288-25
2021-288-10	26	2021-288-26
2021-288-11	27	2021-288-27
2021-288-12	28	2021-288-28
2021-288-13	29	2021-288-29
2021-288-14	30	2021-288-30
2021-288-15	31	2021-288-31
2021-288-16		
	Production No. 2021-288-01 2021-288-02 2021-288-03 2021-288-05 2021-288-06 2021-288-08 2021-288-10 2021-288-11 2021-288-13 2021-288-14 2021-288-15	Production No. Well No.   2021-288-01 17   2021-288-02 18   2021-288-03 19   2021-288-04 20   2021-288-05 21   2021-288-06 22   2021-288-07 23   2021-288-08 24   2021-288-10 26   2021-288-11 27   2021-288-12 28   2021-288-13 29   2021-288-14 30   2021-288-15 31

The negative control primer pairs, Production No. 2020-269-01, can detect contamination with PCR products diluted 10<sup>-7</sup>.

Results of Quality Control: No false positive or false negative amplifications

obtained.

Date of approval: 2021-03-13

Approved by:

**Production Quality Control** 

For In Vitro Diagnostic Use MA099 v03 CoA\_DoC General IVDs Date: March 2021, Rev. No: 00



**HLA-C** low resolution

Certificates

Page 2 of 2

**101.601-24/12 – including** *Taq* **pol.**, IFU-01 **101.601-24u/12u – without** *Taq* **pol.**, IFU-02

Visit <u>www.labproducts.caredx.com</u> for "Instructions for Use" (IFU)

Lot No.: 9L9

Lot-specific information

**Declaration of Conformity** 

Product name:

Olerup SSP® HLA-C low resolution

**Product number:** 

101.601-24/24u, -12/12u

Lot number:

9L9

Intended use:

HLA-C low resolution histocompatibility testing

Manufacturer:

CareDx AB

Franzéngatan 5

SE-112 51 Stockholm, Sweden **Phone:** +46-8-508 939 00 **Fax:** +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden Date: 2021-03-14

0

Quality Assurance

