

Lot No.: **5N0**

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-C*02

Product number: 101.622-12 – including *Taq* polymerase
 101.622-12u – without *Taq* polymerase
Lot number: 5N0
Expiry date: 2025-10-01
Number of tests: 12
Number of wells per test: 44+1

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2020-130-01	17	2020-130-17	33	2020-130-33
2	2020-130-02	18	2020-130-18	34	2020-130-34
3	2020-130-03	19	2020-130-19	35	2020-130-35
4	2020-130-04	20	2020-130-20	36	2020-130-36
5	2020-130-05	21	2020-130-21	37	2020-130-37
6	2020-130-06	22	2020-130-22	38	2021-348-38
7	2020-130-07	23	2020-130-23	39	2020-130-39
8	2020-130-08	24	2020-130-24	40	2021-348-40
9	2020-130-09	25	2020-130-25	41	2020-130-41
10	2020-130-10	26	2020-130-26	42	2020-130-42
11	2020-130-11	27	2020-130-27	43	2020-130-43
12	2020-130-12	28	2020-130-28	44	2021-348-44
13	2020-130-13	29	2020-130-29		
14	2020-130-14	30	2020-130-30		
15	2020-130-15	31	2020-130-31		
16	2020-130-16	32	2021-348-32		

The negative control primer pairs, **Production No. 2020-289-01**, can detect contamination with PCR products diluted 10⁻⁷.

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2021-10-25

Approved by:



Production Quality Control



