

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-C*04 SSP

Product number: 101.612-12 – including *Taq* polymerase
 Lot number: 12M
 Expiry date: 2013-November-01
 Number of tests: 12
 Number of wells per test: 47

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2010-702-01	17	2011-853-17	33	2011-853-33
2	2010-702-02	18	2011-853-18	34	2011-853-34
3	2010-702-03	19	2010-702-19	35	2011-853-35
4	2010-702-04	20	2010-702-20	36	2011-853-36
5	2010-702-05	21	2010-702-21	37	2011-853-37
6	2010-702-06	22	2010-702-22	38	2011-853-38
7	2010-702-07	23	2010-702-23	39	2011-853-39
8	2010-702-08	24	2010-702-24	40	2011-853-40
9	2010-702-09	25	2010-702-25	41	2011-853-41
10	2010-702-10	26	2010-702-26	42	2011-853-42
11	2010-702-11	27	2010-702-27	43	2011-853-43
12	2010-702-12	28	2011-853-28	44	2011-853-44
13	2010-702-13	29	2010-702-29	45	2011-853-45
14	2011-853-14	30	2010-702-30	46	2011-853-46
15	2010-702-15	31	2010-702-31	47	2011-853-47
16	2010-702-16	32	2011-853-32		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 6, 8, 10, 12, 13, 15, 17 to 40 and 42 to 47 were available. The specificity of the primers in primer solutions 8, 10, 12, 13, 15, 17, 18 to 21, 24, 28, 32 to 34, 39 and 42 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer. In primer solutions 6, 25 to 27, 29, 35, 40 and 43 it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solutions 22, 23, 30, 31, 36 to 38 and 44 to 47 it was only possible to test the 3'-primers, the 5'-primers were not possible to test. In primer solutions 9, 15, 18, 19, 24, 28 and 42 one or two of the 5'-primers were not possible to test, and in primer solutions 9, 10, 13, 14, 18, 20, 21, 24, 28, 32, 33 and 39 one or two of the 3'-primers were not possible to test. Additional primers in primer solutions 7, 9, 14 and 16 were tested by separately adding one 5'-primer and/or one 3'-primer.

Lot No.: **12M**


Lot-specific information

www.olerup-ssp.com

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2011-July-15

Approved by:



Production Quality Control

Lot No.: **12M**

Lot-specific information

www.olerup-ssp.com

Declaration of Conformity

Product name: Olerup SSP® HLA-C*04
Product number: 101.612-12
Lot number: 12M

Intended use: HLA-C*04 high resolution histocompatibility testing


Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: Olerup SSP AB.

Stockholm, Sweden
2012-July-20



Asa Olausson
Production Quality Control

