

101.862-12 – including *Taq* polymerase, IFU-01101.862-12u – without *Taq* polymerase, IFU-02Visit www.olerup-ssp.com for

“Instructions for Use” (IFU)

Lot No.: **49S**

Lot-specific information

CERTIFICATE OF ANALYSIS**Olerup SSP® HLA-C*04 Add-on SSP**

Product number: 101.862-12 – including *Taq* polymerase
101.862-12u – without *Taq* polymerase
Lot number: 49S
Expiry date: 2015-December-01
Number of tests: 12
Number of wells per test: 4

Well specifications:

Well No.	Production No.
1	2012-963-01
2	2012-963-02
3	2012-963-03
4	2012-103-04

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 3 and 4 were available. In primer solutions 3 and 4 it was only possible to test the 5'-primer, the 3'-primer was not possible to test.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2013-July-15

Approved by:



Production Quality Control

101.862-12 – including *Taq* polymerase, IFU-01101.862-12u – without *Taq* polymerase, IFU-02Visit www.olerup-ssp.com for

“Instructions for Use” (IFU)

Lot No.: **49S**

Lot-specific information

Declaration of Conformity

Product name: Olerup SSP® HLA-C*04 Add-on**Product number:** 101.862-12/12u**Lot number:** 49S**Intended use:** HLA-C*04 histocompatibility testing

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden
2013-July-15



Åsa Olausson
Production Quality Control