

101.623-12 – including *Taq* polymerase
 101.623-12u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **1S3**

Lot-specific information
CERTIFICATE OF ANALYSIS

Olerup SSP® HLA-C*08

Product number: 101.623-12 – including *Taq* polymerase
 101.623-12u – without *Taq* polymerase
Lot number: 1S3
Expiry date: 2027-05-01
Number of tests: 12
Number of wells per test: 41+1

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2022-433-01	17	2022-433-17	33	2022-433-33
2	2022-433-02	18	2022-433-18	34	2022-433-34
3	2022-433-03	19	2022-433-19	35	2022-433-35
4	2022-433-04	20	2022-433-20	36	2022-433-36
5	2022-433-05	21	2022-433-21	37	2022-433-37
6	2022-433-06	22	2022-433-22	38	2022-433-38
7	2022-433-07	23	2022-433-23	39	2022-433-39
8	2022-433-08	24	2022-433-24	40	2022-433-40
9	2022-433-09	25	2022-433-25	41	2022-433-41
10	2022-433-10	26	2022-433-26		
11	2022-433-11	27	2022-433-27		
12	2022-433-12	28	2022-433-28		
13	2022-433-13	29	2022-433-29		
14	2022-433-14	30	2022-433-30		
15	2022-433-15	31	2022-433-31		
16	2022-433-16	32	2022-433-32		

The negative control primer pairs, **Production No. 2022-404-01**, can detect contamination with PCR products diluted 10⁻⁷.

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval: 2023-06-15

Approved by: 

Production Quality Control



101.623-12 – including *Taq* polymerase
101.623-12u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **1S3**

Lot-specific information

Declaration of Conformity

Product name: *Olerup* SSP® HLA-C*08
Product number: 101.623-12/12u
Lot number: 1S3

Intended use: HLA-C*08 high resolution histocompatibility testing

Manufacturer: *CareDx* AB
Franzéngatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, *CareDx* AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *CareDx* AB.

Stockholm, Sweden

Date:

2023-06-16

Quality Assurance

