

Lot No.: 10L

Lot-specific information

www.olerup-ssp.com

CERTIFICATE OF ANALYSIS

Olerup SSP® KIR HLA Ligand SSP

Product number: 104.201-12 – including *Taq* polymerase
104.201-12u –without *Taq* polymerase
Lot number: 10L
Expiry date: 2013-July-01
Number of tests: 12
Number of wells per test: 5

Well specifications:

| Well No. | Production No. |
|----------|----------------|
| 1 | 2011-827-01 |
| 2 | 2011-827-02 |
| 3 | 2011-827-03 |
| 4 | 2011-827-04 |
| 5 | 2011-827-05 |

The specificity of each primer solution of the kit has been tested against 48 IHWS cell line DNAs.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2011-February-04

Approved by:



Quality Control, Supervisor

Lot No.: 10L

Lot-specific information

www.olerup-ssp.com

Declaration of Conformity

Product name: Olerup SSP® KIR HLA Ligand
Product number: 104.201-12/12u
Lot number: 10L

Intended use: Determination of HLA-C, HLA-B^{Bw4+} and HLA-A^{Bw4+} KIR ligand sequence motifs.

Manufacturer: Olerup SSP AB
Hasselstigen 1
SE-133 33 Saltsjöbaden, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: Olerup SSP AB.

Saltsjöbaden, Sweden
2011-February-04



Olle Olerup
Managing Director