

KIR HLA Ligand
104.201-12 – including *Taq* polymerase
104.201-12u – without *Taq* polymerase
Lot No.: 4E6

Certificates

Lot-specific information

Page 1 of 2
General "Instructions for Use"
IFU-03 can be downloaded from

www.olerup-ssp.com

CERTIFICATE OF ANALYSIS

Olerup SSP® KIR HLA Ligand SSP

Product number: 104.201-12 – including *Taq* polymerase
104.201-12u – without *Taq* polymerase
Lot number: 4E6
Expiry date: 2019-04-01
Number of tests: 12
Number of wells per test: 6+1

Well specifications:


Well No.	Production No.
1	2016-731-01
2	2015-520-02
3	2016-731-03
4	2016-731-04
5	2012-063-05
6	2016-731-06

The negative control primer pairs, Production No. 2016-695-01, can detect contamination with PCR products diluted 10^{-7} .

Results of Quality Control: No false positive or false negative amplifications obtained.

Date of approval:

Approved by:

161031


Production Quality Control

Declaration of Conformity

Product name: Olerup SSP® KIR HLA Ligand

Product number: 104.201-12/12u

Lot number: 4E6

Intended use: Determination of HLA-C, HLA-B^{Bw4+} and HLA-A^{Bw4+} KIR ligand sequence motifs.

Manufacturer: Olerup SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, Olerup SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden

31 Oct 2016 *Andrea Boman*

Andrea Boman
Development manager