

Lot No.: 99Y

Lot-specific information

www.olerup-ssp.com

CERTIFICATE OF ANALYSIS

Olerup SSP® HLA - Negative Control SSP

Product number: 102.102-01 – including *Taq* polymerase
 102.102-01u – without *Taq* polymerase

Lot number: 99Y

Expiry date: 2018-February-01

Number of tests: 96

Number of wells per test: 1

Well specification:

Well No.	Production No.
1	2015-545-01

The negative control primer solution has been tested in a dilution series of the corresponding PCR products, 1 to 10³ down to 1 to 10⁹.

Results: The negative control primer pairs can detect contamination with the corresponding PCR products diluted 1 to 10⁷.

Date of approval: 2015-10-19

Approved by:



Production Quality Control

Lot No.: **99Y**

Lot-specific information

www.olerup-ssp.com

Declaration of Conformity

Product name: *Olerup* SSP® HLA - Negative Control
Product number: 102.102-01/01u
Lot number: 99Y

Intended use: Negative Control in *Olerup* SSP® HLA typings.

Manufacturer: *Olerup* SSP AB
 Franzengatan 5
 SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and EN ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden
 2015-Oct-19



Daniel Malica
 Head of QA and Regulatory Affairs