

Master mix with Taq polymerase

101.803-01/101.804-01/101.805-01

Certificates

Page 1 of 2

Visit www.caredx.com for

"Instructions for Use" (IFU)

Lot No.: QW5

Lot-specific information

CERTIFICATE OF ANALYSIS

Olerup SSP® Master mix with Taq polymerase

Product number:

101.803-01/101.804-01/805-01

Lot number:

QW5

Volume:

Product No. 101.803-01 - 312 μ l Product No. 101.804-01 - 750 μ l Product No. 101.805-01 - 1125 μ l

Expiry date:

2028-12-01

The Master Mix lot is tested against an older approved lot as reference. Typings with both Master Mix lots are performed with an HLA Class I and an HLA Class II kit.

Results:

No false positive or false negative amplifications obtained. All

amplifications are of similar strength and specificity.

Date of approval:

2024-10-04

Approved by:

Production Quality Control



Master mix with Taq polymerase

Certificates

Page 2 of 2

101.803-01/101.804-01/101.805-01

Visit <u>www.caredx.com</u> for "Instructions for Use" (IFU)

Lot No.: QW5

Lot-specific information

Declaration of Conformity

Product name:

Master mix with Taq polymerase

Product number:

101.803-01/101.804-01/101.805-01

Lot number:

QW5

Intended use:

Master mix for Olerup SSP® Typing Kits

Manufacturer:

CareDx AB

Franzéngatan 5

SE-112 51 Stockholm, Sweden **Phone:** +46-8-508 939 00

Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx* AB, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: CareDx AB.

Stockholm, Sweden

2024-10-14

Tu Du

Date:

Quality Assurance