

Olerup SSP® DNA Size Marker

Product number:	103.202-100/500
Lot number:	3K1
Volume:	Product No. 103.202-100 – 2 x 500 µl Product No. 103.202-500 – 10 x 500 µl
Concentration:	20 ng/µl
Format:	Ready-to-load, 10 µl per gel lane
Expiry date:	2021-09-01
Storage:	2-8°C

DESCRIPTION 103.202-100: 2 vials, 500 µl each.
 103.202-500: 10 vials, 500 µl each.
 DNA ladder, 7 bands: 50, 100, 200, 300, 400, 500 and 1 000 bp,
 20 ng/µl
 in 10 mM Tris-HCl (pH 9.0), 50 mM KCl, 5% (v/v) glycerol
 (99.5%), cresol red 100 ng/µl.

The DNA ladder contains 7 fragments of equal intensity:

<u>Fragment</u>	<u>Number of base pairs</u>
1	1000
2	500
3	400
4	300
5	200
6	100
7	50

PRODUCT USE The DNA Size marker can be resolved well in standard LE agarose gels of 1-2%, in NuSieve 3:1 or Metaphor agarose gels up to 4%.
 The DNA Size Marker is supplied in a **ready-to-load format**. Typically load **10 µl per gel lane** of the ready-to-load DNA Size marker, equals 200 ng of DNA Size Marker.

STORAGE Store at 2-8°C.

SHIPPING The DNA Size Marker is shipped at ambient temperature.

STABILITY The DNA Size Marker is stable for 24 months when stored at 2-8°C.

Lot No.: **3K1****CERTIFICATE OF ANALYSIS****Olerup SSP[®] DNA Size Marker****Product number:** 103.202-100/-500**Lot number:** 3K1**Volume:** Product No. 103.202-100 – 2 x 500 µl

Product No. 103.202-500 – 10 x 500 µl

Concentration: 20 ng/µl**Expiry date:** 2021-09-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 20191015

Approved by:



Production Quality Control



Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100/500
Lot number: 3K1

Intended use: DNA Size Marker

Manufacturer: *CareDx AB*
 Franzéngatan 5
 SE-112 51 Stockholm, Sweden
Phone: +46-8-508 939 00
Fax: +46-8-717 88 18

We, CareDx AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) EN ISO 13485:2016, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *CareDx AB*, Franzéngatan 5, SE-112 51 Stockholm, Sweden.

The Authorized Representative located within the Community is: *CareDx AB*.

Stockholm, Sweden

Date: 2019-10-17

Quality Assurance




ADDRESSES:

Manufacturer:

Caredx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-508 939 00

Fax: +46-8-717 88 18

E-mail: orders-se@caredx.com

Web page: <https://labproducts.caredx.com/>

Distributed by:

CareDx GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: orders-at@caredx.com

Web page: <https://labproducts.caredx.com/>

CareDx Lab Solutions Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-653-78171

Fax: 610-344-7989

E-mail: orders-us@caredx.com

Web page: <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx GmbH**.