

103.203-100/500

Lot No.: 59Y

## Olerup SSP® DNA Size Marker for short gel runs

**Product number:** 103.203-100/500  
**Lot number:** 59Y  
**Volume:** Product No. 103.203-100 – 2 x 500 µl  
Product No. 103.203-500 – 10 x 500 µl  
**Concentration:** 20 ng/µl  
**Format:** Ready-to-load, 10 µl per gel lane  
**Expiry date:** 2017-May-01  
**Storage:** 2-8°C

**DESCRIPTION** 103.203-100: 2 vials, 500 µl each.  
103.203-500: 10 vials, 500 µl each.  
DNA ladder, 4 bands: 50, **200**, 500 and 1 000 bp.  
Double intensity of 200 bp band.  
20 ng/µl in 10 mM Tris-HCl (pH 9.0), 50 mM KCl, 5% (v/v)  
glycerol (99.5%), cresol red 100 ng/µl.

The DNA ladder contains 4 fragments, double intensity of 200 bp band:

<u>Fragment</u>	<u>Number of base pairs</u>
1	1000
2	500
<b>3</b>	<b>200</b>
4	50

**PRODUCT USE** The DNA Size marker can be resolved well in standard LE agarose gels of 1-2%, in NuSieve 3:1 or Metaphor agarose gels up to 4%.  
The DNA Size Marker for short gel runs is supplied in a **ready-to-load format**. Typically load **10 µl per gel lane** of the ready-to-load DNA Size Marker for short gel runs, equals 200 ng of DNA Size Marker for short gel runs.

**STORAGE** Store at 2-8°C.

**SHIPPING** The DNA Size Marker for short gel runs is shipped at ambient temperature.

**STABILITY** The DNA Size Marker for short gel runs is stable for 24 months when stored at 2-8°C.

103.203-100/500

Lot No.: **59Y**

## CERTIFICATE OF ANALYSIS

### **Olerup SSP® DNA Size Marker for short gel runs**

<b>Product number:</b>	<b>103.203-100/500</b>
<b>Lot number:</b>	<b>59Y</b>
<b>Volume:</b>	<b>Product No. 103.203-100 – 2 x 500 µl</b> <b>Product No. 103.203-500 – 10 x 500 µl</b>
<b>Concentration:</b>	<b>20 ng/µl</b>
<b>Expiry date:</b>	<b>2017-May-01</b>

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

**Results:** Four fragments of 50, 200, 500 and 1000 bp were visible on a UV transilluminator. Double intensity of 200 bp fragment.

**Date of approval:**

**Approved by:**

**Production Quality Control**

103.203-100/500

Lot No.: 59Y

## Declaration of Conformity

**Product name:** DNA Size Marker for short gel runs

**Product number:** 103.203-100/500

**Lot number:** 59Y

**Intended use:** DNA Size Marker for short gel runs

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2012, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden  
2015-Jun-01

Daniel Malica  
Head of QA and Regulatory Affairs

103.203-100/500

Lot No.: **59Y**

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup-ssp.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup SSP* distributors worldwide, contact **Olerup GmbH**.