

Olerup SSP[®] DNA Size Marker

Product number: 103.202-100/500
Lot number: 91N
Volume: Product No. 103.202-100 – 2 x 500 µl
Product No. 103.202-500 – 10 x 500 µl
Concentration: 20 ng/µl
Format: Ready-to-load, 10 µl per gel lane
Expiry date: 2014-May-01
Storage: 2-8°C

DESCRIPTION 103.202-100: 2 vials, 500 µl each.
103.202-500: 10 vials, 500 µl each.
DNA ladder, 7 bands: 50, 100, 200, 300, 400, 500 and 1 000 bp,
20 ng/µl
in 10 mM Tris-HCl (pH 9.0), 50 mM KCl, 5% (v/v) glycerol
(99.5%), cresol red 100 ng/µl.

The DNA ladder contains 7 fragments of equal intensity:

| <u>Fragment</u> | <u>Number of base pairs</u> |
|-----------------|-----------------------------|
| 1 | 1000 |
| 2 | 500 |
| 3 | 400 |
| 4 | 300 |
| 5 | 200 |
| 6 | 100 |
| 7 | 50 |

PRODUCT USE The DNA Size marker can be resolved well in standard LE agarose gels of 1-2%, in NuSieve 3:1 or Metaphor agarose gels up to 4%.
The DNA Size Marker is supplied in a **ready-to-load format**. Typically load **10 µl per gel lane** of the ready-to-load DNA Size marker, equals 200 ng of DNA Size Marker.

STORAGE Store at 2-8°C.

SHIPPING The DNA Size Marker is shipped at ambient temperature.

STABILITY The DNA Size Marker is stable for 24 months when stored at 2-8°C.

CERTIFICATE OF ANALYSIS

Olerup SSP[®] DNA Size Marker

Product number: 103.202-100/-500
Lot number: 91N
Volume: Product No. 103.202-100 – 2 x 500 µl
Product No. 103.202-500 – 10 x 500 µl
Concentration: 20 ng/ µl
Expiry date: 2014-May-01

10 µl DNA Size Marker was size-separated on a 2% agarose gel stained by ethidium bromide.

Results: Seven fragments of 50, 100, 200, 300, 400, 500 and 1000 bp, were visible on a UV transilluminator.

Date of approval: 2012-May-28

Approved by:

Production Quality Control

Lot No.: **91N**

Declaration of Conformity

Product name: DNA Size Marker
Product number: 103.202-100/500
Lot number: 91N

Intended use: DNA Size Marker

Manufacturer: *Olerup* SSP AB
Franzengatan 5
SE-112 51 Stockholm, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Stockholm, Sweden
2012-May-28

Ann-Cathrin Jareman
Head of QA and Regulatory Affairs

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: info-ssp@olerup.com

Web page: <http://www.olerup-ssp.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: support-at@olerup.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: info.us@olerup.com

Web page: <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.