

101.201-48/12 – including *Taq* pol., IFU-01
101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

Olerup SSP[®] DQ low resolution

Product number:	101.201-48/12 – including <i>Taq</i> pol. 101.201-48u/12u – without <i>Taq</i> pol.
Lot number:	06Y
Expiry date:	2017-August-01
Number of tests:	48 tests – Product No. 101.201-48/48u 12 tests – Product No. 101.201-12/12u
Number of wells per test:	13+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 06Y.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

CHANGES COMPARED TO THE PREVIOUS OLERUP SSP[®] DQ LOW RESOLUTION LOT (68V)

A well containing Negative Control primer pairs has been added.

The format of the Product Insert and Worksheet have been changed.

Six wells have been added to DQ low resolution, wells **9 to 14**.

The 16 well cut PCR plate is marked with ‘DQ LOW’ in silver/gray ink.

The DQ low resolution specificity and interpretation tables have been updated for the HLA-DQB1 alleles described since the previous *Olerup SSP[®]* DQ low resolution lot was made (**Lot No 68V**). The kit design is based on IMGT/HLA database 3.19.0.

As of lot series V, the Specificity Table is included in the lot-specific Product Insert, and the Interpretation Table is included in the Worksheet.

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
2	-	Added	3'-primer added for the DQB1*06:01:13 allele.
5	Added	-	5'-primer added for the DQB1*03:02:16 allele.
7	Moved	Moved	Primer pair moved to well 12.
9	New	New	New primer pair for improved resolution of DQB1*04 alleles, strength of control band has been optimized.
10	New	New	New primer pair for increased redundancy and resolution of DQB1*05 alleles.
11	New	New	New primer pair for improved resolution of the DQ8 serologic group of alleles.
12	Added	Added	Primer pair added from well 12, for improved redundancy of the DQ3 serologic group of alleles.
13	New	New	New primer pairs for improved redundancy for the DQB1*06 alleles.
14	-	-	Negative Control.

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

Well **14** contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup SSP*® HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
 The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

PRODUCT DESCRIPTION

DQ low resolution SSP typing

CONTENT

The primer set contains 5'- and 3'-primers for grouping the DQB1 alleles into the serological groups DQ2 to DQ9.

Please note that DQB1 amplifications usually are somewhat less pronounced than e.g. DRB and DQA1 amplifications even when using the same DNA preparation and exactly the same experimental procedures.

PLATE LAYOUT

Each test consists of 14 PCR reactions in a 16 well PCR plate. Wells 15 and 16 are empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	NC	empty	empty

The 16 well cut PCR plate is marked with ‘DQ LOW’ in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘06Y’.

Wells 1 to 13 – DQ low resolution primers.

Well 14 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 16 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Only the DQB1 alleles will be amplified by the DQ low resolution typing kit. Thus, the interpretation of DQ low resolution typings is not influenced the DQB2 and DQB3 genes.

UNIQUELY IDENTIFIED ALLELES

All the DQB1 alleles, i.e. **DQB1*05:01 to 05:89**, **DQB1*06:01:01 to 06:167**, **DQB1*02:01:01 to 02:53Q**, **DQB1*03:01:01:01 to 03:167** and **DQB1*04*01:01 to 04:27**, recognized by the HLA Nomenclature Committee in January 2015^{1,2} will be amplified by the primers in the DQ low resolution SSP kit. The DQB1 alleles will be grouped into their corresponding serological specificities, i.e.:

DQ5(1) = DQB1*05:01-05:05³
 DQ6(1) = DQB1*06:01-06:33³
 DQ2 = DQB1*02:01-02:05
 DQ3 = DQB1*03:01-03:20³
 DQ4 = DQB1*04:01-04:02³

101.201-48/12 – including *Taq* pol., IFU-01
101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

The DQ3 alleles may be further subdivided into the DQ3, DQ7, DQ8 and DQ9 based upon serology and expert assignment. Thus;

DQ3	=	DQB1*03:06, 03:10, 03:14
DQ7	=	DQB1*03:01:01-03:01:03, 03:04, 03:09, 03:13, 03:16, 03:19
DQ8	=	DQB1*03:02:01, 03:05:01, 03:07, 03:08, 03:11, 03:18
DQ9	=	DQB1*03:03:02, 03:12, 03:15, 03:17, 03:20

¹HLA-DQB1 alleles listed on the IMGT/HLA web page 2015-January-19, release 3.19.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

³The serological split of the DQB1*05:05 to 05:89, DQB1*06:06 to 06:07, 06:10, 06:13, 06:15 to 06:24 and 06:27 to 06:167, the DQB1*02:04 to 02:53Q the DQB1*03:02:02 to 03:02:04, 03:03:03, 03:05:02, 03:07 to 03:09 and 03:11 to 03:167 and the DQB1*04:0301 to 04:27 alleles is not known. The grouping of not serologically defined alleles is taken from the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170.

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information
SPECIFICITY TABLE

DQ low resolution SSP typing

Specificities and sizes of the PCR products of the 13+1 primer mixes used for DQ low resolution SSP typing

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	DQ serology ³	Amplified DQB1 alleles ⁴
1	135 bp, 225 bp	515 bp	5	*05:01:01:01-05:59, 05:61-05:81, 05:83-05:89
2⁷	135 bp 185 bp 220 bp 270 bp	515 bp	1, 5, 6	*06:01:01-06:145, 06:147-06:167
3	210 bp	430 bp	2	*02:01:01-02:53Q
4	220 bp	515 bp	3, 7	*03:01:01:01-03:01:28, 03:04:01-03:04:02, 03:09-03:10:02, 03:13-03:14:02, 03:16, 03:19, 03:21-03:22, 03:24, 03:27-03:29, 03:35-03:36, 03:42, 03:44, 03:46-03:60, 03:69, 03:71, 03:73, 03:75-03:77, 03:80, 03:82-03:84N, 03:92-03:94, 03:101-03:103, 03:108, 03:114-03:116, 03:118N-03:122, 03:127-03:131, 03:133-03:135, 03:138-03:140, 03:142-03:144, 03:147-03:148, 03:150, 03:152, 03:157-03:160, 03:162-03:167
5⁶	130 bp 220 bp	515 bp	6, 8	*03:02:01-03:02:17, 03:05:01-03:05:04, 03:07-03:08, 03:11, 03:18, 03:32, 03:37, 03:45, 03:61, 03:63-03:64, 03:66N-03:68, 03:70, 03:85, 03:104, 03:106-03:107, 03:125, 03:132, 03:146, 03:153, 03:161, 06:29, 06:123, 06:139
6^{6,8}	135 bp	515 bp	2, 3, 6, 9	*02:03, 03:03:02:01-03:03:13, 03:06, 03:12, 03:15, 03:20, 03:25-03:26, 03:30-03:31, 03:33-03:34, 03:38-03:41, 03:43, 03:65, 03:74, 03:79, 03:86-03:91Q, 03:95N-03:99Q, 03:104-03:105, 03:111-03:113, 03:117, 03:123-03:124, 03:126, 03:136-03:137, 03:141, 03:145, 03:149, 03:155-03:156, 04:03:01-04:03:02, 06:03:10, 06:51:01, 06:66, 06:96
7^{5,6,7}	90 bp	515 bp	3, 7, 8, 9	*03:01:01:01-03:01:06, 03:01:08-03:05:04, 03:07, 03:09-03:24, 03:26-03:57, 03:59-03:64, 03:66N-03:103, 03:105-03:106, 03:108-03:136, 03:138-03:167
8^{6,7}	160 bp, 205 bp	430 bp	4	*03:132, 04:01:01-04:02:07, 04:03:01-04:27
9	225 bp	430 bp	4	*04:01:01-04:03:02, 04:06-04:21, 04:22 ^w , 04:23, 04:25N-04:27
10	215 bp	430 bp	5	*05:01:01:01-05:03:09, 05:03:11-05:03:16, 05:05:01-05:33, 05:35-05:43, 05:45-05:51,

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

				05:53, 05:55-05:71, 05:74-05:76, 05:78-05:89
11	185 bp	430 bp	8, 9	*03:03:11, 03:05:01, 03:05:03, 03:17:01, 03:61, 03:72, 03:100
12⁶	185 bp	430 bp	4, 7, 8, 9	*03:01:01:01-03:01:01:03, 03:01:03-03:01:07, 03:01:09-03:02:02, 03:02:04-03:02:12, 03:02:14-03:03:02:03, 03:03:04-03:04:02, 03:05:03-03:17:01, 03:18-03:19, 03:21-03:22, 03:24-03:36, 03:38-03:60, 03:62-03:71, 03:74, 03:76-03:98, 03:101-03:103, 03:106-03:108, 03:110-03:111, 03:113-03:117, 03:119-03:131, 03:133-03:153, 03:155, 03:157-03:161, 03:163-03:167, 04:01:03
13	185 bp	515 bp	6	*04:10, 06:02:01-06:02:23, 06:14:01-06:16, 06:19:01-06:20, 06:23-06:24, 06:33, 06:37, 06:46-06:50, 06:51:02, 06:68, 06:70-06:84, 06:95, 06:97, 06:107, 06:109, 06:111-06:117, 06:122, 06:124-06:127, 06:136-06:138, 06:146-06:147, 06:150-06:152, 06:156, 06:159, 06:161-06:163, 06:166
14^{8,9}	-	-	-	Negative Control

¹ Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DQ low resolution SSP typings. When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers. PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

² The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distributed as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases. In the presence of a specific amplification the intensity of the control band often decreases.

³ The serological split of the DQB1*05:05 to 05:89, DQB1*06:06 to 06:07, 06:10, 06:13, 06:15 to 06:24 and 06:27 to 06:167, the DQB1*02:04 to 02:53Q the DQB1*03:02:02 to 03:02:04, 03:03:03, 03:05:02, 03:07 to 03:09 and 03:11 to 03:167 and the DQB1*04:0301 to 04:27 alleles is not known. The grouping of not serologically defined alleles is taken from the expert-assigned serological grouping in Tissue Antigens (2009) 73:95-170.

⁴ For several DQB1 alleles 1st and/or 3rd exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

101.201-48/12 – including *Taq* pol., IFU-01
101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

⁵HLA-Specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁶Primer mixes 5 to 8 and 12 may give a lower yield of HLA-specific PCR products than the other DQ low resolution primer mixes.

⁷The primer pairs in wells 2, 7 and 8 will in some samples give rise to two HLA-specific PCR fragments.

⁸Primer mixes 6 and 14 may have a tendency to giving rise to primer oligomer formation.

⁹Primer mix 14 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

‘ser’, serological HLA specificity

101.201-48/12 – including *Taq* pol., IFU-01
 101.201-48u/12u – without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
 “Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information
PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13
Length of spec. PCR product	135 230	135 185 220 270	210	220	130 220	135	90	160 205	225	215	185	185	185
Length of int. pos. control ¹	515	515	430	515	515	515	515	430	430	430	430	430	515
5'-primer(s) ²	25(170) 5'-gCA 3'	9(122) 5'-gTT 3'	29(184) 5'-gAg 3'	26(173) 5'-TTA 3'	28(179) 5'-gAC 3'	26(173) 5'-TCT 3'	71(309) 5'-ACC 3'	23(164) 5'-gCT 3'	9(122) 5'-gTT 3'	29(184) 5'-gAC 3'	21(159) 5'-ACC 3'	38(210) 5'-gCA 3'	9(122) 5'-gTT 3'
	26(173) 5'-ggg 3'	24(169) 5'-TgT 3'	30(185) 5'-AAg 3'		28(179) 5'-gAC 3'		71(309) 5'-ACC 3'	38(210) 5'-gCg 3'					
		26(173) 5'-TTA 3'			28(179) 5'-gAC 3'								
		26(173) 5'-TCT 3'											
3'-primer(s) ³	57(266) 5'-CAA 3'	57(266) 5'-CAA 3'	86(353) 5'-gCT 3'	86(353) 5'-gCT 3'	57(266) 5'-Cgg 3'	57(266) 5'-CgT 3'	86(353) 5'-gCT 3'	77(327) 5'-ACg 3'	69(304) 5'-CTC 3'	87(356) 5'-ggT 3'	69(304) 5'-CCT 3'	86(353) 5'-gCT 3'	57(266) 5'-CAT 3'
	87(356) 5'-ggT 3'	86(353) 5'-ACg 3'		86(354) 5'-AgT 3'	57(266) 5'-CAg 3'		86(354) 5'-AgT 3'						58(270) 5'-TCC 3'
	87(356) 5'-ggT 3'	86(353) 5'-ACC 3'			57(266) 5'-Cgg 3'		86(355) 5'-gAC 3'						
	88(361) 5'-CCT 3'	86(354) 5'-TAT 3'			87(356) 5'-ggg 3'		87(358) 5'-gCC 3'						
		86(354) 5'-AAA 3'											
		86(354) 5'-AAg 3'											
		86(354) 5'-AAT 3'											
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.201-48/12– including *Taq* pol., IFU-01
101.201-48u/12u– without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

CELL LINE VALIDATION SHEET																
DQ low resolution primer set ²																
				Well												
				1	2	3	4	5	6	7	8	9	10	11	12	13
				201548901	201548902	201548903	201548904	201548905	201548906	201548907	201548908	201548909	201548910	201548911	201548912	201548913
			Production No.	IHWC cell line ¹				DQB1								
1	9001	SA	*05:01	+	-	-	-	-	-	-	-	-	+	-	-	-
2	9280	LK707	*06:01	*02:02	-	+	+	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*06:01		-	+	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*02:01		-	-	+	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*05:02		+	-	-	-	-	-	-	-	+	-	-	-
6	9353	SM	*03:02	*06:01	-	+	-	-	+	-	+	-	-	-	+	-
7	9020	QBL	*02:01		-	-	+	-	-	-	-	-	-	-	-	-
8	9025	DEU	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
9	9026	YAR	*03:02		-	-	-	-	+	-	+	-	-	-	+	-
10	9107	LKT3	*04:01		-	-	-	-	-	-	+	+	-	-	-	-
11	9051	PTOUT	*02:02		-	-	+	-	-	-	-	-	-	-	-	-
12	9052	DBB	*03:03		-	-	-	-	-	+	+	-	-	-	+	-
13	9004	JESTHOM	*05:01		+	-	-	-	-	-	-	-	+	-	-	-
14	9071	OLGA	*04:02		-	-	-	-	-	-	+	+	-	-	-	-
15	9075	DKB	*03:03		-	-	-	-	-	+	+	-	-	-	+	-
16	9037	SWEIG007	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
17	9282	CTM 3953540	*02:01	*06:03	-	+	+	-	-	-	-	-	-	-	-	-
18	9257	32367	*06:02	*02:02	-	+	+	-	-	-	-	-	-	-	-	+
19	9038	BM16	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
20	9059	SLE005	*06:04		-	+	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
22	9056	KOSE	*05:03	*06:04	+	+	-	-	-	-	-	-	+	-	-	-
23	9124	IHL	*05:03	*06:01	+	+	-	-	-	-	-	-	+	-	-	-
24	9035	JBUSH	*03:01		-	-	+	-	-	+	-	-	-	-	+	-
25	9049	IBW9	*02:02		-	-	+	-	-	-	-	-	-	-	-	-
26	9285	WT49	*02:01		-	-	+	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*04:01	*05:01	+	-	-	-	-	-	-	+	+	-	-	-
28	9320	BEL5GB	*02:02	*03:01	-	-	+	+	-	-	+	-	-	-	+	-
29	9050	MOU	*02:02		-	-	+	-	-	-	-	-	-	-	-	-
30	9021	RSH	*04:02		-	-	-	-	-	-	+	+	-	-	-	-
31	9019	DUCAF	*02:01		-	-	+	-	-	-	-	-	-	-	-	-
32	9297	HAG	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
33	9098	MT14B	*03:02		-	-	-	-	+	-	+	-	-	-	+	-
34	9104	DHIF	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
35	9302	SSTO	*03:05		-	-	-	-	+	-	+	-	-	-	+	-
36	9024	KT17	*03:02		-	-	-	-	+	-	+	-	-	-	+	-
37	9065	HHKB	*06:03		-	+	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
39	9315	CML	*02:01	*03:01	-	-	+	+	-	-	+	-	-	-	+	-
40	9134	WHONP199	*02:02	*03:03	-	-	+	-	-	+	+	-	-	-	+	-
41	9055	H0301	*06:09		-	+	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*06:01		-	+	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*03:03		-	-	-	-	-	+	+	-	-	-	+	-
44	9057	TEM	*05:03		+	-	-	-	-	-	-	-	-	+	-	-
45	9239	SHJO	*02:02		-	-	+	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*06:02		-	+	-	-	-	-	-	-	-	-	-	+
47	9045	TUBO	*03:01		-	-	-	+	-	-	+	-	-	-	+	-
48	9303	TER-ND	*05:01		+	-	-	-	-	-	-	-	+	-	-	-

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.



101.201-48/12– including *Taq* pol., IFU-01
101.201-48u/12u– without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.
In primer solutions 1, 2, 3 and 5 one 5'-primer was not possible to test, and in primer solutions 1, 2, 4, 5, 7 and 13 one or two 3'-primers were not possible to test.
Additional 5'-primers in primer solutions 2 and 7 were tested by separately adding another 3'-primer.
Additional 3'-primers in primer solutions 1 and 2 were tested by separately adding one 5'-primer.

101.201-48/12– including *Taq* pol., IFU-01
101.201-48u/12u– without *Taq* pol., IFU-02

Visit www.olerup-ssp.com for
“Instructions for Use” (IFU)

Lot No.: **06Y**

Lot-specific information

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: info-ssp@olerup.com

Web page: <http://www.olerup-ssp.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: support-at@olerup.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: info.us@olerup.com

Web page: <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.