

Lot No.: **67E**

Lot-specific Information

## **Olerup SSP<sup>®</sup> DQB1\*04**

Product number: 101.215-12u – without *Taq* polymerase  
Lot number: 67E  
Expiry date: 2010-June-01  
Number of tests: 12  
Number of wells per test: 4  
Storage - pre-aliquoted primers: dark at -20°C  
- PCR Master Mix: -20°C  
- Adhesive PCR seals RT  
- Product Insert RT

**This Product Description is only valid for Lot No. 67E.**

### **CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP<sup>®</sup>* DQB1\*04 LOT**

The DQB1\*04 specificity and interpretation tables have been updated with the DQB1 alleles described since the previous *Olerup SSP<sup>®</sup>* DQB1\*04 lot (**Lot No. V70**) was made.

One well has been added to the DQB1\*04 kit,  
well **4**.

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
1	Exchanged	-	Primer exchanged to amplify all DQB1*04 alleles.
4	New	New	New primer pair for the DQB1*0403 allele.

Lot No.: **67E**

Lot-specific Information

## PRODUCT DESCRIPTION

### DQB1\*04 SSP subtyping

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the DQB1\*0401 to DQB1\*0403 alleles.

*Please note that DQB1 amplifications usually are somewhat less pronounced than e.g. DRB and DQA1 amplifications even when using the same DNA preparation and exactly the same experimental procedures.*

#### PLATE LAYOUT

Each test consists of 4 PCR reactions in an 8 well PCR plate. Wells 5 to 8 are empty.

1	2	3	4	empty	empty	empty	empty
---	---	---	---	-------	-------	-------	-------

The 8 well PCR plate is marked with 'DQB1\*04'.

Well No. 1 is marked with the Lot No. '67E'.

The PCR plates are covered with a PCR-compatible foil.

**Please note:** When removing each 8 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

Only the DQB1\*04 alleles will be amplified by the DQB1\*04 subtyping kit, except that the DQB1\*0306 allele will be amplified by primer mix 4. Thus, the interpretation of DQB1\*04 subtypings is not influenced by other groups of the DQB1 alleles or the DQB2 and DQB3 genes.

#### UNIQUELY IDENTIFIED ALLELES

All the DQB1\*04 alleles, i.e. **DQB1\*0401 to DQB1\*0403**, recognized by the HLA Nomenclature Committee in April 2008<sup>1</sup> will give rise to unique amplification patterns by the primers in the DQB1\*04 subtyping kit.

<sup>1</sup>HLA-DQB1 alleles listed on the IMGT/HLA web page 2008-April-08, release 2.21.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

#### RESOLUTION IN HOMO- AND HETEROZYGOTES

The 3 DQB1\*04 alleles can be combined in 6 homozygous and heterozygous combinations. All but two of these genotypes give rise to unique amplification patterns.

+--+ 0402,0403 = 0403,0403

Lot No.: **67E**

Lot-specific Information

## SPECIFICITY TABLE

### DQB1\*04 SSP subtyping

Specificities and sizes of the PCR products of the 4 primer mixes used for DQB1\*04 SSP subtyping

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified DQB1 <sup>3</sup> alleles
1	210	515 bp	0401-0403
2	205	515 bp	0401
3	205	430 bp	0402, 0403
4	195	430 bp	0403, 0306

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DQB1\*04 SSP typings. When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective length of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherit feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DQB1\*04 subtyping.

In addition, well number 2 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>For several DQB1 alleles only partial 2<sup>nd</sup> exon nucleotide sequences are available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. We assume that unknown sequences of the 2<sup>nd</sup> exon of DQB1 alleles are conserved within allelic groups.

Lot No.: **67E**

Lot-specific Information

<b>INTERPRETATION TABLE</b>				
<b>DQB1*04 SSP subtyping</b>				
<b>Amplification patterns of the DQB1*04 alleles</b>				
	<b>Well</b>			
	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>
<b>Length of spec.</b>	<b>210</b>	<b>205</b>	<b>205</b>	<b>195</b>
<b>PCR product</b>				
<b>Length of int.</b>	<b>515</b>	<b>515</b>	<b>430</b>	<b>430</b>
<b>pos. control<sup>1</sup></b>				
<b>5'-primer<sup>2</sup></b>	<b>21</b>	<b>23</b>	<b>23</b>	<b>26</b>
	5'-ACC <sup>3'</sup>	5'-g CT <sup>3'</sup>	5'-g Cg <sup>3'</sup>	5'-T CT <sup>3'</sup>
<b>3'-primer<sup>3</sup></b>	<b>77</b>	<b>77</b>	<b>77</b>	<b>77</b>
	5'-AC g <sup>3'</sup>	5'-AC g <sup>3'</sup>	5'-AC g <sup>3'</sup>	5'-AC g <sup>3'</sup>
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>
<b>DQB1 allele</b>				
<b>*0401</b>	<b>1</b>	<b>2</b>		
<b>*0402</b>	<b>1</b>		<b>3</b>	
<b>*0403</b>	<b>1</b>		<b>3</b>	<b>4</b>
<b>*0306</b>				<b>4</b>
<b>DQB1 allele</b>				
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DQB1\*04 subtyping.

In addition, well number 2 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

<sup>2</sup>The codon, in the 2<sup>nd</sup> exon, matching the specificity-determining 3'-end of the primer is given. Codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given. Empty spaces indicate codon boundaries.

<sup>3</sup>The codon, in the 2<sup>nd</sup> exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given. Empty spaces indicate codon boundaries.

Lot No.: 67E

Lot-specific Information

CELL LINE VALIDATION SHEET								
DQB1*04 SSP subtyping kit								
				Production No.	Well			
					1	2	3	4
					200846901	200846902	200846903	200846904
	IHWC cell line		DQB1					
1	9001 SA		*0501		-	-	-	-
2	9280 LK707		*0601	*0202	-	-	-	-
3	9011 E4181324		*0601		-	-	-	-
4	9275 GU373		*0201		-	-	-	-
5	9009 KAS011		*0502		-	-	-	-
6	9353 SM		*0302	*0601	-	-	-	-
7	9020 QBL		*0201		-	-	-	-
8	9007 DEM		*0302	*0502	-	-	-	-
9	9026 YAR		*0302		-	-	-	-
10	9107 LKT3		*0401		+	+	-	-
11	9051 PITOUT		*0202		-	-	-	-
12	9052 DBB		*0303		-	-	-	-
13	9067 BTB		*0402		+	-	+	-
14	9071 OLGA		*0402		+	-	+	-
15	9075 DKB		*0303		-	-	-	-
16	9037 SWEIG007		*0301		-	-	-	-
17	9008 WILJON		*0602	*0603	-	-	-	-
18	9257 32367		*0602	*0202	-	-	-	-
19	9038 BM16		*0301		-	-	-	-
20	9059 SLE005		*0604		-	-	-	-
21	9064 AMALA		*0301		-	-	-	-
22	9056 KOSE		*0503	*0604	-	-	-	-
23	9124 IHL		*0503	*0601	-	-	-	-
24	9035 JBUSH		*0301		-	-	-	-
25	9049 IBW9		*0202		-	-	-	-
26	9285 WT49		*0201		-	-	-	-
27	9191 CH1007		*0401	*0501	+	+	-	-
28	9320 BEL5GB		*0202	*0301	-	-	-	-
29	9050 MOU		*0202		-	-	-	-
30	9021 RSH		*0402		+	-	+	-
31	9019 DUCAF		*0201		-	-	-	-
32	9297 HAG		*0301		-	-	-	-
33	9098 MT14B		*0302		-	-	-	-
34	9104 DHIF		*0301		-	-	-	-
35	9302 SSTO		*0305		-	-	-	-
36	9024 KT17		*0302		-	-	-	-
37	9065 HHKB		*0603		-	-	-	-
38	9099 LZL		*0301		-	-	-	-
39	9315 CML		*0201	*0301	-	-	-	-
40	9134 WHONP199		*0202	*0303	-	-	-	-
41	9055 H0301		*0609		-	-	-	-
42	9066 TAB089		*0601		-	-	-	-
43	9076 T7526		*0303		-	-	-	-
44	9057 TEM		*0503		-	-	-	-
45	9239 SHJO		*0202		-	-	-	-
46	9013 SCHU		*0602		-	-	-	-
47	9045 TUBO		*0301		-	-	-	-
48	9303 TER-ND		*0501		-	-	-	-



Lot No.: **67E**

Lot-specific Information

## CERTIFICATE OF ANALYSIS

### **Olerup SSP<sup>®</sup> DQB1\*04 SSP**

**Product number:** 101.215-12u – without *Taq* polymerase  
**Lot number:** 67E  
**Expiry date:** 2010-June-01  
**Number of tests:** 12  
**Number of wells per test:** 4

#### **Well specifications:**

Well No.	Production No.
1	2008-469-01
2	2008-469-02
3	2008-469-03
4	2008-469-04

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solution No. 4 were available. The specificities of the primers in this primer solution were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer.

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2009-May-26

**Approved by:**

**Quality Control, Supervisor**

Lot No.: **67E**

Lot-specific Information

## Declaration of Conformity

**Product name:** *Olerup* SSP® DQB1\*04  
**Product number:** 101.215-12u  
**Lot number:** 67E

**Intended use:** DQB1\*04 resolution histocompatibility testing

**Manufacturer:** *Olerup* SSP AB  
Hasselstigen 1  
SE-133 33 Saltsjöbaden, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2000 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Saltsjöbaden, Sweden  
2008-June-24

Olle Olerup  
Managing Director



Lot No.: **67E**

Lot-specific Information

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.