

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information
Olerup SSP® DQB1*06

| | |
|---|--|
| Product number: | 101.212-24/04 – including <i>Taq</i> pol. 101.212-24u/04u – without <i>Taq</i> pol. |
| Lot number: | 1G3 |
| Expiry date: | 2022-05-01 |
| Number of tests: | 24 tests – Product No. 101.212-24/24u 4 tests – Product No. 101.212-04/04u |
| Number of wells per test: | 63+1 |
| Storage - pre-aliquoted primers: | dark at -20°C |
| - PCR Master Mix: | -20°C |
| - Adhesive PCR seals | RT |
| - Product Insert | RT |

This Product Description is only valid for Lot No. 1G3.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP®
DQB1*06 Lot (5E8)**

The DQB1*06 kit is updated to enable separation of:

- Null and Alternatively expressed alleles
- The product documentation has been updated for new alleles of IMGT 3.30

The format of the Worksheet has been changed.

The DQB1*06 primer set, specificity and interpretation tables have been updated for the DQB1 alleles described since the previous *Olerup SSP®* DQB1*06 lot was made (**Lot No. 5E8**).

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

| Well | 5'-primer | 3'-primer | rationale |
|------|-----------|-----------|--|
| 19 | - | Added | 3'-primer added for the DQB1*06:216N allele. |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

Changes in revision R01 compared to R00:

1. Primer mix 7 does not amplify the DQB1*06:69:02 allele.
2. Primer mix 23 does not amplify the DQB1*06:03:07, 06:14:02, 06:206:01 and 06:206:02 alleles.
3. Primer mix 25 does not amplify the DQB1*06:03:07, 06:14:02, 06:206:01 and 06:206:02 alleles.

The corrections above have been implemented in the specificity and interpretation tables.

Change in revision R02 compared to R01:

1. The expiration date has been altered due to extension of shelf-life.

Change in revision R03 compared to R02:

1. Primer mix 43 does not amplify the DQB1*06:103 and DQB1*05:84 alleles. This correction has been implemented in the Specificity and Interpretation tables.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

Well **64** contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup* SSP[®] HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 430 base pairs.

| Length of PCR product | 105 | 200 | 105 | 80 | 75 | 80 | 85 |
|------------------------------|----------------------|-------------------------|----------------------|----------------------|----------------------|----------------------|----------------------|
| 5'-primer¹ | 164 | 340 | 440 | 45 | 45 | 43 | 36 |
| | 5'-CAC ^{3'} | 5'-Agg ^{3'} | 5'-TTA ^{3'} | 5'-Tgg ^{3'} | 5'-Tgg ^{3'} | 5'-Tgg ^{3'} | 5'-TAC ^{3'} |
| | | | | | | | 36 |
| | | | | | | | 5'-TAT ^{3'} |
| 3'-primer² | 231 | 2nd I | 507 | 59 | 58 | 57 | 47 |
| | 5'-TgC ^{3'} | 5'-AAA ^{3'} | 5'-TTg ^{3'} | 5'-CTC ^{3'} | 5'-ggC ^{3'} | 5'-CTC ^{3'} | 5'-ACA ^{3'} |
| | | | | | | | 48 |
| | | | | | | | 5'-gCA ^{3'} |
| | | | | | | | 48 |
| | | | | | | | 5'-gCC ^{3'} |
| | | | | | | | 52 |
| | | | | | | | 5'-TgT ^{3'} |
| A* | + | + | + | | | | |
| B* | + | + | + | | | | |
| C* | + | + | + | | | | |
| DRB1 | | | | + | + | | |
| DRB3 | | | | + | + | | |
| DRB5 | | | | + | | | |
| DQB1 | | | | | + | | |
| DPB1 | | | | | | + | |
| DQA1 | | | | | | | + |

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

PRODUCT DESCRIPTION

DQB1*06 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the DQB1*06:01 to DQB1*06:242 alleles.

Please note that DQB1 amplifications usually are somewhat less pronounced than e.g. DRB and DQA1 amplifications even when using the same DNA preparation and exactly the same experimental procedures.

PLATE LAYOUT

Each test consists of 64 PCR reactions in a 64 well cut PCR plate.

| | | | | | | | |
|----|----|----|----|----|----|----|----|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 |
| 9 | 10 | 11 | 12 | 13 | 14 | 15 | 16 |
| 17 | 18 | 19 | 20 | 21 | 22 | 23 | 24 |
| 25 | 26 | 27 | 28 | 29 | 30 | 31 | 32 |
| 33 | 34 | 35 | 36 | 37 | 38 | 39 | 40 |
| 41 | 42 | 43 | 44 | 45 | 46 | 47 | 48 |
| 49 | 50 | 51 | 52 | 53 | 54 | 55 | 56 |
| 57 | 58 | 59 | 60 | 61 | 62 | 63 | NC |

The 64 well cut PCR plate is marked with ‘DQB1*06’ in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘1G3’.

Wells 1 to 63 – DQB1*06 high resolution primers.

Well 64 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded. The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 64 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between DQB1 alleles non-DQB1*06 alleles will be amplified by some primer mixes. For further details see Specificity Table.

UNIQUELY IDENTIFIED ALLELES

All the DQB1*06 alleles, i.e. **DQB1*06:01 to DQB1*06:242**, recognized by the HLA Nomenclature Committee in October 2017^{1,2} will be amplified by the primers in the DQB1*06 subtyping kit^{1,2}.

The DQB1*06 kit enables separation of the confirmed DQB1*06 alleles as listed in the IMGT/HLA database 3.25. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for DQB1*06 alleles is listed below.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

The DQB1*06 kit also enables identification of null and alternatively expressed alleles.

The following DQB1*06 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

| Alleles | Primer mix | Alleles | Primer mix |
|--------------------|------------|----------------------|------------|
| DQB1*06:33, 06:114 | 27 | DQB1*06:97, 06:124 | 45 |
| DQB1*06:70, 06:75N | 40 | DQB1*06:134, 06:144N | 50 |
| DQB1*06:91, 06:128 | 48 | | |

¹HLA-DQB1 alleles listed on the IMGT/HLA web page 2017-October-27, release 3.30.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in DQB1*06 homo- and heterozygotes is available upon request.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

ALLELE CONFIRMATION STATUS

| Allele | Status ¹ | Allele | Status ¹ | Allele | Status ¹ | Allele | Status ¹ |
|----------------------|---------------------|----------------------|---------------------|----------------------|---------------------|----------------------|---------------------|
| DQB1*06:01:01 | Confirmed | DQB1*06:03:11 | Confirmed | DQB1*06:18:01 | Unconfirmed | DQB1*06:60 | Unconfirmed |
| DQB1*06:01:02 | Confirmed | DQB1*06:03:12 | Unconfirmed | DQB1*06:18:02 | Unconfirmed | DQB1*06:61 | Unconfirmed |
| DQB1*06:01:03 | Unconfirmed | DQB1*06:03:13 | Unconfirmed | DQB1*06:19:01 | Unconfirmed | DQB1*06:62 | Confirmed |
| DQB1*06:01:04 | Unconfirmed | DQB1*06:03:14 | Unconfirmed | DQB1*06:19:02 | Unconfirmed | DQB1*06:63 | Unconfirmed |
| DQB1*06:01:05 | Unconfirmed | DQB1*06:03:15 | Unconfirmed | DQB1*06:20 | Unconfirmed | DQB1*06:64 | Confirmed |
| DQB1*06:01:06 | Unconfirmed | DQB1*06:03:16 | Unconfirmed | DQB1*06:21 | Confirmed | DQB1*06:65 | Unconfirmed |
| DQB1*06:01:07 | Unconfirmed | DQB1*06:03:17 | Unconfirmed | DQB1*06:22:01 | Unconfirmed | DQB1*06:66 | Unconfirmed |
| DQB1*06:01:08 | Confirmed | DQB1*06:03:18 | Unconfirmed | DQB1*06:22:02 | Confirmed | DQB1*06:67 | Unconfirmed |
| DQB1*06:01:09 | Unconfirmed | DQB1*06:03:19 | Unconfirmed | DQB1*06:22:03 | Confirmed | DQB1*06:68 | Confirmed |
| DQB1*06:01:10 | Confirmed | DQB1*06:03:20 | Unconfirmed | DQB1*06:23 | Unconfirmed | DQB1*06:69 | Confirmed |
| DQB1*06:01:11 | Unconfirmed | DQB1*06:03:21 | Unconfirmed | DQB1*06:24 | Unconfirmed | DQB1*06:70 | Unconfirmed |
| DQB1*06:01:12 | Unconfirmed | DQB1*06:03:22 | Unconfirmed | DQB1*06:25 | Confirmed | DQB1*06:71 | Confirmed |
| DQB1*06:01:13 | Confirmed | DQB1*06:04:01 | Confirmed | DQB1*06:26N | Confirmed | DQB1*06:72 | Unconfirmed |
| DQB1*06:01:14 | Confirmed | DQB1*06:04:02 | Unconfirmed | DQB1*06:27:01 | Unconfirmed | DQB1*06:73 | Unconfirmed |
| DQB1*06:01:15 | Unconfirmed | DQB1*06:04:03 | Unconfirmed | DQB1*06:27:02 | Confirmed | DQB1*06:74 | Confirmed |
| DQB1*06:02:01 | Confirmed | DQB1*06:04:04 | Unconfirmed | DQB1*06:28 | Unconfirmed | DQB1*06:75N | Confirmed |
| DQB1*06:02:02 | Unconfirmed | DQB1*06:04:05 | Unconfirmed | DQB1*06:29 | Unconfirmed | DQB1*06:76 | Unconfirmed |
| DQB1*06:02:03 | Unconfirmed | DQB1*06:04:06 | Confirmed | DQB1*06:30 | Unconfirmed | DQB1*06:77N | Confirmed |
| DQB1*06:02:04 | Confirmed | DQB1*06:04:07 | Unconfirmed | DQB1*06:31 | Unconfirmed | DQB1*06:78 | Confirmed |
| DQB1*06:02:05 | Confirmed | DQB1*06:04:08 | Unconfirmed | DQB1*06:32:01 | Unconfirmed | DQB1*06:79:01 | Confirmed |
| DQB1*06:02:06 | Unconfirmed | DQB1*06:04:09 | Unconfirmed | DQB1*06:32:02 | Unconfirmed | DQB1*06:79:02 | Unconfirmed |
| DQB1*06:02:07 | Confirmed | DQB1*06:04:10 | Unconfirmed | DQB1*06:33 | Confirmed | DQB1*06:80 | Unconfirmed |
| DQB1*06:02:08 | Unconfirmed | DQB1*06:05:01 | Confirmed | DQB1*06:34 | Confirmed | DQB1*06:81 | Confirmed |
| DQB1*06:02:09 | Unconfirmed | DQB1*06:05:02 | Unconfirmed | DQB1*06:35 | Unconfirmed | DQB1*06:82 | Unconfirmed |
| DQB1*06:02:10 | Unconfirmed | DQB1*06:06 | Unconfirmed | DQB1*06:36 | Confirmed | DQB1*06:83 | Unconfirmed |
| DQB1*06:02:11 | Confirmed | DQB1*06:07:01 | Confirmed | DQB1*06:37 | Confirmed | DQB1*06:84 | Confirmed |
| DQB1*06:02:12 | Unconfirmed | DQB1*06:07:02 | Confirmed | DQB1*06:38 | Unconfirmed | DQB1*06:85 | Unconfirmed |
| DQB1*06:02:13 | Unconfirmed | DQB1*06:08:01 | Confirmed | DQB1*06:39 | Unconfirmed | DQB1*06:86 | Unconfirmed |
| DQB1*06:02:14 | Confirmed | DQB1*06:08:02 | Unconfirmed | DQB1*06:40 | Unconfirmed | DQB1*06:87 | Unconfirmed |
| DQB1*06:02:15 | Unconfirmed | DQB1*06:08:03 | Unconfirmed | DQB1*06:41 | Confirmed | DQB1*06:88 | Confirmed |
| DQB1*06:02:16 | Unconfirmed | DQB1*06:09:01 | Confirmed | DQB1*06:42 | Unconfirmed | DQB1*06:89 | Unconfirmed |
| DQB1*06:02:17 | Confirmed | DQB1*06:09:02 | Confirmed | DQB1*06:43 | Unconfirmed | DQB1*06:90 | Confirmed |
| DQB1*06:02:18 | Unconfirmed | DQB1*06:09:03 | Unconfirmed | DQB1*06:44 | Confirmed | DQB1*06:91 | Unconfirmed |
| DQB1*06:02:19 | Unconfirmed | DQB1*06:09:04 | Unconfirmed | DQB1*06:45 | Unconfirmed | DQB1*06:92 | Confirmed |
| DQB1*06:02:20 | Confirmed | DQB1*06:09:05 | Unconfirmed | DQB1*06:46 | Confirmed | DQB1*06:93 | Confirmed |
| DQB1*06:02:21 | Unconfirmed | DQB1*06:09:06 | Confirmed | DQB1*06:47 | Unconfirmed | DQB1*06:94 | Confirmed |
| DQB1*06:02:22 | Unconfirmed | DQB1*06:10 | Confirmed | DQB1*06:48 | Confirmed | DQB1*06:95 | Unconfirmed |
| DQB1*06:02:23 | Unconfirmed | DQB1*06:11:01 | Confirmed | DQB1*06:49 | Confirmed | DQB1*06:96 | Unconfirmed |
| DQB1*06:02:24 | Unconfirmed | DQB1*06:11:02 | Confirmed | DQB1*06:50 | Confirmed | DQB1*06:97 | Unconfirmed |
| DQB1*06:02:25 | Unconfirmed | DQB1*06:11:03 | Unconfirmed | DQB1*06:51:01 | Unconfirmed | DQB1*06:98 | Unconfirmed |
| DQB1*06:03:01 | Confirmed | DQB1*06:12 | Unconfirmed | DQB1*06:51:02 | Unconfirmed | DQB1*06:99:01 | Unconfirmed |
| DQB1*06:03:02 | Confirmed | DQB1*06:13:01 | Unconfirmed | DQB1*06:52 | Unconfirmed | DQB1*06:99:02 | Unconfirmed |
| DQB1*06:03:03 | Unconfirmed | DQB1*06:13:02 | Confirmed | DQB1*06:53:01 | Confirmed | DQB1*06:100 | Unconfirmed |
| DQB1*06:03:04 | Unconfirmed | DQB1*06:14:01 | Confirmed | DQB1*06:53:02 | Unconfirmed | DQB1*06:101 | Unconfirmed |
| DQB1*06:03:05 | Confirmed | DQB1*06:14:02 | Unconfirmed | DQB1*06:54N | Unconfirmed | DQB1*06:102N | Unconfirmed |
| DQB1*06:03:06 | Unconfirmed | DQB1*06:14:03 | Confirmed | DQB1*06:55 | Confirmed | DQB1*06:103 | Confirmed |
| DQB1*06:03:07 | Confirmed | DQB1*06:15:01 | Unconfirmed | DQB1*06:56 | Unconfirmed | DQB1*06:104 | Unconfirmed |
| DQB1*06:03:08 | Unconfirmed | DQB1*06:15:02 | Confirmed | DQB1*06:57 | Confirmed | DQB1*06:105 | Unconfirmed |
| DQB1*06:03:09 | Confirmed | DQB1*06:16 | Unconfirmed | DQB1*06:58 | Confirmed | DQB1*06:106 | Confirmed |
| DQB1*06:03:10 | Unconfirmed | DQB1*06:17 | Unconfirmed | DQB1*06:59 | Confirmed | DQB1*06:107 | Confirmed |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| Allele | Status ¹ | Allele | Status ¹ | Allele | Status ¹ | Allele | Status ¹ |
|-----------------------|---------------------|---------------------|---------------------|---------------------|---------------------|-------------|---------------------|
| DQB1*06:108 | Unconfirmed | DQB1*06:137 | Unconfirmed | DQB1*06:166 | Unconfirmed | DQB1*06:196 | Unconfirmed |
| DQB1*06:109 | Unconfirmed | DQB1*06:138 | Unconfirmed | DQB1*06:167 | Unconfirmed | DQB1*06:197 | Unconfirmed |
| DQB1*06:110 | Unconfirmed | DQB1*06:139 | Unconfirmed | DQB1*06:168 | Unconfirmed | DQB1*06:198 | Unconfirmed |
| DQB1*06:111 | Unconfirmed | DQB1*06:140 | Unconfirmed | DQB1*06:169 | Confirmed | DQB1*06:199 | Unconfirmed |
| DQB1*06:112N | Unconfirmed | DQB1*06:141 | Unconfirmed | DQB1*06:170 | Confirmed | DQB1*06:200 | Unconfirmed |
| DQB1*06:113 | Unconfirmed | DQB1*06:142 | Unconfirmed | DQB1*06:171 | Confirmed | DQB1*06:201 | Unconfirmed |
| DQB1*06:114 | Unconfirmed | DQB1*06:143 | Unconfirmed | DQB1*06:172 | Confirmed | DQB1*06:202 | Unconfirmed |
| DQB1*06:115 | Unconfirmed | DQB1*06:144N | Unconfirmed | DQB1*06:173 | Unconfirmed | DQB1*06:203 | Unconfirmed |
| DQB1*06:116 | Unconfirmed | DQB1*06:145 | Unconfirmed | DQB1*06:174 | Unconfirmed | DQB1*06:204 | Unconfirmed |
| DQB1*06:117 | Unconfirmed | DQB1*06:146:01 | Unconfirmed | DQB1*06:175 | Unconfirmed | DQB1*06:205 | Unconfirmed |
| DQB1*06:118:01 | Unconfirmed | DQB1*06:146:02 | Unconfirmed | DQB1*06:176 | Unconfirmed | DQB1*06:206 | Unconfirmed |
| DQB1*06:118:02 | Confirmed | DQB1*06:147 | Unconfirmed | DQB1*06:177 | Unconfirmed | DQB1*06:207 | Unconfirmed |
| DQB1*06:119 | Unconfirmed | DQB1*06:148 | Unconfirmed | DQB1*06:178 | Unconfirmed | DQB1*06:208 | Unconfirmed |
| DQB1*06:120 | Confirmed | DQB1*06:149 | Confirmed | DQB1*06:179N | Confirmed | DQB1*06:209 | Unconfirmed |
| DQB1*06:121 | Unconfirmed | DQB1*06:150 | Unconfirmed | DQB1*06:180 | Unconfirmed | | |
| DQB1*06:122 | Unconfirmed | DQB1*06:151 | Unconfirmed | DQB1*06:181 | Unconfirmed | | |
| DQB1*06:123 | Confirmed | DQB1*06:152 | Unconfirmed | DQB1*06:182 | Unconfirmed | | |
| DQB1*06:124 | Unconfirmed | DQB1*06:153 | Confirmed | DQB1*06:183 | Unconfirmed | | |
| DQB1*06:125 | Confirmed | DQB1*06:154 | Unconfirmed | DQB1*06:184 | Unconfirmed | | |
| DQB1*06:126 | Confirmed | DQB1*06:155 | Confirmed | DQB1*06:185 | Unconfirmed | | |
| DQB1*06:127 | Unconfirmed | DQB1*06:156 | Unconfirmed | DQB1*06:186 | Unconfirmed | | |
| DQB1*06:128 | Unconfirmed | DQB1*06:157 | Unconfirmed | DQB1*06:187 | Unconfirmed | | |
| DQB1*06:129 | Unconfirmed | DQB1*06:158N | Confirmed | DQB1*06:188 | Unconfirmed | | |
| DQB1*06:130 | Unconfirmed | DQB1*06:159 | Unconfirmed | DQB1*06:189 | Unconfirmed | | |
| DQB1*06:131 | Unconfirmed | DQB1*06:160 | Unconfirmed | DQB1*06:190 | Unconfirmed | | |
| DQB1*06:132 | Unconfirmed | DQB1*06:161 | Unconfirmed | DQB1*06:191 | Confirmed | | |
| DQB1*06:133 | Unconfirmed | DQB1*06:162 | Confirmed | DQB1*06:192 | Unconfirmed | | |
| DQB1*06:134 | Unconfirmed | DQB1*06:163 | Confirmed | DQB1*06:193N | Unconfirmed | | |
| DQB1*06:135 | Unconfirmed | DQB1*06:164 | Confirmed | DQB1*06:194 | Unconfirmed | | |
| DQB1*06:136 | Unconfirmed | DQB1*06:165 | Unconfirmed | DQB1*06:195 | Unconfirmed | | |

¹Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2016-July-14, release 3.25.0, www.ebi.ac.uk/imgt/hla.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information
SPECIFICITY TABLE

DQB1*06 SSP subtyping

Specificities and sizes of the PCR products of the 63+1 primer mixes used for DQB1*06 SSP subtyping

| Primer Mix | Size of spec. PCR product ¹ | Size of control band ² | Amplified DQB1*06 alleles ³ | Amplified non-DQB1*06 alleles |
|------------|--|-----------------------------------|--|-------------------------------|
| 1 | 220 bp | 515 bp | *06:01:01-06:01:15, 06:35, 06:43, 06:45, 06:53:01-06:57, 06:82, 06:98-06:105, 06:108, 06:120, 06:132, 06:140, 06:153, 06:157, 06:177, 06:181, 06:194, 06:205, 06:209, 06:214, 06:229, 06:233, 06:239 | |
| 2 | 210 bp | 430 bp | *06:01:01-06:02:28, 06:05:02 [?] -06:06 [?] , 06:10-06:11:03, 06:13:01-06:13:02, 06:16, 06:18:01-06:20, 06:24, 06:29, 06:33, 06:35, 06:37, 06:43, 06:45, 06:47-06:51:02, 06:53:01-06:57, 06:68, 06:70-06:84, 06:95-06:109, 06:111-06:117, 06:120, 06:122-06:127, 06:130-06:132, 06:136-06:140, 06:147, 06:150-06:153, 06:157, 06:159, 06:161, 06:163, 06:166-06:167, 06:173-06:179N, 06:181-06:183, 06:188, 06:192, 06:197-06:198, 06:200-06:201, 06:205, 06:208-06:209, 06:211, 06:213-06:216N, 06:219, 06:224-06:229, 06:232-06:233, 06:235-06:237, 06:240, 06:242 | |
| 3 | 185 bp | 430 bp | *06:02:01:01-06:02:28, 06:14:01-06:16, 06:19:01-06:20, 06:23-06:24, 06:33, 06:37, 06:46-06:50, 06:51:02, 06:68, 06:70-06:84, 06:95, 06:97, 06:107, 06:109, 06:111-06:117, 06:122, 06:124-06:127, 06:136-06:138, 06:146:01-06:147, 06:150-06:152, 06:156, 06:159, 06:161-06:163, 06:166, 06:173-06:175, 06:178-06:179N, 06:182-06:183, 06:188, 06:192, 06:197-06:198, 06:200-06:201, 06:208, 06:211, 06:213, 06:215-06:216N, 06:219, 06:224-06:228, 06:232, 06:235-06:237, 06:240, 06:242 | *04:10 |
| 4 | 130 bp | 430 bp | *06:02:07, 06:03:01:01-06:03:06, 06:03:08-06:03:27, 06:05:02, 06:07:01-06:07:02, 06:11:02-06:11:03, 06:14:01, 06:14:03, 06:25-06:26N, 06:28, 06:30-06:32:02, 06:40-06:41, 06:44, 06:59-06:67, 06:87, 06:90-06:92, 06:110, 06:118:01-06:118:02, 06:128, 06:133-06:134, 06:141, 06:143-06:145, 06:148, 06:154, 06:165, 06:168, 06:172, 06:184-06:185, 06:187, 06:190:01-06:191, 06:195-06:196, 06:199, 06:203, 06:210, 06:218, 06:221-06:223, 06:230, 06:234, 06:238 | |
| 5 | 160 bp | 430 bp | *06:03:01:01-06:03:03, 06:03:05-06:03:07, 06:03:11-06:03:25, 06:03:27, 06:04:02, 06:07:01, 06:08:01, 06:09:02, | *03:23:01, 05:03:12, 05:43:02 |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | | |
|-------------------------|------------------|---------------|--|
| | | | 06:11:01-06:11:03, 06:26N, 06:28, 06:30-06:32:01, 06:40-06:41, 06:44, 06:59-06:62, 06:64-06:65, 06:67, 06:90-06:92, 06:94, 06:110, 06:118:01, 06:118:03, 06:128, 06:133-06:134, 06:141, 06:143-06:145, 06:148, 06:154, 06:165, 06:169-06:170, 06:184-06:185, 06:187, 06:191, 06:195-06:196, 06:199, 06:203, 06:210, 06:218, 06:221-06:223, 06:230, 06:234, 06:238 |
| 6 | 170 bp | 515 bp | *06:02:07, 06:03:01:01-06:03:06, 06:03:08-06:03:18, 06:03:20-06:03:27, 06:08:01-06:08:03, 06:11:02-06:12, 06:14:01, 06:14:03, 06:21, 06:26N, 06:28, 06:31, 06:40-06:41, 06:44, 06:59, 06:61, 06:63-06:65, 06:67, 06:87, 06:90-06:92, 06:110, 06:128, 06:133-06:134, 06:141, 06:143-06:145, 06:148, 06:154, 06:165, 06:170, 06:184-06:185, 06:187, 06:190:01-06:191, 06:195-06:196, 06:199, 06:203, 06:210, 06:218, 06:221-06:223, 06:230, 06:234, 06:238 |
| 7 | 210 bp | 515 bp | *06:04:01-06:04:11, 06:07:01-06:07:02, 06:17, 06:21, 06:25, 06:34, 06:36, 06:38-06:39, 06:52, 06:58, 06:69:01, 06:85-06:86, 06:89, 06:92-06:93, 06:135, 06:149, 06:155, 06:158N, 06:160, 06:164, 06:172, 06:180, 06:193N, 06:202, 06:204, 06:217, 06:231 |
| 8^{4,7} | 90 bp 175 bp | 515 bp | *06:65 *06:04:01-06:07:02, 06:09:01:01-06:09:03, 06:09:05-06:09:07, 06:18:01-06:18:02, 06:22:02, 06:25, 06:27:01-06:27:02, 06:32:01-06:32:02, 06:34, 06:36, 06:38-06:39, 06:52, 06:58, 06:66, 06:69:01, 06:85-06:86, 06:88-06:89, 06:93-06:94, 06:118:01-06:118:02, 06:121, 06:129, 06:135, 06:142, 06:149, 06:155, 06:158N, 06:160, 06:164, 06:168, 06:171-06:172, 06:180, 06:186, 06:189, 06:193N, 06:202, 06:204, 06:207, 06:217, 06:241 |
| 9^{5,7} | 130 bp | 430 bp | *06:04:01-06:05:01, 06:06, 06:08:01-06:09:03, 06:09:05-06:09:07, 06:12, 06:17-06:18:02, 06:21, 06:22:02, 06:27:01-06:27:02, 06:34, 06:36, 06:38-06:39, 06:42, 06:52, 06:58, 06:69:01, 06:85-06:86, 06:88-06:89, 06:93, 06:121, 06:129, 06:135, 06:142, 06:149, 06:155, 06:158N, 06:160, 06:164, 06:171, 06:180, 06:186, 06:189, 06:193N, 06:202, 06:204, 06:207, 06:217, 06:231, 06:241 |
| 10^{7,8} | 260 bp | 515 bp | *06:05:01, 06:05:02 [?] -06:06 [?] , 06:20, 06:31, 06:45, 06:85, 06:156 |
| | | | *05:73, 05:80, 05:98, 05:116 |
| 11 | 155 bp 210 bp | 430 bp | *06:164 *06:05:01, 06:05:02 [?] -06:06 [?] , 06:09:01:01-06:09:07, 06:12, 06:15:01-06:15:02, 06:22:01-06:22:03, 06:42, 06:46, 06:66, 06:88, 06:94, 06:118:01-06:119, 06:121, 06:142, 06:189, 06:207, 06:212 ^w |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | | | |
|---------------------------|--------------------------------------|---------------|---|---|
| 12⁴ | 100 bp 135 bp 180 bp 215 bp | 430 bp | *06:64 *06:94, 06:170 *06:06, 06:149 *06:06 [?] , 06:129, 06:241 | |
| 13 | 185 bp 225 bp | 430 bp | *06:10, 06:130 *06:05:02 [?] , 06:15:01-06:15:02, 06:22:01-06:22:03, 06:37, 06:48, 06:51:01-06:51:02, 06:69:01-06:69:02, 06:139 | *03:30, 03:72, 03:100, 03:215 |
| 14 | 130 bp 215 bp | 430 bp | *06:09:04, 06:13:01-06:13:02, 06:22:01, 06:22:03, 06:55, 06:69:02, 06:119, 06:206:01-06:206:02 *06:146:01-06:146:02 | *03:259 |
| 15^{4,7} | 105 bp 185 bp | 430 bp | *06:14:01-06:14:03, 06:69:01-06:69:02, 06:156, 06:162, 06:206:01-06:206:02 *06:29, 06:123, 06:139 | *04:28, 05:38, 05:62, 05:119 *03:132, 03:215 |
| 16⁵ | 195 bp 215 bp | 515 bp | *06:16 *06:51:01-06:51:02 | *04:01:01:01 ^w , 04:01:02 ^w -04:01:04 ^w , 04:02:01:01 ^w -04:02:10 ^w , 04:03:01 ^w -04:03:02 ^w , 04:06 ^w -04:36 ^{Nw} |
| 17⁴ | 110 bp | 430 bp | *06:23, 06:82, 06:99:01, 06:156, 06:162, 06:169, 06:233 | *05:03:12, 05:43:02 |
| 18 | 145 bp 185 bp 225 bp | 430 bp | *06:193N, 03:66N *06:03:25, 06:17, 06:24, 06:30, 06:42, 06:149 *06:171 | *03:02:23, 03:228 |
| 19⁶ | 140 bp 225 bp | 430 bp | *06:10, 06:25, 06:36, 06:130, 06:193N, 06:216N *06:171 | *03:66N |
| 20^{4,6,7} | 110 bp 210 bp 260 bp | 515 bp | *06:37, 06:125 *06:26N, 06:81 *06:83 | |
| 21 | 155 bp | 430 bp | *06:02:01:01-06:02:04, 06:02:06-06:02:28, 06:05:02 [?] , 06:10, 06:13:01-06:16, 06:20, 06:22:01-06:24, 06:29, 06:33, 06:37, 06:46-06:51:02, 06:68-06:84, 06:96-06:97, 06:106-06:107, 06:109, 06:111-06:117, 06:119, 06:122, 06:124-06:127, 06:130, 06:136-06:138, 06:146:01-06:147, 06:150-06:151, 06:156, 06:159, 06:161-06:162, 06:166, 06:173-06:175, 06:178-06:179N, 06:182-06:183, 06:188, 06:192, 06:198, 06:200-06:201, 06:206:01-06:206:02, 06:208, 06:211, 06:213, 06:215-06:216N, 06:219, 06:224-06:228, 06:232, 06:235-06:237, 06:240, 06:242 | *04:02:07, 04:09, 05:62 |
| 22 | 130 bp 195 bp | 515 bp | *06:07:01-06:07:02, 06:15:01-06:15:02, 06:46, 06:66, 06:92, 06:118:01-06:118:02, 06:172 *06:38, 06:158N | |
| 23 | 160 bp | 515 bp | *06:03:01:01-06:03:06, 06:03:08-06:03:18, 06:03:20-06:03:27, 06:08:01-06:08:03, 06:14:01, 06:14:03, 06:21, 06:28, 06:31, 06:40-06:41, 06:44, 06:59, 06:61, 06:63-06:65, 06:67, 06:87, 06:90-06:92, 06:110, 06:128, 06:133-06:134, 06:141, 06:143-06:145, | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | 06:148, 06:154, 06:170, 06:184-06:185, 06:187, 06:190:01-06:191, 06:195-06:196, 06:199, 06:203, 06:210, 06:218, 06:221- 06:223, 06:230, 06:234, 06:238 | |
|-----------------------|--------------------------------|---|---|
| 24⁵ | 155 bp 430 bp | *06:02:05, 06:19:01-06:19:02, 06:139 | *03:30, 03:72, 03:100, 03:132, 03:215, 04:01:01:01- 04:02:04, 04:02:06, 04:02:08-04:03:03, 04:06-04:08, 04:10, 04:12-04:14, 04:16- 04:36N, 04:38-04:42, 05:38, 05:119 |
| 25 | 210 bp 430 bp | *06:03:01:01-06:03:06, 06:03:08-06:03:27, 06:08:01-06:08:03, 06:14:01, 06:14:03, 06:27:01-06:28, 06:30-06:32:02, 06:40-06:41, 06:44, 06:59-06:65, 06:67, 06:87, 06:90-06:91, 06:98, 06:110, 06:128, 06:133-06:134, 06:141, 06:143-06:145, 06:148, 06:154, 06:168, 06:170, 06:184-06:185, 06:187, 06:190:01- 06:191, 06:195-06:196, 06:210, 06:218, 06:221-06:223, 06:230, 06:234, 06:238 | |
| 26 | 260 bp 165 bp 430 bp | *06:05:02 [?] , 06:06 [?] , 06:49 | |
| 27 | 190 bp 155 bp 430 bp | *06:35, 06:53:01-06:53:02, 06:145, 06:208 | |
| 28 | 180 bp 300 bp 515 bp | *06:28, 06:56, 06:79:01-06:79:02 | |
| 29⁴ | 90 bp 430 bp | *05:73, 05:98, 05:116 | |
| 30⁴ | 115 bp 430 bp | *06:114 *06:40, 06:81, 06:132 *06:57 *06:33 *06:102N *06:50 *06:34 | |
| | | *06:04:01-06:05:01, 06:05:02 [?] -06:06 [?] , 06:07:01-06:07:02, 06:09:01:01-06:09:07, 06:15:01-06:15:02, 06:22:01-06:22:03, 06:25, 06:34, 06:36, 06:38-06:39, 06:52, 06:58, 06:66, 06:69:01-06:69:02, 06:85-06:86, 06:88- 06:89, 06:93-06:94, 06:118:01-06:118:03, 06:121, 06:135, 06:155, 06:158N, 06:160, 06:164, 06:172, 06:180, 06:189, 06:193N, 06:202, 06:204, 06:207, 06:212, 06:217, 06:231 | |
| | | *06:02:01:01-06:02:01:03, 06:02:02-06:02:26, 06:03:01:01, 06:03:02-06:03:26, 06:05:02 [?] - 06:07:01 [?] , 06:08:01 [?] -06:08:03 [?] , 06:09:02 [?] - 06:09:03 [?] , 06:09:05 [?] -06:09:06 [?] , 06:10, 06:11:01 [?] -06:11:03 [?] , 06:13:01 [?] -06:13:02 [?] , 06:14:01, 06:14:02 [?] -06:18:01 [?] , 06:19:01 [?] - 06:20 [?] , 06:22:01 [?] -06:32:02 [?] , 06:33, 06:35 [?] , 06:37, 06:39, 06:40 [?] , 06:44, 06:45 [?] -06:46 [?] , 06:47-06:48, 06:49 [?] -06:50 [?] , 06:51:02 [?] , 06:53:01 [?] -06:54N [?] , 06:56 [?] -06:57 [?] , 06:59 [?] - 06:61 [?] , 06:62, 06:63 [?] -06:69:01 [?] , 06:70 [?] -06:76 [?] , 06:77N-06:79:01, 06:79:02 [?] -06:83 [?] , 06:85 [?] , 06:87-06:88, 06:89 [?] , 06:90, 06:91 [?] -06:97 [?] , 06:106-06:117, 06:118:02 [?] -06:124 [?] , 06:125, | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | | | |
|----------------------------|--------------------------------------|--------|---|--|
| | | | 06:126 [?] , 06:127, 06:128 [?] , 06:130, 06:131 [?] - 06:147 [?] , 06:148, 06:149 [?] -06:184 [?] , 06:185, 06:187-06:188, 06:190:01 [?] -06:199 [?] , 06:200, 06:201 [?] -06:204 [?] , 06:206:01 [?] , 06:207 [?] -06:208 [?] , 06:210 [?] -06:215 [?] , 06:216N, 06:218-06:219, 06:221-06:222, 06:224-06:228, 06:229 [?] - 06:236 [?] | |
| 31 ^{4,6,7} | 100 bp 220 bp | 430 bp | *06:44, 06:47 *06:43 | |
| 32 ^{4,6} | 115 bp | 430 bp | *06:04:01-06:05:01, 06:05:02 [?] -06:07:01 [?] , 06:07:02, 06:08:02 [?] -06:08:03 [?] , 06:09:01:01- 06:09:07, 06:11:02 [?] -06:11:03 [?] , 06:12, 06:13:02 [?] , 06:14:03 [?] -06:18:01 [?] , 06:18:02, 06:19:01 [?] , 06:21, 06:22:01 [?] -06:32:02 [?] , 06:34, 06:35 [?] , 06:36, 06:38, 06:40 [?] , 06:41-06:42, 06:45 [?] , 06:50 [?] , 06:51:02 [?] , 06:52, 06:53:01 [?] - 06:54N [?] , 06:56 [?] -06:57 [?] , 06:58, 06:59 [?] -06:61 [?] , 06:63 [?] -06:76 [?] , 06:79:02 [?] -06:83 [?] , 06:84, 06:85 [?] , 06:86, 06:89 [?] , 06:91 [?] -06:97 [?] , 06:118:01, 06:118:02 [?] -06:124 [?] , 06:126 [?] , 06:128 [?] , 06:129, 06:131 [?] -06:147 [?] , 06:149 [?] -06:184 [?] , 06:186, 06:189, 06:190:01 [?] -06:199 [?] , 06:201 [?] -06:204 [?] , 06:206:01 [?] , 06:206:02, 06:207 [?] -06:208 [?] , 06:210 [?] -06:215 [?] , 06:217, 06:229 [?] -06:236 [?] , 06:238 [?] , 06:241 [?] | |
| 33 | 190 bp | 430 bp | *06:28, 06:56, 06:79:01-06:79:02 | *05:04 [?] , 05:73, 05:98, 05:116 |
| 34 ⁷ | 145 bp | 430 bp | *06:01:01, 06:01:03-06:01:06, 06:01:08- 06:01:11, 06:01:13-06:01:15, 06:43, 06:53:01, 06:54N-06:57, 06:98, 06:99:02-06:100, 06:102N-06:105, 06:108, 06:120, 06:132, 06:140, 06:153, 06:157, 06:167-06:168, 06:181, 06:194, 06:205, 06:209, 06:214, 06:229, 06:239 | *03:01:38, 03:04:02, 03:10:01, 03:12, 03:14:02, 03:70, 03:179, 03:195 |
| 35 | 135 bp 185 bp 260 bp | 430 bp | *06:66, 06:172 *06:54N, 06:135 *06:05:02 [?] , 06:06 [?] , 06:58 | |
| 36 | 165 bp | 430 bp | *06:71, 06:77N-06:78, 06:95, 06:138 | |
| 37 ⁴ | 120 bp 175 bp | 430 bp | *06:80 *06:29, 06:76-06:77N, 06:96, 06:139 | *03:30, 03:72, 03:100, 03:132, 03:215, 04:09 |
| 38 ⁷ | 245 bp 170 bp 245 bp 285 bp | 430 bp | *06:05:02 [?] , 06:137 *06:78, 06:123 *06:137 *06:72-06:73 | |
| 39 ⁴ | 120 bp 155 bp 270 bp | 430 bp | *06:80 *06:138 *06:73-06:74 | |
| 40 ^{4,6} | 105 bp 190 bp | 430 bp | *06:70 *06:75N, 06:106, 06:136 | |
| 41 ⁷ | 215 bp | 430 bp | *06:122, 06:130, 06:148 | *03:194 |
| 42 | 130 bp 165 bp | 430 bp | *06:93-06:94, 06:170 *06:121, 06:142, 06:168 | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | | | |
|---------------------------|--|---------------|---|----------------------------------|
| | 190 bp | | *06:60-06:61 | |
| 43 | 150 bp 180 bp | 430 bp | | *05:14 *05:03:02? |
| 44 | 130 bp 180 bp 220 bp | 430 bp | *06:07:01?, 06:20?, 06:68, 06:131 *06:113 *06:67, 06:174, 06:191 *06:143 | |
| 45 | 150 bp 195 bp 235 bp | 430 bp | *06:97 *06:136 *06:124 | |
| 46 | 170 bp 240 bp | 430 bp | *06:163 *06:86, 06:104, 06:107 | *03:97 |
| 47⁴ | 95 bp | 430 bp | *06:29, 06:59, 06:63, 06:87, 06:96, 06:150 | *03:08, 03:137, 03:194 |
| | 180 bp 220 bp | | *06:90 *06:143 | |
| 48⁴ | 110 bp 205 bp | 430 bp | *06:59, 06:91, 06:145, 06:150, 06:208 *06:128 | *03:194 |
| 49 | 190 bp 230 bp | 430 bp | *06:100, 06:132, 06:140 *06:126 | |
| 50⁴ | 75 bp 150 bp 190 bp 275 bp | 515 bp | *06:134 *06:101 *06:140, 06:144N *06:120 | |
| 51^{4,6,7} | 120 bp | 430 bp | *06:111, 06:189 | *04:17, 05:31, 05:46, 05:108 |
| | 240 bp | | *06:205 | |
| 52⁴ | 95 bp 130 bp 165 bp 195 bp | 430 bp | *06:155 *06:133 *06:188, 06:200 *06:105, 06:185, 06:225 | *02:99, 03:249, 05:47 |
| 53 | 150 bp 195 bp 220 bp | 430 bp | *06:109-06:110, 06:200 *06:105, 06:185, 06:225 *06:147 | *03:115 *02:99, 03:249, 05:47 |
| 54 | 190 bp | 430 bp | *06:07:01?, 06:20?, 06:112N, 06:131 | *05:03:02? |
| 55 | 145 bp | 430 bp | *06:115 | |
| 56 | 130 bp 200 bp | 515 bp | *06:99:01-06:99:02, 06:233 *06:116, 06:187 | |
| 57^{4,7} | 125 bp | 515 bp | *06:07:01-06:07:02, 06:15:01-06:15:02, 06:46, 06:92, 06:118:01-06:118:03 *06:89, 06:135 *06:158N | *05:80 |
| 58 | 185 bp 200 bp 215 bp 275 bp 300 bp | 430 bp | *06:153, 06:233 *06:127, 06:205 *06:52 | *03:52, 03:179 |
| 59 | 170 bp 230 bp | 430 bp | *06:163 *06:117, 06:147 | |
| 60 | 140 bp 180 bp | 430 bp | *06:141, 06:179N *06:191 | |
| 61⁴ | 65 bp | 430 bp | *06:53:01-06:53:02 | |
| 62⁷ | 220 bp | 430 bp | *06:23, 06:156, 06:162, 06:169 | *05:73, 05:98, 05:116 |
| 63 | 155 bp | 515 bp | *06:164 | |
| 64⁹ | - | - | Negative Control | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DQB1*06 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several DQB1 alleles 1st and/or 3rd exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁵Primer mixes 9, 16 and 24 may give rise to a lower yield of HLA-specific PCR product than the other DQB1*06 primer mixes.

⁶Primer mixes 19, 20, 31, 32, 40 and 51 have a tendency to giving rise to primer oligomer formation.

⁷Primer mixes 8, 9, 10, 15, 20, 31, 34, 38, 41, 51, 57 and 62 may have tendencies of unspecific amplifications.

⁸The nucleotide sequence of codon 14 of the DQB1*06:05:02 allele is not yet known. If codon 14 is CTg, then the DQB1*06:05:02 allele will retain its name and will be amplified by the primer pair in well No. 10. If the sequence of codon 14 is ATg, then DQB1*06:05:02 will be renamed to DQB1*06:09:02 (Steven Marsh personal communication), and will not be amplified by the primer pair in well No. 10.

⁹Primer mix 64 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

“?”, nucleotide sequence information is not available for the primer matching sequence.

‘w’, might be weakly amplified.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

PRIMER SPECIFICATION

| Well No. | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 | 11 | 12 |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|
| Length of spec. | 220 | 210 | 185 | 130 | 160 | 170 | 210 | 90 | 130 | 260 | 155 | 100 |
| PCR product | | | | | | | | 175 | | | 210 | 135 |
| | | | | | | | | | | | | 180 |
| | | | | | | | | | | | | 215 |
| Length of int. pos. control ¹ | 515 | 430 | 430 | 430 | 430 | 515 | 515 | 515 | 430 | 515 | 430 | 430 |
| 5'-primer(s) ² | 26(173) 5'-TTA 3' | 30(184) 5'-gAT 3' | 9(122) 5'-gTT 3' | 27(177) 5'-gTA 3' | 9(122) 5'-gTA 3' | 27(177) 5'-gTA 3' | 30(184) 5'-gAC 3' | 27(177) 5'-gTA 3' | 27(177) 5'-gTA 3' | 13(134) 5'-ggA 3' | 30(184) 5'-gAT 3' | 27(177) 5'-gTA 3' |
| | | 30(184) 5'-gAT 3' | | | | | | | | 14(136) 5'-gCC 3' | 48(239) 5'-CCA 3' | |
| 3'-primer(s) ³ | 86(353) 5'-ACg 3' | 86(353) 5'-ACg 3' | 57(266) 5'-CAT 3' | 57(266) 5'-CAT 3' | 48(240) 5'-gCg 3' | 70(304) 5'-CCC 3' | 86(353) 5'-ACC 3' | 43(224) 5'-Cgg 3' | 57(266) 5'-CAA 3' | 86(353) 5'-ACC 3' | 86(353) 5'-ACC 3' | 48(238) 5'-gCA 3' |
| | 86(354) 5'-AAT 3' | 86(354) 5'-AAA 3' | 58(270) 5'-TCC 3' | 57(267) 5'-gCg 3' | | 70(304) 5'-CCC 3' | 86(354) 5'-TAT 3' | 70(304) 5'-CCT 3' | | 87(356) 5'-ggA 3' | | 59(271) 5'-CTT 3' |
| | | 86(354) 5'-AAT 3' | | 58(270) 5'-TCA 3' | | | | 74(317) 5'-CCg 3' | | | | 74(317) 5'-CCg 3' |
| | | | | | | | | | | | | 86(353) 5'-ACg 3' |
| Well No. | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 | 11 | 12 |

| Well No. | 13 | 14 | 15 | 16 | 17 | 18 | 19 | 20 | 21 | 22 | 23 | 24 |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|----------------------|----------------------|-----------------------|----------------------|----------------------|
| Length of spec. | 185 | 130 | 105 | 195 | 110 | 145 | 140 | 110 | 155 | 130 | 160 | 155 |
| PCR product | 225 | 215 | 185 | 215 | | 185 | 225 | 210 | | 195 | | |
| | | | | | | 225 | | 260 | | | | |
| Length of int. pos. control ¹ | 430 | 430 | 430 | 515 | 430 | 430 | 430 | 515 | 430 | 515 | 515 | 430 |
| 5'-primer(s) ² | 9(122) 5'-gTT 3' | 27(177) 5'-gTg 3' | 9(122) 5'-gTT 3' | 9(122) 5'-gTT 3' | 26(173) 5'-ggg 3' | 26(173) 5'-TCT 3' | 26(173) 5'-TCT 3' | 14(136) 5'-gCg 3' | 9(122) 5'-gTT 3' | 34(196) 5'-ACT 3' | 30(184) 5'-gAC 3' | 9(122) 5'-gTT 3' |
| | | | | | 26(173) 5'-TTA 3' | | 154(558) 5'-ACT 3' | 29(181) 5'-CCT 3' | | 57(266) 5'-TgA 3' | | |
| | | | | | | | | 31(187) 5'-ACT 3' | | 133(494) 5'-TCA 3' | | |
| | | | | | | | | 62(282) 5'-AAg 3' | | | | |
| | | | | | | | | 63(285) 5'-Agg 3' | | | | |
| 3'-primer(s) ³ | 57(265) 5'-gCT 3' | 57(266) 5'-CAA 3' | 30(184) 5'-gTg 3' | 60(274) 5'-gTT 3' | 48(240) 5'-gCg 3' | 61(278) 5'-TCT 3' | 57(265) 5'-gCT 3' | 86(353) 5'-ACg 3' | 47(236) 5'-ggT 3' | 86(353) 5'-ACC 3' | 70(304) 5'-CCC 3' | 47(237) 5'-CgA 3' |
| | 70(304) 5'-CCT 3' | 86(353) 5'-ACg 3' | 57(266) 5'-Cgg 3' | 66(294) 5'-ATg 3' | | 71(307) 5'-ggC 3' | 61(278) 5'-TCT 3' | | 48(240) 5'-gCg 3' | 185(650) 5'-Cgg 3' | 70(304) 5'-CCC 3' | 48(240) 5'-gCg 3' |
| | | 86(353) 5'-ACg 3' | | | | 77(327) 5'-ACT 3' | 87(355) 5'-gTg 3' | | 49(241) 5'-TgC 3' | | | |
| | | | | | | 87(355) 5'-gTg 3' | 186(653) 5'-CCg 3' | | | | | |
| | | | | | | | 186(654) 5'-TAA 3' | | | | | |
| Well No. | 13 | 14 | 15 | 16 | 17 | 18 | 19 | 20 | 21 | 22 | 23 | 24 |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| Well No. | 25 | 26 | 27 | 28 | 29 | 30 | 31 | 32 | 33 | 34 | 35 | 36 |
|--|---------------------------------|---------------------------------|---------------------------------|----------------------------------|---------------------------------|----------------------------------|----------------------------------|----------------------------------|---------------------------------|---------------------------------|---------------------------------|---------------------------------|
| Length of spec. | 210 | 165 | 155 | 130 | 90 | 115 | 100 | 115 | 190 | 145 | 135 | 165 |
| PCR product | 260 | 190 | 195 | 180 | | | 220 | | | | 185 | |
| | | | 220 | 300 | | | | | | | 260 | |
| | | | 265 | | | | | | | | | |
| Length of int. pos. control ¹ | 430 | 430 | 430 | 515 | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 |
| 5'-primer(s) ² | 13(134) 5'-ggT ^{3'} | 38(209) 5'-CgT ^{3'} | 11(129) 5'-TTA ^{3'} | 30(184) 5'-gAT ^{3'} | 67(297) 5'-gTA ^{3'} | 130(485) 5'-CCg ^{3'} | 13(134) 5'-ggC ^{3'} | 130(485) 5'-CCA ^{3'} | 38(209) 5'-CgT ^{3'} | 13(134) 5'-ggC ^{3'} | 13(134) 5'-ggC ^{3'} | 9(122) 5'-gTT ^{3'} |
| | 30(184) 5'-gAC ^{3'} | 45(230) 5'-ggA ^{3'} | 26(174) 5'-TAC ^{3'} | 102(400) 5'-TCT ^{3'} | 70(304) 5'-AgA ^{3'} | | 154(558) 5'-ACT ^{3'} | | | | 39(211) 5'-CgT ^{3'} | |
| | 30(185) 5'-ATg ^{3'} | | 31(187) 5'-ACg ^{3'} | | | | | | | | 55(260) 5'-gCC ^{3'} | |
| | | | 31(187) 5'-ACT ^{3'} | | | | | | | | | |
| | | | 38(209) 5'-CAT ^{3'} | | | | | | | | | |
| | | | 48(239) 5'-CCA ^{3'} | | | | | | | | | |
| 3'-primer(s) ³ | 86(353) 5'-ACg ^{3'} | 87(356) 5'-ggA ^{3'} | 86(353) 5'-ACg ^{3'} | 76(322) 5'-gTg ^{3'} | 86(353) 5'-ACC ^{3'} | 154(558) 5'-AAA ^{3'} | 73(314) 5'-CCA ^{3'} | 154(558) 5'-AAA ^{3'} | 87(356) 5'-ggA ^{3'} | 47(237) 5'-CgA ^{3'} | 61(279) 5'-TTT ^{3'} | 48(238) 5'-gCA ^{3'} |
| | 86(354) 5'-AAg ^{3'} | | | 131(488) 5'-ACT ^{3'} | 86(354) 5'-TAT ^{3'} | | 174(618) 5'-ACT ^{3'} | | | | 86(353) 5'-ACC ^{3'} | 50(244) 5'-CAT ^{3'} |
| | | | | 189(661) 5'-CCA ^{3'} | | | | | | | | 51(247) 5'-CgA ^{3'} |
| | | | | | | | | | | | | 53(253) 5'-CTT ^{3'} |
| | | | | | | | | | | | | 53(253) 5'-CTA ^{3'} |
| Well No. | 25 | 26 | 27 | 28 | 29 | 30 | 31 | 32 | 33 | 34 | 35 | 36 |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| Well No. | 37 | 38 | 39 | 40 | 41 | 42 | 43 | 44 | 45 | 46 | 47 | 48 |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|----------------------|----------------------|-----------------------|----------------------|----------------------|
| Length of spec. | 120 | 170 | 120 | 105 | 215 | 130 | 150 | 130 | 150 | 170 | 95 | 110 |
| PCR product | 175 | 245 | 155 | 190 | | 165 | 180 | 180 | 195 | 240 | 180 | 205 |
| | 245 | 285 | 270 | | | 190 | | 220 | 235 | | 220 | |
| Length of int. pos. control ¹ | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 | 430 |
| 5'-primer(s) ² | 9(122) 5'-gTT 3' | 9(122) 5'-gTT 3' | 9(122) 5'-gTT 3' | 9(122) 5'-gTT 3' | 10(125) 5'-CCT 3' | 27(177) 5'-gTA 3' | 8(119) 5'-CgC 3' | 10(126) 5'-CAT 3' | 9(122) 5'-gTT 3' | 9(122) 5'-gTT 3' | 10(126) 5'-CAT 3' | 14(138) 5'-ATC 3' |
| | | | | | 13(134) 5'-ggC 3' | | 14(137) 5'-CAC 3' | 24(167) 5'-CgA 3' | | 102(400) 5'-TCT 3' | 24(166) 5'-gCA 3' | 45(230) 5'-ggC 3' |
| | | | | | | | 133(494) 5'-TCA 3' | 25(169) 5'-TgT 3' | | | 48(238) 5'-ACA 3' | 45(230) 5'-ggA 3' |
| | | | | | | | | 41(217) 5'-TCC 3' | | | 55(260) 5'-gCC 3' | 48(238) 5'-ACA 3' |
| 3'-primer(s) ³ | 36(203) 5'-ACC 3' | 51(247) 5'-CgA 3' | 36(203) 5'-ACC 3' | 31(188) 5'-Agg 3' | 70(304) 5'-CCC 3' | 55(260) 5'-gCA 3' | 57(266) 5'-CAT 3' | 70(304) 5'-CCC 3' | 46(232) 5'-CAA 3' | 52(251) 5'-gCA 3' | 70(304) 5'-CCC 3' | 70(304) 5'-CCC 3' |
| | 53(253) 5'-CTA 3' | 55(260) 5'-gCA 3' | 48(238) 5'-gCA 3' | 59(271) 5'-CTT 3' | | 59(271) 5'-CTT 3' | 170(604) 5'-gAC 3' | | 60(274) 5'-gTg 3' | 74(317) 5'-ACC 3' | | |
| | 55(260) 5'-gCg 3' | 78(328) 5'-CAT 3' | 82(341) 5'-AgC 3' | 60(274) 5'-gTg 3' | | 66(294) 5'-ATg 3' | | | 74(316) 5'-CTC 3' | 172(611) 5'-AgA 3' | | |
| | 56(262) 5'-AgA 3' | 88(359) 5'-CgA 3' | 88(359) 5'-CgA 3' | 60(276) 5'-CAC 3' | | 72(311) 5'-CCg 3' | | | | | | |
| | 78(328) 5'-CAT 3' | 91(369) 5'-TTT 3' | | | | 76(322) 5'-gTg 3' | | | | | | |
| | | | | | | 77(326) 5'-CCA 3' | | | | | | |
| Well No. | 37 | 38 | 39 | 40 | 41 | 42 | 43 | 44 | 45 | 46 | 47 | 48 |

| Well No. | 49 | 50 | 51 | 52 | 53 | 54 | 55 | 56 | 57 | 58 | 59 | 60 |
|--|----------------------|----------------------|-----------------------|-----------------------|-----------------------|----------------------|----------------------|-----------------------|----------------------|-----------------------|-----------------------|----------------------|
| Length of spec. | 190 | 75 | 120 | 95 | 150 | 190 | 145 | 130 | 125 | 215 | 170 | 140 |
| PCR product | 230 | 150 | 240 | 130 | 195 | | | 200 | 185 | 275 | 230 | 180 |
| | | 190 | | 165 | 220 | | | | 200 | 300 | | |
| | | 275 | | 195 | | | | | | | | |
| Length of int. pos. control ¹ | 430 | 515 | 430 | 430 | 430 | 430 | 430 | 515 | 515 | 430 | 430 | 430 |
| 5'-primer(s) ² | 23(163) 5'-AgT 3' | 9(121) 5'-TgA 3' | 104(406) 5'-CAA 3' | 27(177) 5'-gTA 3' | 9(122) 5'-gTT 3' | 8(119) 5'-CgC 3' | 57(266) 5'-TgA 3' | 9(121) 5'-TgC 3' | 34(196) 5'-ACT 3' | 9(121) 5'-TgC 3' | 9(122) 5'-gTT 3' | 24(167) 5'-CgA 3' |
| | 37(206) 5'-gTC 3' | 37(206) 5'-gTC 3' | 141(517) 5'-CCA 3' | 135(500) 5'-TgA 3' | 135(500) 5'-TgA 3' | 10(124) 5'-TCT 3' | | 100(394) 5'-TgT 3' | 38(209) 5'-CgT 3' | 98(389) 5'-CAT 3' | 102(400) 5'-TCT 3' | 34(196) 5'-ACT 3' |
| | 38(209) 5'-CAT 3' | 37(207) 5'-TAA 3' | 146(533) 5'-CCT 3' | | | | | 105(411) 5'-AgC 3' | 39(211) 5'-CgT 3' | 104(406) 5'-CAA 3' | | 39(211) 5'-Cgg 3' |
| | 39(211) 5'-TgA 3' | 49(242) 5'-ggA 3' | | | | | | | 57(267) 5'-gAT 3' | 110(426) 5'-AAA 3' | | |
| | | 75(319) 5'-Agg 3' | | | | | | | | | | |
| 3'-primer(s) ³ | 86(353) 5'-ACg 3' | 86(353) 5'-ACg 3' | 170(604) 5'-gAC 3' | 46(233) 5'-ACC 3' | 69(302) 5'-CCC 3' | 57(266) 5'-CAT 3' | 92(372) 5'-CTA 3' | 38(209) 5'-gCg 3' | 86(353) 5'-ACC 3' | 66(294) 5'-ACT 3' | 52(251) 5'-gCA 3' | 70(304) 5'-CCC 3' |
| | | | | 57(265) 5'-ATT 3' | 168(599) 5'-CTT 3' | | | 154(558) 5'-AAA 3' | | 185(650) 5'-Cgg 3' | 69(302) 5'-CCC 3' | |
| | | | | 176(624) 5'-TgA 3' | 175(619) 5'-CAT 3' | | | | | | 167(596) 5'-CAA 3' | |
| | | | | 179(633) 5'-AgT 3' | 176(624) 5'-TgA 3' | | | | | | | |
| | | | | 186(652) 5'-CAT 3' | 186(652) 5'-CAT 3' | | | | | | | |
| Well No. | 49 | 50 | 51 | 52 | 53 | 54 | 55 | 56 | 57 | 58 | 59 | 60 |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| | | | |
|---|----------------------|----------------------|----------------------|
| Well No. | 61 | 52 | 63 |
| Length of spec. PCR product | 65 | 220 | 155 |
| Length of int. pos. control ¹ | 430 | 430 | 515 |
| 5'-primer(s) ² | 45(230) 5'-ggA 3' | 26(173) 5'-ggg 3' | 48(239) 5'-CCA 3' |
| 3'-primer(s) ³ | 53(254) 5'-CCT 3' | 86(353) 5'-ACg 3' | 86(353) 5'-ACC 3' |
| Well No. | 61 | 62 | 63 |

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 430 or 515 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the longer, 515 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| CELL LINE VALIDATION SHEET | | | | | | | | | | | | | | | | | | | | |
|--|-----------------------------|--------|--------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|
| DQB1*06 SSP subtyping kit ² | | | | | | | | | | | | | | | | | | | | |
| | | | | Well | | | | | | | | | | | | | | | | |
| | | | | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | 10 | 11 | 12 | 13 | 14 | 15 | 16 | |
| | | | | Prod. No. | 201556901 | 201556902 | 201556903 | 201556904 | 201556905 | 201556906 | 201556907 | 201556908 | 201556909 | 201556910 | 201674311 | 201556912 | 201556913 | 201674314 | 201556915 | 201674316 |
| | IHWC cell line ¹ | DQB1 | | | | | | | | | | | | | | | | | | |
| 1 | 9001 SA | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 2 | 9280 LK707 | *06:01 | *02:02 | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 3 | 9011 E4181324 | *06:01 | | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 4 | 9275 GU373 | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 5 | 9009 KAS011 | *05:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 6 | 9353 SM | *03:02 | *06:01 | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 7 | 9020 QBL | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 8 | 9025 DEU | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 9 | 9026 YAR | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 10 | 9107 LKT3 | *04:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 11 | 9051 PITOUT | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 12 | 9052 DBB | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 13 | 9004 JESTHOM | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 14 | 9071 OLGA | *04:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 15 | 9075 DKB | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 16 | 9037 SWEIG007 | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 17 | 9282 CTM3953540 | *02:01 | *06:03 | - | - | - | + | + | + | - | - | - | - | - | - | - | - | - | - | |
| 18 | 9257 32367 | *06:02 | *02:02 | - | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 19 | 9038 BM16 | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 20 | 9059 SLE005 | *06:04 | | - | - | - | - | - | - | + | + | + | - | - | - | - | - | - | - | |
| 21 | 9064 AMALA | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 22 | 9056 KOSE | *05:03 | *06:04 | - | - | - | - | - | - | + | + | + | - | - | - | - | - | - | - | |
| 23 | 9124 IHL | *05:03 | *06:01 | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 24 | 9035 JBUSH | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 25 | 9049 IBW9 | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 26 | 9285 WT49 | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 27 | 9191 CH1007 | *04:01 | *05:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 28 | 9320 BEL5GB | *02:02 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 29 | 9050 MOU | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 30 | 9021 RSH | *04:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 31 | 9019 DUCAF | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 32 | 9297 HAG | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 33 | 9098 MT14B | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 34 | 9104 DHIF | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 35 | 9302 SSTO | *03:05 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 36 | 9024 KT17 | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 37 | 9065 HHKB | *06:03 | | - | - | - | + | + | + | - | - | - | - | - | - | - | - | - | - | |
| 38 | 9099 LZL | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 39 | 9315 CML | *02:01 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 40 | 9134 WHONP199 | *02:02 | *03:03 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 41 | 9055 H0301 | *06:09 | | - | - | - | - | - | - | - | + | + | - | + | - | - | - | - | - | |
| 42 | 9066 TAB089 | *06:01 | | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 43 | 9076 T7526 | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 44 | 9057 TEM | *05:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 45 | 9239 SHJO | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 46 | 9013 SCHU | *06:02 | | - | + | + | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 47 | 9045 TUBO | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 48 | 9303 TER-ND | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| CELL LINE VALIDATION SHEET | | | | | | | | | | | | | | | | | | | | |
|--|------|------------|--------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|
| DQB1*06 SSP subtyping kit ² | | | | | | | | | | | | | | | | | | | | |
| | | | | Well | | | | | | | | | | | | | | | | |
| | | | | 17 | 18 | 19 | 20 | 21 | 22 | 23 | 24 | 25 | 26 | 27 | 28 | 29 | 30 | 31 | 32 | |
| | | | | Prod. No. | 201556917 | 201674318 | 201889419 | 201556920 | 201556921 | 201556922 | 201556923 | 201674324 | 201556925 | 201556926 | 201556927 | 201674328 | 201556929 | 201556930 | 201556931 | 201556932 |
| IHCW cell line ¹ | | DQB1 | | | | | | | | | | | | | | | | | | |
| 1 | 9001 | SA | *05:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 2 | 9280 | LK707 | *06:01 | *02:02 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 3 | 9011 | E4181324 | *06:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 4 | 9275 | GU373 | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 5 | 9009 | KAS011 | *05:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 6 | 9353 | SM | *03:02 | *06:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 7 | 9020 | QBL | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 8 | 9025 | DEU | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 9 | 9026 | YAR | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 10 | 9107 | LKT3 | *04:01 | | - | - | - | - | - | - | - | + | - | - | - | - | - | - | - | |
| 11 | 9051 | PITOUT | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 12 | 9052 | DBB | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 13 | 9004 | JESTHOM | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 14 | 9071 | OLGA | *04:02 | | - | - | - | - | - | - | - | + | - | - | - | - | - | - | - | |
| 15 | 9075 | DKB | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 16 | 9037 | SWEIG007 | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 17 | 9282 | CTM3953540 | *02:01 | *06:03 | - | - | - | - | - | - | + | - | + | - | - | - | - | + | - | |
| 18 | 9257 | 32367 | *06:02 | *02:02 | - | - | - | - | + | - | - | - | - | - | - | - | - | + | - | |
| 19 | 9038 | BM16 | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 20 | 9059 | SLE005 | *06:04 | | - | - | - | - | - | - | - | - | - | - | - | - | + | - | + | |
| 21 | 9064 | AMALA | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 22 | 9056 | KOSE | *05:03 | *06:04 | - | - | - | - | - | - | - | - | - | - | - | - | + | - | + | |
| 23 | 9124 | IHL | *05:03 | *06:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 24 | 9035 | JBUSH | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 25 | 9049 | IBW9 | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 26 | 9285 | WT49 | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 27 | 9191 | CH1007 | *04:01 | *05:01 | - | - | - | - | - | - | - | + | - | - | - | - | - | - | - | |
| 28 | 9320 | BEL5GB | *02:02 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 29 | 9050 | MOU | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 30 | 9021 | RSH | *04:02 | | - | - | - | - | - | - | - | + | - | - | - | - | - | - | - | |
| 31 | 9019 | DUCAF | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 32 | 9297 | HAG | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 33 | 9098 | MT14B | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 34 | 9104 | DHIF | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 35 | 9302 | SSTO | *03:05 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 36 | 9024 | KT17 | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 37 | 9065 | HHKB | *06:03 | | - | - | - | - | - | - | + | - | + | - | - | - | - | + | - | |
| 38 | 9099 | LZL | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 39 | 9315 | CML | *02:01 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 40 | 9134 | WHONP199 | *02:02 | *03:03 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 41 | 9055 | H0301 | *06:09 | | - | - | - | - | - | - | - | - | - | - | - | - | + | - | + | |
| 42 | 9066 | TAB089 | *06:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 43 | 9076 | T7526 | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 44 | 9057 | TEM | *05:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 45 | 9239 | SHJO | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 46 | 9013 | SCHU | *06:02 | | - | - | - | - | + | - | - | - | - | - | - | - | - | + | - | |
| 47 | 9045 | TUBO | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 48 | 9303 | TER-ND | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |

101.212-24/04 – including *Taq* pol., IFU-01
101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

| CELL LINE VALIDATION SHEET | | | | | | | | | | | | | | | | | | | | | |
|--|-----------------------------|--|--------|--------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|-----------|
| DQB1*06 SSP subtyping kit ² | | | | | | | | | | | | | | | | | | | | | |
| | | | | | Well | | | | | | | | | | | | | | | | |
| | | | | | 33 | 34 | 35 | 36 | 37 | 38 | 39 | 40 | 41 | 42 | 43 | 44 | 45 | 46 | 47 | 48 | |
| | | | | | Prod. No. | 201556933 | 201556934 | 201674335 | 201556936 | 201556937 | 201556938 | 201556939 | 201674340 | 201556941 | 201556942 | 201674343 | 201674344 | 201674345 | 201556946 | 201556947 | 201556948 |
| | IHWC cell line ¹ | | DQB1 | | | | | | | | | | | | | | | | | | |
| 1 | 9001 SA | | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 2 | 9280 LK707 | | *06:01 | *02:02 | - | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 3 | 9011 E4181324 | | *06:01 | | - | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 4 | 9275 GU373 | | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 5 | 9009 KAS011 | | *05:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 6 | 9353 SM | | *03:02 | *06:01 | - | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 7 | 9020 QBL | | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 8 | 9025 DEU | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 9 | 9026 YAR | | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 10 | 9107 LKT3 | | *04:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 11 | 9051 PITOUT | | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 12 | 9052 DBB | | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 13 | 9004 JESTHOM | | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 14 | 9071 OLGA | | *04:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 15 | 9075 DKB | | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 16 | 9037 SWEIG007 | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 17 | 9282 CTM3953540 | | *02:01 | *06:03 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 18 | 9257 32367 | | *06:02 | *02:02 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 19 | 9038 BM16 | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 20 | 9059 SLE005 | | *06:04 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 21 | 9064 AMALA | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 22 | 9056 KOSE | | *05:03 | *06:04 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 23 | 9124 IHL | | *05:03 | *06:01 | - | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 24 | 9035 JBUSH | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 25 | 9049 IBW9 | | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 26 | 9285 WT49 | | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 27 | 9191 CH1007 | | *04:01 | *05:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 28 | 9320 BEL5GB | | *02:02 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 29 | 9050 MOU | | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 30 | 9021 RSH | | *04:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 31 | 9019 DUCAF | | *02:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 32 | 9297 HAG | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 33 | 9098 MT14B | | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 34 | 9104 DHIF | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 35 | 9302 SSTO | | *03:05 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 36 | 9024 KT17 | | *03:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 37 | 9065 HHKB | | *06:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 38 | 9099 LZL | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 39 | 9315 CML | | *02:01 | *03:01 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 40 | 9134 WHONP199 | | *02:02 | *03:03 | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 41 | 9055 H0301 | | *06:09 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 42 | 9066 TAB089 | | *06:01 | | - | + | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 43 | 9076 T7526 | | *03:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 44 | 9057 TEM | | *05:03 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 45 | 9239 SHJO | | *02:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 46 | 9013 SCHU | | *06:02 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 47 | 9045 TUBO | | *03:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |
| 48 | 9303 TER-ND | | *05:01 | | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | - | |

101.212-24/04 – including *Taq* pol., IFU-01
 101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: 1G3

Lot-specific information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 12 to 20, 22, 26 to 28, 31, 33 and 35 to 63 were available.

The specificities of the primers in primer solutions 10, 12 to 19, 22, 26, 33, 35, 37, 38, 41, 42, 47, 48, 56 to 58, 61 and 62 were tested by separately adding additional 5'-primers, respectively additional 3'-primers.

In primer solutions 28, 31, 36, 39, 40, 45, 46, 52, 53, 55 and 59 it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solutions 20, 27, 43, 44, 49 to 51, 54, 60 and 63 it was only possible to test the 3'-primers, the 5'-primers were not possible to test.

In primer solution 1 to 4, 6 to 8, 12, 14, 16, 18, 19, 23 to 25, 29, 35, 37, 38 and 42 one or more 3'-primers were not possible to test, and in primer solutions 10, 11, 22, 25, 29, 35, 41, 47, 48 and 56 to 58 one to three 5'-primers were not possible to test. Additional primer in primer solutions 4 and 21 were tested by separately adding one or two 5'-primers or one 3'-primer.

101.212-24/04 – including *Taq* pol., IFU-01
 101.212-24u/04u – without *Taq* pol., IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **1G3**

Lot-specific information

ADDRESSES:

Manufacturer:

CareDx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-508 939 00

Fax: +46-8-717 88 18

E-mail: orders-se@caredx.com

Web page: <https://labproducts.caredx.com/>

Distributed by:

CareDx GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: orders-at@caredx.com

Web page: <https://labproducts.caredx.com/>

CareDx Lab Solutions Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-653-78171

Fax: 610-344-7989

E-mail: orders-us@caredx.com

Web page: <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx GmbH**.