

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information  
**Olerup SSP<sup>®</sup> DQA1**

**Product number:** 101.231-24/04 – including *Taq* pol.  
 101.231-24u/04u – without *Taq* pol.

**Lot number:** 82R

**Expiry date:** 2015-July-01

**Number of tests:** 24 tests – Product No. 101.231-24/24u  
 4 tests – Product No. 101.231-04/04u

**Number of wells per test:** 32

**Storage - pre-aliquoted primers:** dark at -20°C

- PCR Master Mix: -20°C
- Adhesive PCR seals: RT
- Product Insert: RT

**This Product Description is only valid for Lot No. 82R.**

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP<sup>®</sup>  
 DQA1 LOT (58M)**

The Lot-specific information for DQA1 including and without *Taq* polymerase is now described in one common Product Insert.

The DQA1 specificity and interpretation tables have been updated compared the previous *Olerup SSP<sup>®</sup> DQA1* lot (**Lot No. 58M**).

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
5	Exchanged	-	Exchanged 5'-primer for improved allelic resolution.
21	Added	Added, modified	Primer pair added for the DQA1*01:09 allele, increased yield of HLA-specific PCR product.
22	-	Added	3'-primer added for the DQA1*01:10 allele.
27	Added	Added	Primer pair added for the DQA1*01:11 allele.
29	-	Added	3'-primer added for the DQA1*01:08 allele.

Change in revision R01 compared to R00:

1. Primer mix 5 may faintly amplify the DQA1\*01:01:01:01:02:04 and 01:04:01:01:01:04:02 alleles. A footnote has been added in the Specificity Table.

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

## PRODUCT DESCRIPTION

### DQA1 SSP typing

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the DQA1\*01:01 to DQA1\*06:02 alleles.

#### PLATE LAYOUT

Each test consists of 32 PCR reactions in a 32 well cut PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	32

The 32 well cut PCR plate is marked with ‘DQA1’, in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘82R’.

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are heat-sealed with a PCR-compatible foil.

**Please note:** When removing each 32 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

Only DQA1 alleles will be amplified by the DQA1 typing kit. Thus, the interpretation of DQA1 typings is not influenced by the DQA2 gene.

#### UNIQUELY IDENTIFIED ALLELES

All the DQA1 alleles, i.e. **DQA1\*01:01 to 01:11, DQA1\*02:01, DQA1\*03:01 to 03:03, DQA1\*04:01 to DQA1\*04:04, DQA1\*05:01 to 05:11 and DQA1\*06:01 to DQA1\*06:02**, recognized by the HLA Nomenclature Committee in October 2012<sup>1</sup> will give rise to unique amplification patterns by the primers in the DQA1 typing kit.

The DQA1 typing kit cannot distinguish the DQA1\*01:01:01-01:01:02 alleles, the DQA1\*01:02:01-01:01:02:03 alleles, the DQA1\*01:03:01-01:01:03:01:02 alleles, the DQA1\*01:04:01-01:01:04:02 alleles, the DQA1\*03:03:01-03:03:02 alleles, the DQA1\*05:01:01-01:05:01:02 alleles, the DQA1\*05:05:01-01:05:05:01:03 alleles and the DQA1\*06:01:01-06:01:02 alleles.

<sup>1</sup>DQA1 alleles listed on the IMGT/HLA web page 2012-October-17, release 3.10.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

**RESOLUTION IN HOMO- AND HETEROZYGOTES**

A total of 49 alleles generate 34 amplification patterns that can be combined in 595 homozygous and heterozygous combinations. 155 of these genotypes do not give rise to unique amplification patterns. The different lengths of the specific PCR products were not considered in these calculations.

-++-+++	-----+	-----+	-----	*01:06, *04:01:01 = *01:06, *04:01:02
-----+	-----+	-----+	-----	*04:01:01, *04:01:01 = *04:01:01, *04:01:02
-----+	-+++---	-----+	-----	*03:02, *03:02 = *03:02, *03:03:01
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:02:01:01 = *01:02:01:01, *01:02:04
-----+	-----+	-----+	-----	*06:01:01, *06:02 = *06:02, *06:02
-----+	-----+	-----+	-----	*04:01:01, *04:04 = *04:01:02, *04:04 = *04:04, *04:04
-----+	-----+	-----+	-----	*04:01:01, *04:03N = *04:01:02, *04:03N
-----+	-----+	-----+	-----	*04:01:01, *04:02 = *04:01:02, *04:02 = *04:02, *04:02
-----+	-----+	-----+	-----	*04:01:01, *06:01:01 = *04:01:02, *06:01:01
-----+	-+-----	-----+	-----	*03:01:01, *04:01:01 = *03:01:01, *04:01:02
-----+	+-----	-----+	-----	*02:01, *04:01:01 = *02:01, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:08 = *01:02:04, *01:08 = *01:08, *01:08
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:11 = *01:02:04, *01:11
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:09 = *01:02:04, *01:09 = *01:09, *01:09
-++-+++	-+-----	-----+	-----	*01:02:01:01, *03:01:01 = *01:02:04, *03:01:01
-++-+++	+-----	-----+	-----	*01:02:01:01, *02:01 = *01:02:04, *02:01
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:03:01:01 = *01:02:04, *01:03:01:01
++-++-	-----+	-----+	-----	*01:04:01:01, *01:07 = *01:07, *01:07
++-++-	-----+	-----+	-----	*01:01:01, *01:02:01:01 = *01:01:01, *01:02:04
-----+	-----+	-----+	-----	*04:01:01, *06:02 = *04:01:02, *06:02
-----+	-----+	-----+	-----	*04:01:01, *05:10 = *04:01:02, *05:10 = *04:02, *05:10
-----+	-----+	-----+	-----	*04:01:01, *05:04 = *04:01:02, *05:04
-++-+++	-----+	-----+	-----	*01:03:01:01, *06:01:01 = *01:10, *06:01:01
-++-+++	-----+	-----+	-----	*01:03:01:01, *04:01:01 = *01:03:01:01, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *04:03N = *01:02:04, *04:03N = *01:11, *04:03N
-++-+++	-----+	-----+	-----	*01:02:01:01, *06:01:01 = *01:02:04, *06:01:01
-++-+++	-----+	-----+	-----	*01:02:01:01, *04:01:01 = *01:02:01:01, *04:01:02 = *01:02:04, *04:01:01 = *01:02:04, *04:01:02
-++-+++	-+-----	-----+	-----	*01:02:01:01, *03:03:01 = *01:02:04, *03:03:01
-++-+++	-----+	-----+	-----	*01:02:01:01, *01:10 = *01:02:04, *01:10
++-++-	-----+	-----+	-----	*01:01:01, *04:01:01 = *01:01:01, *04:01:02
-----+	-----+	-----+	-----	*05:03, *05:07 = *05:07, *05:07
-----+	-----+	-----+	-----	*05:03, *05:06 = *05:06, *05:06
-----+	-----+	-----+	-----	*04:01:01, *05:03 = *04:01:02, *05:03
-----+	-----+	-----+	-----	*04:01:01, *05:01:01:01 = *04:01:02, *05:01:01:01
-----+	-----+	-----+	-----	*05:05:01:01, *05:08 = *05:08, *05:08
-----+	-----+	-----+	-----	*05:05:01:01, *05:11 = *05:11, *05:11
-----+	-----+	-----+	-----	*05:05:01:01, *05:09 = *05:09, *05:09
-----+	-----+	-----+	-----	*04:01:01, *05:02 = *04:01:02, *05:02
-++-+++	-----+	-----+	-----	*01:03:01:01, *06:02 = *01:10, *06:02
-++-+++	-----+	-----+	-----	*01:10, *04:01:01 = *01:10, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *06:02 = *01:02:04, *06:02 = *01:08, *06:01:01 = *01:08, *06:02
-++-+++	-----+	-----+	-----	*01:08, *04:01:01 = *01:08, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *04:04 = *01:02:04, *04:04
-++-+++	-----+	-----+	-----	*01:11, *04:01:01 = *01:11, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *04:02 = *01:02:04, *04:02
-++-+++	-----+	-----+	-----	*01:09, *04:01:01 = *01:09, *04:01:02
-++-+++	-----+	-----+	-----	*01:02:01:01, *05:10 = *01:02:04, *05:10
-++-+++	-----+	-----+	-----	*01:02:01:01, *05:04 = *01:02:04, *05:04
-++-+++	-+++---	-----+	-----	*01:02:01:01, *03:02 = *01:02:04, *03:02
++-++-	-----+	-----+	-----	*01:04:01:01, *04:01:01 = *01:04:01:01, *04:01:02
++-++-	-----+	-----+	-----	*01:05, *04:01:01 = *01:05, *04:01:02
-----+	-----+	-----+	-----	*04:01:01, *05:07 = *04:01:02, *05:07



101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

**Lot No.: 82R**

**Lot-specific information**

-----+	-----++-	-++-----	-----+--	*04:01:01, *05:06 = *04:01:02, *05:06
-----+	-----++-	+--+-----	-----	*04:01:01, *05:05:01:01 = *04:01:02, *05:05:01:01
-++-----	-----++-	-++-----	-----	*01:02:01:01, *05:03 = *01:02:04, *05:03
-++-----	-----++-	+--+-----	-----	*01:02:01:01, *05:01:01:01 = *01:02:04, *05:01:01:01
-++-----	-----++-	-++-----	-----	*01:02:01:01, *05:02 = *01:02:04, *05:02
+--+-----	-----++-	-----++	+-----	*01:07, *04:01:01 = *01:07, *04:01:02
-----+	-----++-	+--+-----	-----+	*04:01:01, *05:08 = *04:01:02, *05:08
-----+	-----++-	+--+-----	-----+	*04:01:01, *05:11 = *04:01:02, *05:11
-----+	-----++-	+--+-----	+-----	*04:01:01, *05:09 = *04:01:02, *05:09
-----+	-----++-	+--+-----	-----+	*05:03, *05:11 = *05:05:01:01, *05:06 = *05:06, *05:11
-++-----	-----++-	-++-----	-----+	*01:02:01:01, *05:07 = *01:02:04, *05:07
-++-----	-----++-	-++-----	-----+	*01:02:01:01, *05:06 = *01:02:04, *05:06
-++-----	-----++-	+--+-----	-----	*01:02:01:01, *05:05:01:01 = *01:02:04, *05:05:01:01 =
-++-----	-----++-	+--+-----	-----+	*01:09, *05:01:01:01 = *01:09, *05:05:01:01
-++-----	-----++-	+--+-----	-----+	*01:02:01:01, *05:08 = *01:02:04, *05:08 = *01:09, *05:08
-++-----	-----++-	+--+-----	-----+	*01:02:01:01, *05:11 = *01:02:04, *05:11 = *01:09, *05:11
-++-----	-----++-	+--+-----	+-----	*01:02:01:01, *05:09 = *01:02:04, *05:09 = *01:09, *05:09
+--+-----	-----++-	+--+-----	+-----	*01:04:01:01, *05:09 = *01:07, *05:05:01:01 = *01:07, *05:09

\*01:01:01 = \*01:01:01-01:01:02  
 \*01:02:01 = \*01:02:01-01:02:03  
 \*01:03:01:01= \*01:03:01:01-01:03:01:02  
 \*01:04:01 = \*01:04:01-01:04:02  
 \*03:03:01 = \*03:03:01-03:03:02  
 \*05:01:01:01 = \*05:01:01:01-05:01:02  
 \*05:05:01:01= \*05:05:01:01-05:05:01:03  
 \*06:01:01 = \*06:01:01-06:01:02



101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

## SPECIFICITY TABLE

### DQA1 SSP typing

Specificities and sizes of the PCR products of the 32 primer mixes used for DQA1 SSP typing

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified DQA1 alleles <sup>3</sup>
<b>1</b>	145 bp	<b>515 bp</b>	*01:01:01-01:01:02, 01:04:01:01-01:05, 01:07
<b>2</b>	170 bp	<b>515 bp</b>	*01:01:01-01:02:04, 01:04:01:01-01:09, 01:11
<b>3</b>	145 bp	430 bp	*01:02:01:01-01:03:01:02, 01:06, 01:08-01:11
<b>4</b>	170 bp	430 bp	*01:03:01:01-01:03:01:02, 01:10
<b>5<sup>12</sup></b>	205 bp	430 bp	*01:05
<b>6<sup>4</sup></b>	100 bp	430 bp	*01:04:01:01-01:04:02, 01:06 <sup>?</sup> , 01:07
<b>7<sup>4,5</sup></b>	95 bp	430 bp	*01:06
<b>8<sup>4</sup></b>	65 bp	430 bp	*01:01:01-01:02:03, 01:03:01:01-01:03:01:02, 01:06 <sup>?</sup> , 01:08-01:11, 02:01, 03:01:01-03:03:02, 04:01:01-04:04, 05:01:01:01-05:11, 06:01:01-06:02
<b>9</b>	175 bp	430 bp	*02:01
<b>10</b>	185 bp	430 bp	*03:01:01-03:03:02
<b>11</b>	215 bp	430 bp	*03:02
<b>12</b>	225 bp	<b>515 bp</b>	*03:02-03:03:02
<b>13</b>	225 bp	<b>515 bp</b>	*01:01:01-01:11, 02:01, 03:01:01, 04:01:01, 04:02-04:04, 05:01:01:01-05:11, 06:01:01-06:02
<b>14<sup>4,5</sup></b>	125 bp	430 bp	*04:01:01-04:02, 04:04, 05:01:01:01-05:11
<b>15</b>	165 bp	430 bp	*05:01:01:01-05:09, 05:11
<b>16<sup>4</sup></b>	90 bp	430 bp	*05:02
<b>17</b>	200 bp	430 bp	*05:01:01:01-05:01:02, 05:02 <sup>?</sup> , 05:04 <sup>?</sup> , 05:05:01:01-05:05:01:03, 05:08-05:11
<b>18</b>	200 bp	430 bp	*05:02 <sup>?</sup> , 05:03, 05:04 <sup>?</sup> , 05:06-05:07
<b>19</b>	205 bp	430 bp	*05:01:01:01-05:03, 05:05:01:01-05:09, 05:11
<b>20</b>	135 bp	430 bp	*05:04
<b>21<sup>4,7</sup></b>	100 bp, 210 bp	430 bp	*01:09, 05:02 <sup>?</sup> , 05:04 <sup>?</sup> , 05:05:01:01-05:05:01:03, 05:08-05:09, 05:10 <sup>?</sup> , 05:11
<b>22<sup>4,6,8</sup></b>	120 bp, 215 bp	<b>515 bp</b>	*01:10, 06:01:01-06:02
<b>23<sup>4</sup></b>	85 bp	430 bp	*04:01:01-04:04, 06:01:01-06:02

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

## Lot-specific information

Lot No.	Allele(s)	Product Size(s)	Allele(s)
<b>24</b>	220 bp	430 bp	*01:01:01-01:09, 02:01 <sup>w</sup> , 03:01:01-03:03:02, 04:01:01-04:04, 05:01:01-05:01:02, 05:02 <sup>?</sup> , 05:03, 05:04 <sup>?</sup> , 05:06-05:07, 05:10 <sup>?</sup> , 06:01:01-06:02
<b>25<sup>4,9</sup></b>	80 bp, 175 bp	430 bp	*01:07, 05:09
<b>26</b>	160 bp	430 bp	*04:02, 05:10
<b>27<sup>4,6,10</sup></b>	90 bp, 135 bp	430 bp	*01:11, 04:03N
<b>28<sup>4</sup></b>	105 bp	430 bp	*04:04
<b>29</b>	150 bp	430 bp	*01:08, 06:02
<b>30<sup>4,11</sup></b>	115 bp, 215 bp	430 bp	*05:06, 05:11
<b>31<sup>4</sup></b>	100 bp	430 bp	*05:07
<b>32</b>	135 bp	430 bp	*05:08

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of DQA1 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits the respective lengths of the HLA-specific PCR product(s) are given for the alleles amplified by these primer mixes.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers, e.g. the primers in wells 11, 12, 17, 18, 21, 23 and 27.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DQA1 typing.

In addition, wells number 2, 12, 13 and 22 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>For several DQA1 alleles 1<sup>st</sup> and/or 3<sup>rd</sup> exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. We assume that unknown sequences in these regions are conserved within allelic groups.

<sup>4</sup>HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

<sup>5</sup>Primer mixes 7 and 14 may have a tendency to giving rise to primer oligomer formation.

<sup>6</sup>Primer mixes 22 and 27 may have tendencies of unspecific amplifications.

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

**Lot No.: 82R****Lot-specific information**

<sup>7</sup>Primer mix 21: Specific PCR product of 100 bp in the DQA1\*05:02<sup>?</sup>, 05:04<sup>?</sup>, 05:05:01:01-05:05:01:03, 05:08-05:09, 05:10<sup>?</sup> and 05:11 alleles. Specific PCR product of 210 bp in the DQA1\*01:09 allele.

<sup>8</sup>Primer mix 22: Specific PCR product of 120 bp in the DQA1\*06:01:01-06:02 alleles. Specific PCR product of 215 bp in the DQA1\*01:10 allele.

<sup>9</sup>Primer mix 25: Specific PCR product of 80 bp in the DQA1\*05:09 allele. Specific PCR product of 175 bp in the DQA1\*01:07 allele.

<sup>10</sup>Primer mix 27: Specific PCR product of 90 bp in the DQA1\*04:03N allele. Specific PCR product of 135 bp in the DQA1\*01:11 allele.

<sup>11</sup>Primer mix 30: Specific PCR product of 115 bp in the DQA1\*05:11 allele. Specific PCR product of 215 bp in the DQA1\*05:06 allele.

<sup>12</sup>Primer mix 5 may faintly amplify the \*01:01:01-01:02:04 and 01:04:01:01-01:04:02 alleles.

‘w’, may be weakly amplified.

‘?’, nucleotide sequence information not available for the primer matching sequence.

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

<b>INTERPRETATION TABLE</b>																
<b>DQA1 SSP typing</b>																
<b>Amplification patterns of the DQA1*01:01 to *06:02 alleles</b>																
	Well <sup>5</sup>															
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
Length of spec. PCR product(s)	145	170	145	170	205	100	95	65	175	185	215	225	225	125	165	90
Length of int. pos. control <sup>1</sup>	515	515	430	430	430	430	430	430	430	430	430	515	515	430	430	430
5'-primer(s) <sup>2</sup>	33(169)	25(143)	33(169)	25(143)	-4(60)	198(664)	25(143)	-7(49)	7(90)	7(90)	-6(53)	99(366)	99(366)	25(143)	33(169)	59(245)
	5'-Agg <sup>3'</sup>	5'-gTA <sup>3'</sup>	5'-AgC <sup>3'</sup>	5'-gTT <sup>3'</sup>	5'-CCC <sup>3'</sup>	5'-gCA <sup>3'</sup>	5'-gTA <sup>3'</sup>	5'-CCg <sup>3'</sup>	5'-CAC <sup>3'</sup>	5'-CAT <sup>3'</sup>	5'-gAC <sup>3'</sup>	5'-CCC <sup>3'</sup>	5'-CCC <sup>3'</sup>	5'-gTA <sup>3'</sup>	5'-AgC <sup>3'</sup>	5'-CCg <sup>3'</sup>
3'-primer(s) <sup>3</sup>	68(274)	68(274)	68(274)	68(274)	1 <sup>st</sup> I	218(722)	43(199)	2(74)	52(224)	54(232)	1 <sup>st</sup> I	160(548)	160(548)	52(226)	75(293)	75(293)
	5'-TgC <sup>3'</sup>	5'-TgC <sup>3'</sup>	5'-TgC <sup>3'</sup>	5'-TgC <sup>3'</sup>	5'-TTT <sup>3'</sup>	5'-CTT <sup>3'</sup>	5'-AgC <sup>3'</sup>	5'-TgT <sup>3'</sup>	5'-TgT <sup>3'</sup>	5'-TCT <sup>3'</sup>	5'-TTT <sup>3'</sup>	5'-CAT <sup>3'</sup>	5'-Cag <sup>3'</sup>	5'-TTg <sup>3'</sup>	5'-gAC <sup>3'</sup>	5'-gAC <sup>3'</sup>
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
DQA1 allele <sup>4</sup>																
*01:01:01-01:01:02	1	2						8					13			
*01:02:01-01:02:03		2	3					8					13			
*01:02:04		2	3										13			
*01:03:01-01:03:01:02			3	4				8					13			
*01:04:01-01:04:02	1	2				6							13			
*01:05	1	2			5								13			
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16







101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

Length of spec.	145	170	145	170	205	100	95	65	175	185	215	225	225	125	165	90
PCR product(s)																
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
*01:06		2	3			?	7	?					13			
*01:07	1	2				6							13			
*01:08		2	3					8					13			
*01:09		2	3					8					13			
*01:10			3	4				8					13			
*01:11		2	3					8					13			
*02:01								8	9				13			
*03:01:01								8		10			13			
*03:02								8		10	11	12				
*03:03:01-03:03:02								8		10		12				
*04:01:01								8					13	14		
*04:01:02								8						14		
*04:02								8					13	14		
*04:03N								8					13			
*04:04								8					13	14		
*05:01:01:01- 05:01:02								8					13	14	15	
*05:02								8					13	14	15	16
*05:03								8					13	14	15	
*05:04								8					13	14	15	
*05:05:01:01- 05:05:01:03								8					13	14	15	
*05:06								8					13	14	15	
*05:07								8					13	14	15	
*05:08								8					13	14	15	
*05:09								8					13	14	15	
*05:10								8					13	14		
*05:11								8					13	14	15	
*06:01:01-06:01:02								8					13			
*06:02								8					13			
DQA1 allele <sup>4</sup>																
Well No.	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

200	200	205	135	100	120	85	220	80	160	90	105	150	115	100	135	Length of spec. PCR product(s)
17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	Well No.
							24									*01:06
							24	25								*01:07
							24					29				*01:08
				21			24									*01:09
					22											*01:10
										27						*01:11
							w									*02:01
							24									*03:01:01
							24									*03:02
							24									*03:03:01-03:03:02
						23	24									*04:01:01
						23	24									*04:01:02
						23	24		26							*04:02
						23	24			27						*04:03N
						23	24				28					*04:04
17		19					24									*05:01:01-05:01:02
?	?	19		?			?									*05:02
	18	19					24									*05:03
?	?		20	?			?									*05:04
17		19		21												*05:05:01-05:05:01:03
	18	19					24						30			*05:06
	18	19					24							31		*05:07
17		19		21											32	*05:08
17		19		21				25								*05:09
17				?			?		26							*05:10
17		19		21									30			*05:11
					22	23	24									*06:01:01-06:01:02
					22	23	24					29				*06:02
																DQA1 allele <sup>4</sup>
17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	Well No.

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

#### Lot-specific information

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 430 base pairs, for most wells, or a band of 515 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the longer, 515 bp, internal positive control band in order to help in the correct orientation of the DQA1 typing.

In addition, wells number 2, 12, 13 and 22 contain the primer pair giving rise to the longer, 515 bp, internal positive control band in order to allow kit identification.

<sup>2</sup>The codon, and in parenthesis the nucleotide, in the 1<sup>st</sup>, 2<sup>nd</sup>, 3<sup>rd</sup> or 4<sup>th</sup> exon or in the 5' untranslated region, matching the specificity-determining 3'-end of the primer is given. Codon and nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The codon, and in parenthesis the nucleotide, in the 1<sup>st</sup>, 2<sup>nd</sup>, 3<sup>rd</sup> or 4<sup>th</sup> exon or the 1<sup>st</sup> intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Codon and nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>4</sup>The sequence of the DQA1\*03012 allele has been shown to be identical to DQA1\*03:02.

DQA1\*05013 has been renamed to DQA1\*05:05.

<sup>5</sup>Primer mix 21: Specific PCR product of 100 bp in the DQA1\*05:02<sup>?</sup>, 05:04<sup>?</sup>, 05:05:01:01-05:05:01:03, 05:08-05:09, 05:10<sup>?</sup> and 05:11 alleles. Specific PCR product of 210 bp in the DQA1\*01:09 allele.

Primer mix 22: Specific PCR product of 120 bp in the DQA1\*06:01:01-06:02 alleles. Specific PCR product of 215 bp in the DQA1\*01:10 allele.

Primer mix 25: Specific PCR product of 80 bp in the DQA1\*05:09 allele. Specific PCR product of 175 bp in the DQA1\*01:07 allele.

Primer mix 27: Specific PCR product of 90 bp in the DQA1\*04:03N allele. Specific PCR product of 135 bp in the DQA1\*01:11 allele.

Primer mix 30: Specific PCR product of 115 bp in the DQA1\*05:11 allele. Specific PCR product of 215 bp in the DQA1\*05:06 allele.

<sup>6</sup>Primer located upstream of the 1<sup>st</sup> exon.

'w', may be weakly amplified.

'?', nucleotide sequence information not available for the primer matching sequence.

101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

CELL LINE VALIDATION SHEET																				
DQA1 SSP typing kit																				
				Well																
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
				Prod. No.:	200960901	201312902	200960903	200960904	201312905	200960906	200960907	200960908	200960909	200960910	200960911	200960912	200960913	200960914	200960915	200960916
IHC cell line		DQA1*																		
1	9001 SA	*01:01			+	+	-	-	-	-	-	+	-	-	-	-	+	-	-	-
2	9280 LK707	*01:03	*03:03		-	-	+	+	-	-	-	+	-	+	-	+	+	-	-	-
3	9011 E4181324	*01:03			-	-	+	+	-	-	-	+	-	-	-	-	+	-	-	-
4	9275 GU373	*05:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
5	9009 KAS011	*01:02			-	+	+	-	-	-	-	+	-	-	-	-	+	-	-	-
6	9353 SM	*01:03	*03:01		-	-	+	+	-	-	-	+	-	+	-	-	+	-	-	-
7	9020 QBL	*05:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
8	9025 DEU	*03			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
9	9026 YAR	*03:01			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
10	9107 LKT3	*03:03			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
11	9051 PITOUT	*02:01			-	-	-	-	-	-	-	+	+	-	-	-	+	-	-	-
12	9052 DBB	*02:01			-	-	-	-	-	-	-	+	+	-	-	-	+	-	-	-
13	9004 JESTHOM	*01:01			+	+	-	-	-	-	-	+	-	-	-	-	+	-	-	-
14	9071 OLGA	*04:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	-	-
15	9075 DKB	*03:02			-	-	-	-	-	-	-	+	-	+	+	+	-	-	-	-
16	9037 SWEIG007	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
17	9282 CTM3953540	*01:03	*05:01		-	-	+	+	-	-	-	+	-	-	-	-	+	+	+	-
18	9257 32367	*01:02	*03:03		-	+	+	-	-	-	-	+	-	+	-	-	+	+	-	-
19	9038 BM16	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
20	9059 SLE005	*01:02			-	+	+	-	-	-	-	+	-	-	-	-	+	-	-	-
21	9064 AMALA	*05:03			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
22	9056 KOSE	*01:02	*01:04		+	+	+	-	-	+	-	+	-	-	-	-	+	-	-	-
23	9124 IHL	*01:03	*01:04		+	+	+	+	-	+	-	+	-	-	-	-	+	-	-	-
24	9035 JBUSH	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
25	9049 IBW9	*02:01			-	-	-	-	-	-	-	+	+	-	-	-	+	+	+	-
26	9285 WT49	*05:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
27	9191 CH1007	*03:03	*01:05		+	+	-	-	+	-	-	+	-	+	-	+	+	-	-	-
28	9320 BEL5GB	*02:01	*03:03		-	-	-	-	-	-	-	+	+	+	-	+	+	-	-	-
29	9050 MOU	*02:01			-	-	-	-	-	-	-	+	+	-	-	-	+	-	-	-
30	9021 RSH	*04:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	-	-
31	9019 DUCAF	*05:01			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
32	9297 HAG	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
33	9098 MT14B	*03:01			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
34	9104 DHIF	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
35	9302 SSTO	*03:01			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
36	9024 KT17	*03:01			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
37	9065 HHKB	*01:03			-	-	+	+	-	-	-	+	-	-	-	-	+	-	-	-
38	9099 LZL	*05:03			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
39	9315 CML	*03:03	*05:01		-	-	-	-	-	-	-	+	-	+	-	+	+	+	+	-
40	9134 WHONP199	*02:01	*03:02		-	-	-	-	-	-	-	+	+	+	+	+	+	-	-	-
41	9055 H0301	*01:02			-	+	+	-	-	-	-	+	-	-	-	-	+	-	-	-
42	9066 TAB089	*01:03			-	-	+	+	-	-	-	+	-	-	-	-	+	-	-	-
43	9076 T7526	*03:02			-	-	-	-	-	-	-	+	-	+	+	+	-	-	-	-
44	9057 TEM	*01:04			+	+	-	-	-	+	-	-	-	-	-	-	+	-	-	-
45	9239 SHJO	*02:01	*03:03		-	-	+	+	-	-	-	+	-	+	-	+	+	-	-	-
46	9013 SCHU	*01:02			-	+	+	-	-	-	-	+	-	-	-	-	+	-	-	-
47	9045 TUBO	*05:05			-	-	-	-	-	-	-	+	-	-	-	-	+	+	+	-
48	9303 TER-ND	*01:01			+	+	-	-	-	-	-	+	-	-	-	-	+	-	-	-



101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

<b>CELL LINE VALIDATION SHEET</b>																			
<b>DQA1 SSP typing kit</b>																			
				Well															
				17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32
				200960917	200960918	200960919	200960920	201312921	201312922	200960923	200960924	200960925	201189126	201312927	200960928	201312929	201189130	200960931	200960932
	IHWC cell line	DQA1*	Prod. No.:																
1	9001 SA	*01:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
2	9280 LK707	*01:03	*03:03	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
3	9011 E4181324	*01:03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
4	9275 GU373	*05:01		+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
5	9009 KAS011	*01:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
6	9353 SM	*01:03	*03:01	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
7	9020 QBL	*05:01		+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
8	9025 DEU	*03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
9	9026 YAR	*03:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
10	9107 LKT3	*03:03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
11	9051 PITOUT	*02:01		-	-	-	-	-	-	-	-	W	-	-	-	-	-	-	-
12	9052 DBB	*02:01		-	-	-	-	-	-	-	-	W	-	-	-	-	-	-	-
13	9004 JESTHOM	*01:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
14	9071 OLGA	*04:01		-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-
15	9075 DKB	*03:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
16	9037 SWEIG007	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*01:03	*05:01	+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
18	9257 32367	*01:02	*03:03	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
19	9038 BM16	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*01:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
21	9064 AMALA	*05:03		-	+	+	-	-	-	-	-	+	-	-	-	-	-	-	-
22	9056 KOSE	*01:02	*01:04	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
23	9124 IHL	*01:03	*01:04	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
24	9035 JBUSH	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*02:01		-	-	-	-	-	-	-	-	W	-	-	-	-	-	-	-
26	9285 WT49	*05:01		+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
27	9191 CH1007	*03:03	*01:05	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
28	9320 BEL5GB	*02:01	*03:03	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
29	9050 MOU	*02:01		-	-	-	-	-	-	-	-	W	-	-	-	-	-	-	-
30	9021 RSH	*04:01		-	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-
31	9019 DUCAF	*05:01		+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
32	9297 HAG	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*03:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
34	9104 DHIF	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*03:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
36	9024 KT17	*03:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
37	9065 HHKB	*01:03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
38	9099 LZL	*05:03		-	+	+	-	-	-	-	-	+	-	-	-	-	-	-	-
39	9315 CML	*03:03	*05:01	+	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-
40	9134 WHONP199	*02:01	*03:02	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
41	9055 H0301	*01:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
42	9066 TAB089	*01:03		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
43	9076 T7526	*03:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
44	9057 TEM	*01:04		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
45	9239 SHJO	*02:01	*03:03	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
46	9013 SCHU	*01:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
47	9045 TUBO	*05:05		+	-	+	-	+	-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND	*01:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-



101.231-24/04– including *Taq* pol., IFU-01  
 101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
 “Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

## CERTIFICATE OF ANALYSIS

### Olerup SSP® DQA1 SSP

**Product number:** 101.231-24/04 – including *Taq* pol.  
 101.231-24u/04u – without *Taq* pol.  
**Lot number:** 82R  
**Expiry date:** 2015-July-01  
**Number of tests:** 24 tests – Product No. 101.231-24/24u  
 4 tests – Product No. 101.231-04/04u  
**Number of wells per test:** 32

#### Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2009-609-01	13	2009-609-13	25	2009-609-25
2	2013-129-02	14	2009-609-14	26	2011-891-26
3	2009-609-03	15	2009-609-15	27	2013-129-27
4	2009-609-04	16	2009-609-16	28	2009-609-28
5	2013-129-05	17	2009-609-17	29	2013-129-29
6	2009-609-06	18	2009-609-18	30	2011-891-30
7	2009-609-07	19	2009-609-19	31	2009-609-31
8	2009-609-08	20	2009-609-20	32	2009-609-32
9	2009-609-09	21	2013-129-21		
10	2009-609-10	22	2013-129-22		
11	2009-609-11	23	2009-609-23		
12	2009-609-12	24	2009-609-24		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 7, 16, 20, 26 to 29 and 31 were available. In primer solutions 7, 26, 29 and 31 the 5'-primers were tested by adding one additional 3'-primer, the 3'-primers were not possible to test. In primer solutions 16, 20, 27 and 28 the 3'-primers were tested by adding one additional 5'-primer, the 5'-primers were not possible to test. In primer solutions 21 and 30, one 5'-primer was not possible to test, and in primer solution 22 one 3'-primer was not possible to test.

Additional 3'-primers in primer solutions 21 and 30 were tested by separately adding one 5'-primer, and one additional 5'-primer in primer solution 25 was tested by separately adding one 3'-primer.

**Results:** No false positive or false negative amplifications were obtained.

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

*Date of approval:* 2013-February-08

*Approved by:*

Production Quality Control



101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

## Declaration of Conformity

**Product name:** *Olerup* SSP® DQA1  
**Product number:** 101.231-24/04  
**Lot number:** 82R

**Intended use:** HLA-DQA1 high resolution histocompatibility testing

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden  
2014-October-07

Daniel Malica  
Head of QA and Regulatory Affairs

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

101.231-24/04– including *Taq* pol., IFU-01  
101.231-24u/04u– without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **82R**

Lot-specific information

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup-ssp.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.