

Olerup SSP[®] HLA-A low resolution

Product number:	101.401-48/12 – including <i>Taq</i> polymerase
Lot number:	26F
Expiry date:	2010-October-01
Number of tests:	48 tests – Product No. 101.401-48 12 tests – Product No. 101.401-12
Number of wells per test:	23 +1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 26F.

CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP*[®] HLA-A LOW RESOLUTION LOT

The **HLA-A low resolution** specificity and interpretation tables have been updated for the HLA-A alleles described since the previous *Olerup SSP*[®] HLA-A low resolution lot was made (**Lot No. X92**).

One or both primers of the wells detailed below have been exchanged, modified or added compared to the previous lot.

Well	5'-primer	3'-primer	rationale
16	Exchanged	-	To amplify the A*3121 allele.

Lot No.: **26F**

Lot-specific Information

www.olerup.com

Well 24 contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup* SSP[®] HLA Class I, DRB, DQB1 and DPB1 amplicons as well as the amplicons generated by control primer pairs.

PCR product sizes range from 75 to 430 base pairs.
The PCR product generated by the control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80
5'-primer¹	164	340	440	45	45	43
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tg g ^{3'}	5'-Tg g ^{3'}	5'-Tg g ^{3'}
3'-primer²	231	2nd I	507	59	58	57
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}
A*	+	+	+			
B*	+	+	+			
Cw*	+	+	+			
DRB1				+	+	
DRB3				+	+	
DRB5				+		
DQB1					+	
DPB1						+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

PRODUCT DESCRIPTION

HLA-A low resolution

CONTENT

The primer set contains 5'- and 3'-primers for grouping the HLA-A*0101 to A*8001 alleles into the corresponding serological groups A1 to A80.

PLATE LAYOUT

Each test consists of 24 PCR reactions in a 24 well PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24

Wells 1 to 23 – HLA-A low resolution primers.

Well 24 – Negative Control.

The 24 well PCR plate is marked with 'HLA-A low' in silver/gray ink.

Well No. 1 is marked with the Lot No. '26F'.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 24 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Only HLA-A alleles will be amplified by the 23 wells of the HLA-A low resolution primer set except that primer mix 6 will amplify the B*1827 allele. Thus, the interpretation of HLA-A low resolution is only influenced by this HLA-B allele and not by other HLA Class I genes.

UNIQUELY IDENTIFIED ALLELES

All the HLA-A alleles, i.e. **A*0101 to A*8001**, recognized by the HLA Nomenclature Committee in July 2008¹ will be amplified by the primers in the HLA-A low resolution primer set. The HLA-A alleles will be grouped into their corresponding serological specificities².

¹HLA-A, HLA-B and HLA-Cw alleles listed on the IMGT/HLA web page 2008-July-11, release 2.22.0, www.ebi.ac.uk/imgt/hla.

²The A*0318 and A*2619 alleles will give rise to identical amplification patterns. These two alleles can be separated by the respective high resolution SSP primer sets. The A*2314, A*2405, A*241302 and A*2424 alleles will give rise to identical amplification patterns. These four alleles can be separated by the respective high resolution SSP primer sets.

SPECIFICITY TABLE

HLA-A low resolution primer set

Specificities and sizes of the PCR products of the 23 primer mixes used for HLA-A low resolution SSP typing

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	HLA-A serology ³	Amplified HLA-A alleles ^{4,11,12}
1⁵	120, 145 bp	800 bp	A1, A36	01010101-0104N, 0106-0130, 3601-3604
2⁶	135, 200, 255 bp	1070 bp	A2	02010101-020115, 020117-0222, 0224-023501, 023503-0247, 0249-0297, 0299, 9201-9221, 9223-9228, 9230-9248
3⁷	205, 235 bp	1070 bp	A1, A3, A11, A32, A34, A36	0112, 0119, 0121, 03010101-0317, 0319-0339, 1125, 3204, 3402-3404, 3407, 3408, 3602
4⁷	190 bp	800 bp	A1, A2, A3, A11, A26, A30, A36	01010101-0104N, 0106, 0107, 0109-0111N, 0113, 0116N-0118N, 0120-0129, 0278, 0312, 0318, 110101-1127, 1129-1136, 2619, 3008, 3604
5	160, 535 bp	800 bp	A3, A23, A24, A31, A32	0330, 2301-2319Q, 24020101-2411N, 241301-2415, 2417-2464, 2466-2488, 2490N, 2491, 3108, 3205, 3213
6	165, 200 bp	800 bp	A2, A23, A24, A29, A80	9222, 2301-2319Q, 2405, 241302, 2424, 2907, 8001, B*1827
7	175, 205 bp	1070 bp	A2, A23, A24, A26, A33	021701-021702 ^{weakly} , 2314, 24020101-2411N, 241301-241302, 2417-2450, 2454-2456, 2458-2463, 2466-2491, 2616, 3319
8	165, 200 bp	800 bp	A2, A3, A11, A25, A26, A32, A34, A66, A68, A69	0255, 0324, 1110, 250101-2506, 260101-2606, 2608-2615, 2617, 2618, 2620-2637, 3215, 340101-3408, 6601-6607, 680101-6843, 6901
9⁵	75 bp	800 bp	A25, A32	250101-2506, 320101-3202, 3204, 3206-3216
10⁵	85 bp	1070 bp	A2, A26, A33	9246, 260101-2602, 2604, 260701-2618, 2620, 2622-2629, 2632-2637, 3313

Lot No.: **26F**

Lot-specific Information

www.olerup.com

11^{5,7,8}	80, 175, 500 bp	1070 bp	A1, A11, A24, A26, A31, A34, A66	0113, 0128, 110101-1127, 1129-1136, 2419, 2444, 2603, 2606, 2621, 3103, 340101- 3408, 6601, 6604-6607
12	185 bp	800 bp	A11, A25, A26, A31, A34, A43, A66	1111, 2506, 2609, 3103, 3104, 340101-3408, 4301, 6602, 6603
13⁷	175, 225 bp	1070 bp	A1, A2, A3, A25, A26, A34, A43, A66	0113, 0234-023503, 0256, 0262, 9203, 9235, 03010101- 0307, 0309-0311N, 0313-0331, 0333-0335, 0337-0339, 250101- 2505, 260101-260106, 2602 ^{weakly} , 2603, 2605-2608, 2610-2633, 2635-2637, 3408, 4301, 6601, 6604-6607
14	200, 240 bp	1070 bp	A29, A33	29010101-2917, 3313
15⁵	90, 135 bp	1070 bp	A1, A30	0113, 0128, 300101-3004, 3006-3020, 3022-3025
16	370, 400 bp	1070 bp	A31, A32	2914, 310102-3122, 3205
17	180 bp	1070 bp	A29, A32	2913, 320101-3203, 3205-3216
18	200 bp	1070 bp	A33, A68	3301, 330301-3319, 6829
19	160, 200 bp	800 bp	A74	7401-7412N
20⁹	155, 240 bp	800 bp	A2, A210, A25, A30, A68	0210, 0234-023503, 0239, 0246, 0248, 0250, 0252, 0256, 0262, 0270, 0273, 0278, 0293, 0295, 9203, 9210, 9214, 9217, 9222, 9229, 2505, 3008, 680101-6843
21	240, 375 bp	800 bp	A2, A24, A26, A68, A69	0255, 2482, 2622, 6829, 6901
22^{5,10}	85, 240 bp	800 bp	A2, A30, A36	0234-023503, 0246, 0248, 0256, 0262, 0270, 0278, 9203, 9229, 3008, 3601-3604
23⁵	75, 160 bp	800 bp	A2, A26, A68, A80	0255, 2603, 2605, 2606, 2621, 2630, 6805, 6815, 6820, 8001

Lot No.: **26F**

Lot-specific Information

www.olerup.com

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-A low resolution SSP typings.

When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective length of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-A low resolution typing.

In addition, wells number 4, 5, 6, 8, 9, 11, 12, 19 to 23 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to allow kit identification.

In the presence of a specific amplification the intensity of the control band often decreases.

³The serological reactivity of all HLA-A alleles is not known. In this table we use the information in the HLA Dictionary 2004 on the www.ebi.ac.uk/imgt/hla web site and the information available at the www.worldmarrow.org web site and have also inferred the serological grouping from the naming of the sequence-defined allele.

⁴For several HLA-A alleles only partial 1st exon nucleotide sequences are available. We assume that unknown sequences are conserved within allelic groups.

Nucleotide sequence information is available for only exons 2 and 3 of many HLA Class I alleles and for only exon 2 of many HLA Class II alleles and not for other exons or for the introns of these alleles. We assume that unknown sequences in these exons and in the introns are conserved within loci and within allelic groups.

⁵Specific PCR fragments shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR bands.

⁶The primer pairs in vial 2 will in most samples give rise to two specific PCR fragments.

⁷Primer mixes 3, 4, 11 and 13 may give rise to nonspecific amplifications.

⁶The primer pairs in vial 11 will in most samples give rise to two to three specific PCR fragments.

⁹Primer mix 20 will give rise to nonspecific amplifications.

¹⁰Primer mix 22 might faintly amplify most A*11 alleles.

¹¹The A*0318 and A*2619 alleles will give rise to identical amplification patterns. These two alleles can be separated by the respective high resolution SSP primer sets.

The A*2314, A*2405, A*241302 and A*2424 alleles will give rise to identical amplification patterns. These four alleles can be separated by the respective high resolution SSP primer sets.

¹²Primer mix 6 will amplify the B*1827 allele.

‘w’, might be weakly amplified.

INTERPRETATION TABLE																								
HLA-A low resolution SSP typing																								
Amplification patterns of the A*0101 to A*8001 alleles																								
		Well ⁶																						
		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
Length of spec.		120	135	205	190	160	165	175	165	75	85	80	185	175	200	90	370	180	200	160	155	240	85	75
PCR product(s)		145	200	235		535	200	205	200			175		225	240	135	400			200	240	375	240	160
			255									500												
Length of int.		800	1070	1070	800	800	800	1070	800	800	1070	1070	800	1070	1070	1070	1070	1070	1070	800	800	800	800	800
pos. control ¹																								
5'-primer(s) ²		98	78	363	98	144	106	98	98	266	257	302	103	98	98	203	41	180	98	180	78	28	78	176
		5'-CTT ^{3'}	5'-TCT ^{3'}	5'-ATA ^{3'}	5'-CTA ^{3'}	5'-gCC ^{3'}	5'-CCA ^{3'}	5'-CTC ^{3'}	5'-CTA ^{3'}	5'-ACg ^{3'}	5'-Cgg ^{3'}	5'-gga ^{3'}	5'-CCT ^{3'}	5'-CTT ^{3'}	5'-CAC ^{3'}	5'-gAA ^{3'}	5'-CTT ^{3'}	5'-TTT ^{3'}	5'-CAC ^{3'}	5'-TTT ^{3'}	5'-TCT ^{3'}	5'-TCg ^{3'}	5'-TCT ^{3'}	5'-gCA ^{3'}
		123	106		413	317	176	368	102	266	261	385	423	423		363					391	261	527	261
		5'-AgT ^{3'}	5'-CCA ^{3'}		5'-CCg ^{3'}	5'-gCT ^{3'}	5'-gCA ^{3'}	5'-gTT ^{3'}	5'-ACA ^{3'}	5'-ACg ^{3'}	5'-AAC ^{3'}	5'-ggC ^{3'}	5'-gCT ^{3'}	5'-gCT ^{3'}	5'-gCT ^{3'}	5'-ATA ^{3'}					5'-ACg ^{3'}	5'-AAC ^{3'}	5'-TgC ^{3'}	5'-AAC ^{3'}
			144				368		413												391			
			5'-gCA ^{3'}				5'-gTT ^{3'}		5'-CCg ^{3'}												5'-ACg ^{3'}			
			144																					
			5'-TCA ^{3'}																					
3'-primer(s) ³		203	240	527	256	265	256	259	259	302	299	341	257	282	257	299	238	317	256	299	265	97	265	292
		5'-TCT ^{3'}	5'-gga ^{3'}	5'-CCA ^{3'}	5'-CTg ^{3'}	5'-CCC ^{3'}	5'-CCC ^{3'}	5'-gTT ^{3'}	5'-gTT ^{3'}	5'-ggC ^{3'}	5'-TCg ^{3'}	5'-CgT ^{3'}	5'-gCA ^{3'}	5'-gAC ^{3'}	5'-gCA ^{3'}	5'-CCA ^{3'}	5'-CCT ^{3'}	5'-gga ^{3'}	5'-CCC ^{3'}	5'-CCA ^{3'}	5'-CCC ^{3'}	5'-ggT ^{3'}	5'-CCC ^{3'}	5'-gTT ^{3'}
			292	527	559	570	292	502	538			521	559	559	299	411	265		259	299	282	355	282	292
			5'-gTg ^{3'}	5'-CCT ^{3'}	5'-CCg ^{3'}	5'-CCg ^{3'}	5'-gTT ^{3'}	5'-CTT ^{3'}	5'-CCA ^{3'}			5'-ggg ^{3'}	5'-CTC ^{3'}	5'-CCg ^{3'}	5'-TCg ^{3'}	5'-TCA ^{3'}	5'-CCC ^{3'}		5'-gTT ^{3'}	5'-CCA ^{3'}	5'-gAC ^{3'}	5'-gAC ^{3'}	5'-gAC ^{3'}	5'-gTg ^{3'}
				555		502	539					559								341	506	570	299	
				5'-CCA ^{3'}		5'-CTg ^{3'}	5'-TCT ^{3'}					5'-CgT ^{3'}								5'-CgT ^{3'}	5'-TgT ^{3'}	5'-CAC ^{3'}	5'-TCT ^{3'}	
				555		538														5'-CgT ^{3'}	5'-TgT ^{3'}	5'-CAC ^{3'}	5'-TCT ^{3'}	
			5'-gCA ^{3'}			5'-Cag ^{3'}																		
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
HLA-A allele ⁴	ser ⁵																							
*01010101-0104N, 0106, 0107, 0109-0111N, 0116N-0118N, 0120, 0122N-0127N, 0129	A1, null	1			4																			
*0108, 0114, 0115N, 0130	1	1																						
*0112, 0119	-	1	3																					
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23

Lot No.: **26F**

Lot-specific Information

www.olerup.com

Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
*0113	-	1			4							11		13		15								
*0121	-	1		3	4																			
*0128	-	1			4							11				15								
*02010101-020115, 020117-0209, 0211-0216, 0218-0222, 0224-0233, 0236-0238, 0240-0245, 0247, 0249, 0251, 0253N, 0254, 0257-0261, 0263- 0269, 0271, 0272, 027401- 0277, 0279-0292, 0294N, 0296, 0297, 0299, 9201, 9202, 9204-9209, 9211- 9213N, 9215, 9216, 9218- 9221, 9223-9228, 9230- 9234, 9236-9245, 9247, 9248	A2, A203 , null		2																					
*0210, 0239, 0250, 0252, 0273, 0293, 0295, 9210, 9214, 9217	A210		2																		20			
*021701-021702	A2		2					w																
*0234, 023501, 023503, 0256, 0262, 9203	-		2											13							20		22	
*023502	-													13							20		22	
*0246, 0270	-		2																		20		22	
*0248, 9229	-																				20		22	
*0255	-		2						8													21		23
*0278	-		2		4																20		22	
*9222	-						6														20			
*9235	-		2											13										
*9246	-		2								10													
*03010101-0307, 0309- 0311N, 0313-0317, 0319- 0323, 0325-0329, 0331, 0333-0335, 0337-0339	A3			3										13										
*0308, 0332, 0336N	-			3																				
*0312	-			3	4																			
*0318, 2619 ⁷	-				4									13										
*0324	-			3					8					13										
*0330	-			3		5								13										
*110101-1109, 1112- 112402, 1126, 1127, 1129- 1136	A11				4							11												
*1110	-				4				8			11												
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23

Lot No.: **26F**

Lot-specific Information

www.olerup.com

Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
*1111	-				4							11	12											
*1125	-			3	4							11												
*2301-2313, 2315-2319Q	A23					5	6																	
*2314, 2405, 241302, 2424 ^b	-					5	6	7																
*24020101-2404, 2406-2411N, 241301, 2417, 2418, 2420-2423, 2425-2443, 2445N-2450, 2454-2456, 2458-2463, 2466-2481, 2483N-2488, 2490, 2491	A24					5		7																
*2414, 2415, 2451-2453, 2457, 2464	A24					5																		
*2419, 2444	-					5		7				11												
*2482	-					5		7														21		
*2489	-							7																
*250101-2504	A25								8	9				13										
*2505	-								8	9				13								20		
*2506	-								8	9			12											
*260101-260106, 2608, 2610-2615, 2617, 2618, 2620, 2623-2629, 2632, 2633, 2635-2637	A26, null								8			10		13										
*2602	A26								8		10			w										
*2603, 2606, 2621	A26								8			11		13										23
*2604, 2634	A26								8		10													
*2605, 2630	A26								8					13										23
*260701-260702	A26										10			13										
*2609	A26								8		10		12											
*2616	-							7			10			13										
*2622	-								8		10			13									21	
*2631	-								8					13										
*29010101-2906, 2908N-2912, 2915-2917	A29														14									
*2907	-							6							14									
*2913	-														14				17					
*2914	-														14		16							
*300101-3004, 3006, 3007, 3009-3020, 3022-3025	A30															15								
*3008	-				4											15						20	22	
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23

Lot No.: **26F**

Lot-specific Information

www.olerup.com

Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
*310102-3102, 3105-3107, 3109-3122	A31																16							
*3103	-											11	12				16							
*3104	A31												12				16							
*3108	-					5											16							
*320101-3202, 3206-3212, 3214, 3216	A32									9								17						
*3203	-																	17						
*3204	A3			3						9														
*3205	-					5											16	17						
*3213	-					5				9								17						
*3215	-								8	9								17						
*3301, 330301-3312, 3314-3318	A33																	18						
*3313	-									10					14			18						
*3319	-							7										18						
*340101-340102, 3405, 3406	A34								8			11	12											
*3402-3404, 3407	-			3					8			11	12											
*3408	-			3					8			11	12	13										
*3601, 3603	A36	1																					22	
*3602	-	1		3																			22	
*3604	-	1			4																		22	
*4301	A43											12	13											
*6601, 6604-6607	A66							8			11		13											
*6602, 6603	A66							8				12												
*680101-6804, 6806-6814, 6816-6819, 6821-6828, 6830-6843	A68								8												20			
*6805, 6815, 6820	-								8												20			23
*6829	-								8									18	20	21				
*6901	A69								8														21	
*7401-7412N	A74																			19				
*8001	A80					6																		23
HLA-A allele ⁴	ser ⁵																							
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23
B*1827						6																		
Well No.		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23

Lot No.: **26F**

Lot-specific Information

www.olerup.com

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-A low resolution SSP typing.

In addition, wells number 4, 5, 6, 8, 9, 11, 12, 19 to 23 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to allow kit identification.

²The nucleotide position, in the 1st, 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position, in the 2nd, or 3rd exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

⁴The sequence of the A*0105N has been shown to be identical to A*0104N.

The A*020116 allele has been renamed to A*9234.

The A*020120 has been renamed to A*020118.

The sequence of the A*0223 allele has been shown to be identical to A*0222.

The sequence of the A*0298 allele has been shown to be identical to A*0296.

The A*1128 allele has been renamed to A*111502.

The sequence of the A*2401 allele has been shown to be in error.

The sequence of the A*2412 allele has been shown to be identical to A*2408.

The A*2416 allele has been renamed to A*3108.

The A*2465 allele has been renamed to A*241302.

The sequence of the A*3005 allele has been shown to be identical to A*3004.

The A*3021 allele has been renamed to A*301102.

The sequence of the A*31011 allele has been shown to be identical to A*310102.

The sequence of the A*3302 allele has been shown to be identical to A*3303.

⁵The serological reactivity of all HLA-A alleles is not known. In this table we use the information in the HLA Dictionary 2004 on the www.ebi.ac.uk/imgt/hla web site and the information available at the www.worldmarrow.org web site and have also inferred the serological grouping from the naming of the sequence-defined allele.

⁶The primer pairs in vials 2 and 11 will in most samples give rise to two specific PCR fragments.

⁷The A*0318 and A*2619 alleles will give rise to identical amplification patterns. These two alleles can be separated by the respective high resolution SSP primer sets.

⁸The A*2314, A*2405, A*2424 and A*2465 alleles will give rise to identical amplification patterns. These two alleles can be separated by the respective high resolution SSP primer sets. 'w', may be weakly amplified.

CELL LINE VALIDATION SHEET																				
HLA-A low resolution primer set																				
				Lot No.:	Well															
					1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
					200739101	200739102	200739103	200739104	200847005	200739106	200739107	200739108	200739109	200739110	200739111	200739112	200739113	200739114	20847015	200847016
	IHWC cell line	A*	A*																	
1	9001 SA	*2402			-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-
2	9280 LK707	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324	*0101			+	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*3001			+	-	-	-	-	-	-	-	-	-	-	-	-	-	+	+
5	9009 KAS011	*0101			+	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*0201	*2603		-	+	-	-	-	-	-	+	-	-	+	-	+	-	-	-
7	9020 QBL	*2601			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
8	9007 DEM	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*2601			-	-	-	-	-	-	-	+	-	+	-	-	+	-	-	-
10	9107 LKT3	*2402			-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-
11	9051 PITOUT	*2902			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-
12	9052 DBB	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9067 BTB	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*3101			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*2402			-	-	-	-	+	-	+	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*2902			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-
17	9008 WILJON	*2501			-	-	-	-	-	-	-	+	+	-	-	-	+	-	-	-
18	9257 32367	*3303	*7401		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*0217			-	+	-	-	-	-	W	-	-	-	-	-	-	-	-	-
22	9056 KOSE	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*0201	*3401		-	+	-	-	-	-	-	+	-	-	+	+	-	-	-	-
24	9035 JBUSH	*3201			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
25	9049 IBW9	*3301			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*0205			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*2410	*2901		-	-	-	-	+	-	+	-	-	-	-	-	-	+	-	-
28	9320 BEL5GB	*0201	*2902		-	+	-	-	-	-	-	-	-	-	-	-	-	+	-	-
29	9050 MOU	*2902			-	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-
30	9021 RSH	*3001	*6802		-	-	-	-	-	-	-	+	-	-	-	-	-	-	+	+
31	9019 DUCAF	*3002			-	-	-	-	-	-	-	-	-	-	-	-	-	-	+	+
32	9297 HAG	*0201			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*3101			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*3101			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*3201			-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
36	9024 KT17	*0206	*1101		-	+	-	+	-	-	-	-	-	-	+	-	-	-	-	-
37	9065 HHKB	*0301			-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
38	9099 LZL	*0217			-	+	-	-	-	-	W	-	-	-	-	-	-	-	-	-
39	9315 CML	*0101	*0301		+	-	+	+	-	-	-	-	-	-	-	-	+	-	-	-
40	9134 WHONP199	*0207	*3001		-	+	-	-	-	-	-	-	-	-	-	-	-	-	+	+
41	9055 H0301	*0301			-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
42	9066 TAB089	*0207			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*0207			-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*6601			-	-	-	-	-	-	-	+	-	-	+	-	+	-	-	-
45	9239 SHJO	*2301	*2402		-	-	-	-	+	+	+	-	-	-	-	-	-	-	-	-
46	9013 SCHU	*0301			-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
47	9045 TUBO	*0216	*0301		-	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-
48	9303 TER-ND	*0201	*1101		-	+	-	+	-	-	-	-	-	-	+	-	-	-	-	-

CELL LINE VALIDATION SHEET												
HLA-A low resolution primer set												
					Well							
					17	18	19	20	21	22	23	
					Lot No.:	200739717	200739718	200739719	200852920	200739721	200739722	200739723
	IHC cell line	A*	A*									
1	9001 SA	*2402			-	-	-	-	-	-	-	-
2	9280 LK707	*0201			-	-	-	-	-	-	-	-
3	9011 E4181324	*0101			-	-	-	-	-	-	-	-
4	9275 GU373	*3001			-	-	-	-	-	-	-	-
5	9009 KAS011	*0101			-	-	-	-	-	-	-	-
6	9353 SM	*0201	*2603		-	-	-	-	-	-	-	+
7	9020 QBL	*2601			-	-	-	-	-	-	-	-
8	9007 DEM	*0201			-	-	-	-	-	-	-	-
9	9026 YAR	*2601			-	-	-	-	-	-	-	-
10	9107 LKT3	*2402			-	-	-	-	-	-	-	-
11	9051 PITOUT	*2902			-	-	-	-	-	-	-	-
12	9052 DBB	*0201			-	-	-	-	-	-	-	-
13	9067 BTB	*0201			-	-	-	-	-	-	-	-
14	9071 OLGA	*3101			-	-	-	-	-	-	-	-
15	9075 DKB	*2402			-	-	-	-	-	-	-	-
16	9037 SWEIG007	*2902			-	-	-	-	-	-	-	-
17	9008 WILJON	*2501			-	-	-	-	-	-	-	-
18	9257 32367	*3303	*7401		-	+	+	-	-	-	-	-
19	9038 BM16	*0201			-	-	-	-	-	-	-	-
20	9059 SLE005	*0201			-	-	-	-	-	-	-	-
21	9064 AMALA	*0217			-	-	-	-	-	-	-	-
22	9056 KOSE	*0201			-	-	-	-	-	-	-	-
23	9124 IHL	*0201	*3401		-	-	-	-	-	-	-	-
24	9035 JBUSH	*3201			+	-	-	-	-	-	-	-
25	9049 IBW9	*3301			-	+	-	-	-	-	-	-
26	9285 WT49	*0205			-	-	-	-	-	-	-	-
27	9191 CH1007	*2410	*2901		-	-	-	-	-	-	-	-
28	9320 BEL5GB	*0201	*2902		-	-	-	-	-	-	-	-
29	9050 MOU	*2902			-	-	-	-	-	-	-	-
30	9021 RSH	*3001	*6802		-	-	-	+	-	-	-	-
31	9019 DUCAF	*3002			-	-	-	-	-	-	-	-
32	9297 HAG	*0201			-	-	-	-	-	-	-	-
33	9098 MT14B	*3101			-	-	-	-	-	-	-	-
34	9104 DHIF	*3101			-	-	-	-	-	-	-	-
35	9302 SSTO	*3201			+	-	-	-	-	-	-	-
36	9024 KT17	*0206	*1101		-	-	-	-	-	-	-	-
37	9065 HHKB	*0301			-	-	-	-	-	-	-	-
38	9099 LZL	*0217			-	-	-	-	-	-	-	-
39	9315 CML	*0101	*0301		-	-	-	-	-	-	-	-
40	9134 WHONP199	*0207	*3001		-	-	-	-	-	-	-	-
41	9055 H0301	*0301			-	-	-	-	-	-	-	-
42	9066 TAB089	*0207			-	-	-	-	-	-	-	-
43	9076 T7526	*0207			-	-	-	-	-	-	-	-
44	9057 TEM	*6601			-	-	-	-	-	-	-	-
45	9239 SHJO	*2301	*2402		-	-	-	-	-	-	-	-
46	9013 SCHU	*0301			-	-	-	-	-	-	-	-
47	9045 TUBO	*0216	*0301		-	-	-	-	-	-	-	-
48	9303 TER-ND	*0201	*1101		-	-	-	-	-	-	-	-

CERTIFICATE OF ANALYSIS

Olerup SSP[®] HLA-A low resolution

Product number: 101.401-48/12 – including *Taq* polymerase
Lot number: 26F
Expiry date: 2010-October-01
Number of tests: 48 tests – Product No. 101.401-48
12 tests – Product No. 101.401-12
Number of wells per test: 23 + 1

Well specifications:

Well No.	Production No.	Well No.	Production No.	Well No.	Production No.
1	2007-397-01	9	2007-397-09	17	2007-397-17
2	2007-397-02	10	2007-397-10	18	2007-397-18
3	2007-397-03	11	2007-332-11	19	2007-397-29
4	2007-397-04	12	2007-397-12	20	2008-529-20
5	2008-470-05	13	2007-397-13	21	2007-397-21
6	2007-332-06	14	2007-397-14	22	2007-397-22
7	2007-397-07	15	2008-470-15	23	2007-397-23
8	2007-397-08	16	2008-470-16		

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

Additional 5'-primers in primer solutions 8 and 10 were tested by separately adding one 3'-primer. Additional 3'-primers in primer solutions 3, 4, 8, 14, 16, 18, 19 and 23 were tested by separately adding one 5'-primer. Additional 5'- and 3'-primers in primer solutions 5, 7, 12 and 22 were tested by separately adding one 3'-primer, respectively one 5'-primer.

The negative control primer pairs, **Production No. 2008-417-01**, can detect contamination with PCR products diluted 10^{-7} .

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2009-May-29

Approved by:

Quality Control, Supervisor

Lot No.: **26F**

Lot-specific Information

www.olerup.com

Declaration of Conformity

Product name: *Olerup* SSP® HLA-A low resolution
Product number: 101.401-48/12
Lot number: 26F

Intended use: HLA-A low resolution histo-compatibility testing

Manufacturer: *Olerup* SSP AB
Hasselstigen 1
SE-133 33 Saltsjöbaden, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2000 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Saltsjöbaden, Sweden
2009-May-29

Olle Olerup
Managing Director

Lot No.: **26F**

Lot-specific Information

www.olerup.com

ADDRESSES:

Manufacturer:

Olerup SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: info-ssp@olerup.com

Web page: <http://www.olerup.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: support-at@olerup.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: info.us@olerup.com

Web page: <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.