

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information
Olerup SSP® HLA-A*23

Product number: 101.421-06 – including *Taq* polymerase
 101.421-06u – without *Taq* polymerase
Lot number: 0M5
Expiry date: 2025-04-01
Number of tests: 6
Number of wells per test: 37+1
Storage - pre-aliquoted primers: dark at -20°C
 - PCR Master Mix: -20°C
 - Adhesive PCR seals RT
 - Product Insert RT

This Product Description is only valid for Lot No. 0M5.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP®
 HLA-A*23 LOT (7K4)**

- The product documentation has been updated for new alleles of IMGT 3.43.0
- The kit resolution focuses on common and well documented (CWD) alleles¹.

¹As described in section Uniquely Identified Alleles.

The HLA-A*23 specificity and interpretation tables have been updated for the HLA alleles described since the previous *Olerup SSP®* HLA-A*23 lot was made (**Lot No. 7K4**).

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
10	Removed	-	5'-primer removed for improved HLA-specific amplification.
12	Removed	Removed	Primer pair removed for improved HLA-specific amplification.
14	Modified	-	5'-primer modified for improved HLA-specific amplification.
36	Added	Added	Negative control moved to well 38. Primer pair added for the A*23:104 allele.
37	New	New	New primer pair added for the A*23:105 allele.
38	-	-	Negative control added from vial 36.

¹S. J. Mack, P. Cano, J. A. Hollenbach et al.
 Common and well-documented HLA alleles: 2012 update to the CWD catalogue. *Tissue Antigens*, 2013, 81, 194–203



0197

For *In Vitro* Diagnostic Use
 MA123 v01 SSP PI Template
 Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

Well **38** contains Negative Control primer pairs, that will amplify a majority of the *Olerup* SSP[®] HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 200 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information
PRODUCT DESCRIPTION

HLA-A*23 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the A*23:01 to A*23:110 alleles.

PLATE LAYOUT

Each test consists of 38 PCR reactions in a 48 well cut PCR plate. Wells 39 to 48 are empty.

Note: This lot was manufactured using white plastic trays

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	32
33	34	35	36	37	NC	empty	empty
empty	empty	empty	empty	empty	empty	empty	empty

The 48 well cut PCR plate is marked with ‘HLA-A*23’ in silver/gray ink.

Well No. 1 is marked with the Lot Number ‘0M5’.

Wells 1 to 37 – HLA-A*23 high resolution primers.

Well 38 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded. The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 48 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between HLA-A alleles non-HLA-A*23 alleles will be amplified by some primer mixes. For further details see Specificity Table.

UNIQUELY IDENTIFIED ALLELES

All the HLA-A*23 alleles, i.e. **A*23:01 to A*23:110 alleles**, recognized by the HLA Nomenclature Committee in January 2021^{1,2} will be amplified by the primers in the HLA-A*23 subtyping kit.

The HLA-A*23 kit enables separation of the confirmed HLA-A*23 alleles as listed in the IMGT/HLA database 3.30.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-A*23 alleles is listed below.



0197

For *In Vitro* Diagnostic Use
 MA123 v01 SSP PI Template
 Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

The HLA-A*23 kit also enables identification of many null and alternatively expressed alleles.

The following HLA-A*23 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix	Alleles	Primer mix
A*23:08N, 23:22	7	A*23:16, 23:29	20
A*23:09, 23:26	8	A*23:38N, 23:56	30
A*23:10, 23:23	10	A*23:41, 23:42	25

¹HLA-A alleles listed on the IMGT/HLA web page 2021-January-18, release 3.43.0, www.ebi.ac.uk/imgt/hla.p

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

ALLELE CONFIRMATION STATUS

Allele	Status ¹	Allele	Status ¹	Allele	Status ¹	Allele	Status ¹
A*23:01:01:01	Confirmed	A*23:08N	Confirmed	A*23:37:01	Unconfirmed	A*23:66	Unconfirmed
A*23:01:01:02	Confirmed	A*23:09	Unconfirmed	A*23:37:02	Unconfirmed	A*23:67	Unconfirmed
A*23:01:01:03	Unconfirmed	A*23:10	Unconfirmed	A*23:38N	Unconfirmed	A*23:68	Confirmed
A*23:01:01:04	Unconfirmed	A*23:11N	Confirmed	A*23:39	Unconfirmed	A*23:70	Unconfirmed
A*23:01:02	Confirmed	A*23:12	Confirmed	A*23:40	Unconfirmed	A*23:71	Unconfirmed
A*23:01:03	Confirmed	A*23:13	Unconfirmed	A*23:41	Unconfirmed	A*23:72	Unconfirmed
A*23:01:04	Confirmed	A*23:14:01	Unconfirmed	A*23:42	Unconfirmed	A*23:73	Unconfirmed
A*23:01:05	Unconfirmed	A*23:14:02	Unconfirmed	A*23:43	Confirmed	A*23:74	Unconfirmed
A*23:01:06	Confirmed	A*23:15	Confirmed	A*23:44	Unconfirmed	A*23:75	Unconfirmed
A*23:01:07	Unconfirmed	A*23:16	Unconfirmed	A*23:45	Unconfirmed	A*23:76	Unconfirmed
A*23:01:08	Unconfirmed	A*23:17	Confirmed	A*23:46	Unconfirmed	A*23:77	Unconfirmed
A*23:01:09	Confirmed	A*23:18	Confirmed	A*23:47	Confirmed	A*23:78	Unconfirmed
A*23:01:10	Confirmed	A*23:19N	Confirmed	A*23:48	Confirmed	A*23:79	Unconfirmed
A*23:01:11	Unconfirmed	A*23:20	Unconfirmed	A*23:49	Confirmed	A*23:80	Unconfirmed
A*23:01:12	Unconfirmed	A*23:21	Confirmed	A*23:50	Unconfirmed	A*23:81	Unconfirmed
A*23:01:13	Unconfirmed	A*23:22	Unconfirmed	A*23:51	Unconfirmed	A*23:82	Unconfirmed
A*23:01:14	Unconfirmed	A*23:23	Unconfirmed	A*23:52	Unconfirmed	A*23:83	Unconfirmed
A*23:01:15	Unconfirmed	A*23:24	Confirmed	A*23:53	Confirmed		
A*23:01:16	Unconfirmed	A*23:25	Unconfirmed	A*23:54	Unconfirmed		
A*23:01:17	Unconfirmed	A*23:26	Unconfirmed	A*23:55	Unconfirmed		
A*23:01:18	Unconfirmed	A*23:27	Confirmed	A*23:56	Confirmed		
A*23:01:19	Unconfirmed	A*23:28	Confirmed	A*23:57	Unconfirmed		
A*23:01:20	Unconfirmed	A*23:29	Unconfirmed	A*23:58	Unconfirmed		
A*23:02	Unconfirmed	A*23:30	Confirmed	A*23:59	Confirmed		
A*23:03:01	Unconfirmed	A*23:31	Confirmed	A*23:60	Unconfirmed		
A*23:03:02	Unconfirmed	A*23:32	Unconfirmed	A*23:61	Unconfirmed		
A*23:04	Confirmed	A*23:33	Unconfirmed	A*23:62	Unconfirmed		
A*23:05	Confirmed	A*23:34	Unconfirmed	A*23:63	Unconfirmed		
A*23:06	Confirmed	A*23:35	Unconfirmed	A*23:64	Unconfirmed		
A*23:07N	Unconfirmed	A*23:36	Unconfirmed	A*23:65	Confirmed		

¹Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2017-October-27, release 3.30.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-A*23 homo- and heterozygotes is available upon request.



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information
SPECIFICITY TABLE
HLA-A*23 SSP subtyping

Specificities and sizes of the PCR products of the 37+1 primer mixes used for
HLA-A*23 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified HLA-A*23 alleles ³	Other amplified HLA Class I alleles
1	210 bp	800 bp	*23:01:01:01-23:01:32, 23:03:01-23:65, 23:67-23:68, 23:70-23:104, 23:106N-23:110	*02:17:02:01-02:17:04, 02:108, 02:110, 02:268, 02:300, 02:303, 02:617, 02:657, 03:347, 24:13:01-24:13:02, 24:18, 24:24, 24:94, 24:188, 24:207:01-24:207:02, 24:228, 24:315, 24:355, 24:392, 24:515, 29:07, 29:49, 31:29, 33:187, B*18:27
2	160 bp	1070 bp	*23:01:01:01-23:02, 23:05-23:42, 23:44-23:68, 23:71-23:78, 23:80-23:82, 23:84N-23:107Q, 23:109	*01:88:02-01:88:03, 01:236, 02:19, 02:36-02:37, 02:54, 02:255, 02:417, 02:723, 02:804, 02:848, 02:883, 02:949, 11:209, 24:02:01:01-24:02:41, 24:02:43-24:02:69, 24:02:71-24:02:141, 24:04-24:09N, 24:11N, 24:13:01-24:15, 24:17:01:01-24:17:01:02, 24:19-24:20:01:02, 24:24-24:32, 24:34-24:64, 24:66-24:74:02, 24:76-24:93, 24:95-24:109, 24:111-24:124, 24:126-24:137, 24:139-24:157, 24:159-24:166, 24:168-24:203, 24:205-24:206, 24:209, 24:212-24:218, 24:220-24:282, 24:284-24:288, 24:290-24:298, 24:301-24:314, 24:316-24:338, 24:341-24:348, 24:350-24:356, 24:358-24:373, 24:375-24:386, 24:388N-24:389N, 24:391-24:440, 24:442-24:458, 24:460:01:01-24:460:01:02, 24:462-24:470, 24:472-24:482, 24:484-24:487, 24:490-24:493, 24:496-24:498, 24:500-24:505, 24:507-24:513Q, 24:516-24:519, 33:19, 68:26, 68:65, 68:115, 68:131
3⁴	125 bp	800 bp	*23:03:01-23:03:02:02, 23:70, 23:83	*01:301Q, 03:347, 11:139, 24:21:01, 24:21:03, 24:208:01-24:208:02:02, 24:489, 24:499, 29:07, 29:49, 31:29, 33:187, B*18:27
	270 bp		*23:25	*24:313:02
4	200 bp	1070 bp	*23:28	*24:77
5	230 bp	800 bp	*23:06	*31:36
6⁵	470 bp	1070 bp	*23:07N	*01:04:01:01N-01:04:01:02N, 03:21N, 11:21N, 24:11N, 30:132N
7⁴	95 bp 145 bp 205 bp	800 bp	*23:08N *23:39 *23:22	*02:82N, 02:773N *30:22, 30:104
8⁶	140 bp 170 bp 215 bp	1070 bp	*23:26, 23:39 *23:48, 24:403 *23:09	*30:22, 30:104 *01:02:01-01:02:02, 01:20, 01:188, 01:190, 01:307, 24:129
9	235 bp	1070 bp	*23:05	*24:25, 24:308



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

				24:292-24:295, 24:297-24:298, 24:300-24:303N, 24:305-24:314, 24:317-24:323N, 24:325-24:326, 24:328-24:338, 24:340-24:354, 24:356-24:380, 24:382-24:411, 24:413-24:457, 24:459-24:470, 24:472-24:477, 24:479Q-24:480, 24:482-24:514N, 24:516-24:519, 26:16, 30:163, 33:19, 33:119, 68:45, 68:117
17⁴	125 bp 225 bp	1070 bp	*23:33 *23:13	*24:386, 30:102 *03:72, 11:88, 24:07:01:01-24:07:04, 24:19, 24:24, 24:131, 24:288, 24:290, 24:294Q, 24:339, 24:347:01-24:347:02, 24:387, 24:406, 24:453, 24:457, 24:477, 24:489, 24:510, 29:37, 29:56, 30:01:01:01-30:01:17, 30:11:01-30:11:02, 30:14L-30:20, 30:23-30:26, 30:30-30:31, 30:35-30:43, 30:48-30:49, 30:52-30:54, 30:56, 30:58-30:60, 30:62-30:63, 30:65, 30:72-30:75, 30:78N, 30:81-30:83, 30:86-30:89, 30:91-30:98, 30:102, 30:104, 30:106, 30:109-30:116, 30:118, 30:120, 30:122-30:123N, 30:125-30:126, 30:128-30:132N, 30:134-30:138, 30:140-30:143, 30:145N, 30:147-30:148, 30:154, 30:159, 30:161, 30:163-30:165, 30:167-30:168, 30:170-30:173, 30:176, 30:178N-30:181, 68:45, 68:117
18⁴	120 bp	1070 bp	*23:01:01:01-23:01:32, 23:02 ^w , 23:04-23:23, 23:25-23:33, 23:35-23:56, 23:58-23:65, 23:67-23:68, 23:71-23:82, 23:84N-23:104, 23:106N-23:110	*02:01:22, 02:06:04, 02:40:01-02:40:02, 02:51, 02:99:02, 02:130, 02:171:01, 24:13:02, 24:24, 24:207:02, 24:315, 24:392, 29:101:01, 31:67-31:68, 32:28, 32:66, 33:32:01, 68:08:01:01-68:08:01:02, 68:51 ^w , 68:168, B*40:323:02, B*51:136, B*53:60, C*07:204:01, C*07:482
19^{4,7}	120 bp 195 bp	1070 bp	*23:15 *23:46	*24:66, 24:308, 30:75
20⁵	130 bp 195 bp 230 bp 270 bp	800 bp	*23:29 *23:46 *23:16 *23:32	*24:128 *24:66, 24:308, 30:75 *24:297 B*51:293
21⁴	90 bp	800 bp	*23:01:01:01-23:02, 23:03:01 [?] , 23:03:02:01-23:04, 23:06-23:07N, 23:08N [?] , 23:09, 23:10 [?] , 23:11N-23:13, 23:14:01 [?] -23:14:02 [?] , 23:15, 23:16 [?] , 23:18-23:19N, 23:20 [?] -23:25 [?] , 23:26, 23:27 [?] -23:29 [?] , 23:30, 23:31 [?] -23:37:02 [?] , 23:38N, 23:39 [?] -23:55 [?] , 23:56, 23:57 [?] -23:64 [?] , 23:65, 23:66 [?] -23:67 [?] , 23:68, 23:70 [?] -23:75 [?] ,	*24:463, 24:515



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

Lot No.: 0M5		Lot-specific Information		
			23:76-23:80, 23:81 [?] - 23:83 [?] , 23:85-23:95, 23:97-23:103N, 23:105-23:106N, 23:108N, 23:109 [?] , 23:110	
22 ^{4,5}	90 bp	1070 bp	*23:31, 23:45, 23:53, 23:70	*01:301Q, 03:347, 11:139, 24:17:01:01- 24:17:01:02, 24:41, 24:62, 24:106, 24:208:01-24:208:02:02, 24:330, 24:488, 24:499, 24:503, 29:07, 29:49, 31:29, 33:187
	290 bp		*23:19N	
23 ^{4,7}	80 bp	1070 bp	*23:45	*02:41, 02:80, 02:117, 02:289:01, 02:304, 02:454, 02:872, 24:62, 26:10, 31:67-31:68, 32:28, 32:66, 33:32:01
	170 bp		*23:20, 23:58	
24 ⁴	105 bp 180 bp 240 bp	800 bp	*23:49 *23:21 *23:40	*24:444, 30:06, C*04:361 C*04:344 *24:302
25 ⁴	95 bp 205 bp	800 bp	*23:41 *23:42	*02:221, 29:15, 30:162, 31:78
26 ⁴	80 bp	1070 bp	*23:36	*24:32, 32:05, 32:79
27 ^{4,5}	85 bp	1070 bp	*23:53, 23:70	*01:301Q, 03:347, 11:139, 24:17:01:01- 24:17:01:02, 24:41, 24:208:01- 24:208:02:02, 24:488, 24:499, 24:503, 29:07, 29:49
	200 bp		*23:34, 23:57	*02:17:02:01-02:17:04, 02:108, 02:110, 02:268, 02:300, 02:303, 02:617, 02:657, 24:13:01, 24:94, 24:188, 24:207:01, 24:228, 24:515
	285 bp		*23:35	
28	385 bp	800 bp	*23:03:01 [?] , 23:05, 23:08N [?] , 23:10 [?] , 23:14:01 [?] -23:14:02 [?] , 23:16 [?] , 23:17:01:01- 23:17:02, 23:20 [?] -23:25 [?] , 23:27 [?] -23:29 [?] , 23:31 [?] - 23:37:02 [?] , 23:39 [?] -23:55 [?] , 23:57 [?] -23:64 [?] , 23:66 [?] - 23:67 [?] , 23:70 [?] -23:75 [?] , 23:81 [?] -23:83 [?] , 23:84N, 23:96, 23:104, 23:107Q	*02:19, 02:804, 24:02:01:01-24:02:09, 24:02:11-24:02:96, 24:02:97 [?] , 24:02:98- 24:02:122, 24:02:124-24:05:02, 24:06 [?] , 24:07:01:01-24:10:01:02, 24:10:02 [?] , 24:11N, 24:13:01 [?] , 24:13:02- 24:14:01:04, 24:15 [?] , 24:17:01:01-24:18, 24:19 [?] , 24:20:01:01-24:20:01:02, 24:21:01 [?] -24:21:02 [?] , 24:21:03-24:23, 24:24 [?] , 24:25, 24:26 [?] -24:27 [?] , 24:28- 24:32, 24:34 [?] , 24:35-24:37, 24:38 [?] - 24:40N [?] , 24:41, 24:42 [?] -24:43 [?] , 24:44, 24:45N [?] -24:47 [?] , 24:48N, 24:49 [?] , 24:50, 24:51 [?] -24:52 [?] , 24:53-24:54, 24:55 [?] , 24:56:01-24:58, 24:59 [?] -24:60N [?] , 24:61, 24:62 [?] -24:64 [?] , 24:66 [?] -24:67 [?] , 24:68, 24:69 [?] -24:70 [?] , 24:71-24:72, 24:73 [?] - 24:78 [?] , 24:79, 24:80 [?] , 24:81, 24:82 [?] - 24:85 [?] , 24:86N, 24:87 [?] -24:89 [?] , 24:90:01N, 24:90:02N [?] -24:94 [?] , 24:95, 24:96 [?] , 24:97, 24:98 [?] -24:105 [?] , 24:106, 24:107 [?] -24:115 [?] , 24:116, 24:117 [?] - 24:119 [?] , 24:120, 24:121 [?] , 24:122, 24:123 [?] -24:127 [?] , 24:128-24:129, 24:130 [?] -24:131 [?] , 24:132N, 24:133 [?] - 24:134 [?] , 24:135:01, 24:135:02 [?] , 24:136, 24:137 [?] -24:141 [?] , 24:142:01, 24:142:02 [?] ,



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

				24:143, 24:144 [?] -24:147 [?] , 24:148, 24:149 [?] -24:151 [?] , 24:152, 24:153 [?] -24:162 [?] , 24:163N, 24:164 [?] -24:172:01 [?] , 24:172:02, 24:173 [?] -24:182 [?] , 24:183N, 24:184 [?] -24:193 [?] , 24:194, 24:195 [?] -24:206 [?] , 24:207:01-24:207:02, 24:208:01 [?] , 24:208:02:01-24:208:02:02, 24:209 [?] -24:210 [?] , 24:212 [?] -24:214 [?] , 24:215, 24:216 [?] -24:224 [?] , 24:225:01-24:225:02, 24:226:01 [?] , 24:226:02, 24:227 [?] -24:230 [?] , 24:231, 24:232N [?] , 24:233-24:235, 24:236 [?] -24:237 [?] , 24:238, 24:239 [?] -24:247 [?] , 24:248-24:252N, 24:253 [?] , 24:254-24:256, 24:257 [?] -24:258 [?] , 24:259-24:273, 24:274 [?] -24:277 [?] , 24:278N, 24:279 [?] , 24:280, 24:281 [?] -24:285 [?] , 24:286, 24:287 [?] -24:292 [?] , 24:293-24:294Q, 24:295 [?] -24:301 [?] , 24:302, 24:303N [?] -24:308 [?] , 24:309, 24:310:01 [?] , 24:310:02, 24:311 [?] -24:312N [?] , 24:313:01, 24:313:02 [?] , 24:314, 24:315 [?] -24:320 [?] , 24:321, 24:322 [?] -24:330 [?] , 24:331, 24:332 [?] -24:340 [?] , 24:341, 24:342 [?] -24:351 [?] , 24:352-24:353, 24:354 [?] , 24:355-24:359N, 24:360 [?] , 24:361-24:364, 24:365 [?] -24:366 [?] , 24:367-24:373, 24:374 [?] -24:386 [?] , 24:388N [?] , 24:389N, 24:390 [?] -24:393 [?] , 24:394, 24:395 [?] , 24:396N-24:412, 24:413 [?] , 24:414, 24:415 [?] -24:416 [?] , 24:417-24:420, 24:421 [?] , 24:422-24:431, 24:432 [?] , 24:433N-24:444, 24:445N [?] -24:446 [?] , 24:447Q-24:451, 24:452 [?] , 24:453-24:459, 24:461-24:462, 24:464, 24:466-24:470, 24:472-24:514N, 24:516-24:519, 33:19 [?]
29	170 bp 200 bp 240 bp 290 bp	1070 bp	*23:48 *23:37:01-23:37:02 *23:40 *23:47	*24:403 *24:273, 24:506 *24:302 *24:234, 24:339, 24:347:02
30⁴	110 bp 245 bp	1070 bp	*23:56 *23:38N	*24:61
31	185 bp	1070 bp	*23:44	
32⁴	85 bp	1070 bp	*23:59	*01:203, 02:671, 11:166, 30:56, 31:85, 32:72, 80:01:01:01-80:08N
33	165 bp	1070 bp	*23:65	*24:298
34⁴	90 bp 170 bp	1070 bp	*23:18 *23:58	*24:267
35⁴	105 bp 260 bp	800 bp	*23:49 *23:27	*24:444, 30:06, C*04:361 *24:255
36	200 bp	1070 bp	*23:104	
37⁵	195 bp	1070 bp	*23:01:01:01-23:33, 23:35-23:56, 23:58-23:65, 23:67-23:68, 23:70-23:98,	*24:24, 24:71, 24:315, 24:392, 29:07, 29:49, 31:29, 33:187, B*18:27



0197

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

	23:100-23:103N, 23:105-23:110
38⁸	Negative Control

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-A*23 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings. Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁵Primer mixes 6, 20, 22, 27 and 37 may give rise to a lower yield of HLA-specific PCR product than the other A*23 primer mixes.

⁶Primer mix 8 may have tendency of unspecific amplification.

⁷Primer mixes 19 and 23 have tendencies to giving rise to primer oligomer formation.

⁸Primer mix 38 contains a negative control, which will amplify a majority of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 200 base pairs.

Abbreviations

‘w’, may be weakly amplified.

‘?’, nucleotide sequence information not available for the primer matching sequence.



0197

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information
PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec. PCR product	210	160	125	200	230	470	95	140	235	135	200	190
			270				145	170		240	270	
							205	215				
Length of int. pos. control ¹	800	1070	800	1070	800	1070	800	1070	1070	800	800	1070
5'-primer(s) ²	368	453	368	98	144	3 rd I	98	98	28	368	90	144
	5'-gTT 3'	5'-AAA 3'	5'-gTT 3'	5'-CTC 3'	5'-gCC 3'	5'-ATA 3'	5'-CTC 3'	5'-CTC 3'	5'-TCg 3'	5'-gTT 3'	5'-AgT 3'	5'-gCC 3'
							564				160	
							5'-TgA 3'				5'-ACg 3'	
3'-primer(s) ³	539	570	453	256	331	621	200	193	92	463	317	292
	5'-TCA 3'	5'-CCg 3'	5'-TCg 3'	5'-CTg 3'	5'-CTC 3'	5'-ggg 3'	5'-TCC 3'	5'-CgA 3'	5'-AAC 3'	5'-gCT 3'	5'-ggA 3'	5'-gTg 3'
			595				262	200		559		
			5'-CCg 3'				5'-TgC 3'	5'-TCC 3'		5'-CCg 3'		
							616	227		571		
							5'-CgT 3'	5'-CTg 3'		5'-CCT 3'		
								271				
								5'-CAT 3'				
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Length of spec. PCR product	90	85	210	175	125	120	120	130	90	90	80	105
	210	190	290	205	225		195	195		290	170	180
		245						230				240
								270				
Length of int. pos. control ¹	800	1070	1070	800	1070	1070	1070	800	800	1070	1070	800
5'-primer(s) ²	368	368	368	98	98	453	418	90	920	368	414	98
	5'-gTT 3'	5'-gTT 3'	5'-gTT 3'	5'-CTC 3'	5'-CTC 3'	5'-AAA 3'	5'-AgC 3'	5'-AgT 3'	5'-CCA 3'	5'-gTT 3'	5'-CAg 3'	5'-CTC 3'
				368			493	228			678	
				5'-gTT 3'			5'-CTg 3'	5'-ATg 3'			5'-AgA 3'	
								379				
								5'-ACA 3'				
								418				
								5'-AgC 3'				
3'-primer(s) ³	419	413	538	259	181	524	570	317	968	418	453	163
	5'-CgC 3'	5'-gCC 3'	5'-CAg 3'	5'-gTT 3'	5'-gTA 3'	5'-CAC 3'	5'-CCg 3'	5'-ggA 3'	5'-CAg 3'	5'-gTC 3'	5'-TCT 3'	5'-CgC 3'
	530	518	616	502	282	538		570		619	806	238
	5'-CCA 3'	5'-CCT 3'	5'-CgC 3'	5'-CTT 3'	5'-gAC 3'	5'-CAg 3'		5'-CCg 3'		5'-gTT 3'	5'-CTA 3'	5'-CCT 3'
	539	570		539	282						809	299
	5'-TCC 3'	5'-CAC 3'		5'-TCT 3'	5'-gAC 3'						5'-CAA 3'	5'-TCg 3'
Well No.	13	14	15	16	17	18	19	20	21	22	23	24



0197

For *In Vitro* Diagnostic Use
MA123 v01 SSP PI Template
Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

Well No.	25	26	27	28	29	30	31	32	33	34	35	36	37
Length of spec.	95	80	85	385	170	110	185	85	165	90	105	200	195
PCR product	205		200		200	245				170	260		
			285		240								
					290								
Length of int. pos. control ¹	800	1070	1070	800	1070	1070	1070	1070	1070	1070	800	1070	1070
5'-primer(s) ²	376	265	368	678	98	98	395	176	453	678	98	368	368
	5'-gCT 3'	5'-CAg 3'	5'-gTT 3'	5'-AgA 3'	5'-CTC 3'	5'-CTC 3'	5'-gCA 3'	5'-gCA 3'	5'-AAA 3'	5'-AgA 3'	5'-CTC 3'	5'-gTT 3'	5'-gTT 3'
	484				368								
	5'-ACg 3'				5'-gTT 3'								
3'-primer(s) ³	538	302	413	920	227	167	538	218	577	728	163	526	521
	5'-CAA 3'	5'-ggC 3'	5'-gCC 3'	5'-Tgg 3'	5'-CTg 3'	5'-ACC 3'	5'-CAA 3'	5'-gCg 3'	5'-gCC 3'	5'-CCT 3'	5'-CgC 3'	5'-CCg 3'	5'-ggg 3'
			524		255	303				809	316		
			5'-CAT 3'		5'-TCC 3'	5'-AgT 3'				5'-CAA 3'	5'-gAA 3'		
			534		255								
			5'-CgT 3'		5'-TCT 3'								
			614		299								
			5'-Tgg 3'		5'-TCg 3'								
					616								
					5'-CgC 3'								
Well No.	25	26	27	28	29	30	31	32	33	34	35	36	37

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

101.421-06 – including Taq polymerase, IFU-01
101.421-06u – without Taq polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

CELL LINE VALIDATION SHEET																				
HLA-A*23 SSP subtyping kit ²																				
				Well																
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
				Prod. No.:	202129601	202129602	202129603	202129604	202129605	202129606	202129607	202129608	202129609	202129610	202129611	202129612	202129613	202129614	202129615	202129616
	IHWC cell line ¹	A*	A*																	
1	9001 SA	*24:02		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
2	9280 LK707	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324	*01:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*30:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009 KAS011	*01:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*02:01	*26:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL	*26:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025 DEU	*31:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*26:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*24:02		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
11	9051 PITOUT	*29:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*31:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*24:02		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
16	9037 SWEIG007	*29:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*03:01	*80:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257 32367	*33:03	*74:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*02:17		+	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	W
22	9056 KOSE	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*02:01	*34:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035 JBUSH	*32:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*33:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*02:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*24:10	*29:01	-	-	-	-	-	-	-	-	-	W	-	-	-	+	-	-	+
28	9320 BEL5GB	*02:01	*29:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU	*29:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*30:01	*68:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019 DUCAF	*30:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG	*02:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*31:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*31:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*32:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17	*02:06	*11:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*03:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099 LZL	*02:17		+	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	W
39	9315 CML	*01:01	*03:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199	*02:07	*30:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301	*03:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089	*02:07		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*02:06	*02:07	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*66:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO	*23:01	*24:02	+	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
46	9013 SCHU	*03:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*02:16	*03:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND	*02:01	*11:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-



0197

101.421-06 – including Taq polymerase, IFU-01
101.421-06u – without Taq polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
“Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

CELL LINE VALIDATION SHEET																				
HLA-A*23 SSP subtyping kit ²																				
				Prod. No.:	Well															
					17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32
	IHWC cell line ¹	A*	A*		202129617	202129618	202129619	202129620	202129621	202129622	202129623	202129624	202129625	202129626	202129627	202129628	202129629	202129630	202129631	202129632
1	9001 SA	*24:02			-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
2	9280 LK707	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324	*01:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*30:01			+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009 KAS011	*01:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*02:01	*26:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL	*26:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025 DEU	*31:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*26:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*24:02			-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
11	9051 PITOUT	*29:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*31:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*24:02			-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
16	9037 SWEIG007	*29:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*03:01	*80:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
18	9257 32367	*33:03	*74:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*02:17			-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
22	9056 KOSE	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*02:01	*34:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035 JBUSH	*32:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*33:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*02:05			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*24:10	*29:01		-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
28	9320 BEL5GB	*02:01	*29:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU	*29:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*30:01	*68:02		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019 DUCAF	*30:02			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG	*02:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*31:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*31:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*32:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17	*02:06	*11:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*03:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099 LZL	*02:17			-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
39	9315 CML	*01:01	*03:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199	*02:07	*30:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301	*03:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089	*02:07			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*02:06	*02:07		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*66:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO	*23:01	*24:02		-	+	-	-	+	-	-	-	-	-	-	+	-	-	-	-
46	9013 SCHU	*03:01			-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*02:16	*03:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND	*02:01	*11:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-



0197

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

CELL LINE VALIDATION SHEET									
HLA-A*23 SSP subtyping kit ²									
				Prod. No.:	Well				
					33	34	35	36	37
				202129633	202129634	202129635	202129636	202129637	
	IHWC cell line ¹	A*	A*						
1	9001 SA	*24:02		-	-	-	-	-	
2	9280 LK707	*02:01		-	-	-	-	-	
3	9011 E4181324	*01:01		-	-	-	-	-	
4	9275 GU373	*30:01		-	-	-	-	-	
5	9009 KAS011	*01:01		-	-	-	-	-	
6	9353 SM	*02:01	*26:03	-	-	-	-	-	
7	9020 QBL	*26:01		-	-	-	-	-	
8	9025 DEU	*31:01		-	-	-	-	-	
9	9026 YAR	*26:01		-	-	-	-	-	
10	9107 LKT3	*24:02		-	-	-	-	-	
11	9051 PITOUT	*29:02		-	-	-	-	-	
12	9052 DBB	*02:01		-	-	-	-	-	
13	9004 JESTHOM	*02:01		-	-	-	-	-	
14	9071 OLGA	*31:01		-	-	-	-	-	
15	9075 DKB	*24:02		-	-	-	-	-	
16	9037 SWEIG007	*29:02		-	-	-	-	-	
17	9282 CTM3953540	*03:01	*80:01	-	-	-	-	-	
18	9257 32367	*33:03	*74:01	-	-	-	-	-	
19	9038 BM16	*02:01		-	-	-	-	-	
20	9059 SLE005	*02:01		-	-	-	-	-	
21	9064 AMALA	*02:17		-	-	-	-	-	
22	9056 KOSE	*02:01		-	-	-	-	-	
23	9124 IHL	*02:01	*34:01	-	-	-	-	-	
24	9035 JBUSH	*32:01		-	-	-	-	-	
25	9049 IBW9	*33:01		-	-	-	-	-	
26	9285 WT49	*02:05		-	-	-	-	-	
27	9191 CH1007	*24:10	*29:01	-	-	-	-	-	
28	9320 BEL5GB	*02:01	*29:02	-	-	-	-	-	
29	9050 MOU	*29:02		-	-	-	-	-	
30	9021 RSH	*30:01	*68:02	-	-	-	-	-	
31	9019 DUCAF	*30:02		-	-	-	-	-	
32	9297 HAG	*02:01		-	-	-	-	-	
33	9098 MT14B	*31:01		-	-	-	-	-	
34	9104 DHIF	*31:01		-	-	-	-	-	
35	9302 SSTO	*32:01		-	-	-	-	-	
36	9024 KT17	*02:06	*11:01	-	-	-	-	-	
37	9065 HHKB	*03:01		-	-	-	-	-	
38	9099 LZL	*02:17		-	-	-	-	-	
39	9315 CML	*01:01	*03:01	-	-	-	-	-	
40	9134 WHONP199	*02:07	*30:01	-	-	-	-	-	
41	9055 H0301	*03:01		-	-	-	-	-	
42	9066 TAB089	*02:07		-	-	-	-	-	
43	9076 T7526	*02:06	*02:07	-	-	-	-	-	
44	9057 TEM	*66:01		-	-	-	-	-	
45	9239 SHJO	*23:01	*24:02	-	-	-	-	+	
46	9013 SCHU	*03:01		-	-	-	-	-	
47	9045 TUBO	*02:16	*03:01	-	-	-	-	-	
48	9303 TER-ND	*02:01	*11:01	-	-	-	-	-	



0197

For *In Vitro* Diagnostic Use
 MA123 v01 SSP PI Template
 Date: June 2021, Rev. No: 00

101.421-06 – including *Taq* polymerase, IFU-01
 101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 3 to 5, 9, 11, 13, 15, 19, 20, 23 to 26, 29 to 31 and 33 to 36 were available.

The specificities of the primers in primer solutions 3, 13, 15, 19, 20, 23, 26 and 36 were tested by separately adding one or two additional 5'-primers, respectively one or two additional 3'-primers.

In primer solutions 4, 5, 9, 24, 29, 30 and 33 to 35 it was only possible to test the 5'-primers, the 3'-primers were not possible to test.

In primer solutions 11, 25 and 31 it was only possible to test the 3'-primers, the 5'-primers was not possible to test.

In primer solutions 3, 7, 8, 10, 13 to 15, 17, 22, 23 and 27 one or more 3'-primers were not possible to test, and in primer solutions 19 and 20 one or more 5'-primer were not possible to test. In addition, one or more 3'-and/or 5'-primers in primer solutions 7, 8, 10, 12, 14, 16, 18 and 27 were tested by separately adding one additional 5'-primer and/or one additional 3'-primer.



0197

101.421-06 – including *Taq* polymerase, IFU-01
101.421-06u – without *Taq* polymerase, IFU-02

Visit <https://labproducts.caredx.com> for
 “Instructions for Use” (IFU)

Lot No.: **0M5**

Lot-specific Information

ADDRESSES:

Manufacturer:

CareDx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-508 939 00

Fax: +46-8-717 88 18

E-mail: orders-se@caredx.com

Web page: <https://labproducts.caredx.com/>

Distributed by:

CareDx Lab Solutions Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-653-78171

Fax: 610-344-7989

E-mail: orders-us@caredx.com

Web page: <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx AB**.



0197

For *In Vitro* Diagnostic Use
 MA123 v01 SSP PI Template
 Date: June 2021, Rev. No: 00