

## **Olerup SSP<sup>®</sup> HLA-A\*80**

Product number:	101.434-06 – including <i>Taq</i> polymerase
Lot number:	70M
Expiry date:	2014-March-01
Number of tests:	6
Number of wells per test:	3
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

**This Product Description is only valid for Lot No. 70M.**

### **CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP*<sup>®</sup> HLA-A\*80 LOT**

The HLA-A\*80 primer set as well as the specificity and interpretation tables are unchanged compared to the previous *Olerup SSP*<sup>®</sup> HLA-A\*80 lot (**Lot No. 06K**).

## PRODUCT DESCRIPTION

### HLA-A\*80 SSP subtyping

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the HLA-A\*80:01 to A\*80:02 alleles.

#### PLATE LAYOUT

Each test consists of 3 PCR reactions in an 8 well cut PCR plate. Wells 4 to 8 are empty.

1	2	3	empty	empty	empty	empty	empty
---	---	---	-------	-------	-------	-------	-------

The 8 well cut PCR plate is marked with 'A80' in silver/gray ink.

Well No. 1 is marked with the Lot No. '70M'.

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are heat-sealed with a PCR-compatible foil.

**Please note:** When removing each 8 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

Only the HLA-A\*80 alleles will be amplified by the HLA-A\*80 subtyping kit. Thus, the interpretation of HLA-A\*80 subtypings is not influenced by other HLA genes.

#### UNIQUELY IDENTIFIED ALLELES

The HLA-A\*80 alleles, i.e. **HLA-A\*80:01** and **HLA-A\*80:02**, recognized by the HLA Nomenclature Committee in July 2011<sup>1</sup> will give rise to unique amplification patterns by the primers in the HLA-A\*80 subtyping kit.

<sup>1</sup>HLA-A alleles listed on the IMGT/HLA web page 2011-July-14, release 3.5.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

#### RESOLUTION IN HOMO- AND HETEROZYGOTES

The two HLA-A\*80 alleles can be combined in 3 homozygous and heterozygous combinations. All these genotypes give rise to unique amplification patterns.

## SPECIFICITY TABLE

### HLA-A\*80 SSP subtyping

Specificities and sizes of the PCR products of the 3 primer mixes used for HLA-A\*80 SSP subtyping

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified HLA-A alleles
<b>1</b> <sup>3</sup>	135 bp	<b>800 bp</b>	*80:01-80:02
<b>2</b>	155 bp	1070 bp	*80:01
<b>3</b>	165 bp	1070 bp	*80:02

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-A\*80 SSP typings.

When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective lengths of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-A\*80 SSP subtyping.

In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>Short specific PCR fragments are less intense and not as sharp as longer specific bands.

<b>INTERPRETATION TABLE</b>			
<b>HLA-A*80 SSP typing</b>			
	<b>Well</b>		
	<b>1</b>	<b>2</b>	<b>3</b>
<b>Length of spec.</b>	<b>135</b>	<b>155</b>	<b>165</b>
<b>PCR product</b>			
<b>Length of int.</b>	<b>800</b>	<b>1070</b>	<b>1070</b>
<b>pos. control<sup>1</sup></b>			
<b>5'-primer<sup>2</sup></b>	<b>176</b>	<b>176</b>	<b>176</b>
	5' -gCA 3'	5' -gCA 3'	5' -gCA 3'
<b>3'-primer<sup>3</sup></b>	<b>270</b>	<b>292</b>	<b>299</b>
	5' -ACA 3'	5' -gTT 3'	5' -CCA 3'
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>
<b>HLA-A allele</b>			
<b>*80:01</b>	<b>1</b>	<b>2</b>	
<b>*80:02</b>	<b>1</b>		<b>3</b>
<b>HLA-A allele</b>			
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-A\*80 subtyping. .

<sup>2</sup>The nucleotide position, in the 2<sup>nd</sup> exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The nucleotide position, in the 2<sup>nd</sup> exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

CELL LINE VALIDATION SHEET							
HLA-A*80 SSP kit							
					Well		
					1	2	2
				Lot No.:	201072301	201072302	201072303
	IHC cell line	A*	A*				
1	9001 SA	*24:02			-	-	-
2	9280 LK707	*02:01			-	-	-
3	9011 E4181324	*01:01			-	-	-
4	9275 GU373	*30:01			-	-	-
5	9009 KAS011	*01:01			-	-	-
6	9353 SM	*02:01	*26:03		-	-	-
7	9020 QBL	*26:01			-	-	-
8	9025 DEU	*31:01			-	-	-
9	9026 YAR	*26:01			-	-	-
10	9107 LKT3	*24:02			-	-	-
11	9051 PITOUT	*29:02			-	-	-
12	9052 DBB	*02:01			-	-	-
13	9004 JESTHOM	*02:01			-	-	-
14	9071 OLGA	*31:01			-	-	-
15	9075 DKB	*24:02			-	-	-
16	9037 SWEIG007	*29:02			-	-	-
17	9282 CTM3953540	*03:01	*80:01		+	+	-
18	9257 32367	*33:03	*74:01		-	-	-
19	9038 BM16	*02:01			-	-	-
20	9059 SLE005	*02:01			-	-	-
21	9064 AMALA	*02:17			-	-	-
22	9056 KOSE	*02:01			-	-	-
23	9124 IHL	*02:01	*34:01		-	-	-
24	9035 JBUSH	*32:01			-	-	-
25	9049 IBW9	*33:01			-	-	-
26	9285 WT49	*02:05			-	-	-
27	9191 CH1007	*24:10	*29:01		-	-	-
28	9320 BEL5GB	*02:01	*29:02		-	-	-
29	9050 MOU	*29:02			-	-	-
30	9021 RSH	*30:01	*68:02		-	-	-
31	9019 DUCAF	*30:02			-	-	-
32	9297 HAG	*02:01			-	-	-
33	9098 MT14B	*31:01			-	-	-
34	9104 DHIF	*31:01			-	-	-
35	9302 SSTO	*32:01			-	-	-
36	9024 KT17	*02:06	*11:01		-	-	-
37	9065 HHKB	*03:01			-	-	-
38	9099 LZL	*02:17			-	-	-
39	9315 CML	*01:01	*03:01		-	-	-
40	9134 WHONP199	*02:07	*30:01		-	-	-
41	9055 H0301	*03:01			-	-	-
42	9066 TAB089	*02:07			-	-	-
43	9076 T7526	*02:06	*02:07		-	-	-
44	9057 TEM	*66:01			-	-	-
45	9239 SHJO	*23:01	*24:02		-	-	-
46	9013 SCHU	*03:01			-	-	-
47	9045 TUBO	*02:16	*03:01		-	-	-
48	9303 TER-ND	*02:01	*11:01		-	-	-

## CERTIFICATE OF ANALYSIS

### **Olerup SSP® HLA-A\*80 SSP**

Product number: 101.434-06 – including *Taq* polymerase  
Lot number: 70M  
Expiry date: 2014-March-01  
Number of tests: 6  
Number of wells per test: 3

#### Well specifications:

Well No.	Production No.
1	2010-723-01
2	2010-723-02
3	2010-723-03

The specificity of each primer solution of the kit has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solution 3 were available. The specificities of the primers in primer solution 3 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer.

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2011-October-06

**Approved by:**

**Quality Control, Supervisor**

Lot No.: **70M**

Lot-specific Information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## Declaration of Conformity

**Product name:** *Olerup* SSP® HLA-A\*80  
**Product number:** 101.434-06  
**Lot number:** 70M

**Intended use:** HLA-A\*80 high resolution histocompatibility testing

**Manufacturer:** *Olerup* SSP AB  
Franzengatan 5  
SE-112 51 Stockholm, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Stockholm, Sweden  
2011-October-06

Ann-Cathrin Jareman  
Head of QA and Regulatory Affairs

Lot No.: **70M**

Lot-specific Information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup-ssp.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.