

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information  
**Olerup SSP® HLA-B\*49**

[www.olerup-ssp.com](http://www.olerup-ssp.com)

Product number:	101.547-06 – including <i>Taq</i> polymerase 101.547-06u – without <i>Taq</i> polymerase
Lot number:	88V
Expiry date:	2016-December-01
Number of tests:	6
Number of wells per test:	18+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

**This Product Description is only valid for Lot No. 88V.**

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP®  
HLA-B\*49 LOT (56S).**

The HLA-B\*49 kit is updated for new alleles to enable separation of:

- Confirmed<sup>1</sup> alleles as listed in the IMGT/HLA database
- Polymorphisms in exons outside of the region encoding the peptide binding domain
- Null and Alternatively expressed alleles

A well containing Negative Control primer pairs has been added.

The format of the Product Insert and Worksheet have been changed.

Three wells have been added to HLA-B\*49, wells **17 to 19**.

<sup>1</sup>As described in section Uniquely Identified Alleles.

The HLA-B\*49 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP®* HLA-B\*49 lot was made (**Lot No. 56S**).

As of lot series V, the Specificity Table is included in the lot-specific Product Insert, and the Interpretation Table is included in the Worksheet.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
17	New	New	New primer pair for the B*49:23 allele.
18	New	New	New primer pair for the B*49:27 allele.
19	New	New	Negative Control

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

Well **19** contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup* SSP® HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.

The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
<b>5'-primer<sup>1</sup></b>	<b>164</b>	<b>340</b>	<b>440</b>	<b>45</b>	<b>45</b>	<b>43</b>	<b>36</b>
	5'-CAC <sup>3'</sup>	5'-Agg <sup>3'</sup>	5'-TTA <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-TAC <sup>3'</sup>
							<b>36</b>
							5'-TAT <sup>3'</sup>
<b>3'-primer<sup>2</sup></b>	<b>231</b>	<b>2<sup>nd</sup> I</b>	<b>507</b>	<b>59</b>	<b>58</b>	<b>57</b>	<b>47</b>
	5'-TgC <sup>3'</sup>	5'-AAA <sup>3'</sup>	5'-TTg <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ggC <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ACA <sup>3'</sup>
							<b>48</b>
							5'-gCA <sup>3'</sup>
							<b>48</b>
							5'-gCC <sup>3'</sup>
							<b>52</b>
							5'-TgT <sup>3'</sup>
<b>A*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>B*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>C*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>DRB1</b>				<b>+</b>	<b>+</b>		
<b>DRB3</b>				<b>+</b>	<b>+</b>		
<b>DRB5</b>				<b>+</b>			
<b>DQB1</b>					<b>+</b>		
<b>DPB1</b>						<b>+</b>	
<b>DQA1</b>							<b>+</b>

<sup>1</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>2</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon or the 2<sup>nd</sup> intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## PRODUCT DESCRIPTION

### HLA-B\*49 SSP typing

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the B\*49:01 to B\*49:31 alleles.

#### PLATE LAYOUT

Each HLA-B\*49 test consists of 19 PCR reactions in a 24 well cut PCR plate. Wells 20 to 24 are empty.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	NC	empty	empty	empty	empty	empty

The 24 well PCR plate is marked with 'HLA-B\*49' in silver/gray ink.

Well No. 1 is marked with the Lot No. '88V'.

Wells 1 to 18– HLA-B\*49 high resolution primers.

Well 19 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are heat-sealed with a PCR-compatible foil.

**Please note:** When removing each 24 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

Due to the sharing of sequence motifs between HLA-B alleles, non-HLA-B\*49 alleles will be amplified by primer mixes 1 to 14 and 16. In addition, a few HLA-C alleles will be amplified by primer mixes 3 and 16.

For further details see Specificity Table.

#### UNIQUELY IDENTIFIED ALLELES

All the HLA-B\*49 alleles, i.e. **B\*49:01 to B\*49:31**, recognized by the HLA Nomenclature Committee in April 2014<sup>1,2</sup> will be amplified by the primers in the HLA-B\*49 SSP kit<sup>3</sup>.

The HLA-B\*49 kit enables separation of the confirmed HLA-B\*49 alleles as listed in the IMGT/HLA database. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-B\*49 alleles is listed below.

The HLA-B\*49 kit also enables identification of polymorphisms in exons outside of the region encoding the peptide binding domain and of null and alternatively expressed alleles.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

The HLA-B\*49 subtyping kit cannot distinguish the silent mutations in the B\*49:01:01 to B\*49:01:06 alleles.

The following HLA-B\*49 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix
B*49:07, B*49:21	9
B*49:08, B*49:16	10
B*49:13, B*49:14	15

<sup>1</sup>HLA-B alleles listed on the IMGT/HLA web page 2014-April-14, release 3.16.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

<sup>2</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

<sup>3</sup>The HLA-B\*49 primer set cannot separate the B\*49:03 and the B\*14:37 alleles. These alleles can be distinguished by the HLA-B low resolution kit.

## ALLELE CONFIRMATION STATUS

Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>
<b>B*49:01:01</b>	<b>Confirmed</b>	B*49:16	Unconfirmed
B*49:01:02	Unconfirmed	<b>B*49:17</b>	<b>Confirmed</b>
B*49:01:03	Unconfirmed	<b>B*49:18</b>	<b>Confirmed</b>
B*49:01:04	Unconfirmed	<b>B*49:19N</b>	<b>Confirmed</b>
B*49:01:05	Unconfirmed	B*49:20	Unconfirmed
B*49:01:06	Unconfirmed	<b>B*49:21</b>	<b>Confirmed</b>
<b>B*49:02</b>	<b>Confirmed</b>	B*49:22	Unconfirmed
B*49:03	Unconfirmed	<b>B*49:23</b>	<b>Confirmed</b>
B*49:04	Unconfirmed	B*49:24	Unconfirmed
B*49:05	Unconfirmed	B*49:25	Unconfirmed
B*49:06	Unconfirmed	B*49:26	Unconfirmed
B*49:07	Unconfirmed	<b>B*49:27</b>	<b>Confirmed</b>
B*49:08	Unconfirmed	B*49:28	Unconfirmed
<b>B*49:09</b>	<b>Confirmed</b>	B*49:29	Unconfirmed
B*49:10	Unconfirmed	B*49:30	Unconfirmed
<b>B*49:11</b>	<b>Confirmed</b>	B*49:31	Unconfirmed
B*49:12	Unconfirmed		
B*49:13	Unconfirmed		
B*49:14	Unconfirmed		
<b>B*49:15</b>	<b>Confirmed</b>		

<sup>1</sup>Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2014-April-14, release 3.16.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

## RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-B\*49 homo- and heterozygotes is available upon request.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## SPECIFICITY TABLE

### HLA-B\*49 SSP subtyping

Specificities and sizes of the PCR products of the 18+1 primer mixes used for HLA-B\*49 SSP subtyping

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified HLA-B*49 alleles <sup>3</sup>	Other amplified HLA-B alleles <sup>4</sup>
1	145 bp	800 bp	*49:01:01-49:01:06, 49:04-49:25, 49:27-49:31	*40:13, 40:19, 40:109, 40:117, 44:18, 44:25, 44:50:01, 44:95, 51:112, 53:30, 57:45, 57:51, 57:69
2	145 bp 190 bp 255 bp	800 bp	*49:06 *49:26 *49:02	*58:13 *39:31 *13:23, 13:55, 18:09, 27:01, 37:10, 40:47, 40:96, 40:157, 40:201, 44:15, 44:55, 44:103, 44:188
3 <sup>5</sup>	115 bp 145 bp	800 bp	*49:22 *49:02	*15:33, 15:248, 51:126, 57:54, <b>C*03:87, C*05:27, C*05:39</b> *40:47, 40:96, 40:110, 40:157, 40:201, 44:02:01:01-44:02:12, 44:02:14-44:02:18, 44:02:20-44:05:02, 44:05:04, 44:07, 44:10, 44:12-44:17, 44:19N-44:24, 44:26-44:39, 44:41:01-44:43:02, 44:45, 44:47-44:49, 44:51-44:56N, 44:58N-44:59:02, 44:61N-44:74, 44:76-44:82, 44:84:01-44:89, 44:91-44:94, 44:96-44:128, 44:132-44:133, 44:135-44:151, 44:153-44:155, 44:157-44:169, 44:171N-44:193, 47:04
4	170 bp	1070 bp	*49:03	*15:32:02, 18:04, 40:01:11, 40:58, 44:02:10, 44:34:01, 51:07:01, 52:01:02, 52:01:04, 52:01:09, 52:02:01, 52:03, 52:09, 52:10:03, 52:21, 52:31:02
5 <sup>5</sup>	205 bp 105 bp	1070 bp	*49:09 *49:01:01-49:03, 49:06-49:17, 49:19N-49:31	*13:15, 14:37, 51:62, 51:106, 52:25, 56:34 *07:78, 13:02:01-13:02:14, 13:02:16-13:03, 13:08-13:09, 13:14-13:16, 13:18-13:19, 13:27, 13:30-13:34, 13:37-13:38, 13:40-13:42, 13:44-13:45, 13:48-13:49N, 13:53-13:56N, 13:58, 13:65-13:70, 13:74-13:75, 14:37, 15:42, 15:303, 35:60, 44:15, 44:18, 45:01:01-45:01:02, 45:03-45:08, 45:10-45:14, 46:11, 46:18, 50:01:01-50:02, 50:04-50:08, 50:10-50:11, 50:13, 50:15, 50:18-50:19, 50:31-50:32, 50:34-50:38, 51:15, 51:157, 52:25, 54:01:01-54:03, 54:05N, 54:07-54:08N, 54:10, 54:12-54:13, 54:16-54:30, 54:32, 55:01:01-55:01:06, 55:01:08-55:03, 55:05, 55:07, 55:09-55:12, 55:15-55:16, 55:18-55:19, 55:21-55:22, 55:24-55:26, 55:29-55:31, 55:33-55:38, 55:40-55:41, 55:43, 55:45-55:48, 55:50, 55:52, 55:54-55:57, 55:59-55:60, 55:62-55:64, 56:01:01-56:01:07, 56:07-56:08, 56:13-56:14, 56:16-56:17, 56:19N-56:20:02, 56:23-56:30, 56:33-56:43, 59:01:01:01-59:01:01:02, 59:04-59:08
6 <sup>6</sup>	160 bp	1070 bp	*49:01:01-49:03, 49:06-49:17, 49:19N-49:31	*13:01:01-13:03, 13:06-13:09, 13:11-13:12:02, 13:14-13:17, 13:19-13:20, 13:22:01-13:23, 13:25, 13:27-13:30, 13:32-13:34, 13:36-13:40,

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

Lot No.: <b>88V</b>		Lot-specific information		<a href="http://www.olerup-ssp.com">www.olerup-ssp.com</a>
				13:42-13:45, 13:47-13:50, 13:52-13:58, 13:60-13:61, 13:63N-13:70, 13:74-13:78, 14:37, 15:42, 15:86, 15:224, 35:60, 39:17, 40:48, 40:71, 44:10, 46:11, 46:18, 50:01:01-50:02, 50:04-50:08, 50:10-50:13, 50:15-50:19, 50:31-50:32, 50:34-50:38, 51:15, 51:62, 51:106, 51:157, 52:25, 54:01:01-54:03, 54:05N, 54:07-54:08N, 54:12-54:13, 54:17-54:19, 54:21-54:32, 55:01:01-55:03, 55:05, 55:07, 55:10-55:12, 55:15-55:16, 55:18-55:19, 55:22, 55:24-55:26, 55:29-55:31, 55:33-55:36, 55:38-55:45, 55:47-55:48, 55:50, 55:53-55:55N, 55:57, 55:59-55:64, 56:01:01-56:01:04, 56:01:06-56:02, 56:04, 56:07-56:08, 56:10, 56:14, 56:16-56:17, 56:19N-56:20:02, 56:23-56:30, 56:33-56:42, 59:01:01:01-59:01:01:02, 59:04-59:08
<b>7<sup>5</sup></b>	105 bp	1070 bp	*49:04-49:05	*13:04, 13:46, 13:72, 15:04:01-15:04:02, 15:16:01-15:16:03, 15:67, 15:95, 15:137, 15:155, 15:222, 15:254, 15:293, 15:310, 35:37, 35:235, 39:06:01-39:06:04, 39:33-39:34, 39:50, 39:57, 39:62, 39:64, 39:83, 39:90, 40:86, 40:159, 41:30, 45:02, 45:09, 46:32, 50:09, 50:33, 51:13:01-51:13:02, 51:37, 51:63, 51:92, 51:97, 51:172, 52:14, 52:39, 54:14, 55:13, 55:23, 55:32, 55:58, 56:22, 59:03, 73:01-73:02
	255 bp		*49:20	*44:15, 44:18, 44:20, 44:47, 44:100, 45:01:01-45:03, 45:05-45:07, 45:09, 45:11-45:14, 50:02, 51:23, 51:108, 52:15
<b>8<sup>5.7</sup></b>	70 bp	1070 bp	*49:05	*07:78, 07:84, 08:09, 08:84, 13:02:01-13:02:05, 13:02:07-13:04, 13:08-13:09, 13:14-13:16, 13:18-13:19, 13:27, 13:30-13:35, 13:37-13:38, 13:40-13:42, 13:44-13:49N, 13:53-13:56N, 13:58-13:59, 13:62, 13:65-13:70, 13:75, 15:04:01-15:04:02, 15:16:01-15:16:02, 15:42, 15:67, 15:83, 15:95, 15:137, 15:155, 15:222, 15:254, 15:293, 15:303, 15:310, 27:14, 27:81, 35:60, 39:06:01-39:06:04, 39:33-39:34, 39:50, 39:57, 39:62, 39:64, 39:83, 39:90, 40:06:01:01-40:06:10, 40:44, 40:53, 40:70:01-40:70:02, 40:75, 40:83, 40:86, 40:93, 40:95-40:96, 40:103, 40:109-40:110, 40:127, 40:131, 40:148, 40:159, 40:161-40:162, 40:165, 40:177, 40:190, 40:230, 40:244, 40:256N, 40:268-40:270, 40:275, 41:01, 41:05-41:07, 41:09, 41:12, 41:14, 41:16-41:17, 41:20-41:22, 41:25-41:26, 41:28-41:30, 41:32, 42:04, 44:20, 44:47, 44:100, 45:14, 46:11, 46:18, 46:32, 50:20, 51:01:01-51:01:18, 51:01:20-51:03, 51:05, 51:07:01-51:16, 51:18-51:24:05, 51:26-51:41N, 51:43-51:44N, 51:48-51:55, 51:57-51:58, 51:60-51:61:02, 51:63, 51:65-51:80, 51:82-51:93, 51:95-51:134, 51:136-51:138, 51:140-51:147, 51:149N-51:161, 51:163-51:173Q, 52:01:01:01-52:39, 54:01:01-54:05N, 54:07-54:08N, 54:10-54:14, 54:16-54:32, 55:01:01-55:01:06, 55:01:08-55:03, 55:05, 55:07, 55:09-55:13, 55:15-55:26, 55:28-55:48, 55:50, 55:52-55:64, 56:01:01-56:01:07, 56:05:01-56:08, 56:13-56:17, 56:19N-56:30, 56:33-56:43,

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

Lot No.: <b>88V</b>		Lot-specific information		<a href="http://www.olerup-ssp.com">www.olerup-ssp.com</a>
	155 bp		*49:19N	58:08:01-58:08:02, 59:01:01:01-59:08, 73:01-73:02, 78:01:01-78:07
<b>9</b>	155 bp 240 bp	<b>800 bp</b>	*49:21 *49:07	*45:10, 50:31
<b>10<sup>5</sup></b>	80 bp 250 bp	<b>800 bp</b>	*49:16 *49:08	*52:33
<b>11</b>	505 bp	1070 bp		*08:03, 08:52, 08:78, 37:34, 38:30, 44:06, 44:18, 44:25, 44:95, 51:08:01-51:08:02, 51:20, 51:36, 51:44N, 51:97, 51:141, 51:153, 52:19, 53:22, 57:09, 57:24
<b>12</b>	135 bp	1070 bp	*49:10, 49:24, 49:31	*15:46, 15:53, 15:106, 35:47, 35:154, 40:10:01-40:10:02, 40:40, 40:49, 40:58, 40:110, 40:129, 40:164, 40:210, 40:219, 40:229, 44:02:01:01-44:03:15, 44:03:17-44:05:04, 44:09-44:14, 44:16-44:17, 44:19N-44:39, 44:41:01-44:43:02, 44:45-44:48, 44:50:01-44:54, 44:56N, 44:58N-44:59:02, 44:61N-44:80, 44:82, 44:84:01-44:102, 44:104-44:110, 44:112-44:129, 44:132-44:133, 44:135-44:144, 44:146-44:155, 44:157-44:187, 44:189-44:193, 47:01:01:01-47:09, 50:13, 53:30, 57:45, 57:69
<b>13<sup>5,6</sup></b>	150 bp	1070 bp	*49:11	*13:01:01-13:03, 13:06-13:09, 13:11-13:12:02, 13:14-13:15, 13:17, 13:18 <sup>w</sup> , 13:19, 13:22:01-13:23, 13:25, 13:27-13:30, 13:32-13:34, 13:36-13:40, 13:42-13:45, 13:47, 13:49N-13:50, 13:52-13:58, 13:60-13:61, 13:63N-13:70, 13:73 <sup>w</sup> , 13:74-13:78, 15:86, 15:224, 35:60, 40:48, 40:71, 44:10, 46:11, 46:18, 51:15, 51:157, 54:01:01-54:03, 54:05N, 54:07-54:08N, 54:10 <sup>w</sup> , 54:12-54:13, 54:17-54:19, 54:21-54:32, 55:02:01-55:02:08, 55:07, 55:10, 55:12, 55:16, 55:18-55:19, 55:22, 55:26, 55:30, 55:34-55:35, 55:37 <sup>w</sup> , 55:39, 55:41-55:43, 55:47-55:48, 55:50, 55:61-55:63, 56:01:01-56:01:04, 56:01:06-56:02, 56:04, 56:07-56:08, 56:10, 56:14, 56:16-56:17, 56:19N-56:20:02, 56:23-56:24, 56:26-56:29, 56:33-56:36, 56:38N-56:42, 59:01:01:01-59:01:01:02, 59:04-59:08
<b>14</b>	155 bp 190 bp	1070 bp	*49:19N *49:12, 49:17	*51:130
<b>15<sup>6</sup></b>	160 bp 190 bp 290 bp	1070 bp	*49:13 *49:17 *49:14	
<b>16<sup>5</sup></b>	95 bp  135 bp	1070 bp	*49:18  *49:15	*35:01:10, 35:04:02, 40:28, 50:14, 51:56:01-51:56:02, 53:28, 55:01:04, <b>C*02:02:23</b>
<b>17</b>	180 bp	1070 bp	*49:23	
<b>18</b>	175 bp	1070 bp	*49:27	
<b>19<sup>8</sup></b>	-	-	<b>Negative Control</b>	

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B\*49 SSP typings.



101.547-06 – including *Taq pol.*, IFU-01  
101.547-06u – without *Taq pol.*, IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

**Lot No.: 88V**

**Lot-specific information**

**www.olerup-ssp.com**

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>For several HLA Class I alleles 1<sup>st</sup> and/or 4<sup>th</sup> exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

<sup>4</sup>Due to the sharing of sequence motifs between HLA-B alleles, non-HLA-B\*49 alleles will be amplified by primer mixes 1 to 14 and 16. In addition, a few HLA-C alleles will be amplified by primer mixes 3 and 16.

<sup>5</sup>HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

<sup>6</sup>Primer mixes 6, 13 and 15 may have tendencies of unspecific amplifications.

<sup>7</sup>Primer mix 8 may give rise to a lower yield of HLA-specific PCR product than the other HLA-B\*49 primer mixes.

<sup>8</sup>Primer mix 19 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by control primer pairs. PCR product sizes range from 75 to 200 base pairs. The PCR product generated by the control primer pair is 430 base pairs.

'w', may be weakly amplified.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

## PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec. PCR product	145	145	115	170	105	160	105	70	155	80	505	135
		190	145	205			255	155	240	250		
		255										
Length of int. pos. control <sup>1</sup>	800	800	800	1070	1070	1070	1070	1070	800	800	1070	1070
5'-primer(s) <sup>2</sup>	206	97	206	141	357	420	357	357	206	357	317	97
	5'-gAA 3'	5'-TCC 3'	5'-gAA 3'	5'-ATT 3'	5'-Tgg 3'	5'-TTA 3'	5'-Tgg 3'	5'-Tgg 3'	5'-gAA 3'	5'-Tgg 3'	5'-gCT 3'	5'-TCT 3'
		454	485	420					362			
		5'-ACA 3'	5'-CAA 3'	5'-TTA 3'					5'-gAg 3'			
3'-primer(s) <sup>3</sup>	309	246	309	272	420	538	419	387	319	394	538	193
	5'-ATC 3'	5'-TAT 3'	5'-gTg 3'	5'-Tgg 3'	5'-gCT 3'	5'-CAG 3'	5'-Cgg 3'	5'-TCC 3'	5'-gCA 3'	5'-gCA 3'	5'-gTC 3'	5'-CgT 3'
		309	559	583			419	470	559	565		
		5'-gTg 3'	5'-CAG 3'	5'-gTg 3'			5'-CgA 3'	5'-TCT 3'	5'-CTC 3'	5'-CAG 3'		
		559					572					
		5'-CAG 3'					5'-gCg 3'					
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

Well No.	13	14	15	16	17	18
Length of spec. PCR product	150	155	160	95	180	175
		190	190	135		
			290			
Length of int. pos. control <sup>1</sup>	1070	1070	1070	1070	1070	1070
5'-primer(s) <sup>2</sup>	420	357	357	176	357	140
	5'-TTA 3'	5'-Tgg 3'	5'-Tgg 3'	5'-gAC 3'	5'-Tgg 3'	5'-CAC 3'
				379		
				5'-ACC 3'		
3'-primer(s) <sup>3</sup>	527	470	476	272	497	272
	5'-CCA 3'	5'-TCT 3'	5'-CCC 3'	5'-Tgg 3'	5'-Tgg 3'	5'-Tgg 3'
		505	506	435		
		5'-gCT 3'	5'-Tgg 3'	5'-TCT 3'		
		506	608			
		5'-Tgg 3'	5'-gCT 3'			
Well No.	13	14	15	16	17	18

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

<sup>2</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

<b>CELL LINE VALIDATION SHEET</b>																			
<b>HLA-B*49 SSP primer set<sup>2</sup></b>																			
				Well															
				Prod. No.:															
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
				201191101	201321002	201207703	201191104	201191105	201191106	201191107	201191108	201207709	201436810	201191111	201191112	201436813	201436814	201436815	201191116
IHC cell line <sup>1</sup>		B*																	
1	9001 SA	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280 LK707	*52:01	*73:01	-	-	-	-	-	-	+	+	-	-	-	-	-	-	-	-
3	9011 E4181324	*52:01		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
4	9275 GU373	*15:10	*53:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009 KAS011	*37:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*39:01	*51:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
7	9020 QBL	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025 DEU	*35:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*54:01		-	-	-	-	+	+	-	+	-	-	-	-	+	-	-	-
11	9051 PITOUT	*44:03		-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	-
12	9052 DBB	*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9025 JESTHOM	*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*08:01	*55:01	-	-	-	-	+	+	-	+	-	-	-	-	-	-	-	-
18	9257 32367	*14:01	*56:01	-	-	-	-	+	+	-	+	-	-	-	+	-	-	-	-
19	9038 BM16	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056 KOSE	*35:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*40:02	*56:02	-	-	-	-	-	+	-	-	-	-	-	+	-	-	-	-
24	9035 JBUSH	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*07:05	*51:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
28	9320 BEL5GB	*44:02	*44:03	-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	-
29	9050 MOU	*44:03		-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	-
30	9021 RSH	*42:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019 DUCAF	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG	*41:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*44:02		-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	-
36	9024 KT17	*15:01	*35:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099 LZL	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315 CML	*08:01	*27:05	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199	*13:02	*46:01	-	-	-	-	+	+	-	+	-	-	-	-	+	-	-	-
41	9055 H0301	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO	*42:01	*50:01	-	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-
46	9013 SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*51:01		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
48	9303 TER-ND	*35:01	*44:03	-	-	+	-	-	-	-	-	-	-	-	+	-	-	-	-

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

<b>CELL LINE VALIDATION SHEET</b>					
<b>HLA-B*49 SSP primer set<sup>2</sup></b>					
					Well
					17 18
					Prod. No.:
					201436817
					201436818
	<b>IHWC cell line<sup>1</sup></b>		<b>B*</b>		
1	9001	SA	*07:02		- -
2	9280	LK707	*52:01	*73:01	- -
3	9011	E4181324	*52:01		- -
4	9275	GU373	*15:10	*53:01	- -
5	9009	KAS011	*37:01		- -
6	9353	SM	*39:01	*51:01	- -
7	9020	QBL	*18:01		- -
8	9025	DEU	*35:01		- -
9	9026	YAR	*38:01		- -
10	9107	LKT3	*54:01		- -
11	9051	PITOUT	*44:03		- -
12	9052	DBB	*57:01		- -
13	9025	JESTHOM	*27:05		- -
14	9071	OLGA	*15:01	*15:20	- -
15	9075	DKB	*40:01		- -
16	9037	SWEIG007	*40:02		- -
17	9282	CTM3953540	*08:01	*55:01	- -
18	9257	32367	*14:01	*56:01	- -
19	9038	BM16	*18:01		- -
20	9059	SLE005	*40:01		- -
21	9064	AMALA	*15:01		- -
22	9056	KOSE	*35:03		- -
23	9124	IHL	*40:02	*56:02	- -
24	9035	JBUSH	*38:01		- -
25	9049	IBW9	*14:02		- -
26	9285	WT49	*58:01		- -
27	9191	CH1007	*07:05	*51:01	- -
28	9320	BEL5GB	*44:02	*44:03	- -
29	9050	MOU	*44:03		- -
30	9021	RSH	*42:01		- -
31	9019	DUCAF	*18:01		- -
32	9297	HAG	*41:02		- -
33	9098	MT14B	*40:01		- -
34	9104	DHIF	*38:01		- -
35	9302	SSTO	*44:02		- -
36	9024	KT17	*15:01	*35:01	- -
37	9065	HHKB	*07:02		- -
38	9099	LZL	*15:01		- -
39	9315	CML	*08:01	*27:05	- -
40	9134	WHONP199	*13:02	*46:01	- -
41	9055	H0301	*14:02		- -
42	9066	TAB089	*46:01		- -
43	9076	T7526	*46:01		- -
44	9057	TEM	*38:01		- -
45	9239	SHJO	*42:01	*50:01	- -
46	9013	SCHU	*07:02		- -
47	9045	TUBO	*51:01		- -
48	9303	TER-ND	*35:01	*44:03	- -

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

<sup>1</sup>The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

<sup>2</sup>The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 2, 4, 9 to 11 and 14 to 18 were available. The specificities of the primers in primer solutions 2, 4, 9, 11 and 16 were tested by separately adding additional 3'-primers, respectively additional 5'-primers. In primer solution 10, 14, 15 and 17 it was only possible to test the 5'-primer, the 3'-primer was not possible to test. In primer solution 18 it were only possible to test the 3'-primer, the 5'-primer were not possible to test. In primer solutions 2 and 16, one 5'-primer was not possible to test, and in primer solutions 8 and 9, one 3'-primer was not possible to test. One additional 5'-primer and one additional 3'-primer in primer solution 3 were tested by separately adding one 3'-primer respectively one 3'-primer.

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

101.547-06 – including *Taq* pol., IFU-01  
101.547-06u – without *Taq* pol., IFU-02

Visit [www.olerup-ssp.com](http://www.olerup-ssp.com) for  
“Instructions for Use” (IFU)

Lot No.: **88V**

Lot-specific information

[www.olerup-ssp.com](http://www.olerup-ssp.com)

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup-ssp.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.