

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information
Olerup SSP[®] HLA-B*55

Product number:	101.570-06 – including <i>Taq</i> polymerase 101.570-06u – without <i>Taq</i> polymerase
Lot number:	4R6
Expiry date:	2026-12-01
Number of tests:	6
Number of wells per test:	31+1
Storage - pre-aliquoted primers:	dark, between -15°C and -25°C
- PCR Master Mix:	between -15°C and -25°C
- Adhesive PCR seals	RT

This Product Description is only valid for Lot No. 4R6.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP[®]
 HLA-B*55 Lot (4L7)**

- The product documentation has been updated for new alleles of IMGT 3.50.0
- The kit resolution focuses on common and well documented (CWD) alleles¹.

The HLA-B*55 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP[®]* HLA-B*55 lot was made (**Lot No. 4L7**).

The HLA-B*55 primer set is unchanged compared to the previous *Olerup SSP[®]* HLA-B*55 (**Lot No. 4L7**).

¹S. J. Mack, P. Cano, J. A. Hollenbach et al.
 Common and well-documented HLA alleles: 2012 update to
 the CWD catalogue. *Tissue Antigens*, 2013, 81, 194–203



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

Well **32** contains Negative Control primer pairs, that will amplify a majority of the *Olerup* SSP[®] HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.

The PCR product generated by the positive control primer pair is 200 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

101.570-06 – including *Taq* polymerase
101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

PRODUCT DESCRIPTION

HLA-B*55 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the B*55:01 to B*55:131 alleles.

PLATE LAYOUT

Each test consists of 32 PCR reactions in a 32 well PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	NC

The 32 well PCR plate is marked with 'HLA-B*55' in silver/gray ink.

Well No. 1 is marked with the Lot No. '4R6'.

Wells 1 to 31 – HLA-B*55 high resolution primers.

Well 32 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 32 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between HLA-B alleles non-HLA-B*55 alleles will be amplified by some primer mixes. For further details see Specificity Table.

UNIQUELY IDENTIFIED ALLELES

All the HLA-B*55 alleles, i.e. **B*55:01 to B*55:131**, recognized by the HLA Nomenclature Committee in October 2022^{1,2} will be amplified by the primers in the HLA-B*55 subtyping kit³.

The HLA-B*55 kit enables separation of the confirmed HLA-B*55 alleles as listed in the IMGT/HLA database 3.26.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-B*55 alleles is listed below.

The HLA-B*55 kit also enables identification of many null and alternatively expressed alleles.



0197

For *In Vitro* Diagnostic Use
MA123 v02 SSP PI Template
Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

The following HLA-B*55 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix
B*55:26, 55:62	25

¹HLA-B alleles listed on the IMGT/HLA web page 2022-October-12, release 3.50.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

³The HLA-B*55 primer set cannot separate the following alleles. These alleles can be distinguished by the HLA-B low resolution kit and/or the HLA-B*15, B*51 and B*56 high resolution kit.

Alleles

B*55:22, 56:60
B*55:78, 55:118N, 55:121, 15:42, 51:300



0197

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information
ALLELE CONFIRMATION STATUS

Allele	Status ¹	Allele	Status ¹	Allele	Status ¹	Allele	Status ¹
B*55:01:01	Confirmed	B*55:07	Confirmed	B*55:37	Confirmed	B*55:67	Unconfirmed
B*55:01:02	Unconfirmed	B*55:08	Confirmed	B*55:38	Unconfirmed	B*55:68	Unconfirmed
B*55:01:03	Confirmed	B*55:09	Unconfirmed	B*55:39	Unconfirmed	B*55:69	Unconfirmed
B*55:01:04	Confirmed	B*55:10	Confirmed	B*55:40	Unconfirmed	B*55:70	Unconfirmed
B*55:01:05	Confirmed	B*55:11	Confirmed	B*55:41	Unconfirmed	B*55:71	Unconfirmed
B*55:01:06	Confirmed	B*55:12	Confirmed	B*55:42	Unconfirmed	B*55:72	Confirmed
B*55:01:07	Unconfirmed	B*55:13	Unconfirmed	B*55:43	Unconfirmed	B*55:73	Unconfirmed
B*55:01:08	Unconfirmed	B*55:14	Unconfirmed	B*55:44	Unconfirmed	B*55:74	Unconfirmed
B*55:01:09	Unconfirmed	B*55:15	Unconfirmed	B*55:45	Confirmed	B*55:75	Unconfirmed
B*55:01:10	Unconfirmed	B*55:16	Confirmed	B*55:46	Unconfirmed	B*55:76	Unconfirmed
B*55:01:11	Unconfirmed	B*55:17	Confirmed	B*55:47	Unconfirmed	B*55:77	Unconfirmed
B*55:01:12	Unconfirmed	B*55:18	Confirmed	B*55:48	Unconfirmed	B*55:78	Unconfirmed
B*55:01:13	Unconfirmed	B*55:19	Unconfirmed	B*55:49	Unconfirmed	B*55:80	Unconfirmed
B*55:01:14	Unconfirmed	B*55:20	Confirmed	B*55:50	Confirmed		
B*55:01:15	Unconfirmed	B*55:21	Unconfirmed	B*55:51	Unconfirmed		
B*55:01:16	Unconfirmed	B*55:22	Unconfirmed	B*55:52	Unconfirmed		
B*55:02:01:01	Confirmed	B*55:23	Confirmed	B*55:53	Unconfirmed		
B*55:02:01:02	Unconfirmed	B*55:24	Unconfirmed	B*55:54	Confirmed		
B*55:02:02	Unconfirmed	B*55:25	Confirmed	B*55:55N	Unconfirmed		
B*55:02:03	Confirmed	B*55:26	Unconfirmed	B*55:56	Confirmed		
B*55:02:04	Unconfirmed	B*55:27	Unconfirmed	B*55:57	Confirmed		
B*55:02:05	Unconfirmed	B*55:28	Unconfirmed	B*55:58	Unconfirmed		
B*55:02:06	Unconfirmed	B*55:29	Confirmed	B*55:59	Unconfirmed		
B*55:02:07	Confirmed	B*55:30	Unconfirmed	B*55:60	Unconfirmed		
B*55:02:08	Unconfirmed	B*55:31	Confirmed	B*55:61	Unconfirmed		
B*55:02:09	Unconfirmed	B*55:32	Unconfirmed	B*55:62	Unconfirmed		
B*55:02:10	Unconfirmed	B*55:33	Confirmed	B*55:63	Unconfirmed		
B*55:03	Unconfirmed	B*55:34	Confirmed	B*55:64	Unconfirmed		
B*55:04	Confirmed	B*55:35	Unconfirmed	B*55:65	Unconfirmed		
B*55:05	Unconfirmed	B*55:36	Confirmed	B*55:66	Confirmed		

¹Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2016-October-14, release 3.26.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-B*55 homo- and heterozygotes is available upon request.



0197

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: 4R6

Lot-specific information
SPECIFICITY TABLE

HLA-B*55 SSP subtyping

Specificities and sizes of the PCR products of the 31+1 primer mixes used for HLA-B*55 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified HLA-B*55 alleles ³	Other amplified HLA Class I alleles
1	400 bp	800 bp	*55:01:01:01-55:05, 55:07-55:17, 55:19-55:33, 55:35-55:46, 55:48-55:77, 55:79-55:117N, 55:119-55:120, 55:122-55:132N	*07:65, 07:209, 07:271, 27:157, 35:76, 35:561, 40:166 [?] , 44:90, 44:356, 54:01:01:01-54:23, 54:25-54:29, 54:31-54:46, 56:01:01:01-56:13, 56:15-56:16, 56:18-56:22, 56:24-56:32, 56:34-56:41, 56:43, 56:45-56:53, 56:55:01:02-56:66, 56:69-56:88, 81:01:01:01-81:01:05, 81:03 [?] -81:04N [?] , 81:06 [?] -81:08 [?] , 81:09-81:10, 82:01:01:01-82:04, 83:01
2 ⁴	125 bp	1070 bp	*55:01:01:01-55:03, 55:05, 55:07, 55:09-55:13, 55:15-55:16, 55:18-55:19, 55:21-55:37, 55:39-55:48, 55:50, 55:52-55:71, 55:73-55:85, 55:87-55:109, 55:111-55:127, 55:129-55:132N	*07:235, 13:01:01:01-13:02:30, 13:02:32-13:04, 13:06-13:09, 13:11-13:20, 13:22:01-13:23, 13:25-13:34, 13:36-13:61, 13:63N-13:64, 13:66-13:70, 13:72-13:78, 13:80-13:107, 13:109-13:165, 13:167-13:173, 14:22, 14:37, 14:55, 14:102, 15:16:01:01-15:16:05, 15:20, 15:34:01-15:34:02, 15:42, 15:62, 15:67, 15:85-15:86, 15:95, 15:137, 15:194, 15:222, 15:242:02, 15:254, 15:303, 15:362, 15:393, 15:407-15:408, 15:485-15:486, 15:550, 15:552, 15:587, 15:599, 15:602, 15:613, 15:623, 18:22, 18:52, 18:105, 27:14, 35:01:01:01-35:01:42, 35:01:44-35:05:02, 35:05:04-35:08:15, 35:10-35:11:03, 35:13-35:15:02, 35:19-35:21, 35:23-35:30:01:02, 35:33-35:38, 35:40N-35:42:02, 35:45-35:50, 35:52, 35:54-35:58, 35:60, 35:62-35:63, 35:65Q-35:66, 35:68:01-35:72, 35:74-35:78, 35:80-35:85, 35:89-35:98, 35:100-35:101:02, 35:103-35:112:02, 35:115-35:116, 35:119-35:124, 35:126-35:134N, 35:136-35:148, 35:150:01-35:163, 35:166-35:173:02N, 35:175-35:178, 35:180-35:184, 35:186-35:195, 35:197-35:198, 35:200, 35:202-35:204, 35:206-35:207, 35:209-35:214, 35:216N-35:229, 35:231-35:232, 35:235-35:241, 35:243-35:250, 35:252-35:259, 35:261-35:264, 35:267-35:275, 35:277-35:316, 35:318-35:320, 35:322-35:330, 35:332-35:345, 35:347-35:371, 35:373-35:408, 35:410-35:413, 35:415-35:424, 35:426-35:470, 35:472-35:473, 35:477-35:483, 35:485-35:524, 35:526-35:533, 35:536-35:538, 35:540-35:546, 35:548, 35:550-35:555, 35:557-35:568, 37:101, 40:28, 44:10, 44:15:01:01-44:15:01:02, 44:18:01:01-44:18:01:02, 44:48, 44:256, 44:276, 44:339-44:340, 44:355, 44:481, 45:01:01:01-45:25, 45:28N-45:30, 46:11, 46:18, 46:33, 46:53, 46:67, 48:02:01-48:02:03, 48:25, 49:01:01:01-49:01:18,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				49:01:20-49:14, 49:16-49:39, 49:41-49:60N, 49:62-49:78, 50:01:01:01-50:01:01:13, 50:01:03-50:01:13, 50:01:15-50:02:01:02, 50:05-50:19, 50:21-50:22, 50:31-50:50, 50:52-50:59, 50:62-50:79, 51:15, 51:56:01-51:56:03, 51:157, 51:158:02, 51:225, 51:300, 51:372, 53:01:01:01-53:06, 53:08:01-53:43:02, 53:45-53:61, 53:63-53:72, 54:01:01:01-54:05N, 54:07-54:08N, 54:10, 54:12-54:14, 54:16-54:36, 54:39-54:41, 54:43-54:44, 54:46, 56:01:01:01-56:01:03, 56:01:05-56:02:03, 56:07-56:11, 56:13-56:17, 56:19N, 56:22-56:26, 56:28N-56:30, 56:33-56:47, 56:50-56:56, 56:58-56:62, 56:64-56:70, 56:72-56:89, 57:05, 58:01:01:01-58:01:02, 58:01:04-58:02:02, 58:04-58:07, 58:09, 58:11-58:13, 58:15-58:19, 58:21-58:26, 58:28:01-58:29, 58:31N-58:45:02, 58:47-58:58, 58:60, 58:62-58:87, 58:89-58:113, 58:115-58:126, 58:128N-58:139, 59:01:01:01-59:01:01:02, 59:03-59:12, 82:01:01:01-82:04, C*02:56, C*04:180:01-04:180:02, C*05:217, C*06:20, C*07:81, C*07:168, C*07:450:01-C*07:450:02, C*07:896, C*07:959, C*08:123, C*08:139, C*15:126, C*16:98, C*16:102, C*16:110
3	150 bp	1070 bp	*55:01:01:01-55:01:29, 55:03, 55:05, 55:09 ^w , 55:11, 55:15, 55:24-55:25, 55:29, 55:31, 55:33, 55:36, 55:38, 55:40, 55:44-55:45, 55:52 ^w , 55:53-55:55N, 55:59-55:60, 55:64, 55:66, 55:73-55:76, 55:78-55:79, 55:84-55:85, 55:87, 55:90-55:95, 55:97N-55:106, 55:108-55:109, 55:111, 55:113-55:116, 55:118N-55:119, 55:121-55:122, 55:124-55:127	*07:78 ^w , 13:16, 13:20, 13:48, 13:128 ^w , 15:42, 51:62, 51:106:01-51:106:02, 51:223 ^w , 51:225, 51:300, 52:25:01-52:25:02, 54:20 ^w , 56:25, 56:37, 56:66, 59:12, C*01:90^w, C*08:207^w, C*15:91, C*15:128
4	235 bp	1070 bp	*55:03, 55:49	*08:15, 15:289, 35:74, 35:186, 40:73, 46:01:01:01, 46:01:02-46:01:09, 46:02-46:04, 46:06-46:32:01, 46:33-46:40:01, 46:41N-46:43, 46:52, 46:66
5⁴	95 bp	800 bp	*55:50	*08:83, 13:06, 13:53, 13:131, 14:43, 14:50, 18:74N, 18:78, 38:38, 41:37, A*31:147, C*04:467, C*05:110, C*12:120
	235 bp		*55:04, 55:49, 55:86, 55:110	*07:19, 07:31:01-07:31:02, 07:34, 07:43, 08:01:01:01-08:01:55, 08:01:57-08:05, 08:07-08:08N, 08:10, 08:13-08:15, 08:17-08:20:02, 08:22-08:24, 08:26:01-08:39, 08:41-08:48, 08:50-08:59:02, 08:61-08:78, 08:80-08:83, 08:85-08:88, 08:90-08:106, 08:108-08:124, 08:126-08:128, 08:130-08:155, 08:157-08:164, 08:166-08:170,



0197

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				08:172-08:174, 08:176-08:180, 08:182-08:183, 08:185, 08:187-08:204, 08:206-08:231, 08:233-08:253, 08:256-08:290, 08:292-08:303, 14:05, 14:13, 14:53, 18:14, 18:22 ^w , 35:87, 37:09, 38:19, 39:03:01:01-39:03:02, 39:14:01:01-39:14:01:03, 39:24:01-39:24:03, 39:29, 39:37:01:01-39:37:01:02, 39:76, 39:120, 39:144, 39:152, 39:182, 39:184-39:185, 40:39, 41:02:01:01-41:02:11, 41:04, 41:10-41:11, 41:13, 41:15, 41:18, 41:23-41:24, 41:27, 41:31, 41:36, 41:38-41:43, 41:45N-41:47, 41:49, 41:51-41:52, 41:55, 41:58, 41:65, 41:68-41:69, 41:71-41:72, 41:74, 42:01:01:01-42:02:02, 42:05:01-42:12, 42:14-42:32, 44:166, 48:37, C*04:443, C*06:264, C*07:294, C*07:526:01, C*07:530:01, C*07:681, C*08:08:01-08:08:02
6^{4,5}	100 bp	1070 bp	*55:05	*15:390, C*03:349, C*16:152
7	180 bp	1070 bp	*55:08, 55:51	*07:20, 07:24, 07:60, 07:100, 07:131, 07:220, 07:223, 07:358, 07:464, 08:21, 08:25, 08:175, 13:62, 13:86, 15:09:01-15:10:08, 15:30:01:01-15:30:01:03, 15:37, 15:45, 15:48, 15:63, 15:90, 15:99, 15:150, 15:243, 15:248, 15:287, 15:312-15:313, 15:324, 15:329, 15:338, 15:361, 15:427, 15:435, 15:439, 15:445, 15:455, 15:503, 15:506, 15:530, 15:541, 15:563, 15:611, 15:620, 27:76, 27:153, 35:02:01:01-35:02:19, 35:04:01:01-35:04:03, 35:09:01:01-35:09:04, 35:12:01:01-35:12:04, 35:18, 35:22:01:01-35:22:01:02, 35:34, 35:39, 35:44, 35:81, 35:83, 35:95, 35:129N, 35:149, 35:154, 35:162, 35:172, 35:182-35:184, 35:199, 35:201, 35:211, 35:220, 35:233, 35:251:01-35:251:02, 35:258, 35:266, 35:270, 35:273, 35:285, 35:292, 35:309, 35:311, 35:316, 35:321, 35:323:01-35:323:02, 35:335, 35:339, 35:357, 35:361, 35:366-35:367, 35:372, 35:374, 35:377-35:379, 35:384, 35:387, 35:391, 35:396, 35:403, 35:407, 35:410, 35:419, 35:425, 35:443, 35:463, 35:474, 35:477, 35:483, 35:487, 35:492, 35:501, 35:505, 35:508N, 35:522, 35:535, 35:539, 35:550, 35:553, 35:555, 35:557, 35:562, 35:565, 40:05:01:01-40:05:02, 40:26, 40:28, 40:63, 40:92, 40:174, 40:324, 40:343, 40:496, 40:523, 41:17, 41:62, 44:62, 44:77, 44:82, 44:107, 44:123, 44:184, 44:213, 44:459, 45:29, 46:74, 46:81, 49:18:01-49:18:02, 50:14, 51:01:01:01-51:01:14, 51:01:16-51:01:31, 51:01:33-51:01:38, 51:01:40-51:01:58, 51:01:60-51:09:03, 51:11N-51:12, 51:18-51:20, 51:22-51:24:05, 51:26-51:30, 51:32-51:33, 51:35, 51:38-51:41N, 51:43-51:44N, 51:46, 51:48-51:61:02, 51:64-51:80, 51:83-51:84, 51:86-51:91, 51:94-51:96, 51:98N-51:100, 51:102-51:105, 51:107-51:142, 51:145-51:152, 51:154-51:156, 51:158:01-51:171, 51:173Q-51:181, 51:183-51:191, 51:193:01-51:196, 51:198-51:213, 51:215, 51:217-51:219,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				51:221-51:222, 51:224, 51:226, 51:228-51:250, 51:252-51:260, 51:262-51:273N, 51:275-51:279, 51:282-51:294, 51:296-51:297, 51:299, 51:301-51:318N, 51:320-51:339, 51:341-51:351, 51:353-51:358, 51:360-51:375, 52:01:01:01-52:01:20, 52:01:22-52:07:02, 52:09:01-52:13, 52:15-52:24, 52:27-52:36, 52:38, 52:40-52:61, 52:64-52:66, 52:68-52:70, 52:72-52:74, 52:76-52:83, 52:85-52:87, 52:89N-52:96N, 52:98-52:110N, 53:19, 53:36, 53:42, 56:05:01-56:06, 56:15, 56:21, 57:02:01:01-57:03:05, 57:05, 57:07, 57:09, 57:12, 57:17, 57:28N, 57:39, 57:42, 57:46, 57:57, 57:63, 57:66, 57:70, 57:80, 57:82, 57:84, 57:94, 57:96, 57:101, 57:124, 57:137, 57:155, 58:08:01-58:08:02, 58:28:01-58:28:02, 58:65, 78:01:01:02-78:06, 78:09-78:10, C*03:81, C*03:152, C*03:317, C*03:382, C*03:388, C*03:526
8	230 bp 425 bp	800 bp	*55:66 *55:12	*56:41
9	215 bp 375 bp	1070 bp	*55:25 *55:11, 55:14	*07:225, 07:227
10^{4,6}	70 bp	1070 bp	*55:07, 55:124	*07:13, 07:247, 07:270, 08:188, 15:436, 27:198, 39:121, 46:86, 54:01:02, 59:06, 67:02:01:01-67:02:01:02, C*01:02:48, C*02:02:61, C*03:02:18, C*03:03:42, C*03:04:50, C*07:01:22, C*07:02:31, C*08:16:02, C*14:02:27, C*15:02:06
	220 bp		*55:17, 55:38, 55:72	*08:09, 08:84, 15:83:01:01-15:83:01:02, 39:06:02:01-39:06:08, 39:34, 39:57, 39:62, 39:64, 39:83:02-39:83:03, 39:90, 39:109, 39:129, 39:132, 39:140-39:141, 39:154, 39:178, 39:186-39:187, 41:66, 42:04, 51:21, 51:36, 51:101, 51:143, 51:153:01-51:153:02, 51:197, 51:280, 51:295, 51:340, 51:376, 52:37, 52:71, 52:84, 54:42, 78:07, C*14:92
11⁴	90 bp 125 bp 180 bp	1070 bp	*55:48 *55:15 *55:09, 55:22, 55:24	*07:78, 13:01:01:01-13:02:37, 13:07N-13:09, 13:11, 13:14-13:20, 13:22:01-13:23, 13:25, 13:27-13:34, 13:36-13:45, 13:47, 13:49N-13:50, 13:52, 13:54-13:58, 13:60-13:61, 13:63N-13:70, 13:73-13:78, 13:80-13:85, 13:87-13:107, 13:109-13:119, 13:121-13:126, 13:128-13:130, 13:132-13:139N, 13:141-13:145, 13:147-13:159, 13:161N-13:163, 13:165-13:173, 40:48, 40:377, 45:10, 49:07, 50:31, 54:26, 56:60, 59:09
12	210 bp	1070 bp	*55:10, 55:82	*07:02:01:01-07:02:05, 07:02:07-07:02:27, 07:02:29-07:02:97, 07:04:01-07:07:03, 07:09:01-07:09:02, 07:11-07:12, 07:14:01-07:15, 07:17-07:26, 07:28, 07:30-07:31:02, 07:33:01-07:36, 07:39-07:49N, 07:51-07:64, 07:66-07:68:03, 07:73-07:82, 07:84, 07:87-07:124, 07:126-07:142, 07:144-07:179, 07:181N-07:182N, 07:185, 07:187-07:203, 07:205-07:206, 07:208-07:209,



0197

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

			Lot-specific information	
				07:211-07:218, 07:220-07:227, 07:229:01-07:258, 07:260-07:291, 07:293-07:305, 07:307-07:312, 07:314-07:351N, 07:353-07:367, 07:369-07:382, 07:384-07:399, 07:401-07:414, 07:416-07:427, 07:429-07:452, 07:454-07:469N, 38:26, 42:01:01:01-42:01:04, 42:01:06, 42:04-42:06, 42:08, 42:10, 42:12-42:16:02, 42:19-42:25, 42:28-42:32, 56:84, 67:01:01-67:01:03, 67:03-67:07, 81:01:01:01-81:04N, 81:06-81:10
13	150 bp	1070 bp	*55:02:01:01-55:02:16, 55:07, 55:10, 55:12, 55:16, 55:18-55:19, 55:22, 55:26, 55:30, 55:34-55:35, 55:37 ^w , 55:39, 55:41-55:43, 55:47-55:48, 55:50, 55:54, 55:61-55:63, 55:65, 55:67, 55:69-55:72, 55:77, 55:80, 55:82-55:83N, 55:88-55:89N, 55:96, 55:107, 55:112, 55:117N, 55:120, 55:128-55:132N	*13:01:01:01-13:02:33, 13:02:35-13:03, 13:06-13:09, 13:11-13:12:02, 13:14-13:15, 13:17, 13:18 ^w , 13:19, 13:22:01-13:23, 13:25, 13:27-13:30, 13:32-13:34, 13:36-13:40, 13:42-13:45, 13:47, 13:49N-13:50, 13:52-13:58, 13:60-13:61, 13:63N-13:70, 13:73 ^w , 13:74-13:78, 13:80-13:85, 13:87-13:88, 13:90-13:107, 13:109, 13:111-13:126, 13:129-13:130, 13:132-13:139N, 13:141-13:144, 13:146-13:153, 13:155-13:158, 13:161N-13:163, 13:165-13:173, 15:86, 15:224, 15:485, 15:552, 18:191, 35:60, 35:432, 35:446, 40:48, 40:71, 40:377, 44:10, 44:339-44:340, 46:11, 46:18, 49:11, 49:55, 50:70, 51:15, 51:157, 54:01:01:01-54:03, 54:05N, 54:07-54:08N, 54:10 ^w , 54:12-54:13, 54:17-54:19, 54:21-54:32, 54:34-54:39, 54:41, 54:43-54:46, 56:01:01:01-56:01:04, 56:01:06-56:01:15, 56:01:17-56:02:03, 56:04:01:01-56:04:04, 56:07-56:08, 56:10, 56:14, 56:16-56:17, 56:19N-56:20:02, 56:23-56:24, 56:26-56:29, 56:33-56:36, 56:38N-56:42, 56:44-56:49, 56:51 ^w , 56:52-56:58, 56:60-56:62, 56:63 ^w , 56:64-56:65, 56:67-56:87, 56:89, 58:73, 59:01:01:01-59:01:01:02, 59:04-59:10N
14 ⁴	95 bp	1070 bp	*55:07, 55:124	*07:247, 07:270, 54:01:01:01-54:01:06, 54:01:08-54:23, 54:25 ^w , 54:26-54:34, 54:36-54:46
15	245 bp	1070 bp	*55:01:01:01-55:03, 55:05, 55:07, 55:10-55:13, 55:15-55:21, 55:23, 55:25-55:48, 55:50, 55:52-55:85, 55:87-55:109, 55:111-55:132N	*08:09, 08:84, 08:254, 13:46, 13:53, 13:131, 14:37, 15:42, 15:83:01:01-15:83:01:02, 15:485, 35:60, 35:276, 39:06:01:01-39:06:08, 39:34, 39:57, 39:62, 39:64, 39:83:01-39:83:03, 39:90, 39:109, 39:127, 39:129, 39:132, 39:140-39:141, 39:154, 39:178, 39:186-39:187, 41:01:01:01-41:01:08, 41:05-41:07, 41:09, 41:12, 41:14, 41:16, 41:20-41:22, 41:25-41:26, 41:28-41:30, 41:32-41:35, 41:37, 41:44, 41:50, 41:53-41:54, 41:57, 41:59-41:61, 41:64, 41:66-41:67, 41:70, 41:73, 41:75-41:79, 42:04, 45:08, 46:18, 49:76, 50:45, 50:49, 51:21, 51:36, 51:101, 51:143, 51:153:01-51:153:02, 51:157, 51:197, 51:280, 51:295, 51:300, 51:340, 51:376, 52:37, 52:71, 52:84, 54:01:01:01-54:02, 54:04-54:05N, 54:07-54:08N, 54:10-54:25, 54:27-54:32, 54:34-54:37, 54:39, 54:41-54:46, 56:23, 56:54, 59:01:01:01-59:03, 59:05-59:08, 59:10N-59:11, 78:07, C*05:152, C*07:463, C*07:758, C*12:183, C*14:92
16	245 bp	1070 bp		*13:03-13:04, 13:48, 13:62, 13:120, 13:164, 15:04:01:01-15:04:04, 15:16:01:01-15:16:05,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

15:67, 15:95, 15:137, 15:155, 15:222, 15:254, 15:293, 15:303, 15:310, 15:362, 15:408, 15:499, 15:517, 15:531, 15:550, 15:552, 15:602, 15:609, 15:613, 27:153, 35:37, 35:235, 35:265, 35:273, 39:33, 41:17, 44:15:01:01-44:15:01:02, 44:18:01:01-44:18:01:02, 44:20, 44:47, 44:100, 44:197, 44:339, 44:355, 44:459, 45:01:01:01-45:07, 45:09, 45:11-45:23, 45:24 ^w , 45:25, 45:27-45:30, 46:11, 46:32:01-46:32:02, 49:01:01:01-49:01:07, 49:01:09-49:06, 49:08-49:14, 49:16-49:46, 49:48-49:60N, 49:62-49:75, 49:77-49:78, 50:01:01:01-50:02:01:02, 50:04:01-50:11, 50:13-50:16, 50:18-50:22, 50:32-50:44, 50:46-50:48, 50:50, 50:52-50:79, 51:01:01:01-51:01:31, 51:01:33-51:01:38, 51:01:40-51:01:58, 51:01:60-51:03, 51:05:01:01-51:05:01:02, 51:07:01-51:09:03, 51:11N-51:15, 51:17-51:20, 51:22-51:24:05, 51:26-51:30, 51:32-51:33, 51:35, 51:37-51:41N, 51:43-51:44N, 51:48-51:55, 51:57-51:58, 51:60-51:61:02, 51:63:01-51:63:02, 51:65-51:80, 51:83-51:92:02, 51:94-51:100, 51:102-51:134, 51:136-51:138, 51:140-51:142, 51:145-51:147, 51:149N-51:152, 51:154-51:156, 51:158:01-51:193:02, 51:195-51:196, 51:198-51:212, 51:214-51:219, 51:221-51:226, 51:228-51:279, 51:282-51:292, 51:294, 51:296-51:299, 51:301-51:309, 51:311-51:318N, 51:320-51:339, 51:342-51:351, 51:354-51:358, 51:360-51:375, 52:01:01:01-52:01:20, 52:01:22-52:07:02, 52:09:01-52:36, 52:38-52:66, 52:68-52:70, 52:72-52:83, 52:85-52:87, 52:89N-52:110N, 54:03, 54:33, 54:40, 56:01:01:01-56:01:12, 56:01:14-56:01:18, 56:05:01-56:08, 56:13-56:17, 56:20:01-56:22, 56:24-56:30, 56:33-56:50, 56:52-56:53, 56:55:01:02, 56:58-56:59, 56:61-56:62, 56:64-56:65, 56:67-56:87, 56:89, 58:08:01-58:08:02, 58:122, 59:04, 59:12, 78:01:01:02-78:06, 78:09-78:10, C*03:464				
17⁴	105 bp	1070 bp	*55:17, 55:20, 55:27-55:28:02, 55:39, 55:81	*07:84, 07:429, 08:09, 08:84, 08:254, 13:35, 13:59, 13:62, 15:83:01:01-15:83:01:02, 27:14, 27:81, 27:130, 27:153, 35:273, 35:534, 40:06:01:01-40:06:08, 40:06:10-40:06:17, 40:06:19-40:06:30, 40:44, 40:53, 40:70:01-40:70:02, 40:75, 40:83, 40:93, 40:95-40:96, 40:103, 40:109-40:110, 40:127, 40:131, 40:148, 40:161-40:162, 40:165, 40:167, 40:177, 40:190, 40:230, 40:244, 40:256N, 40:268-40:270, 40:275, 40:294-40:295, 40:298:01-40:298:02, 40:306-40:307, 40:311, 40:313, 40:317-40:318, 40:340-40:342, 40:348-40:350, 40:354, 40:361N-40:364, 40:366, 40:373-40:374, 40:376, 40:387-40:389, 40:408, 40:411, 40:423, 40:425, 40:427, 40:434, 40:445, 40:449, 40:451, 40:453-40:454, 40:459-40:460, 40:464, 40:468, 40:471, 40:473,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				40:476, 40:478, 40:482, 40:484, 40:487N-40:488N, 40:492, 40:494, 40:519-40:520, 40:526, 41:01:01:01-41:01:08, 41:05-41:07, 41:09, 41:12, 41:14, 41:16-41:17, 41:20-41:22, 41:25-41:26, 41:28-41:29, 41:32-41:35, 41:37, 41:53-41:54, 41:56Q-41:57, 41:59-41:61, 41:64, 41:66-41:67, 41:70, 41:73, 41:75-41:79, 42:04, 44:20, 44:47, 44:100, 44:197, 44:459, 45:29, 49:18:01-49:18:02, 50:14, 50:20, 50:42, 51:01:01:01-51:01:47, 51:01:49-51:03, 51:05:01:01-51:05:01:02, 51:07:01-51:12, 51:14, 51:16-51:24:05, 51:26-51:34, 51:36, 51:38-51:41N, 51:43-51:44N, 51:48-51:55, 51:57-51:58, 51:60-51:61:02, 51:65-51:80, 51:82-51:91, 51:93-51:96, 51:98N, 51:100-51:105, 51:107-51:130, 51:132-51:134, 51:136-51:138, 51:140-51:147, 51:149N-51:156, 51:158:01-51:171, 51:173Q-51:193:02, 51:195-51:212, 51:214-51:215, 51:217-51:219, 51:221-51:222, 51:224, 51:226, 51:228-51:250, 51:252-51:280, 51:282-51:292, 51:294-51:299, 51:301-51:309, 51:311-51:318N, 51:320-51:340, 51:342-51:351, 51:354-51:358, 51:360-51:376, 52:01:01:01-52:13, 52:15-52:24, 52:26-52:38, 52:40-52:74, 52:76-52:87, 52:89N-52:96N, 52:98-52:110N, 54:04, 54:11, 54:15, 54:42, 56:05:01-56:06, 56:15, 56:21, 58:08:01-58:08:02, 59:02, 59:11, 78:01:01:02-78:10, C*03:464, C*14:92
18	215 bp 310 bp 415 bp	800 bp	*55:25 *55:19 *55:18, 55:31	*13:125, 18:163, 35:559, 40:10:01:01-40:10:02, 40:476, 44:344, 54:30, 56:23, 56:42, 56:44, 56:54, 56:67, 56:89, 59:01:01:01-59:12
19	205 bp	1070 bp	*55:30, 55:36	*07:234, 15:244, 15:315, 18:126, 39:54, 39:138, 50:56, C*02:131, C*03:70, C*03:127, C*03:179, C*03:274
20	220 bp	1070 bp	*55:20, 55:56	*08:09, 08:84, 08:254, 13:46, 13:154, 15:83:01:01-15:83:01:02, 41:01:01:01-41:01:08, 41:05-41:07, 41:09, 41:12, 41:14, 41:16-41:17, 41:20-41:22, 41:25-41:26, 41:28-41:30, 41:32-41:35, 41:37, 41:44, 41:50, 41:53-41:54, 41:56Q-41:57, 41:59-41:61, 41:64, 41:66-41:67, 41:70, 41:73, 41:75-41:79, 42:04, 44:15:01:01-44:15:01:02, 44:18:01:01-44:18:01:02, 44:20, 44:100, 44:355, 45:01:01:01-45:25, 45:27-45:30, 51:08:01:01-51:08:06, 51:20, 51:36, 51:44N, 51:97, 51:141, 51:153:01-51:153:02, 51:263, 51:274-51:275, 51:342, 51:354, 51:362, 51:371, 52:19, 56:13
21	170 bp 270 bp	800 bp	*55:83N *55:10	A*01:411N *07:209, 07:271, 08:267, 15:506, 18:01:01:01-18:03:02, 18:05:01:01-18:15, 18:18:01:01-18:28, 18:30-18:79, 18:81-18:97, 18:99-18:101, 18:103-18:106Q, 18:108-18:111, 18:113-18:119, 18:121-18:124, 18:126-18:226, 27:75, 35:25, 35:124-35:125, 35:142, 35:162, 35:267, 35:320, 35:487,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				35:494, 35:557, 37:01:01:01-37:21, 37:23-37:40, 37:42N-37:94, 37:96-37:106, 39:189, 40:149 [?] , 44:130, 44:156, 44:221, 44:224, 44:258, 51:116, 51:176, 51:179, 51:218, 51:354, 52:16, 52:27, 52:77, 53:43:01-53:43:02, 54:21, 56:16, 57:58, 57:80, 58:12, 81:01:01:01-81:01:05, 81:03 [?] -81:04N [?] , 81:05, 81:06 [?] -81:08 [?] , 81:10, 82:01:01:01-82:04, 83:01
22⁴	95 bp	800 bp	*55:23, 55:27, 55:68	*07:84, 07:429, 13:04, 13:35, 13:72, 13:127, 15:04:01:01-15:04:04, 15:16:01:01-15:16:05, 15:67, 15:95, 15:155, 15:222, 15:254, 15:293, 15:310, 15:362, 15:408, 15:499, 15:517, 15:531, 15:550, 15:602, 15:609, 15:613, 35:37, 35:235, 35:265, 40:44, 40:159, 41:21, 44:20, 44:47, 44:100, 44:197, 45:09, 46:32:01-46:32:02, 49:04:01-49:05:02, 50:33, 51:37, 51:90, 51:92:01-51:92:02, 51:197, 51:216, 51:261, 51:274, 51:298, 52:12, 52:26, 52:63, 52:75, 54:14-54:15, 56:50, 58:122, C*03:464, C*07:463, C*07:758, C*12:183
	225 bp		*55:09, 55:21, 55:37, 55:52	*07:78, 07:84, 07:429, 13:18, 13:31, 13:41, 13:128, 13:131, 13:145, 13:164, 15:04:01:01-15:04:04, 15:137, 15:303, 15:499, 15:531, 15:609, 35:265, 39:33, 40:161, 46:32:01-46:32:02, 49:41, 50:47, 51:05:01:01-51:05:01:02, 51:29, 51:54, 51:61:01-51:61:02, 51:82, 51:197, 51:217, 51:279, 51:376, 52:21:01-52:21:02, 52:24, 52:97, 54:10, 54:20, 54:33, 54:40, 56:43, C*05:152, C*12:183, C*14:92
23	165 bp 220 bp	1070 bp	*55:29 *55:16, 55:21	*13:105, 51:174:02 *07:143, 13:31, 13:41, 13:131, 13:145, 13:164, 15:04:01:01-15:04:04, 15:137, 15:303, 15:499, 15:531, 15:609, 27:157, 35:265, 40:166, 42:02:01:01-42:02:02, 42:09, 42:17-42:18, 42:27, 45:06, 46:32:01-46:32:02, 50:47, 51:61:01-51:61:02, 52:21:01-52:21:02, 52:97, 54:02, 54:33, 54:40, 56:35, 56:43, C*12:183
24⁴	100 bp	1070 bp	*55:13, 55:23, 55:27, 55:32, 55:58, 55:68	*07:84, 07:429, 13:04, 13:35, 13:46, 13:72, 13:127, 13:140, 13:160, 15:04:01:01-15:04:04, 15:16:01:01-15:16:05, 15:67, 15:95, 15:137, 15:155, 15:222, 15:254, 15:293, 15:310, 15:362, 15:408, 15:499, 15:517, 15:531, 15:550, 15:602, 15:609, 15:613, 35:37, 35:235, 35:265, 35:276, 39:06:01:01-39:06:08, 39:33-39:34, 39:50, 39:57, 39:62, 39:64, 39:83:01-39:83:03, 39:90, 39:102, 39:109, 39:127, 39:129, 39:132, 39:140-39:141, 39:154, 39:178, 39:186-39:187, 40:44, 40:86, 40:159, 40:330, 41:21, 41:30, 41:50, 44:20, 44:47, 44:100, 44:197, 45:02, 45:09, 46:32:01-46:32:02, 49:04:01-49:05:02, 49:50, 50:09, 50:33, 51:13:01-51:13:02, 51:37, 51:63:01-51:63:02, 51:90, 51:92:01-51:92:02, 51:97, 51:172, 51:197, 51:216, 51:251, 51:261, 51:274, 51:298, 52:12, 52:14:01-52:14:02, 52:26, 52:39, 52:63, 52:75, 52:97, 54:14-54:15, 56:22, 56:50, 58:122, 59:03,



0197

For *In Vitro* Diagnostic Use
 MA123 v02 SSP PI Template
 Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

				73:01:01:01-73:03, C*03:464, C*05:152, C*07:463, C*07:758, C*12:183
	235 bp		*55:33	
25⁵	155 bp 280 bp 380 bp	1070 bp	*55:62 *55:26 *55:18, 55:34, 55:40	*13:125, 27:196, 40:10:01:01-40:10:02, 40:476, 44:344, 54:35, 56:67
26^{4,5}	75 bp 310 bp 400 bp	1070 bp	*55:42 *55:45 *55:34, 55:47	*13:88, 18:215, 44:362
27⁴	95 bp 170 bp	1070 bp	*55:46 *55:43	*50:39, 54:16 *13:155, 51:127
28^{4,5}	95 bp	800 bp	*55:41, 55:44, 55:58	*13:46, 15:137, 40:330, 49:54, 51:172, 51:223, 73:01:01:01-73:03
	215 bp		*55:57	
29	160 bp	1070 bp	*55:83N	*08:101, 13:132, 15:434, 35:501, 41:55, 44:24, 44:181, 46:86, 51:39, 54:01:01:01-54:01:10, 54:03-54:04, 54:06-54:38, 54:40-54:46, A*01:411N, C*03:610, C*05:30, C*06:125, C*07:441:01-07:441:02, C*08:249, C*16:86, C*16:144, C*16:147
30⁴	100 bp	1070 bp	*55:35, 55:50	*07:409, 08:83, 13:06, 13:53, 13:131, 14:43, 14:50, 15:76, 15:101, 18:74N, 18:78, 38:38, 41:37, 56:39, A*31:147, C*04:467, C*05:110, C*12:120
31⁴	105 bp	800 bp	*55:13, 55:58	*13:46, 13:140, 15:137, 35:235, 35:276, 39:06:01:01-39:06:08, 39:33-39:34, 39:50, 39:57, 39:62, 39:64, 39:83:01-39:83:03, 39:90, 39:102, 39:109, 39:127, 39:129, 39:132, 39:140-39:141, 39:154, 39:178, 39:186-39:187, 40:86, 40:330, 41:50, 45:02, 49:50, 50:09, 51:13:01-51:13:02, 51:92:01-51:92:02, 51:172, 52:14:01-52:14:02, 52:75, 56:22, 59:03, 73:01:01:01-73:03, C*05:152
	170 bp 235 bp		*55:43 *55:55N	*13:155, 51:127
32⁷			Negative Control	

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B*55 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.



0197

For *In Vitro* Diagnostic Use
MA123 v02 SSP PI Template
Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: 4R6

Lot-specific information

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁵Primer mixes 6, 25, 26 and 28 may have tendencies of unspecific amplifications.

⁶Primer mix 10 may give rise to a long fragment of approximately 650 bp in some HLA-B alleles. This band should not be considered in the interpretation of HLA-B*55 typings.

⁷Primer mix 32 contains a negative control, which will amplify a majority of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 200 base pairs.

Abbreviations

'?', nucleotide sequence of the primer matching region is not known.

'w', may be weakly amplified.



0197

For *In Vitro* Diagnostic Use
MA123 v02 SSP PI Template
Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: 4R6

Lot-specific information

PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec.	400	125	150	235	95	100	180	230	215	70	90	210
PCR product					235			425	375	220	125	180
Length of int.	800	1070	1070	1070	800	1070	1070	800	1070	1070	1070	1070
pos. control ¹												
5'-primer(s) ²	41 5'-CTg 3'	379 5'-ACC 3'	420 5'-TTA 3'	106 5'-CCA 3'	363 5'-AgC 3'	245 5'-ggC 3'	419 5'-gTA 3'	48 5'-gCC 3'	48 5'-gCC 3'	175 5'-TgA 3'	186 5'-AgA 3'	103 5'-CCT 3'
		379 5'-ACC 3'			506 5'-gCT 3'				272 5'-CTA 3'	379 5'-ACg 3'	420 5'-TTA 3'	
3'-primer(s) ³	272 5'-TgT 3'	463 5'-gCT 3'	527 5'-CCT 3'	299 5'-TCA 3'	559 5'-CgT 3'	302 5'-ggC 3'	559 5'-Cag 3'	107 5'-ACT 3'	94 5'-gAC 3'	206 5'-CCC 3'	272 5'-TgT 3'	272 5'-TgT 3'
	275 5'-CCA 3'							302 5'-ggT 3'	362 5'-TCA 3'	559 5'-CgT 3'	469 5'-CCg 3'	559 5'-CTC 3'
Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Length of spec.	150	95	245	245	105	215	205	220	170	95	165	100
PCR product						310			270	225	220	235
						415						
Length of int.	1070	1070	1070	1070	1070	800	1070	1070	800	800	1070	1070
pos. control ¹												
5'-primer(s) ²	420 5'-TTA 3'	206 5'-Agg 3'	357 5'-Tgg 3'	357 5'-Tgg 3'	357 5'-Tgg 3'	48 5'-ggC 3'	134 5'-CCA 3'	357 5'-Tgg 3'	41 5'-CTg 3'	357 5'-Tgg 3'	97 5'-TCC 3'	357 5'-Tgg 3'
			357 5'-Tgg 3'				141 5'-ATg 3'		668 5'-ggC 3'		357 5'-Tgg 3'	
3'-primer(s) ³	523 5'-ACA 3'	259 5'-gTT 3'	559 5'-CgT 3'	559 5'-CAg 3'	419 5'-CgT 3'	94 5'-gAC 3'	299 5'-TCT 3'	538 5'-gTC 3'	142 5'-TgA 3'	412 5'-gTC 3'	272 5'-TgT 3'	412 5'-gTC 3'
	527 5'-CCA 3'				424 5'-gTg 3'	187 5'-gTT 3'			799 5'-CTA 3'	539 5'-TCC 3'	479 5'-CCA 3'	419 5'-Cgg 3'
						292 5'-gTA 3'					538 5'-CCA 3'	419 5'-CgA 3'
						292 5'-gTA 3'						550 5'-CAT 3'
						292 5'-gTA 3'						
Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Well No.	25	26	27	28	29	30	31					
Length of spec.	155	75	95	95	160	100	105					
PCR product	280	310	170	215			170					
	380	400					235					
Length of int.	1070	1070	1070	800	1070	1070	800					
pos. control ¹												
5'-primer(s) ²	48 5'-gCC 3'	41 5'-CTg 3'	357 5'-Tgg 3'	362 5'-gAC 3'	106 5'-CCA 3'	209 5'-ggC 3'	357 5'-Tgg 3'					
	3 rd I	431			668	506						
	5'-Agg 3'	5'-CgA 3'			5'-ggC 3'	5'-gCT 3'						
	160 5'-gTT 3'	181 5'-gTT 3'	412 5'-gTC 3'	409 5'-ATA 3'	226 5'-CAC 3'	272 5'-TgT 3'	419 5'-CgA 3'					
3'-primer(s) ³	259 5'-CTC 3'	272 5'-Tgg 3'	487 5'-CgT 3'	427 5'-gTg 3'	799 5'-CTA 3'	559 5'-CgT 3'	487 5'-CgT 3'					
	259 5'-CTC 3'	463 5'-gCT 3'		535 5'-CTC 3'			549 5'-AgT 3'					
	682 5'-CCT 3'											
Well No.	25	26	27	28	29	30	31					



0197

For *In Vitro* Diagnostic Use
MA123 v02 SSP PI Template
Date: December 2022, Rev. No: 00

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: 4R6

Lot-specific information

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-B*55 SSP subtyping kit²																				
				Well																
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
				Prod. No.:	201777501	201777502	201777503	201777504	201777505	201777506	201777507	201777508	201777509	202023210	201777511	201777512	201777513	201777514	202023215	201777516
	IHCW cell line ¹	B*																		
1	9001 SA	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
2	9280 LK707	*52:01	*73:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	+
3	9011 E4181324	*52:01		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	+
4	9275 GU373	*15:10	*53:01	-	+	-	-	-	-	-	+	-	-	-	-	-	-	-	-	-
5	9009 KAS011	*37:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*39:01	*51:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	+
7	9020 QBL	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025 DEU	*35:01		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*54:01		+	+	-	-	-	-	-	-	-	-	-	-	-	+	+	+	-
11	9051 PITOUT	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB	*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9025 JESTHOM	*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*15:01	*15:20	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*08:01	*55:01	+	+	+	-	-	-	-	-	-	-	-	-	-	-	-	+	-
18	9257 32367	*14:01	*56:01	+	+	-	-	-	-	-	-	-	-	-	-	-	+	-	-	+
19	9038 BM16	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056 KOSE	*35:03		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*40:02	*56:02	+	+	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
24	9035 JBUSH	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*58:01		-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*07:05	*51:01	-	-	-	-	-	-	-	+	-	-	-	-	+	-	-	-	+
28	9320 BEL5GB	*44:02	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*42:01		-	-	-	-	+	-	-	-	-	-	-	-	+	-	-	-	-
31	9019 DUCAF	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG	*41:02		-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*44:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17	*15:01	*35:01	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
38	9099 LZL	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315 CML	*08:01	*27:05	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199	*13:02	*46:01	-	+	-	+	-	-	-	-	-	-	-	+	-	+	-	-	-
41	9055 H0301	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089	*46:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*46:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO	*42:01	*50:01	-	+	-	-	+	-	-	-	-	-	-	-	+	-	-	-	+
46	9013 SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-
47	9045 TUBO	*51:01		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-	+
48	9303 TER-ND	*35:01	*44:03	-	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-



101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

CELL LINE VALIDATION SHEET																			
HLA-B*55 SSP subtyping kit ²																			
				Well															
				17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
				201777517	201777518	201777519	201777520	201897921	201777522	201777523	201777524	201777525	201777526	201777527	201777528	201897929	202023230	201777531	
	IHWC cell line ¹		B*	Prod. No.:															
1	9001 SA		*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
2	9280 LK707		*52:01	*73:01	+	-	-	-	-	-	-	+	-	-	-	+	-	-	+
3	9011 E4181324		*52:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	
4	9275 GU373		*15:10	*53:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
5	9009 KAS011		*37:01		-	-	-	-	+	-	-	-	-	-	-	-	-	-	
6	9353 SM		*39:01	*51:01	+	-	-	-	-	-	-	-	-	-	-	-	-	-	
7	9020 QBL		*18:01		-	-	-	-	+	-	-	-	-	-	-	-	-	-	
8	9025 DEU		*35:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
9	9026 YAR		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
10	9107 LKT3		*54:01		-	-	-	-	-	-	-	-	-	-	-	+	-	-	
11	9051 PITOUT		*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
12	9052 DBB		*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
13	9025 JESTHOM		*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
14	9071 OLGA		*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
15	9075 DKB		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
16	9037 SWEIG007		*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
17	9282 CTM3953540		*08:01	*55:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
18	9257 32367		*14:01	*56:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
19	9038 BM16		*18:01		-	-	-	-	+	-	-	-	-	-	-	-	-	-	
20	9059 SLE005		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
21	9064 AMALA		*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
22	9056 KOSE		*35:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
23	9124 IHL		*40:02	*56:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
24	9035 JBUSH		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
25	9049 IBW9		*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
26	9285 WT49		*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
27	9191 CH1007		*07:05	*51:01	+	-	-	-	-	-	-	-	-	-	-	-	-	-	
28	9320 BEL5GB		*44:02	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
29	9050 MOU		*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
30	9021 RSH		*42:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
31	9019 DUCAF		*18:01		-	-	-	-	+	-	-	-	-	-	-	-	-	-	
32	9297 HAG		*41:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
33	9098 MT14B		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
34	9104 DHIF		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
35	9302 SSTO		*44:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
36	9024 KT17		*15:01	*35:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
37	9065 HHKB		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
38	9099 LZL		*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
39	9315 CML		*08:01	*27:05	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
40	9134 WHONP199		*13:02	*46:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
41	9055 H0301		*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
42	9066 TAB089		*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
43	9076 T7526		*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
44	9057 TEM		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
45	9239 SHJO		*42:01	*50:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
46	9013 SCHU		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	
47	9045 TUBO		*51:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	
48	9303 TER-ND		*35:01	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	

101.570-06 – including *Taq* polymerase
 101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
 “Instructions for Use” (IFU)

Lot No.: 4R6

Lot-specific information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 6, 8 to 10, 19, 26, 27 and 30 were available. The specificities of the primers in primer solutions 8, 10, 19, 26, 27 and 30 were tested by separately adding one or two additional 5'-primers, respectively one or three additional 3'-primers.

In primer solution 6 it was only possible to test the 3'-primer, the 5'-primer was not possible to test.

In primer solution 9 it was only possible to test the 5'-primers, the 3'-primers were not possible to test.

In primer solutions 2, 11, 15, 19 and 26 one 5'-primer was not possible to test.

In primer solutions 1, 8, 11, 13, 17, 18, 21, 23 to 29 and 31 one or more 3'-primers were not possible to test.

Additional primers in primer solutions 5, 11, 18, 21, 25 and 29 were tested by separately adding additional 3'-primers and/or additional 5'-primers.



0197

101.570-06 – including *Taq* polymerase
101.570-06u – without *Taq* polymerase

Visit www.caredx.com for
“Instructions for Use” (IFU)

Lot No.: **4R6**

Lot-specific information

ADDRESSES:

Manufacturer:

Caredx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-508 939 00

Fax: +46-8-717 88 18

E-mail: orders-se@caredx.com

Web page: www.caredx.com

CareDx Lab Solutions Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-653-78171

Fax: 610-344-7989

E-mail: orders-us@caredx.com

Web page: www.caredx.com

For information on CareDx distributors worldwide, contact **CareDx AB**.



0197

For *In Vitro* Diagnostic Use

MA123 v02 SSP PI Template

Date: December 2022, Rev. No: 00