

Olerup SSP[®] HLA-B*67

| | |
|----------------------------------|---|
| Product number: | 101.550-06u – without <i>Taq</i> polymerase |
| Lot number: | 16K |
| Expiry date: | 2012-May-01 |
| Number of tests: | 6 |
| Number of wells per test: | 4 |
| Storage - pre-aliquoted primers: | dark at -20°C |
| - PCR Master Mix: | -20°C |
| - Adhesive PCR seals | RT |
| - Product Insert | RT |

This Product Description is only valid for Lot No. 16K.

CHANGES COMPARED TO THE PREVIOUS *OLERUP SSP*[®] HLA-B*67 LOT.

The HLA-B*67 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP*[®] HLA-B*67 lot was made (**Lot No. 45F**).

The HLA-B*67 primer set is unchanged compared to the previous lot.

PRODUCT DESCRIPTION

HLA-B*67 SSP typing

CONTENT

The primer set contains 5'- and 3'-primers for identifying the B*67:01 to B*67:02 alleles.

PLATE LAYOUT

Each HLA-B*67 test consists of 4 PCR reactions in an 8 well cut PCR plate. Wells 5 to 8 are empty.

| | | | | | | | |
|---|---|---|---|-------|-------|-------|-------|
| 1 | 2 | 3 | 4 | empty | empty | empty | empty |
|---|---|---|---|-------|-------|-------|-------|

The 8 well PCR plate is marked with 'B67' in silver/gray ink.

Well No. 1 is marked with the Lot No. '16K'.

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded.

The PCR plates are heat-sealed with a PCR-compatible foil.

Please note: When removing each 8 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

The interpretation of HLA-B*67 SSP subtypings will be influenced by most B*07, two B*15, the B*35:76, the B*38, most B*39, most B*42, the B*46, the B*54:02, two B*55, two B*56, the B*58:20, the B*73:01, the B*81 and the B*83:01 alleles when present on the other haplotype.

UNIQUELY IDENTIFIED ALLELES

All the HLA-B*67, i.e. **B*67:01 to B*67:02**, recognized by the HLA Nomenclature Committee in April 2010¹ will be amplified by the primers in the HLA-B*67 SSP kit. The HLA-B*67 subtyping kit cannot distinguish the B*67:01:01 and B*67:01:02 alleles.

¹HLA-B alleles listed on the IMGT/HLA web page 2010-April-01, release 3.0.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

The two HLA-B*67 alleles can be combined in 3 homozygous and heterozygous combinations. All these genotypes give rise to unique amplification patterns.

SPECIFICITY TABLE

HLA-B*67 SSP subtyping

Specificities and sizes of the PCR products of the 4 primer mixes used for HLA-B*67 SSP subtyping

| Primer Mix | Size of spec. PCR product ¹ | Size of control band ² | Amplified HLA-B*67 alleles | Other amplified HLA-B alleles ³ |
|----------------|--|-----------------------------------|----------------------------|---|
| 1 ⁴ | 210 bp | 800 bp | *67:01:01-67:01:02 | *07:02:01-07:02:19, 07:04-07:07, 07:09, 07:11-07:12, 07:14-07:15, 07:17-07:26, 07:28, 07:30-07:31, 07:33-07:36, 07:39-07:49N, 07:51-07:64, 07:66-07:68, 07:73-07:82, 07:84, 07:87-07:99, 42:01:01-42:01:02, 42:02 ^w , 42:04-42:06, 42:08, 42:09 ^w , 42:10, 42:12-42:13, 54:02 ^w , 55:10, 81:01-81:04N |
| 2 | 325 bp | 1070 bp | *67:02 | |
| 3 ⁵ | 165 bp | 800 bp | *67:01:01-67:02 | *38:01:01, 38:01:03-38:07, 38:09, 38:11-38:21, 39:01:01:01-39:01:01:02L, 39:01:03-39:06:02, 39:08-39:13:02, 39:15-39:16, 39:18-39:20, 39:22-39:24, 39:26-39:28, 39:31, 39:35, 39:37-39:42, 39:44-39:46, 39:48-39:49, 39:51-39:57, 58:20 |
| 4 | 160 bp | 1070 bp | | *07:02:01-07:02:02, 07:02:04-07:02:19, 07:04-07:05:03, 07:06-07:07, 07:09-07:15, 07:17-07:26, 07:28-07:31, 07:33-07:36, 07:39-07:46, 07:48-07:49N, 07:51-07:55, 07:57-07:68, 07:70-07:72, 07:74-07:84, 07:86-07:99, 15:76, 15:101, 35:76, 42:01:01-42:02, 42:04-42:06, 42:08-42:10, 42:12-42:13, 46:01:01, 46:01:02 ^w , 46:01:03-46:01:04, 46:02 ^w , 46:03-46:22, 55:35, 56:03, 56:06, 73:01 ^w , 81:01-81:04N, 83:01 |

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B*67 SSP typings.

When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective lengths of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Wells number 1 and 3 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band.

In the presence of a specific amplification the intensity of the control band often decreases.

³Due to the sharing of sequence motifs between HLA-B alleles non-HLA-B*67 alleles will be amplified by primer mixes 1, 3 and 4.

⁴Primer mix 1 has a tendency of giving rise to nonspecific amplifications.

⁵Primer mix 3 may give a lower yield of specific PCR product than other HLA-B*67 primer mixes. “w”, may be weakly amplified.

| INTERPRETATION TABLE | | | | |
|---|----------------------|----------------------|----------------------|----------------------|
| HLA-B*67 SSP subtyping | | | | |
| Amplification patterns of the B*67 alleles | | | | |
| | Well | | | |
| | 1 | 2 | 3 | 4 |
| Length of spec. | 210 | 325 | 165 | 160 |
| PCR product | | | | |
| Length of int. | 800 | 1070 | 800 | 1070 |
| pos. control ¹ | | | | |
| 5'-primer ² | 103 | 44 | 419 | 165 |
| | 5'-CCT ^{3'} | 5'-ggC ^{3'} | 5'-gTT ^{3'} | 5'-ACC ^{3'} |
| 3'-primer ³ | 272 | 201 | 544 | 282 |
| | 5'TgT ^{3'} | 5'-CTT ^{3'} | 5'-ggT ^{3'} | 5'gCC ^{3'} |
| Well No. | 1 | 2 | 3 | 4 |
| HLA-B allele | | | | |
| *67:01:01-67:01:02 | 1 | | 3 | |
| *67:02 | | 2 | 3 | |
| *07:02:01-07:02:02, 07:02:04-07:02:19, 07:04-07:05:03, 07:06-07:07, 07:09, 07:11-07:12, 07:14-07:15, 07:17-07:26, 07:28, 07:30-07:31, 07:33-07:36, 07:39-07:46, 07:48-07:49N, 07:51-07:55, 07:57-07:64, 07:66-07:68, 07:74-07:82, 07:84, 07:87-07:99, 42:01:01-42:01:02, 42:04-42:06, 42:08, 42:10, 42:12-42:13, 81:01-81:04N | 1 | | | 4 |
| *07:02:03, 07:05:04, 07:47, 07:56, 07:73, 55:10 | 1 | | | |
| *07:10, 07:13, 07:29, 07:65, 07:70-07:72, 07:83, 07:86, 15:76, 15:101, 35:76, 46:01:01, 46:01:03-46:01:04, 46:03-46:22, 55:35, 56:03, 56:06, 83:01 | | | | 4 |
| *38:01:01, 38:01:03-38:07, 38:09, 38:11-38:21, 39:01:01:01-39:01:01:02L, 39:01:03-39:06:02, 39:08-39:13:02, 39:15-39:16, 39:18-39:20, 39:22-39:24, 39:26-39:28, 39:31, 39:35, 39:37-39:42, 39:44-39:46, 39:48-39:49, 39:51-39:57, 58:20 | | | 3 | |
| *42:02, 42:09 | w | | | 4 |
| *46:01:02, 46:02, 73:01 | | | | w |
| *54:02 | w | | | |
| HLA-B allele | | | | |
| Well No. | 1 | 2 | 3 | 4 |

Lot No.: **16K**

Lot-specific information

www.olerup-ssp.com

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Wells number 1 and 3 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band.

²The nucleotide position, in the 1st, 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

“w”, may be weakly amplified.

| CELL LINE VALIDATION SHEET | | | | | | | | | |
|----------------------------|------|------------|--------|--------|------------|-----------|-----------|-----------|--|
| HLA-B*67 SSP primer set | | | | | | | | | |
| | | | | | Well | | | | |
| | | | | | 1 | 2 | 3 | 4 | |
| | | | | | 200854801 | 200854802 | 200854803 | 200854804 | |
| | | | | | Prod. No.: | | | | |
| IHC cell line | | | B* | | | | | | |
| 1 | 9001 | SA | *07:02 | | + | - | - | + | |
| 2 | 9280 | LK707 | *52:01 | *73:01 | - | - | - | w | |
| 3 | 9011 | E4181324 | *52:01 | | - | - | - | - | |
| 4 | 9275 | GU373 | *15:10 | *53:01 | - | - | - | - | |
| 5 | 9009 | KAS011 | *37:01 | | - | - | - | - | |
| 6 | 9353 | SM | *39:01 | *51:01 | - | - | + | - | |
| 7 | 9020 | QBL | *18:01 | | - | - | - | - | |
| 8 | 9025 | DEU | *35:01 | | - | - | - | - | |
| 9 | 9026 | YAR | *38:01 | | - | - | + | - | |
| 10 | 9107 | LKT3 | *54:01 | | - | - | - | - | |
| 11 | 9051 | PITOUT | *44:03 | | - | - | - | - | |
| 12 | 9052 | DBB | *57:01 | | - | - | - | - | |
| 13 | 9004 | JESTHOM | *27:05 | | - | - | - | - | |
| 14 | 9071 | OLGA | *15:01 | *15:20 | - | - | - | - | |
| 15 | 9075 | DKB | *40:01 | | - | - | - | - | |
| 16 | 9037 | SWEIG007 | *40:02 | | - | - | - | - | |
| 17 | 9282 | CTM3953540 | *08:01 | *55:01 | - | - | - | - | |
| 18 | 9257 | 32367 | *14:01 | *56:01 | - | - | - | - | |
| 19 | 9038 | BM16 | *18:01 | | - | - | - | - | |
| 20 | 9059 | SLE005 | *40:01 | | - | - | - | - | |
| 21 | 9064 | AMALA | *15:01 | | - | - | - | - | |
| 22 | 9056 | KOSE | *35:03 | | - | - | - | - | |
| 23 | 9124 | IHL | *40:02 | *56:02 | - | - | - | - | |
| 24 | 9035 | JBUSH | *38:01 | | - | - | + | - | |
| 25 | 9049 | IBW9 | *14:02 | | - | - | - | - | |
| 26 | 9285 | WT49 | *58:01 | | - | - | - | - | |
| 27 | 9191 | CH1007 | *07:05 | *51:01 | + | - | - | + | |
| 28 | 9320 | BEL5GB | *44:02 | *44:03 | - | - | - | - | |
| 29 | 9050 | MOU | *44:03 | | - | - | - | - | |
| 30 | 9021 | RSH | *42:01 | | + | - | - | + | |
| 31 | 9019 | DUCAF | *18:01 | | - | - | - | - | |
| 32 | 9297 | HAG | *41:02 | | - | - | - | - | |
| 33 | 9098 | MT14B | *40:01 | | - | - | - | - | |
| 34 | 9104 | DHIF | *38:01 | | - | - | + | - | |
| 35 | 9302 | SSTO | *44:02 | | - | - | - | - | |
| 36 | 9024 | KT17 | *15:01 | *35:01 | - | - | - | - | |
| 37 | 9065 | HHKB | *07:02 | | + | - | - | + | |
| 38 | 9099 | LZL | *15:01 | | - | - | - | - | |
| 39 | 9315 | CML | *08:01 | *27:05 | - | - | - | - | |
| 40 | 9134 | WHONP199 | *13:02 | *46:01 | - | - | - | + | |
| 41 | 9055 | H0301 | *14:02 | | - | - | - | - | |
| 42 | 9066 | TAB089 | *46:01 | | - | - | - | + | |
| 43 | 9076 | T7526 | *46:01 | | - | - | - | + | |
| 44 | 9057 | TEM | *38:01 | | - | - | + | - | |
| 45 | 9239 | SHJO | *42:01 | *50:01 | + | - | - | + | |
| 46 | 9013 | SCHU | *07:02 | | + | - | - | + | |
| 47 | 9045 | TUBO | *51:01 | | - | - | - | - | |
| 48 | 9303 | TER-ND | *35:01 | *44:03 | - | - | - | - | |

CERTIFICATE OF ANALYSIS

Olerup SSP[®] HLA-B*67 SSP

Product number: 101.550-06u – without *Taq* polymerase
Lot number: 16K
Expiry date: 2012-May-01
Number of tests: 6
Number of wells per test: 4

Well specifications:

| Well No. | Production No. |
|----------|----------------|
| 1 | 2008-548-01 |
| 2 | 2008-548-02 |
| 3 | 2008-548-03 |
| 4 | 2008-548-04 |

The specificity of each primer solution of the HLA-B*67 primer set has been tested against 48 well characterized IHWC cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solution 2 were available. The specificities of the primers in primer solution 2 was tested by separately adding one additional 5'-primer, respectively one additional 3'-primer.

Results: No false positive or false negative amplifications were obtained.

Date of approval: 2010-May-12

Approved by:

Quality Control, Supervisor

Lot No.: **16K**

Lot-specific information

www.olerup-ssp.com

Declaration of Conformity

Product name: *Olerup* SSP® HLA-B*67
Product number: 101.550-06u
Lot number: 16K

Intended use: HLA-B*67 high resolution histocompatibility testing

Manufacturer: *Olerup* SSP AB
Hasselstigen 1
SE-133 33 Saltsjöbaden, Sweden
Phone: +46-8-717 88 27
Fax: +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2008 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex II List B, conformity assessed using Annex IV, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Notified Body: Lloyd's Register Quality Assurance Limited, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry CV3 4FJ, United Kingdom. (Notified Body number: 0088.)

Saltsjöbaden, Sweden
2010-May-12

Olle Olerup
Managing Director

Lot No.: **16K**

Lot-specific information

www.olerup-ssp.com

ADDRESSES:

Manufacturer:

Olerup SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: info-ssp@olerup.com

Web page: <http://www.olerup-ssp.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: support-at@olerup.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: info.us@olerup.com

Web page: <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.