

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information
Olerup SSP® HLA-B*08

Product number:	101.513-24/04 – including <i>Taq</i> pol. 101.513-24u/04u – without <i>Taq</i> pol.
Lot number:	4G9
Expiry date:	2021-02-01
Number of tests:	24 tests – Product No. 101.513-24/24u 4 tests – Product No. 101.513-04/04u
Number of wells per test:	47+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 4G9.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP®
HLA-B*08 LOT (3E6)**

The HLA-B*08 kit is updated for new alleles to enable separation of:

- Null and Alternatively expressed alleles
- The product documentation has been updated for new alleles of IMGT 3.31.0.

The format of the Worksheet has been changed.

The HLA-B*08 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP®* HLA-B*08 lot was made (**Lot No. 3E6**).

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
6	-	Added	3'-primer added for the B*08:97 allele.
25	Added	-	5'-primer added from primer mix 45.
27	-	Modified	3'-primer modified for improved HLA-specific amplification, 3'-primer modified for the B*08:30N allele.
30	-	Modified	3'-primer modified for improved HLA-specific amplification.
37	Added	-	5'-primer added for the B*08:165 allele.
44	-	Added	3'-primer added for the B*08:97 allele.
45	Moved	-	5'-primer moved to primer mix 25 for improved HLA-specific amplification.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

Well **48** contains Negative Control primer pairs, that will amplify more than 95% of the *Olerup SSP*® HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.
The PCR product generated by the positive control primer pair is 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

PRODUCT DESCRIPTION

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

HLA-B*08 SSP subtyping

CONTENT

The primer set contains 5'- and 3'-primers for identifying the B*08:01 to B*08:190 alleles.

PLATE LAYOUT

Each test consists of 48 PCR reactions in a 48 well cut PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16
17	18	19	20	21	22	23	24
25	26	27	28	29	30	31	32
33	34	35	36	37	38	39	40
41	42	43	44	45	46	47	NC

The 48 well cut PCR plate is marked with ‘HLA-B*08’ in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘4G9’.

Wells 1 to 47 – HLA-B*08 high resolution primers.

Well 48 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded. The PCR plates are covered with a PCR-compatible foil.

Please note: When removing each 48 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between HLA-B alleles, non-HLA-B*08 alleles will be amplified by primer mixes. For further details see Specificity Table.

UNIQUELY IDENTIFIED ALLELES

All the HLA-B*08 alleles, i.e. **B*08:01 to B*08:190**, recognized by the HLA Nomenclature Committee in January 2018^{1,2} will be amplified by the primers in the HLA-B*08 subtyping kit^{3,4}.

The HLA-B*08 kit enables separation of the confirmed HLA-B*08 alleles as listed in the IMGT/HLA database 3.24.0. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-B*08 alleles is listed below.

The HLA-B*08 kit also enables identification of polymorphisms in exons outside of the region encoding the peptide binding domain and of null and alternatively expressed alleles.

The following HLA-B*08 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

Alleles	Primer mix	Alleles	Primer mix
B*08:11, 08:76	13	B*08:45, 08:59:01-08:59:02	34
B*08:19N, 08:109	22	B*08:51, 08:57	35
B*08:21, 08:47	23	B*08:58, 08:61	38
B*08:33, 08:53:01-08:53:02	36	B*08:68, 08:69	45
B*08:41, 08:48	37	B*08:101, 08:128	24
B*08:44, 08:145	19	B*08:104, 08:140	46

¹HLA-B alleles listed on the IMGT/HLA web page 2018-january-19, release 3.31.0, www.ebi.ac.uk/imgt/hla.

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

³The HLA-B*08 primer set cannot separate the B*08:156 and the B*15:180 alleles. These alleles can be distinguished by the HLA-B low resolution kit and/or the HLA-B*35 or the HLA-B*15 high resolution kit.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

ALLELE CONFIRMATION STATUS

Allele	Status ¹	Allele	Status ¹	Allele	Status ¹	Allele	Status ¹
B*08:01:01	Confirmed	B*08:17	Unconfirmed	B*08:62	Confirmed	B*08:112	Unconfirmed
B*08:01:02	Confirmed	B*08:18	Confirmed	B*08:63	Confirmed	B*08:113	Confirmed
B*08:01:03	Unconfirmed	B*08:19N	Unconfirmed	B*08:64	Unconfirmed	B*08:114	Unconfirmed
B*08:01:04	Confirmed	B*08:20	Confirmed	B*08:65	Unconfirmed	B*08:115	Unconfirmed
B*08:01:05	Confirmed	B*08:21	Unconfirmed	B*08:66	Unconfirmed	B*08:116	Unconfirmed
B*08:01:06	Unconfirmed	B*08:22	Unconfirmed	B*08:67N	Unconfirmed	B*08:117	Unconfirmed
B*08:01:07	Confirmed	B*08:23	Confirmed	B*08:68	Unconfirmed	B*08:118	Unconfirmed
B*08:01:08	Confirmed	B*08:24	Confirmed	B*08:69	Unconfirmed	B*08:119	Unconfirmed
B*08:01:09	Unconfirmed	B*08:25	Confirmed	B*08:70	Confirmed	B*08:120	Unconfirmed
B*08:01:10	Confirmed	B*08:26:01	Unconfirmed	B*08:71	Confirmed	B*08:121	Unconfirmed
B*08:01:11	Unconfirmed	B*08:26:02	Unconfirmed	B*08:72N	Unconfirmed	B*08:122	Unconfirmed
B*08:01:12	Unconfirmed	B*08:26:03	Unconfirmed	B*08:73	Unconfirmed	B*08:123	Unconfirmed
B*08:01:13	Confirmed	B*08:27	Unconfirmed	B*08:74	Unconfirmed	B*08:124	Unconfirmed
B*08:01:14	Unconfirmed	B*08:28	Confirmed	B*08:75	Unconfirmed	B*08:125	Unconfirmed
B*08:01:15	Confirmed	B*08:29	Unconfirmed	B*08:76	Unconfirmed	B*08:126	Unconfirmed
B*08:01:16	Unconfirmed	B*08:30N	Unconfirmed	B*08:77	Unconfirmed	B*08:127	Unconfirmed
B*08:01:17	Confirmed	B*08:31	Unconfirmed	B*08:78	Confirmed	B*08:128	Confirmed
B*08:01:18	Unconfirmed	B*08:32	Unconfirmed	B*08:79	Confirmed	B*08:129	Unconfirmed
B*08:01:19	Unconfirmed	B*08:33	Confirmed	B*08:80	Unconfirmed	B*08:130	Unconfirmed
B*08:01:20	Unconfirmed	B*08:34	Unconfirmed	B*08:81	Unconfirmed	B*08:131	Unconfirmed
B*08:01:21	Confirmed	B*08:35	Confirmed	B*08:82N	Unconfirmed	B*08:132	Confirmed
B*08:01:22	Unconfirmed	B*08:36	Unconfirmed	B*08:83	Unconfirmed	B*08:133	Unconfirmed
B*08:01:23	Unconfirmed	B*08:37	Confirmed	B*08:84	Confirmed	B*08:134	Unconfirmed
B*08:01:24	Unconfirmed	B*08:38	Unconfirmed	B*08:85	Unconfirmed	B*08:135	Confirmed
B*08:01:25	Unconfirmed	B*08:39	Confirmed	B*08:86N	Unconfirmed	B*08:136	Confirmed
B*08:01:26	Unconfirmed	B*08:40	Confirmed	B*08:87	Unconfirmed	B*08:137	Confirmed
B*08:01:27	Unconfirmed	B*08:41	Confirmed	B*08:88	Unconfirmed	B*08:138	Unconfirmed
B*08:01:28	Unconfirmed	B*08:42	Unconfirmed	B*08:89	Unconfirmed	B*08:139	Unconfirmed
B*08:01:29	Confirmed	B*08:43	Unconfirmed	B*08:90	Unconfirmed	B*08:140	Confirmed
B*08:01:30	Confirmed	B*08:44	Unconfirmed	B*08:91	Unconfirmed	B*08:141	Unconfirmed
B*08:01:31	Unconfirmed	B*08:45	Unconfirmed	B*08:92	Unconfirmed	B*08:142	Unconfirmed
B*08:01:32	Unconfirmed	B*08:46	Unconfirmed	B*08:93	Unconfirmed	B*08:143	Unconfirmed
B*08:01:33	Unconfirmed	B*08:47	Confirmed	B*08:94	Confirmed	B*08:144	Unconfirmed
B*08:01:34	Unconfirmed	B*08:48	Unconfirmed	B*08:95	Unconfirmed	B*08:145	Confirmed
B*08:02	Confirmed	B*08:49	Unconfirmed	B*08:96	Unconfirmed	B*08:146	Unconfirmed
B*08:03	Confirmed	B*08:50	Unconfirmed	B*08:97	Unconfirmed	B*08:147	Unconfirmed
B*08:04	Confirmed	B*08:51	Unconfirmed	B*08:98	Unconfirmed	B*08:148N	Unconfirmed
B*08:05	Unconfirmed	B*08:52	Confirmed	B*08:99	Unconfirmed	B*08:149	Unconfirmed
B*08:07	Unconfirmed	B*08:53:01	Confirmed	B*08:100	Confirmed	B*08:150	Confirmed
B*08:08N	Confirmed	B*08:53:02	Unconfirmed	B*08:101	Confirmed	B*08:151	Unconfirmed
B*08:09	Confirmed	B*08:54	Unconfirmed	B*08:102	Unconfirmed	B*08:152	Unconfirmed
B*08:10	Unconfirmed	B*08:55	Unconfirmed	B*08:103	Unconfirmed	B*08:153	Unconfirmed
B*08:11	Unconfirmed	B*08:56:01	Confirmed	B*08:104	Confirmed	B*08:154	Unconfirmed
B*08:12:01	Confirmed	B*08:56:02	Confirmed	B*08:105	Confirmed	B*08:155	Unconfirmed
B*08:12:02	Unconfirmed	B*08:57	Unconfirmed	B*08:106	Unconfirmed	B*08:156	Unconfirmed
B*08:12:03	Confirmed	B*08:58	Unconfirmed	B*08:107	Unconfirmed	B*08:157	Unconfirmed
B*08:13	Confirmed	B*08:59:01	Confirmed	B*08:108	Confirmed	B*08:158	Unconfirmed
B*08:14	Unconfirmed	B*08:59:02	Unconfirmed	B*08:109	Unconfirmed	B*08:159	Unconfirmed
B*08:15	Unconfirmed	B*08:60	Confirmed	B*08:110	Confirmed	B*08:160	Unconfirmed
B*08:16	Confirmed	B*08:61	Unconfirmed	B*08:111	Unconfirmed	B*08:161	Unconfirmed

¹Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2016-April-15, release 3.24.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-B*08 homo- and heterozygotes is available upon request.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information
SPECIFICITY TABLE

HLA-B*08 SSP subtyping

Specificities and sizes of the PCR products of the 47+1 primer mixes used for HLA-B*08 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified HLA-B*08 alleles ³	Other amplified HLA Class I alleles ⁴
1	215 bp	800 bp	*08:01:01:01-08:05, 08:07-08:08N, 08:10-08:11, 08:14-08:15, 08:17-08:19N, 08:21-08:24, 08:26:01-08:39, 08:41-08:48, 08:50-08:54:02, 08:56:01-08:59:02, 08:61-08:69, 08:71-08:78, 08:80-08:83, 08:85-08:88, 08:90-08:106, 08:108-08:128, 08:130-08:142, 08:144-08:155, 08:157-08:170, 08:172-08:180, 08:182-08:183, 08:185-08:190	*07:04, 07:19, 07:25, 07:146, 35:87, 37:09, 41:02:01-41:02:07, 41:04, 41:10-41:11, 41:13, 41:15, 41:18-41:19, 41:23-41:24, 41:27, 41:31, 41:36, 41:38-41:43, 41:45N-41:47, 41:49, 41:51-41:52, 42:01:01-42:02:01:02, 42:05:01-42:12, 42:14-42:18, 42:20-42:24, 44:106, 44:158, 44:166, 44:213
2	215 bp	800 bp	*08:01:01:01-08:03, 08:05, 08:07-08:16, 08:18-08:25, 08:27-08:49, 08:51-08:53:02, 08:55-08:61, 08:63-08:64, 08:66-08:84, 08:86N-08:93, 08:95-08:109, 08:111-08:122, 08:124-08:131, 08:133-08:145, 08:147-08:190	*15:180, 35:218, 35:256, 51:68, 51:176, C*06:147
3	245 bp	1070 bp	*08:01:01:01-08:01:36, 08:04:01-08:05, 08:07-08:14, 08:16-08:25, 08:27-08:49, 08:51, 08:53:01-08:61, 08:63-08:77, 08:79-08:93, 08:95-08:116, 08:118-08:125, 08:127-08:130, 08:132-08:138, 08:140-08:190	*07:29, 07:186, 14:28, 15:142, 15:180, 15:429, 35:218, 35:256, 39:12, 40:328, C*06:147, C*07:124, C*07:155
4	245 bp	1070 bp	*08:02, 08:113, 08:117	*13:93, 44:49, 44:156
5	205 bp 245 bp	1070 bp	*08:148N *08:03, 08:23, 08:52, 08:54:01-08:54:02, 08:78, 08:110, 08:133	*15:142, 15:429, 35:218, 35:256, 40:328, 49:25, 51:68, 51:176, 57:49, C*06:147
6	205 bp	1070 bp	*08:04:01-08:04:02, 08:17, 08:24, 08:54:01-08:54:02, 08:97, 08:110	*13:93, 15:142, 15:429, 37:62, 40:328, 44:49, 44:156, 49:25, C*07:01:06

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

7⁵	215 bp 250 bp	1070 bp	*08:05 *08:03, 08:36, 08:78	*49:25, 51:68, 51:176, 57:49, C*07:377
8⁶	170 bp	1070 bp	*08:20, 08:40, 08:70, 08:79	*07:05:01:01-07:06:02, 07:32, 07:40, 07:53, 07:69, 07:78, 07:90, 07:97, 07:105, 07:112, 07:123, 07:137- 07:138, 07:140, 07:176, 07:182N, 07:201N, 07:206-07:207, 07:210, 07:213-07:214, 07:222, 07:249, 07:258, 07:262, 07:264, 07:269- 07:270, 07:278, 07:283, 07:287, 07:293, 07:304, 13:18, 13:31, 13:41, 13:73, 15:30:01:01-15:30:01:02, 15:58, 15:73, 15:150, 15:303, 15:324, 35:18, 35:44, 39:08, 39:11, 39:18, 39:33, 39:36, 40:15-40:16, 40:23, 40:128, 40:161, 40:198, 40:324, 46:61, 48:05, 48:19, 51:05, 51:29, 51:54, 51:61:01-51:61:02, 51:82, 51:217, 52:21:01-52:21:02, 52:24, 54:10, 54:20, 54:33, 55:09, 55:21, 55:37, 55:52, 56:43, 56:51, 57:02:01-57:02:02, 57:12, 57:28N, 57:42, 57:63, C*01:90, C*01:116, C*02:68
9⁴	115 bp	1070 bp	*08:49, 08:71	*13:46, 15:51, 15:83, 15:179:01- 15:179:02, 15:199, 15:218Q, 18:56, 18:113, 27:47, 27:127, 35:38, 35:115, 35:169, 37:01:01:01- 37:01:11, 37:03N-37:06:02, 37:08, 37:10-37:11, 37:13-37:38, 37:41- 37:65, 37:67-37:70, 38:30, 41:24, 42:10, 42:17, 44:02:01:01-44:02:48, 44:05:01-44:06, 44:08-44:09, 44:11- 44:12, 44:14-44:25, 44:27:01- 44:27:02, 44:33-44:34:02, 44:41:01- 44:42, 44:44, 44:48-44:49, 44:51- 44:52N, 44:55, 44:58N-44:59:02, 44:62-44:63, 44:66-44:68, 44:70- 44:73, 44:75, 44:78, 44:80, 44:83- 44:84:02, 44:86-44:91, 44:93, 44:95, 44:97, 44:99-44:102, 44:104, 44:106, 44:112-44:113, 44:116- 44:119, 44:123, 44:126:01-44:127, 44:131-44:134, 44:136-44:140, 44:142, 44:145, 44:148-44:149N, 44:151-44:152, 44:158, 44:162, 44:168-44:173, 44:176-44:177, 44:179, 44:185, 44:187, 44:190- 44:191, 44:195N-44:196, 44:200-44:201, 44:206, 44:208, 44:211-44:214, 44:216-44:221, 44:225-44:226, 44:229-44:230, 44:235-44:236, 44:238,

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

	165 bp		*08:07, 08:37, 08:89, 08:107	44:240-44:244, 44:249, 44:253-44:255, 44:257, 44:260-44:265, 44:267N, 44:269-44:270, 44:273-44:274, 45:01:01-45:21, 51:08:01-51:08:02, 51:20, 51:36, 51:44N, 51:97, 51:141, 51:153, 52:19, 53:22, 55:56, 56:13, 57:09, 57:24, 82:01-82:03, 83:01 *07:04, 07:19, 07:25, 07:146, 15:51, 15:179:01-15:179:02, 15:199, 15:218Q, 18:56, 18:113, 35:38, 35:115, 35:169, 35:260, 37:09, 37:35, 37:56, 41:04, 41:18, 41:21, 41:51, 42:09, 42:16, 44:17, 44:51, 44:123, 44:230, 45:09, 53:22, 57:24
10⁴	85 bp 170 bp	800 bp	*08:40, 08:165, 08:171 *08:08N	*27:83, 41:19
11	290 bp	800 bp	*08:09, 08:27, 08:84	*41:01:01-41:01:05, 41:05, 41:07, 41:12, 41:14, 41:16-41:17, 41:20-41:22, 41:25-41:26, 41:28, 41:30, 41:33-41:35, 41:37, 41:44, 41:50, 41:53-41:54, 42:04, 55:20
12	385 bp	1070 bp	*08:10	*39:78
13⁴	85 bp 290 bp	1070 bp	*08:11 *08:49, 08:60, 08:76, 08:129, 08:181	*41:48, 42:13, 53:15
14⁴	115 bp 280 bp	1070 bp	*08:43 *08:12:01-08:12:03, 08:16, 08:49, 08:89, 08:107, 08:129	*41:03:01-41:03:02, 41:48, 42:13, 53:15
15⁴	75 bp	1070 bp	*08:01:01:01-08:01:34, 08:01:36-08:05, 08:07-08:25, 08:27-08:49, 08:50 ^w , 08:51-08:55, 08:57-08:61, 08:62 ^w , 08:63-08:65, 08:67N-08:93, 08:96-08:101, 08:103-08:121, 08:123-08:150, 08:152-08:179, 08:181-08:189	*07:29, 07:86 ^w , 14:11 ^w , 14:28, 15:03:01:01 ^w -15:03:06 ^w , 15:09:01 ^w -15:10:04 ^w , 15:18:01:01 ^w -15:18:06 ^w , 15:23 ^w , 15:29 ^w , 15:37 ^w , 15:47:01 ^w -15:47:02 ^w , 15:49 ^w , 15:51 ^w -15:52 ^w , 15:54 ^w , 15:61 ^w -15:62 ^w , 15:64:01 ^w -15:64:02 ^w , 15:69 ^w , 15:72 ^w , 15:74 ^w , 15:80 ^w , 15:90 ^w -15:91 ^w , 15:93 ^w , 15:98 ^w -15:99 ^w , 15:103 ^w , 15:108 ^w , 15:114 ^w -15:115 ^w , 15:119 ^w , 15:123 ^w -15:124 ^w , 15:127 ^w , 15:131 ^w -15:134 ^w , 15:151 ^w , 15:153 ^w , 15:156 ^w , 15:158 ^w , 15:161 ^w , 15:173 ^w , 15:176 ^w , 15:180, 15:186 ^w , 15:197:01 ^w -15:198 ^w , 15:200 ^w , 15:210 ^w , 15:220 ^w -15:221 ^w , 15:226N ^w , 15:229 ^w , 15:235 ^w , 15:238 ^w , 15:242:01 ^w -15:243 ^w , 15:251 ^w -15:253 ^w , 15:255 ^w , 15:263 ^w , 15:266 ^w , 15:274 ^w -15:275:02 ^w , 15:281 ^w -15:282 ^w , 15:290 ^w , 15:292 ^w -15:294N ^w , 15:306 ^w -15:307 ^w , 15:311 ^w -15:314 ^w , 15:323 ^w , 15:329 ^w , 15:335 ^w , 15:337 ^w -15:338 ^w , 15:351 ^w , 15:354 ^w , 15:369 ^w , 15:376 ^w , 15:379 ^w -15:380N ^w , 15:388 ^w -15:389 ^w ,

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

				15:397 ^w , 15:414 ^w , 15:426 ^w -15:427 ^w , 15:429, 15:433 ^w -15:435 ^w , 18:25 ^w , 18:68 ^w , 35:25 ^w , 35:124 ^w -35:125 ^w , 35:142 ^w , 35:320 ^w , 37:55 ^w , 38:33 ^w , 39:04 ^w , 40:12 ^w , 41:22 ^w , 44:156, 44:221 ^w , 44:258 ^w , 48:22 ^w , 49:25, 51:176, 51:179 ^w , 52:16 ^w , 52:27 ^w , 53:43 ^w , 54:21 ^w , 56:16 ^w , 57:58 ^w , 57:80 ^w , 58:12 ^w , 83:01 ^w , C*06:125 , C*07:441:01-07:441:02
16⁴	105 bp	800 bp	*08:13, 08:25, 08:55, 08:143, 08:171	*27:83, 42:19, 53:15
17⁴	95 bp	1070 bp	*08:14, 08:87	*07:28, 07:118, 27:32, 40:180, 44:54, 44:106, 44:135
18	425 bp	800 bp	*08:15	*07:13, 07:15, 07:92, 07:253 ^w , 07:304, 39:27, 40:73, 67:02, 73:01- 73:02
	470 bp		*08:39	*07:93, 38:27
19	245 bp 415 bp	1070 bp	*08:44 *08:16, 08:123, 08:145	*07:13, 07:15, 07:160, 42:18, 67:02
20	140 bp 205 bp	1070 bp	*08:22 *08:17, 08:38, 08:54:01- 08:54:02, 08:148N	*13:93, 15:142, 15:429, 37:62, 40:328, 44:49, 44:156, 49:25
21	465 bp	1070 bp	*08:18	*07:102, 73:01-73:02
22	130 bp 215 bp	800 bp	*08:19N *08:109	
23^{4,6}	85 bp 235 bp	800 bp	*08:21, 08:25 *08:17, 08:47	*41:17, 53:15 *13:39, 13:67, 15:57, 15:135, 35:27, 35:56, 35:185, 35:203, 35:355, 44:75, 44:254, 55:12, 56:41, 78:06, 78:09, C*05:30 , C*06:125
24⁴	85 bp 210 bp	1070 bp	*08:101 *08:26:01-08:26:03, 08:50, 08:94, 08:128	*51:39, C*06:147 *07:08:01-07:08:02, 07:32, 07:38, 07:207, 07:228:01-07:228:02, 14:42, 15:08:01-15:08:02, 15:29, 15:56, 15:191, 15:236, 35:01:01:01-35:19, 35:21-35:25, 35:27, 35:29:01-35:49, 35:51-35:62, 35:64:01-35:71, 35:74- 35:75, 35:77, 35:79-35:81, 35:83, 35:86-35:94, 35:96-35:134N, 35:136-35:161, 35:163-35:185, 35:187-35:216N, 35:219-35:225, 35:227-35:228, 35:230, 35:232- 35:254, 35:257-35:279, 35:281- 35:289, 35:291-35:293, 35:295, 35:297-35:316, 35:318-35:322, 35:324-35:360, 38:06-38:07, 39:58, 39:83, 44:06, 44:232, 48:06, 48:31, 51:01:01:01-51:06:03, 51:08:01- 51:21, 51:23-51:24:05, 51:26-51:46, 51:48-51:67, 51:69-51:102, 51:104- 51:115, 51:117-51:127, 51:129- 51:144, 51:146-51:150,

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

				51:152-51:166, 51:168-51:175, 51:177-51:188, 51:190-51:213, 51:215-51:221, 51:223-51:229, 52:06:01-52:06:03, 53:01:01- 53:08:02, 53:10-53:16, 53:18-53:27, 53:29-53:37, 53:39, 53:41-53:49, 59:01:01:01-59:07, 59:09-59:10N, 78:01:01:01-78:02:02, 78:04, 78:07- 78:09
25	180 bp 225 bp 250 bp	1070 bp	*08:49, 08:71 *08:28, 08:35, 08:37, 08:89, 08:107 *08:150	*27:83, 41:24, 42:10, 42:17, 53:15 *41:24, 41:30, 41:51, 42:09, 53:15
26	210 bp	1070 bp	*08:131, 08:136, 08:139	A*01:199, A*02:42, A*02:310, A*11:133, A*33:60
27⁴	105 bp 140 bp 215 bp	1070 bp	*08:30N *08:22 *08:29	*07:195, 40:281
28⁴	85 bp 210 bp	1070 bp	*08:31 *08:66	
29	150 bp 190 bp 250 bp	1070 bp	*08:32, 08:133, 08:180 *08:137 *08:36	*35:218, 35:256, 37:62, 51:68, 51:176, C*07:209 C*07:377
30	250 bp	1070 bp	*08:82N, 08:86N, 08:135	*35:89
31	380 bp	1070 bp	*08:34	
32	230 bp	1070 bp	*08:01:01:01-08:05, 08:08N-08:13, 08:15- 08:27, 08:29-08:30N, 08:32-08:36, 08:38-08:48, 08:50-08:73, 08:75-08:78, 08:80-08:86N, 08:88, 08:90-08:102, 08:104- 08:106, 08:108-08:110, 08:112, 08:114-08:115, 08:117-08:133, 08:135- 08:155, 08:157-08:174, 08:176-08:186, 08:188- 08:190	*35:87, 41:01:01-41:03:02, 41:05, 41:07, 41:11-41:14, 41:16-41:17, 41:19-41:20, 41:22-41:28, 41:30- 41:31, 41:33-41:48, 41:49 ^w , 41:50, 41:53-41:54, 42:01:01-42:02:01:02, 42:04, 42:06-42:07, 42:10-42:13, 42:15, 42:17-42:19, 42:21-42:24, 44:166, 55:20
33	220 bp 255 bp 285 bp	800 bp	*08:55 *08:135 *08:42	*38:19, 39:03:01:01-39:03:01:02, 39:14:01:01-39:14:01:02, 39:24:01- 39:24:02, 39:29, 39:37, 39:76, 39:120 *35:89
34	210 bp 305 bp	1070 bp	*08:45 *08:59:01-08:59:02	*40:228 [?]
35⁴	120 bp 205 bp 235 bp	1070 bp	*08:51 *08:57 *08:113	

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

36⁴	125 bp 270 bp	800 bp	*08:20, 08:53:01-08:53:02, 08:84 *08:33, 08:100	*42:19
37⁴	85 bp 155 bp 190 bp 270 bp	1070 bp	*08:40, 08:165, 08:171 *08:41 *08:105 *08:48	*27:83, 41:19 *35:192, 35:240, 51:141
38	180 bp 265 bp	1070 bp	*08:61, 08:105 *08:58	
39⁷	190 bp 225 bp	1070 bp	*08:46, 08:137 *08:132	C*06:195, C*07:506 *07:29, 07:186, C*02:14:01-02:14:02, C*02:107, C*03:39, C*03:67, C*03:344, C*05:43, C*06:05, C*06:44, C*06:124, C*06:138, C*06:198, C*07:39-07:40, C*07:177, C*07:210, C*07:238, C*07:328, C*07:335, C*07:563, C*08:37, C*15:23, C*15:63, C*15:138, C*16:21, C*16:80
40	265 bp	1070 bp	*08:26:01-08:26:03, 08:56:01-08:56:02, 08:180	*07:56:01-07:56:02, 07:72, 07:206, 14:54, 38:28, 39:63, 40:21 [?] , 40:58 [?] , 40:137 [?] , 40:248 [?] , 40:282 [?] , 40:325 [?] , 40:351 [?] , 40:364 [?] -40:365 [?] , 41:12, 41:36, 42:21, 49:03, 49:48
41	180 bp 210 bp	1070 bp	*08:108 *08:62, 08:94	*07:08:01-07:08:02, 07:32, 07:38, 07:207, 07:228:01-07:228:02, 18:07:01-18:07:02, 18:29, 18:79, 18:136, 27:23, 35:50, 35:84, 35:162, 35:197, 35:217, 35:231, 35:267, 35:280, 35:323, 37:21, 37:67, 38:06- 38:07, 39:58, 39:83, 40:07-40:08, 40:13, 40:68, 40:106, 40:162, 40:220, 40:232, 41:46, 48:06, 48:31, 51:57, 51:103, 51:116, 51:151, 51:218, 53:21
42	255 bp	1070 bp	*08:63, 08:100, 08:150	
43⁷	185 bp 215 bp	800 bp	*08:64 *08:70	*07:113 *07:55, 07:100, 15:07:01:01- 15:07:03, 15:45, 15:68, 15:126, 15:207, 15:324, 15:331, 15:405, 15:431, 46:12, 48:19, C*02:60
44⁵	170 bp 215 bp	1070 bp	*08:67N, 08:97 *08:65	C*07:01:06 *14:28, 39:12
45⁴	120 bp 225 bp	1070 bp	*08:68 *08:49, 08:69	*41:50, 42:06, 42:20
46	175 bp 295 bp	1070 bp	*08:140 *08:104	
47⁸	240 bp	1070 bp	*08:03, 08:72N, 08:86N	*49:25, 51:68, 51:176, 57:49
48⁹	-	-	Negative Control	

101.513-24/04 – including Taq pol., IFU-01
101.513-24u/04u – without Taq pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: 4G9

Lot-specific information

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B*08 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings. Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

⁵Primer mixes 7 and 44 may have a tendency of giving rise to primer oligomer formation.

⁶Primer mixes 8 and 23 may give rise to a lower yield of HLA-specific PCR product than the other B*08 primer mixes.

⁷Primer mixes 39 and 43 may have tendencies of unspecific amplifications.

⁸Primer mix 47 may give rise to a lower yield of HLA-specific PCR product than the other HLA-B*08 primer mixes in the B*08:03 alleles

⁹Primer mix 48 contains a negative control, which will amplify more than 95% of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 430 base pairs.

'w', might be weakly amplified.

'?', nucleotide sequence information not available for the primer matching sequence.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec. PCR product	215	215	245	245	205	205	215	170	115	85	290	385
Length of int. pos. control ¹	800	800	1070	1070	1070	1070	1070	1070	1070	800	800	1070
5'-primer(s) ²	363 5'-AgC 3'	97 5'-TCg 3'	97 5'-TCg 3'	97 5'-TCg 3'	97 5'-TCg 3'	97 5'-TCg 3'	97 5'-TCg 3'	412 5'-ATA 3'	412 5'-ATg 3'	475 5'-Cg 3'	355 5'-TCC 3'	44 5'-ggC 3'
				412 5'-ATA 3'					463 5'-TgA 3'	560 5'-CgA 3'	357 5'-Tgg 3'	
3'-primer(s) ³	538 5'-gTC 3'	272 5'-TgA 3'	299 5'-TCT 3'	309 5'-gTg 3'	259 5'-gTT 3'	226 5'-TAC 3'	269 5'-Agg 3'	539 5'-TCC 3'	538 5'-gTC 3'	603 5'-gTg 3'	603 5'-gTg 3'	259 5'-gTC 3'
			299 5'-TCT 3'	604 5'-CgC 3'	292 5'-gTA 3'	251 5'-CCg 3'	302 5'-ggT 3'	539 5'-TCC 3'				
					309 5'-ATC 3'	272 5'-Tgg 3'	307 5'-CCA 3'					
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Length of spec. PCR product	85	115	75	105	95	425	245	140	465	130	85	85
Length of int. pos. control ¹	1070	1070	1070	800	1070	800	1070	1070	1070	800	800	1070
5'-primer(s) ²	355 5'-TCA 3'	363 5'-Agg 3'	106 5'-CCA 3'	539 5'-gCT 3'	363 5'-AgC 3'	44 5'-ggC 3'	44 5'-ggC 3'	97 5'-TCg 3'	44 5'-ggC 3'	636 5'-CAC 3'	106 5'-CCA 3'	97 5'-TCT 3'
	560 5'-CgC 3'	363 5'-Agg 3'						363 5'-AgC 3'			560 5'-CCT 3'	107 5'-CAC 3'
		530 5'-ggT 3'										228 5'-gTg 3'
3'-primer(s) ³	603 5'-gTg 3'	603 5'-gTg 3'	142 5'-TgA 3'	603 5'-gTg 3'	418 5'-gTC 3'	299 5'-TCA 3'	119 5'-ggg 3'	259 5'-gTT 3'	341 5'-CgT 3'	724 5'-CTA 3'	302 5'-ggT 3'	272 5'-TgA 3'
						343 5'-T 3'	289 5'-AgC 3'	261 5'-gTC 3'		809 5'-CAA 3'	302 5'-ggT 3'	
							289 5'-AgC 3'	461 5'-gCg 3'			603 5'-gTg 3'	
Well No.	13	14	15	16	17	18	19	20	21	22	23	24

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

Well No.	25	26	27	28	29	30	31	32	33	34	35	36
Length of spec.	180	210	105	85	150	250	380	230	220	210	120	125
PCR product	225		140	210	190				255	305	205	270
	250		215		250				285		235	
Length of int. pos. control ¹	1070	1070	1070	1070	1070	1070	1070	1070	800	1070	1070	800
5'-primer(s) ²	395 5'-gCC 3'	142 5'-TCT 3'	317 5'-gCA 3'	103 5'-CCA 3'	97 5'-TCg 3'	363 5'-AgC 3'	319 5'-gCT 3'	412 5'-ATA 3'	363 5'-AgC 3'	44 5'-ggC 3'	108 5'-ATC 3'	364 5'-gCT 3'
	419 5'-gTC 3'		363 5'-AgC 3'	610 5'-TgA 3'				412 5'-ATA 3'			194 5'-Cgg 3'	385 5'-ggT 3'
	463 5'-TgA 3'										412 5'-ATA 3'	511 5'-AgC 3'
												527 5'-TgA 3'
3'-primer(s) ³	603 5'-gTg 3'	311 5'-ggg 3'	2 nd 5'-TCg 3'	272 5'-TgA 3'	206 5'-CCg 3'	573 5'-AgT 3'	412 5'-gTT 3'	603 5'-gTg 3'	544 5'-ggT 3'	85 5'-CAA 3'	272 5'-TgA 3'	603 5'-gTg 3'
			426 5'-TCT 3'	3 rd 5'-TAT 3'	247 5'-ATg 3'	577 5'-gCA 3'		603 5'-gTg 3'	577 5'-gCA 3'	178 5'-gAg 3'	604 5'-CgC 3'	
			461 5'-gCg 3'		307 5'-CCA 3'	589 5'-CTA 3'			608 5'-CCT 3'	180 5'-TCC 3'		
Well No.	25	26	27	28	29	30	31	32	33	34	35	36

Well No.	37	38	39	40	41	42	43	44	45	46	47
Length of spec.	85	180	190	265	180	255	185	170	120	175	240
PCR product	155	265	225		210		215	215	225	295	
	190										
	270										
Length of int. pos. control ¹	1070	1070	1070	1070	1070	1070	800	1070	1070	1070	1070
5'-primer(s) ²	160 5'-ACT 3'	139 5'-TCg 3'	97 5'-TCg 3'	44 5'-ggC 3'	97 5'-TCC 3'	385 5'-ggT 3'	363 5'-AgC 3'	97 5'-TCg 3'	419 5'-gTT 3'	350 5'-TCC 3'	97 5'-TCg 3'
	376 5'-gCA 3'	379 5'-ACA 3'			106 5'-CCg 3'	388 5'-CgA 3'			419 5'-gTT 3'	467 5'-CTA 3'	363 5'-AgC 3'
	454 5'-ACC 3'	454 5'-ACC 3'			397 5'-gCT 3'	395 5'-gCC 3'			523 5'-CCT 3'		
	560 5'-CgA 3'										
3'-primer(s) ³	272 5'-TgA 3'	272 5'-TgA 3'	245 5'-ACC 3'	142 5'-TgC 3'	272 5'-TgA 3'	603 5'-gTg 3'	506 5'-Tgg 3'	225 5'-ATT 3'	603 5'-gTg 3'	603 5'-gTg 3'	292 5'-gTA 3'
	603 5'-gTg 3'	603 5'-gTg 3'	247 5'-ATg 3'		538 5'-gTC 3'		538 5'-CCA 3'	226 5'-TAC 3'			295 5'-TCA 3'
			280 5'-CTg 3'					272 5'-TgC 3'			573 5'-AgT 3'
Well No.	37	38	39	40	41	42	43	44	45	46	47

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-B*08 SSP subtyping kit²																				
				Well																
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
				Prod. No.:	201893001	201893002	201893003	201893004	201893005	201893006	201893007	201893008	201893009	201893010	201893011	201893012	201893013	201893014	201893015	201893016
IHWC cell line¹			B*																	
1	9001	SA	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*52:01	*73:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*52:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*15:10	*53:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*37:01		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
6	9353	SM	*39:01	*51:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*35:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107	LKT3	*54:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051	PITOUT	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004	JESTHOM	*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075	DKB	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*08:01	*55:01	+	+	+	-	-	-	-	-	-	-	-	-	-	-	+	-
18	9257	32367	*14:01	*56:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038	BM16	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*35:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124	IHL	*40:02	*56:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035	JBUSH	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049	IBW9	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*07:05	*51:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
28	9320	BEL5GB	*44:02	*44:03	-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
29	9050	MOU	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*42:01		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019	DUCAF	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*41:02		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098	MT14B	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302	SSTO	*44:02		-	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-
36	9024	KT17	*15:01	*35:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315	CML	*08:01	*27:05	+	+	+	-	-	-	-	-	-	-	-	-	-	-	+	-
40	9134	WHONP199	*13:02	*46:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055	H0301	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057	TEM	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*42:01	*50:01	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*51:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303	TER-ND	*35:01	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

CELL LINE VALIDATION SHEET																					
HLA-B*08 SSP subtyping kit ²																					
				Well																	
				17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32		
				Prod. No.:																	
				201893017	201893018	201893019	201893020	201893021	201893022	201893023	201893024	201893025	201893026	201893027	201893028	201893029	201893030	201893031	201893032		
	IHC cell line ¹			B*																	
1	9001	SA		*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
2	9280	LK707		*52:01	*73:01	-	+	-	-	+	-	-	-	-	-	-	-	-	-		
3	9011	E4181324		*52:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
4	9275	GU373		*15:10	*53:01	-	-	-	-	-	-	+	-	-	-	-	-	-	-		
5	9009	KAS011		*37:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
6	9353	SM		*39:01	*51:01	-	-	-	-	-	-	+	-	-	-	-	-	-	-		
7	9020	QBL		*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
8	9007	DEM		*35:01		-	-	-	-	-	-	+	-	-	-	-	-	-	-		
9	9026	YAR		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
10	9107	LKT3		*54:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
11	9051	PITOUT		*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
12	9052	DBB		*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
13	9004	JESTHOM		*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
14	9071	OLGA		*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
15	9075	DKB		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
16	9037	SWEIG007		*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
17	9282	CTM3953540		*08:01	*55:01	-	-	-	-	-	-	-	-	-	-	-	-	-	+		
18	9257	32367		*14:01	*56:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
19	9038	BM16		*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
20	9059	SLE005		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
21	9064	AMALA		*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
22	9056	KOSE		*35:03		-	-	-	-	-	-	+	-	-	-	-	-	-	-		
23	9124	IHL		*40:02	*56:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
24	9035	JBUSH		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
25	9049	IBW9		*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
26	9285	WT49		*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
27	9191	CH1007		*07:05	*51:01	-	-	-	-	-	-	+	-	-	-	-	-	-	-		
28	9320	BEL5GB		*44:02	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
29	9050	MOU		*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
30	9021	RSH		*42:01		-	-	-	-	-	-	-	-	-	-	-	-	-	+		
31	9019	DUCAF		*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
32	9297	HAG		*41:02		-	-	-	-	-	-	-	-	-	-	-	-	-	+		
33	9098	MT14B		*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
34	9104	DHIF		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
35	9302	SSTO		*44:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
36	9024	KT17		*15:01	*35:01	-	-	-	-	-	-	+	-	-	-	-	-	-	-		
37	9065	HHKB		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
38	9099	LZL		*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
39	9315	CML		*08:01	*27:05	-	-	-	-	-	-	-	-	-	-	-	-	-	+		
40	9134	WHONP199		*13:02	*46:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-		
41	9055	H0301		*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
42	9066	TAB089		*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
43	9076	T7526		*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
44	9057	TEM		*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
45	9239	SHJO		*42:01	*50:01	-	-	-	-	-	-	-	-	-	-	-	-	-	+		
46	9013	SCHU		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-		
47	9045	TUBO		*51:01		-	-	-	-	-	-	+	-	-	-	-	-	-	-		
48	9303	TER-ND		*35:01	*44:03	-	-	-	-	-	-	+	-	-	-	-	-	-	-		

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

CELL LINE VALIDATION SHEET																		
HLA-B*08 SSP subtyping kit²																		
			Prod. No.:	Well														
				33	34	35	36	37	38	39	40	41	42	43	44	45	46	47
				201893033	201893034	201893035	201893036	201893037	201893038	201893039	201893040	201893041	201893042	201893043	201893044	201893045	201893046	201893047
	IHWC cell line¹	B*																
1	9001 SA	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280 LK707	*52:01	*73:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324	*52:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373	*15:10	*53:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009 KAS011	*37:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM	*39:01	*51:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9007 DEM	*35:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3	*54:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051 PITOUT	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB	*57:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM	*27:05		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA	*15:01	*15:20	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007	*40:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540	*08:01	*55:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257 32367	*14:01	*56:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056 KOSE	*35:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL	*40:02	*56:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035 JBUSH	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49	*58:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007	*07:05	*51:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
28	9320 BEL5GB	*44:02	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU	*44:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH	*42:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019 DUCAF	*18:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG	*41:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B	*40:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO	*44:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17	*15:01	*35:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099 LZL	*15:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315 CML	*08:01	*27:05	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199	*13:02	*46:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301	*14:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526	*46:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM	*38:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO	*42:01	*50:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013 SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO	*51:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND	*35:01	*44:03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 5 to 7, 10, 12 to 14, 16, 17, 19, 20, 25 to 31, 33 to 40, 42 and 44 to 47 were available. The specificities of the primers in primer solutions 5 to 7, 10, 13, 14, 16, 17, 19, 20, 23, 25 to 27, 29, 33, 35, 36, 37, 39 to 41, 44, 45 and 47 were tested by separately adding one or two additional 5'-primers, respectively one or two additional 3'-primers. In primer solutions 12, 30 and 34 it was only possible to test the 5'-primer, the 3'-primers were not possible to test. In primer solutions 28, 31, 38, 42 and 46 it was only possible to test the 3'-primers, the 5'-primers were not possible to test. In primer solutions 4 to 7, 19, 20, 22, 27, 29, 33, 35, 39, 43, 44 and 47 one or two of the 3'-primers were not possible to test, and in primer solutions 10, 13, 14, 24, 25, 27, 35 to 37 and 45 one, two or three of the 5'-primers were not possible to test. Additional primers in primer solutions 4, 9, 11, 18 and 24 were tested by separately adding either one 3'-primer or one 5'-primer.

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

101.513-24/04 – including *Taq* pol., IFU-01
101.513-24u/04u – without *Taq* pol., IFU-02

Visit www.olerup.com for
“Instructions for Use” (IFU)

Lot No.: **4G9**

Lot-specific information

ADDRESSES:

Manufacturer:

Olerup SSP AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-717 88 27

Fax: +46-8-717 88 18

E-mail: olerup-se@caredx.com

Web page: <http://www.olerup.com>

Distributed by:

Olerup GmbH, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

Tel: +43-1-710 15 00

Fax: +43-1-710 15 00 10

E-mail: olerup-at@caredx.com

Web page: <http://www.olerup.com>

Olerup Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-OLERUP1

Fax: 610-344-7989

E-mail: olerup-us@caredx.com

Web page: <http://www.olerup.com>

For information on *Olerup* distributors worldwide, contact **Olerup GmbH**.