

Olerup SSP® HLA-B*82

Product number:	101.552-06 – including <i>Taq</i> polymerase 101.552-06u – without <i>Taq</i> polymerase
Lot number:	9H6
Expiry date:	2023-08-01
Number of tests:	6
Number of wells per test:	5+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

This Product Description is only valid for Lot No. 9H6

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP®
HLA-B*82 LOT (4H2).**

The HLA-B*82 kit is updated for new alleles to enable separation of:

- Confirmed¹ alleles as listed in the IMGT/HLA database
- Polymorphisms in exons outside of the region encoding the peptide binding domain
- Null and Alternatively expressed alleles

¹As described in section Uniquely Identified Alleles.

The HLA-B*82 specificity and interpretation tables have been updated for the HLA-B alleles described since the previous *Olerup SSP®* HLA-B*82 lot was made (**Lot No. 4H2**). The kit design is based on IMGT/HLA database 3.35.0.

The HLA-B*82 primer set is unchanged compared to the previous *Olerup SSP®* HLA-B*82 (**Lot No. 4H2**).

Changes in revision R01 compared to R00:

1. The expiration date has been altered due to shelf-life extension.



0197

Well 6 contains Negative Control primer pairs, that will amplify most *Olerup* SSP[®] HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs. The PCR products generated by the positive control primer pair are 200 and 430 base pairs.

Length of PCR product	105	200	105	80	75	80	85
5'-primer¹	164	340	440	45	45	43	36
	5'-CAC ^{3'}	5'-Agg ^{3'}	5'-TTA ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-Tgg ^{3'}	5'-TAC ^{3'}
							36
							5'-TAT ^{3'}
3'-primer²	231	2nd I	507	59	58	57	47
	5'-TgC ^{3'}	5'-AAA ^{3'}	5'-TTg ^{3'}	5'-CTC ^{3'}	5'-ggC ^{3'}	5'-CTC ^{3'}	5'-ACA ^{3'}
							48
							5'-gCA ^{3'}
							48
							5'-gCC ^{3'}
							52
							5'-TgT ^{3'}
A*	+	+	+				
B*	+	+	+				
C*	+	+	+				
DRB1				+	+		
DRB3				+	+		
DRB5				+			
DQB1					+		
DPB1						+	
DQA1							+

¹The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

²The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2nd or 3rd exon or the 2nd intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

PRODUCT DESCRIPTION

HLA-B*82 SSP typing

CONTENT

The primer set contains 5'- and 3'-primers for identifying the B*82:01 to B*82:03 alleles.

PLATE LAYOUT

Each HLA-B*82 test consists of 6 PCR reactions in an 8 well cut PCR plate. Wells 7 to 8 are empty.

1	2	3	4	5	NC	empty	empty
---	---	---	---	---	----	-------	-------

The 8 well PCR plate is marked with 'B82' in silver/gray ink.

Well No. 1 is marked with the Lot No. '9H6'.

Wells 1 to 5 – HLA-B*82 high resolution primers.

Well 6 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded. The PCR plates are heat-sealed with a PCR-compatible foil.

Please note: When removing each 8 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

INTERPRETATION

Due to the sharing of sequence motifs between HLA-A alleles non-HLA-B*82 alleles will be amplified by some primer mixes. For further details see Specificity Table.

UNIQUELY IDENTIFIED ALLELES

All the HLA-B*82, i.e. **B*82:01 to B*82:03**, recognized by the HLA Nomenclature Committee in January 2019^{1,2} will be amplified by the primers in the HLA-B*82 SSP kit.

The HLA-B*82 kit enables separation of the confirmed HLA-B*82 alleles as listed in the IMGT/HLA database. An HLA allele is listed as confirmed by IMGT/HLA if it has been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-B*82 alleles is listed below.

The HLA-B*82 kit also enables identification of polymorphisms in exons outside of the region encoding the peptide binding domain and of null and alternatively expressed alleles

¹HLA-B alleles listed on the IMGT/HLA web page 2019-January-23, release 3.35.0, www.ebi.ac.uk/imgt/hla.



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

²Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

ALLELE CONFIRMATION STATUS

Allele	Status ¹
B*82:01:01:01	Confirmed
B*82:01:01:02	Unconfirmed
B*82:01:02	Unconfirmed
B*82:02:01:01	Confirmed
B*82:02:01:02	Unconfirmed
B*82:02:02	Unconfirmed
B*82:03	Unconfirmed

¹Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2019-January-23, release 3.35.0, www.ebi.ac.uk/imgt/hla.

RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-B*82 homo- and heterozygotes is available upon request.



0197

SPECIFICITY TABLE

HLA-B*82 SSP subtyping

Specificities and sizes of the PCR products of the 5+1 primer mixes used for HLA-B*82 SSP subtyping

Primer Mix	Size of spec. PCR product ¹	Size of control band ²	Amplified HLA-B*82 alleles ³	Other amplified HLA Class I alleles
1	195 bp	800 bp	*82:01:01:01-82:03	*44:10, 44:15:01:01-44:15:01:02, 44:18, 44:140, 44:339-44:340, 45:01:01:01-45:01:07, 45:05-45:07, 45:11-45:23, 45:25, 49:20, 50:02, 50:56
2 ^{4,5,6}	140 bp	800 bp	*82:01:01:01-82:01:02, 82:03	
3	230 bp	1070 bp	*82:02:01:01-82:02:02	*15:06, 15:27:01-15:27:04, 15:84, 15:109, 15:195, 15:327, 15:344, 15:398, C*03:89, C*03:271, C*03:338, C*04:08, C*04:34, C*04:147, C*04:212, C*18:08
4	210 bp	1070 bp	*82:01:01:01-82:02:02	
5	230 bp	1070 bp	*82:03	
6 ⁷			Negative Control	

¹Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-B*82 SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

²The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

³For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

Lot No.: 9H6

Lot-specific information

SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

⁴Primer mix 2 may have a tendency of unspecific amplifications.

⁵Primer mix 2 may have a tendency to giving rise to primer oligomer formation.

⁶Primer mix 2 may give rise to a lower yield of HLA-specific PCR product than the other B*82 primer mixes.

⁷Primer mix 6 contains a negative control, which will amplify most HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR products generated by the HGH positive control primer pair are 200 and 430 base pairs.



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

PRIMER SPECIFICATION

Well No.	1	2	3	4	5
Length of spec.	195	140	230	210	230
PCR product					
Length of int. pos. control ¹	800	800	1070	1070	1070
5'-primer(s) ²	420	557	368	368	369
	5' -TTA 3' 5' -ggA 3' 5' -gTT 3' 5' -gTT 3' 5' -TAT 3'				
3'-primer(s) ³	572	3 rd I	557	538	557
	5' -gCg 3' 5' -TAT 3' 5' -ggC 3' 5' -gTC 3' 5' -ggT 3'				
Well No.	1	2	3	4	5

¹The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

²The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.

³The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the www.ebi.ac.uk/imgt/hla web site. The sequence of the 3 terminal nucleotides of the primer is given.



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

CELL LINE VALIDATION SHEET										
HLA-B*82 SSP primer set ²										
					Well					
					1	2	3	4	5	
					Prod. No.:	201891601	201555602	201670903	201670904	201902505
IHCW cell line ¹			B*							
1	9001	SA	*07:02		-	-	-	-	-	
2	9280	LK707	*52:01	*73:01	-	-	-	-	-	
3	9011	E4181324	*52:01		-	-	-	-	-	
4	9275	GU373	*15:10	*53:01	-	-	-	-	-	
5	9009	KAS011	*37:01		-	-	-	-	-	
6	9353	SM	*39:01	*51:01	-	-	-	-	-	
7	9020	QBL	*18:01		-	-	-	-	-	
8	9025	DEU	*35:01		-	-	-	-	-	
9	9026	YAR	*38:01		-	-	-	-	-	
10	9107	LKT3	*54:01		-	-	-	-	-	
11	9051	PITOUT	*44:03		-	-	-	-	-	
12	9052	DBB	*57:01		-	-	-	-	-	
13	9004	JESTHOM	*27:05		-	-	-	-	-	
14	9071	OLGA	*15:01	*15:20	-	-	-	-	-	
15	9075	DKB	*40:01		-	-	-	-	-	
16	9037	SWEIG007	*40:02		-	-	-	-	-	
17	9282	CTM3953540	*08:01	*55:01	-	-	-	-	-	
18	9257	32367	*14:01	*56:01	-	-	-	-	-	
19	9038	BM16	*18:01		-	-	-	-	-	
20	9059	SLE005	*40:01		-	-	-	-	-	
21	9064	AMALA	*15:01		-	-	-	-	-	
22	9056	KOSE	*35:03		-	-	-	-	-	
23	9124	IHL	*40:02	*56:02	-	-	-	-	-	
24	9035	JBUSH	*38:01		-	-	-	-	-	
25	9049	IBW9	*14:02		-	-	-	-	-	
26	9285	WT49	*58:01		-	-	-	-	-	
27	9191	CH1007	*07:05	*51:01	-	-	-	-	-	
28	9320	BEL5GB	*44:02	*44:03	-	-	-	-	-	
29	9050	MOU	*44:03		-	-	-	-	-	
30	9021	RSH	*42:01		-	-	-	-	-	
31	9019	DUCAF	*18:01		-	-	-	-	-	
32	9297	HAG	*41:02		-	-	-	-	-	
33	9098	MT14B	*40:01		-	-	-	-	-	
34	9104	DHIF	*38:01		-	-	-	-	-	
35	9302	SSTO	*44:02		-	-	-	-	-	
36	9024	KT17	*15:01	*35:01	-	-	-	-	-	
37	9065	HHKB	*07:02		-	-	-	-	-	
38	9099	LZL	*15:01		-	-	-	-	-	
39	9315	CML	*08:01	*27:05	-	-	-	-	-	
40	9134	WHONP199	*13:02	*46:01	-	-	-	-	-	
41	9055	H0301	*14:02		-	-	-	-	-	
42	9066	TAB089	*46:01		-	-	-	-	-	
43	9076	T7526	*46:01		-	-	-	-	-	
44	9057	TEM	*38:01		-	-	-	-	-	
45	9239	SHJO	*42:01	*50:01	-	-	-	-	-	
46	9013	SCHU	*07:02		-	-	-	-	-	
47	9045	TUBO	*51:01		-	-	-	-	-	
48	9303	TER-ND	*35:01	*44:03	-	-	-	-	-	



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

¹The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

²The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

The specificities of the primers in primer solution 5 were tested by separately adding one additional 5'-primer, respectively one 3'-primer.

In addition, one of the 3'-primers in primer solution 3 was tested by adding an additional 5'-primer.



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01

101.552-06 – including *Taq* pol., IFU-01

Visit <https://labproducts.caredx.com> for

101.552-06u – without *Taq* pol., IFU-02

“Instructions for Use” (IFU)

Lot No.: **9H6**

Lot-specific information

ADDRESSES:

Manufacturer:

CareDx AB, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Tel: +46-8-508 939 00

Fax: +46-8-717 88 18

E-mail: orders-se@caredx.com

Web page: <https://labproducts.caredx.com/>

Distributed by:

CareDx Lab Solutions Inc., 901 S. Bolmar St., Suite R, West Chester, PA 19382

Tel: 1-877-653-78171

Fax: 610-344-7989

E-mail: orders-us@caredx.com

Web page: <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx AB**.



0197

For *In Vitro* Diagnostic Use

MA123 v01 SSP PI Template

Date: August 2021, Rev. No: 01