

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

**Olerup SSP<sup>®</sup> HLA-C\*12**

Product number:	101.624-12 – including <i>Taq</i> polymerase 101.624-12u – without <i>Taq</i> polymerase
Lot number:	7K1
Expiry date:	2024-02-01
Number of tests:	12
Number of wells per test:	47+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

**This Product Description is only valid for Lot No. 7K1.**

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

**CHANGES COMPARED TO THE PREVIOUS OLERUP SSP<sup>®</sup>  
HLA-C\*12 LOT (1H9)**

- The product documentation has been updated for new alleles of IMGT 3.38.0
- The kit resolution focuses on common and well documented (CWD) alleles<sup>1</sup>.

<sup>1</sup>As described in section Uniquely Identified Alleles.

The HLA-C\*12 primer set, specificity and interpretation tables have been updated for the HLA-C alleles described since the previous *Olerup SSP<sup>®</sup> HLA-C\*12* lot was made (Lot No. 1H9).

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

**Lot-specific information**

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	Rationale
10	Removed	-	5'-primer removed for improved HLA-specific amplification.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

Well **48** contains Negative Control primer pairs, that will amplify a majority of the *Olerup SSP<sup>®</sup>* HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.  
The PCR product generated by the positive control primer pair is 200 base pairs.

Length of PCR product	105	200	105	80	75	80	85
<b>5'-primer<sup>1</sup></b>	<b>164</b>	<b>340</b>	<b>440</b>	<b>45</b>	<b>45</b>	<b>43</b>	<b>36</b>
	5'-CAC <sup>3'</sup>	5'-Agg <sup>3'</sup>	5'-TTA <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-TAC <sup>3'</sup>
							<b>36</b>
							5'-TAT <sup>3'</sup>
<b>3'-primer<sup>2</sup></b>	<b>231</b>	<b>2<sup>nd</sup> I</b>	<b>507</b>	<b>59</b>	<b>58</b>	<b>57</b>	<b>47</b>
	5'-TgC <sup>3'</sup>	5'-AAA <sup>3'</sup>	5'-TTg <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ggC <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ACA <sup>3'</sup>
							<b>48</b>
							5'-gCA <sup>3'</sup>
							<b>48</b>
							5'-gCC <sup>3'</sup>
							<b>52</b>
							5'-TgT <sup>3'</sup>
<b>A*</b>	+	+	+				
<b>B*</b>	+	+	+				
<b>C*</b>	+	+	+				
<b>DRB1</b>				+	+		
<b>DRB3</b>				+	+		
<b>DRB5</b>				+			
<b>DQB1</b>					+		
<b>DPB1</b>						+	
<b>DQA1</b>							+

<sup>1</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>2</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon or the 2<sup>nd</sup> intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

## PRODUCT DESCRIPTION

### HLA-C\*12 SSP typing

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the C\*12:02 to C\*12:304 alleles.

#### PLATE LAYOUT

Each HLA-C\*12 test consists of 48 PCR reactions in a 48 well cut PCR plate.

<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>
<b>9</b>	<b>10</b>	<b>11</b>	<b>12</b>	<b>13</b>	<b>14</b>	<b>15</b>	<b>16</b>
<b>17</b>	<b>18</b>	<b>19</b>	<b>20</b>	<b>21</b>	<b>22</b>	<b>23</b>	<b>24</b>
<b>25</b>	<b>26</b>	<b>27</b>	<b>28</b>	<b>29</b>	<b>30</b>	<b>31</b>	<b>32</b>
<b>33</b>	<b>34</b>	<b>35</b>	<b>36</b>	<b>37</b>	<b>38</b>	<b>39</b>	<b>40</b>
<b>41</b>	<b>42</b>	<b>43</b>	<b>44</b>	<b>45</b>	<b>46</b>	<b>47</b>	<b>NC</b>

The 48 well PCR plate is marked with ‘HLA-C\*12’ in silver/gray ink.

Well No. 1 is marked with the Lot No. ‘7K1’.

Wells 1 to 47 – HLA-C\*12 high resolution primers.

Well 48 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays and should be disregarded. The PCR plates are covered with a PCR-compatible foil.

The PCR plates are heat-sealed with a PCR-compatible foil.

**Please note:** When removing each 48 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

Due to the sharing of sequence motifs between HLA-C alleles non-HLA-C\*12 alleles will be amplified by some primer mixes. For further details see Specificity Table.

#### UNIQUELY IDENTIFIED ALLELES

All the HLA-C\*12 alleles, i.e. **C\*12:02 to C\*12:304**, recognized by the HLA Nomenclature Committee in October 2019<sup>1,2</sup> will be amplified by the primers in the HLA-C\*12 subtyping kit.

The HLA-C\*12 kit enables separation of the confirmed HLA-C\*12 alleles as listed in the IMGT/HLA database 3.27.0. An HLA allele is listed as confirmed by IMGT/HLA if it has

101.624-12 – including **Taq pol.**, IFU-01  
101.624-12u – without **Taq pol.**, IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

**Lot No.: 7K1**

**Lot-specific information**

been sequenced by more than a single laboratory or from multiple sources. Current allele confirmation status for HLA-C\*12 alleles is listed below.

The HLA-C\*12 kit also enables identification of many null and alternatively expressed alleles.

The following HLA-C\*12 alleles can be distinguished by the different sizes of the HLA-specific PCR product:

Alleles	Primer mix	Alleles	Primer mix
C*12:03:09, 12:159	17	C*12:32, 12:34	30
C*12:06, 12:48, 12:81	6	C*12:39N, 12:167	25
C*12:11, 12:227	14	C*12:45, 12:50, 12:232N	28
C*12:23, 12:203	22	C*12:46N, 12:139	20
C*12:29, 12:38	29	C*12:86, 12:104N	29
C*12:30, 12:36	27	C*12:109, 12:171	34

The following alleles give rise to identical amplification patterns with the HLA-C\*12 primer set. These alleles can be distinguished by the HLA-C low resolution kit and the HLA-C\*05, HLA-C\*07, HLA-C\*08 or HLA-C\*16 kit, respectively.

**Alleles**

C\*12:09, C\*05:16, C\*05:85, C\*05:107, C\*16:88

C\*12:33, C\*16:91

C\*12:181, C\*07:723

C\*12:208, C\*12:222, C\*12:233, C\*12:251, C\*12:258, C\*08:05, C\*08:25, C\*16:15:01

<sup>1</sup>HLA-C alleles listed on the IMGT/HLA web page 2019-October-17, release 3.38.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

<sup>2</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from the web page <http://hla.alleles.org/alleles/deleted.html>.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

**ALLELE CONFIRMATION STATUS**

Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>
C*12:02:01	Unconfirmed	<b>C*12:03:29</b>	<b>Confirmed</b>	<b>C*12:35</b>	<b>Confirmed</b>	C*12:84N	Unconfirmed
<b>C*12:02:02:01</b>	<b>Confirmed</b>	C*12:03:30	Unconfirmed	C*12:36	Unconfirmed	C*12:85	Unconfirmed
C*12:02:02:02	Unconfirmed	C*12:03:31	Unconfirmed	<b>C*12:37</b>	<b>Confirmed</b>	<b>C*12:86</b>	<b>Confirmed</b>
C*12:02:03	Unconfirmed	C*12:03:32	Unconfirmed	<b>C*12:38</b>	<b>Confirmed</b>	C*12:87	Unconfirmed
<b>C*12:02:04</b>	<b>Confirmed</b>	C*12:03:33	Unconfirmed	<b>C*12:39N</b>	<b>Confirmed</b>	C*12:88	Unconfirmed
C*12:02:05	Unconfirmed	C*12:03:34:01	Unconfirmed	C*12:40	Unconfirmed	C*12:89	Unconfirmed
C*12:02:06	Unconfirmed	C*12:03:34:02	Unconfirmed	<b>C*12:41</b>	<b>Confirmed</b>	<b>C*12:90</b>	<b>Confirmed</b>
C*12:02:07	Unconfirmed	<b>C*12:03:35</b>	<b>Confirmed</b>	<b>C*12:42Q</b>	<b>Confirmed</b>	C*12:91	Unconfirmed
C*12:02:08	Unconfirmed	C*12:03:36	Unconfirmed	<b>C*12:43</b>	<b>Confirmed</b>	C*12:92	Unconfirmed
C*12:02:09	Unconfirmed	<b>C*12:03:37</b>	<b>Confirmed</b>	<b>C*12:44</b>	<b>Confirmed</b>	C*12:93	Unconfirmed
C*12:02:10	Unconfirmed	C*12:03:38	Unconfirmed	<b>C*12:45</b>	<b>Confirmed</b>	<b>C*12:94</b>	<b>Confirmed</b>
C*12:02:11	Unconfirmed	C*12:03:39	Unconfirmed	C*12:46N	Unconfirmed	C*12:95	Unconfirmed
<b>C*12:02:12</b>	<b>Confirmed</b>	C*12:03:40	Unconfirmed	<b>C*12:47</b>	<b>Confirmed</b>	C*12:96	Unconfirmed
<b>C*12:02:13</b>	<b>Confirmed</b>	<b>C*12:03:41</b>	<b>Confirmed</b>	C*12:48	Unconfirmed	C*12:97	Unconfirmed
C*12:02:14	Unconfirmed	C*12:03:42	Unconfirmed	C*12:49	Unconfirmed	C*12:98	Unconfirmed
C*12:02:15	Unconfirmed	C*12:04:01	Unconfirmed	<b>C*12:50</b>	<b>Confirmed</b>	<b>C*12:99:01</b>	<b>Confirmed</b>
<b>C*12:03:01:01</b>	<b>Confirmed</b>	<b>C*12:04:02</b>	<b>Confirmed</b>	C*12:51	Unconfirmed	C*12:99:02	Unconfirmed
C*12:03:01:02	Unconfirmed	<b>C*12:05</b>	<b>Confirmed</b>	C*12:52	Unconfirmed	<b>C*12:100</b>	<b>Confirmed</b>
C*12:03:01:03	Unconfirmed	C*12:06	Unconfirmed	C*12:53	Unconfirmed	<b>C*12:101</b>	<b>Confirmed</b>
<b>C*12:03:01:04</b>	<b>Confirmed</b>	<b>C*12:07</b>	<b>Confirmed</b>	<b>C*12:54</b>	<b>Confirmed</b>	<b>C*12:102</b>	<b>Confirmed</b>
C*12:03:01:05	Unconfirmed	C*12:08	Unconfirmed	C*12:55	Unconfirmed	C*12:103	Unconfirmed
C*12:03:01:06	Unconfirmed	<b>C*12:09</b>	<b>Confirmed</b>	C*12:56	Unconfirmed	C*12:104N	Unconfirmed
C*12:03:01:07	Unconfirmed	C*12:10:01	Unconfirmed	C*12:57:01	Unconfirmed	C*12:105N	Unconfirmed
C*12:03:02	Unconfirmed	C*12:10:02	Unconfirmed	C*12:57:02	Unconfirmed	C*12:106	Unconfirmed
C*12:03:03	Unconfirmed	C*12:11	Unconfirmed	<b>C*12:58</b>	<b>Confirmed</b>	C*12:107	Unconfirmed
<b>C*12:03:04</b>	<b>Confirmed</b>	<b>C*12:12</b>	<b>Confirmed</b>	<b>C*12:59</b>	<b>Confirmed</b>	C*12:108	Unconfirmed
<b>C*12:03:05</b>	<b>Confirmed</b>	<b>C*12:13</b>	<b>Confirmed</b>	<b>C*12:60</b>	<b>Confirmed</b>	<b>C*12:109</b>	<b>Confirmed</b>
<b>C*12:03:06</b>	<b>Confirmed</b>	C*12:14:01	Unconfirmed	C*12:61	Unconfirmed	C*12:110	Unconfirmed
C*12:03:07	Unconfirmed	C*12:14:02	Unconfirmed	<b>C*12:62</b>	<b>Confirmed</b>	C*12:111	Unconfirmed
<b>C*12:03:08</b>	<b>Confirmed</b>	C*12:15	Unconfirmed	<b>C*12:63</b>	<b>Confirmed</b>	C*12:112	Unconfirmed
<b>C*12:03:09</b>	<b>Confirmed</b>	<b>C*12:16</b>	<b>Confirmed</b>	C*12:64	Unconfirmed	<b>C*12:113</b>	<b>Confirmed</b>
<b>C*12:03:10</b>	<b>Confirmed</b>	C*12:17	Unconfirmed	C*12:65	Unconfirmed	C*12:114	Unconfirmed
<b>C*12:03:11</b>	<b>Confirmed</b>	<b>C*12:18:01</b>	<b>Confirmed</b>	C*12:66	Unconfirmed	C*12:115	Unconfirmed
<b>C*12:03:12</b>	<b>Confirmed</b>	C*12:18:02	Unconfirmed	C*12:67	Unconfirmed	C*12:116	Unconfirmed
<b>C*12:03:13</b>	<b>Confirmed</b>	<b>C*12:19</b>	<b>Confirmed</b>	C*12:68	Unconfirmed	C*12:117	Unconfirmed
<b>C*12:03:14</b>	<b>Confirmed</b>	C*12:20	Unconfirmed	C*12:69	Unconfirmed	C*12:118	Unconfirmed
<b>C*12:03:15</b>	<b>Confirmed</b>	<b>C*12:21</b>	<b>Confirmed</b>	C*12:70	Unconfirmed	C*12:119	Unconfirmed
C*12:03:16	Unconfirmed	C*12:22	Unconfirmed	C*12:71	Unconfirmed	C*12:120	Unconfirmed
<b>C*12:03:17</b>	<b>Confirmed</b>	<b>C*12:23</b>	<b>Confirmed</b>	<b>C*12:72</b>	<b>Confirmed</b>	C*12:121	Unconfirmed
<b>C*12:03:18</b>	<b>Confirmed</b>	<b>C*12:24</b>	<b>Confirmed</b>	<b>C*12:73</b>	<b>Confirmed</b>	<b>C*12:122</b>	<b>Confirmed</b>
<b>C*12:03:19</b>	<b>Confirmed</b>	<b>C*12:25</b>	<b>Confirmed</b>	C*12:74	Unconfirmed	<b>C*12:123</b>	<b>Confirmed</b>
C*12:03:20	Unconfirmed	C*12:26	Unconfirmed	C*12:75	Unconfirmed	C*12:124	Unconfirmed
C*12:03:21	Unconfirmed	<b>C*12:27</b>	<b>Confirmed</b>	C*12:76	Unconfirmed	C*12:125	Unconfirmed
<b>C*12:03:22</b>	<b>Confirmed</b>	<b>C*12:28</b>	<b>Confirmed</b>	C*12:77	Unconfirmed	C*12:126	Unconfirmed
<b>C*12:03:23</b>	<b>Confirmed</b>	C*12:29	Unconfirmed	C*12:78	Unconfirmed	C*12:127	Unconfirmed
C*12:03:24	Unconfirmed	<b>C*12:30</b>	<b>Confirmed</b>	C*12:79	Unconfirmed	C*12:128	Unconfirmed
C*12:03:25	Unconfirmed	C*12:31	Unconfirmed	C*12:80N	Unconfirmed	C*12:129	Unconfirmed
C*12:03:26	Unconfirmed	C*12:32	Unconfirmed	C*12:81	Unconfirmed	C*12:130	Unconfirmed
<b>C*12:03:27</b>	<b>Confirmed</b>	C*12:33	Unconfirmed	<b>C*12:82</b>	<b>Confirmed</b>	C*12:131	Unconfirmed
C*12:03:28	Unconfirmed	<b>C*12:34</b>	<b>Confirmed</b>	C*12:83	Unconfirmed	<b>C*12:132</b>	<b>Confirmed</b>

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: <b>7K1</b>		Lot-specific information					
Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>	Allele	Status <sup>1</sup>
C*12:133	Unconfirmed	<b>C*12:153</b>	<b>Confirmed</b>	C*12:173	Unconfirmed	C*12:193	Unconfirmed
C*12:134	Unconfirmed	C*12:154	Unconfirmed	C*12:174	Unconfirmed	C*12:194	Unconfirmed
C*12:135	Unconfirmed	C*12:155Q	Unconfirmed	C*12:175	Unconfirmed	C*12:195	Unconfirmed
<b>C*12:136</b>	<b>Confirmed</b>	<b>C*12:156</b>	<b>Confirmed</b>	C*12:176	Unconfirmed	C*12:196	Unconfirmed
C*12:137	Unconfirmed	C*12:157	Unconfirmed	C*12:177	Unconfirmed	C*12:197	Unconfirmed
C*12:138	Unconfirmed	<b>C*12:158</b>	<b>Confirmed</b>	C*12:178	Unconfirmed	C*12:198	Unconfirmed
<b>C*12:139</b>	<b>Confirmed</b>	<b>C*12:159</b>	<b>Confirmed</b>	C*12:179	Unconfirmed	C*12:199	Unconfirmed
C*12:140	Unconfirmed	C*12:160	Unconfirmed	C*12:180	Unconfirmed	C*12:200	Unconfirmed
C*12:141	Unconfirmed	C*12:161	Unconfirmed	C*12:181	Unconfirmed	C*12:201	Unconfirmed
C*12:142	Unconfirmed	<b>C*12:162</b>	<b>Confirmed</b>	C*12:182	Unconfirmed	C*12:202	Unconfirmed
<b>C*12:143</b>	<b>Confirmed</b>	<b>C*12:163</b>	<b>Confirmed</b>	C*12:183	Unconfirmed	C*12:203	Unconfirmed
<b>C*12:144</b>	<b>Confirmed</b>	<b>C*12:164</b>	<b>Confirmed</b>	C*12:184	Unconfirmed	C*12:204	Unconfirmed
C*12:145	Unconfirmed	<b>C*12:165</b>	<b>Confirmed</b>	C*12:185	Unconfirmed	<b>C*12:205</b>	<b>Confirmed</b>
C*12:146	Unconfirmed	C*12:166	Unconfirmed	C*12:186	Unconfirmed	C*12:206	Unconfirmed
C*12:147	Unconfirmed	C*12:167	Unconfirmed	C*12:187	Unconfirmed		
C*12:148N	Unconfirmed	C*12:168	Unconfirmed	<b>C*12:188</b>	<b>Confirmed</b>		
<b>C*12:149</b>	<b>Confirmed</b>	C*12:169	Unconfirmed	C*12:189	Unconfirmed		
C*12:150	Unconfirmed	C*12:170	Unconfirmed	C*12:190	Unconfirmed		
C*12:151	Unconfirmed	C*12:171	Unconfirmed	C*12:191	Unconfirmed		
C*12:152	Unconfirmed	C*12:172	Unconfirmed	C*12:192	Unconfirmed		

<sup>1</sup>Allele status “confirmed” or “unconfirmed” as listed on the IMGT/HLA web page 2017-January-20, release 3.27.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

### RESOLUTION IN HOMO- AND HETEROZYGOTES

Results file with resolution in HLA-C\*12 homo- and heterozygotes is available upon request.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

**SPECIFICITY TABLE**

**HLA-C\*12 SSP subtyping**

Specificities and sizes of the PCR products of the 47+1 primer mixes used for HLA-C\*12 SSP subtyping

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified HLA-C*12 alleles <sup>3</sup>	Other amplified HLA Class I alleles
1	235 bp	800 bp	*12:02:01-12:03:12, 12:03:13 <sup>w</sup> , 12:03:14-12:04:02:02, 12:06-12:08, 12:10:01-12:15, 12:17-12:20, 12:22-12:32, 12:34-12:48, 12:50-12:54, 12:56-12:97, 12:99:01-12:146, 12:148N-12:187, 12:189-12:194, 12:196-12:216, 12:218-12:268, 12:270N-12:300, 12:302-12:304	*02:12, 02:49, 02:55:01-02:55:02, 02:115, 03:15, 03:27, 03:38:01-03:38:02, 03:53, 03:69, 03:130, 03:136, 03:163, 03:246, 03:274, 03:297, 03:431, 03:461, 04:03:01:01-04:03:01:02, 04:03:03-04:03:07, 04:06:01-04:06:03, 04:16, 04:80, 04:103, 04:107, 04:147, 04:160, 04:171, 04:190, 04:256, 04:286, 04:294, 04:335, 04:337, 04:351:01-04:351:02, 04:357, 04:363, 04:381, 04:383, 04:393, 04:400, 04:402, 05:42, 05:46, 06:03:01-06:03:02, 06:76:01-06:76:02, 06:132:01-06:132:02, 07:26:01:01-07:26:03, 07:92, 07:96:01-07:96:02, 07:314:01-07:314:03, 07:317, 07:351, 07:578, 07:583, 07:723, 07:756, 07:821N, 08:05, 08:21, 08:25, 08:137, 15:03, 15:16, 15:25, 16:15:01-16:15:02, 16:25, 16:64, 17:01:01:02-17:19, 17:21-17:26, 17:27N <sup>w</sup> , 17:28-17:39, 17:41-17:51:01:02, 18:09, <b>A*25:55, A*68:226, B*07:13, B*07:15, B*07:160, B*67:02:01:01-67:02:01:02</b>
2 <sup>4</sup>	100 bp	1070 bp	*12:02:01-12:02:05, 12:02:07, 12:02:09-12:02:11, 12:02:13, 12:02:16-12:02:18, 12:02:20-12:02:29, 12:02:31, 12:02:33, 12:03:19, 12:03:32, 12:08, 12:10:01-12:10:02, 12:14:02, 12:16:01-12:18:01, 12:22, 12:27, 12:30, 12:36, 12:40-12:41, 12:44, 12:49, 12:56, 12:64, 12:67-12:69, 12:72-12:74, 12:80N, 12:83-12:86, 12:96, 12:103-12:106,	*01:04, 01:21, 14:02:21, 16:02:13



101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

**Lot-specific information**

			12:112, 12:114, 12:117, 12:123-12:124, 12:126- 12:128, 12:130, 12:132, 12:134, 12:136-12:137, 12:142, 12:145-12:146, 12:148N, 12:151, 12:155Q, 12:161-12:162, 12:164, 12:166, 12:168- 12:169, 12:177, 12:179, 12:183, 12:188, 12:193, 12:196, 12:198, 12:200, 12:204, 12:207, 12:212, 12:214, 12:217, 12:219N, 12:221, 12:226, 12:228, 12:231, 12:234, 12:236N, 12:239-12:241, 12:243, 12:247, 12:250, 12:252, 12:255, 12:261, 12:263, 12:268, 12:270N, 12:275, 12:279-12:281, 12:285, 12:287, 12:294, 12:296, 12:303-12:304
<b>3</b>	<b>220 bp</b>	<b>800 bp</b>	*12:03:01:01-12:07, 12:11-12:13:01:02, 12:15, 12:23, 12:25-12:26, 12:28-12:29, 12:31-12:35, 12:37-12:39N, 12:42Q- 12:43, 12:45-12:48, 12:50-12:55, 12:57:01- 12:63, 12:65-12:66, 12:70-12:71, 12:75-12:79, 12:81-12:82, 12:87-12:95, 12:97-12:102, 12:107- 12:111, 12:113, 12:115- 12:116, 12:119-12:122, 12:125, 12:129, 12:131, 12:133, 12:135, 12:138- 12:141, 12:143-12:144, 12:147, 12:149-12:150, 12:152, 12:154, 12:156- 12:160, 12:163, 12:165, 12:167, 12:170-12:174, 12:176, 12:178, 12:180, 12:182, 12:184-12:187, 12:189-12:192, 12:194- 12:195:03, 12:197, 12:199, 12:201-12:203, 12:205-12:206, 12:209- 12:211, 12:213, 12:215- 12:216, 12:218, 12:220, 12:223, 12:225, 12:227, 12:229-12:230, 12:232N, 12:235, 12:237-12:238,
			*01:04, 01:09:01-01:09:02, 02:05:01- 02:05:03, 02:17, 04:360, 06:02:01:01- 06:02:01:43, 06:02:03-06:02:42, 06:02:44-06:03:02, 06:07-06:13, 06:15-06:34:02, 06:36-06:39, 06:41- 06:71, 06:73-06:78, 06:80, 06:82- 06:100, 06:102:01-06:117, 06:119- 06:122, 06:124-06:126, 06:128N- 06:135, 06:137-06:142, 06:145- 06:152N, 06:154-06:196, 06:198- 06:202, 06:205-06:263N, 06:265- 06:270, 06:272-06:292, 14:16, 16:04:01:01-16:04:01:03, 16:04:03- 16:04:05, 16:29, 16:33, 16:42, 16:55, 16:61, 16:66, 16:78, 16:82, 16:91, 16:109, 16:124, 16:149-16:150, <b>B*15:510</b>

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

**Lot-specific information**

			Lot-specific information	
			12:242, 12:244-12:246, 12:248-12:249, 12:253- 12:254, 12:256-12:257, 12:259-12:260, 12:262, 12:264-12:267, 12:269, 12:271-12:274N, 12:276- 12:278, 12:282-12:284, 12:286, 12:288-12:293, 12:295N, 12:297, 12:299N-12:302	
<b>4</b>	340 bp	1070 bp	*12:04:01-12:05, 12:09, 12:21, 12:33, 12:41, 12:54, 12:60, 12:146, 12:188	*01:14, 01:59, 01:118, 02:02:01- 02:02:03, 02:02:06-02:02:11, 02:02:13-02:11, 02:13-02:26:03, 02:28-02:40:02, 02:42-02:86, 02:88- 02:114, 02:116-02:125, 02:127- 02:128, 02:130, 02:132-02:186, 03:07:01:01-03:07:02, 03:15, 03:45, 03:130, 03:140, 03:163, 03:243, 03:268, 03:297, 03:461, 04:01:01:01- 04:01:01:29, 04:01:01:31-04:01:28, 04:01:30-04:01:76, 04:01:78- 04:01:120, 04:03:01:01-04:10, 04:12- 04:20, 04:23-04:28, 04:30-04:35:02, 04:37-04:54:02, 04:56-04:171, 04:173N-04:213, 04:215N-04:248, 04:250-04:298, 04:300N-04:405, 05:01:01:01-05:01:50, 05:03-05:128N, 05:130-05:227, 06:02:01:01- 06:02:01:43, 06:02:03-06:02:11, 06:02:13-06:02:64, 06:02:66-06:10, 06:12-06:51, 06:53:01-06:121, 06:123, 06:125-06:146, 06:148-06:216, 06:218-06:247, 06:249-06:251, 06:253-06:279, 06:281N-06:292, 07:07, 07:09, 07:49, 07:76:01- 07:76:02, 07:210, 07:238, 07:247, 07:315, 07:328, 07:403, 07:406, 07:598, 08:10, 14:04, 14:12, 14:49, 14:64, 14:77, 14:108, 15:02:01:01- 15:05:10, 15:05:13-15:06:03, 15:08:01-15:13:01:02, 15:15-15:19, 15:22-15:24, 15:26-15:42, 15:44:01- 15:70, 15:72-15:115N, 15:117-15:143, 15:145N-15:207, 16:02:01:01- 16:02:14, 16:02:16-16:02:17, 16:09, 16:12, 16:19, 16:25, 16:46-16:48, 16:57, 16:60, 16:63, 16:69-16:70, 16:74, 16:77N, 16:84, 16:88-16:91, 16:99, 16:101-16:104, 16:107-16:108, 16:115, 16:120-16:121, 16:123N, 16:132N-16:133, 16:136, 16:140, 16:143-16:145, 16:153,

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

				16:155-16:156, 16:163, 17:01:01:02-17:21, 17:23-17:51:01:02, 18:01:01:01-18:14
<b>5<sup>5</sup></b>	130 bp	1070 bp	*12:05, 12:09, 12:21, 12:33, 12:49, 12:98, 12:188	*01:02:34, 02:02:01-02:02:03, 02:02:06-02:02:07, 02:02:09-02:02:12, 02:02:14-02:02:28, 02:02:30-02:11, 02:13-02:40:02, 02:42-02:48, 02:50-02:54, 02:56-02:62, 02:64-02:89, 02:91-02:114, 02:116-02:130, 02:132-02:186, 03:04:25, 04:10-04:11, 04:36, 04:55, 04:153, 04:169, 04:210, 04:214-04:215N, 04:347, 05:01:01:01-05:01:11, 05:01:13-05:01:23, 05:01:24 <sup>w</sup> , 05:01:25-05:01:32, 05:01:33 <sup>w</sup> , 05:01:34-05:01:50, 05:03-05:29:01, 05:30-05:41, 05:43-05:45, 05:47-05:108, 05:110-05:128N, 05:129 <sup>w</sup> , 05:130-05:136, 05:138-05:227, 06:05, 08:01:01:01-08:02:02, 08:02:04-08:02:10, 08:02:12-08:04:03, 08:06-08:20, 08:22:01:01-08:24, 08:26N-08:63, 08:65-08:72:02, 08:74-08:94, 08:95 <sup>w</sup> , 08:96-08:136, 08:138-08:203, 08:205-08:206, 14:02:03, 14:03:01:01-14:03:01:02, 14:03:03, 14:08, 14:10, 14:22, 14:35N, 14:38, 14:41, 14:53-14:54, 14:61, 14:70, 14:79, 14:86, 14:90, 14:112-14:113, 14:116, 15:02:01:01-15:02:09, 15:02:11-15:02:16, 15:02:18-15:02:44, 15:04:01:01-15:05:11, 15:05:13-15:13:01:02, 15:15, 15:17-15:19, 15:21-15:24, 15:26-15:42, 15:44:01-15:44:02, 15:45 <sup>w</sup> , 15:46-15:66, 15:67 <sup>w</sup> , 15:68-15:82, 15:84Q-15:98, 15:100-15:183, 15:185N-15:207, 16:01:01:01-16:01:01:15, 16:01:03-16:01:15, 16:01:17-16:02:16, 16:04:01:01-16:04:01:03, 16:04:03-16:04:05, 16:06-16:14, 16:16Q-16:24, 16:26-16:63, 16:65-16:68, 16:70, 16:71 <sup>w</sup> , 16:72-16:84, 16:86-16:125, 16:127-16:164
<b>6<sup>4</sup></b>	75 bp 150 bp 415 bp	1070 bp	*12:48, 12:102 *12:06 *12:08, 12:81, 12:188, 12:287	*01:118, 01:128, 03:08, 03:29, 03:31, 03:246, 04:112, 04:169, 05:36, 06:44, 06:252, 14:73, 16:64, 16:70, 16:87
<b>7</b>	225 bp	1070 bp	*12:04:01-12:05, 12:09, 12:21, 12:33, 12:41, 12:54, 12:60, 12:146, 12:188	*02:26:01, 02:107, 02:134, 02:152, 02:167, 04:42:01-04:42:02, 04:103, 04:220, 05:01:01:01-05:01:25, 05:01:27-05:01:28, 05:01:30-05:01:50, 05:03-05:12, 05:14-05:25,

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

				05:27-05:29:02, 05:31-05:57, 05:58:02, 05:58:04-05:77, 05:79- 05:128N, 05:130-05:159, 05:161- 05:195, 05:197-05:199, 05:201, 05:203-05:227, 06:02:01:01- 06:02:01:43, 06:02:03-06:02:08, 06:02:10-06:02:11, 06:02:13-06:02:20, 06:02:22-06:02:64, 06:02:66-06:02:82, 06:03:02-06:10, 06:12-06:17, 06:19- 06:32, 06:34:01-06:51, 06:53:01- 06:70:01, 06:71-06:103, 06:105- 06:118, 06:121, 06:123, 06:126- 06:131, 06:133-06:146, 06:148- 06:187, 06:189-06:216, 06:218- 06:247, 06:249-06:251, 06:253, 06:255-06:279, 06:281N-06:292, 07:07, 07:09, 07:49, 07:76:01- 07:76:02, 07:210, 07:238, 07:247, 07:315, 07:328, 07:403, 07:406, 07:598, 08:10, 15:02:25, 15:05:15, 15:10:01, 15:10:04, 15:23:01, 15:63, 15:107, 15:138, 16:02:01:01-16:02:07, 16:02:09-16:02:14, 16:02:16-16:02:17, 16:09, 16:12, 16:19, 16:25, 16:46- 16:48, 16:57, 16:60, 16:63, 16:69, 16:74, 16:77N, 16:84, 16:88-16:91, 16:99, 16:101-16:104, 16:107-16:108, 16:115, 16:120, 16:123N, 16:132N- 16:133, 16:136, 16:140, 16:143, 16:145, 16:153, 16:155-16:156, 16:163, 17:01:01:02-17:01:08, 17:01:10-17:21, 17:23-17:51:01:02, 18:01:01:01-18:11, 18:13-18:14
<b>8<sup>4</sup></b>	95 bp 155 bp 195 bp 245 bp	1070 bp	*12:15 *12:40 *12:80N *12:07	*15:02:14 *16:14  <b>B*35:310</b>
<b>9<sup>4</sup></b>	95 bp	1070 bp	*12:02:01-12:04:02:02, 12:06-12:08, 12:10:01- 12:20, 12:22-12:27, 12:29-12:32, 12:34-12:48, 12:50-12:57:02, 12:59- 12:62, 12:64-12:97, 12:99:01-12:107, 12:109- 12:131, 12:133-12:134, 12:136-12:145, 12:147- 12:187, 12:189-12:251, 12:253-12:286, 12:288- 12:300, 12:302-12:304	*01:17, 01:21, 01:128, 01:152, 01:157, 02:12, 02:55:01-02:55:02, 03:27, 03:38:01-03:38:02, 03:130, 03:163, 03:246, 03:431, 04:33, 04:107, 04:172, 04:231, 05:46, 06:30, 06:204, 06:210, 07:07, 07:16, 07:51, 07:181, 07:367, 07:598, 07:723, 08:05, 08:21, 08:25, 14:04, 14:73, 14:77, 15:03, 15:16, 16:15:01-16:15:02, 16:25, 17:01:01:02-17:01:14, 17:02-17:10, 17:12-17:14, 17:16:01-17:25, 17:27N- 17:39, 17:41-17:51:01:02, <b>B*07:13</b> , <b>B*67:02:01:01-67:02:01:02</b>
<b>10</b>	155 bp	1070 bp	*12:03:01:01-12:03:01:32, 12:03:03-12:03:07,	*01:02:18, 06:02:38, 06:76:02, 14:02:08, 16:01:17, 16:15:02

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

**Lot-specific information**

			12:03:09-12:03:15, 12:03:17-12:03:18, 12:03:21-12:03:29, 12:03:31, 12:03:33, 12:03:35-12:03:46, 12:03:48-12:03:61, 12:04:02:01-12:07, 12:11- 12:13:01:02, 12:15, 12:20, 12:23-12:25, 12:28-12:29, 12:31-12:32, 12:34-12:35, 12:37-12:39N, 12:42Q- 12:43, 12:45-12:48, 12:50-12:55, 12:57:01- 12:58, 12:60-12:63, 12:65-12:66, 12:70-12:71, 12:75-12:79, 12:81-12:82, 12:87-12:95, 12:97- 12:102, 12:107-12:111, 12:113, 12:115-12:116, 12:119-12:122, 12:125, 12:129, 12:131, 12:133, 12:135, 12:138-12:141, 12:143, 12:147, 12:149- 12:150, 12:152-12:154, 12:156-12:160, 12:163, 12:165, 12:167, 12:170- 12:176, 12:178, 12:180, 12:182, 12:184, 12:186- 12:187, 12:189-12:192, 12:194-12:195:02, 12:197, 12:199, 12:201- 12:203, 12:205-12:206, 12:209-12:211, 12:213, 12:215-12:216, 12:218, 12:220, 12:223, 12:225, 12:227, 12:229-12:230, 12:232N, 12:235, 12:237- 12:238, 12:242, 12:244- 12:246, 12:248-12:249, 12:253-12:254, 12:256- 12:257, 12:259-12:260, 12:262, 12:264-12:267, 12:269, 12:271-12:274N, 12:276-12:278, 12:282- 12:284, 12:286, 12:288- 12:293, 12:295N, 12:297, 12:299N-12:302	
11	230 bp	1070 bp	*12:09, 12:24, 12:158	*01:02:01:01-01:03, 01:06-01:07:01, 01:08, 01:10-01:20, 01:23-01:34, 01:37N-01:48, 01:51-01:54, 01:56N- 01:78, 01:80-01:119, 01:121Q-01:130, 01:132-01:157, 01:159, 01:161- 01:175, 01:177-01:187, 03:58, 03:86,

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

				03:94, 04:37, 04:230, 04:263, 05:16, 05:85, 05:107, 06:05-06:06, 06:204, 06:271, 08:12:01:01-08:12:01:02, 14:02:01:01-14:02:31, 14:02:33-14:05, 14:07N, 14:10-14:14, 14:17-14:27, 14:29-14:52, 14:55-14:62, 14:64-14:76, 14:78-14:86, 14:88-14:89, 14:91, 14:93N-14:104, 14:106-14:116, 15:102, 16:53, 16:68, 16:88, <b>B*07:239, B*14:03</b>
12	135 bp	1070 bp	*12:02:01-12:02:22, 12:02:23 <sup>w</sup> , 12:02:24-12:02:27, 12:02:28 <sup>w</sup> , 12:02:30-12:03:03, 12:03:05-12:03:23, 12:03:24 <sup>w</sup> , 12:03:25-12:03:33, 12:03:35-12:03:61, 12:06-12:08, 12:10:01-12:16:01, 12:17-12:20, 12:22-12:26, 12:28-12:32, 12:34-12:40, 12:42Q-12:53, 12:55-12:59, 12:61-12:71, 12:72 <sup>w</sup> , 12:73-12:122, 12:124-12:134, 12:135 <sup>w</sup> , 12:136-12:145, 12:147-12:153, 12:154 <sup>w</sup> , 12:155Q-12:187, 12:189-12:194, 12:195:02, 12:196-12:304	*01:02:34, 01:21, 02:12 <sup>w</sup> , 02:27:01-02:27:02, 02:87, 02:115, 02:126 <sup>w</sup> , 02:131, 03:04:25, 04:11, 04:29, 04:36, 04:55, 04:172, 04:214 <sup>w</sup> , 06:11, 06:122 <sup>w</sup> , 06:124 <sup>w</sup> , 06:147, 06:217, 06:248, 06:252, 07:01:13, 07:01:74, 07:02:09, 07:04:01:01-07:04:11, 07:04:13-07:04:19, 07:11-07:12, 07:45, 07:63, 07:68, 07:101, 07:139, 07:142, 07:181, 07:199:01-07:199:02, 07:272, 07:302, 07:323-07:324, 07:329N, 07:338, 07:354-07:355, 07:358, 07:364-07:365, 07:378, 07:394-07:395, 07:420, 07:426, 07:428, 07:447, 07:459, 07:466-07:467, 07:480, 07:487, 07:501, 07:523, 07:534-07:535, 07:552, 07:562-07:563, 07:569, 07:583, 07:585-07:586, 07:600:01N-07:600:02N, 07:622, 07:626, 07:651, 07:655, 07:656 <sup>w</sup> , 07:664, 07:672N, 07:674, 07:693, 07:723, 07:742, 07:751N, 07:780, 07:797N, 07:831, 08:01:01:01-08:02:10, 08:02:12-08:09, 08:11-08:63, 08:65-08:94, 08:95 <sup>w</sup> , 08:96-08:203, 08:205-08:206, 14:02:03, 14:03:01:01-14:03:01:02, 14:03:03, 14:08, 14:10, 14:22, 14:35N, 14:38, 14:41, 14:53-14:54, 14:61, 14:70, 14:79, 14:86, 14:90, 14:112-14:113, 14:116, 15:07:01:01-15:07:01:02, 15:21 <sup>w</sup> , 15:25, 15:116 <sup>w</sup> , 15:144, 16:01:01:01-16:01:01:15, 16:01:03-16:01:25, 16:01:27-16:01:36, 16:04:01:01-16:04:01:03, 16:04:03-16:04:05, 16:06-16:08, 16:10-16:11, 16:13-16:18, 16:20-16:24, 16:26-16:36, 16:37 <sup>w</sup> , 16:38-16:45, 16:49-16:56, 16:58-16:59, 16:61-16:62, 16:64-16:68, 16:71 <sup>w</sup> , 16:72-16:73, 16:75-16:76, 16:78-16:83, 16:86-16:87, 16:92-16:93, 16:95-16:98, 16:100, 16:105, 16:109-16:114,

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

				16:116-16:119, 16:122, 16:124-16:125, 16:127-16:131, 16:134-16:135, 16:137-16:139, 16:141-16:142, 16:146-16:152, 16:154, 16:157-16:162, 16:164, <b>B*07:02:77, B*08:01:36, B*15:01:50, B*15:434, B*15:436, B*18:03:02, B*35:08:02, B*35:08:05, B*67:02:01:01-67:02:01:02</b>
<b>13<sup>4</sup></b>	105 bp	1070 bp	*12:31	*04:01:05, 04:01:75, 05:106:02, 08:01:19, 08:160
	150 bp		*12:10:01-12:10:02, 12:155-12:156	*04:01:05, 04:01:75, 14:02:08, 14:02:21
<b>14<sup>4</sup></b>	100 bp	1070 bp	*12:16:01, 12:147, 12:227, 12:279	*01:02:34, 01:21, 02:42, 02:107, 02:152, 04:140, 04:166:01, 04:220, 05:98, 05:197, 06:02:72, 06:05 <sup>w</sup> , 07:01:74, 07:02:09, 08:14, 08:80, 08:103, 15:63, 15:113, 16:80, <b>B*15:436, B*67:02:01:01-67:02:01:02</b>
	150 bp		*12:11, 12:60, 12:118	
<b>15</b>	140 bp	1070 bp	*12:12, 12:301	*02:131, 08:73, <b>B*07:02:77, B*08:01:36, B*15:01:50, B*15:436, B*18:03:02, B*35:08:02, B*35:08:05</b>
<b>16</b>	185 bp 225 bp	1070 bp	*12:13:01:01-12:13:01:02 *12:14:01-12:14:02, 12:176	*01:60, 02:180, 04:58, 04:160, 04:368, 05:23, 05:62, 05:134, 05:143, 05:151, 06:118, 08:07, 08:47, 08:104, 08:188, 14:17, 15:65, 17:01:01:02-17:31, 17:33-17:34, 17:36-17:51:01:02
<b>17</b>	130 bp 565 bp	1070 bp	*12:99:02, 12:159 *12:03:04, 12:03:09, 12:195:01	<b>B*40:02:21<sup>w</sup></b> *01:02:18, 03:03:40, 06:02:38, 07:447, 14:02:08, <b>B*27:05:27, B*57:01:24</b>
<b>18</b>	145 bp 175 bp 245 bp 270 bp	1070 bp	*12:17, 12:27 *12:35, 12:201 *12:17, 12:27 *12:35	*04:12  *03:53, <b>A*68:226</b> <b>A*68:166, B*07:297, B*38:60, B*55:92</b>
<b>19<sup>4</sup></b>	100 bp	1070 bp	*12:14:02, 12:18:01, 12:25, 12:83, 12:169	*03:104, 07:01:20, 07:01:27, 07:02:38, 07:17:04, 07:581, 08:122, 15:02:39, <b>A*03:01:64, A*30:01:14, A*33:188, B*07:365, B*08:144, B*15:492, B*18:77, B*27:148, B*35:373, B*39:66, B*40:29, B*44:32</b>
<b>20<sup>4</sup></b>	105 bp 175 bp	1070 bp	*12:46N *12:22, 12:58, 12:94 12:252	*01:129, 04:52, 04:55, 04:405, 05:55, 14:10, 14:48, 15:12, 15:144
	230 bp		*12:19, 12:139, 12:158	*01:31, 14:38
<b>21</b>	250 bp	1070 bp	*12:14:01-12:14:02, 12:18:01-12:18:02, 12:20, 12:83, 12:169, 12:175	*01:22, 01:35, 01:131, 01:160, 05:11, 05:17, 05:27, 05:68, 05:79, 05:115, 05:134, 05:151, 05:184, 06:04:01-06:04:02, 06:118, 06:153, 06:197, 08:01:01:01-08:01:29,

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

				08:03:01-08:04:03, 08:06, 08:08:01-08:11, 08:13-08:14, 08:16:01-08:16:02, 08:20-08:22:01:02, 08:24, 08:26N, 08:36N, 08:38-08:42, 08:44, 08:46, 08:50, 08:54, 08:56-08:61, 08:65-08:66, 08:72:01-08:72:02, 08:78-08:89N, 08:91, 08:93, 08:95-08:99, 08:101-08:102, 08:104-08:106, 08:109, 08:113, 08:117, 08:119, 08:121N-08:122, 08:124, 08:127N-08:131, 08:133, 08:135-08:139, 08:141Q, 08:143-08:145, 08:147-08:148, 08:153-08:155, 08:157, 08:162-08:165, 08:168, 08:173N-08:178, 08:182, 08:186-08:190, 08:192-08:194, 08:196-08:197, 08:199, 08:203, 08:205, 14:06, 14:15, 14:53, 14:77, 14:87, 15:02:01:01-15:07:01:02, 15:09:01:01-15:13:01:02, 15:15-15:19, 15:21-15:24, 15:26-15:50, 15:52:01-15:73, 15:76-15:83, 15:85-15:101, 15:103-15:142, 15:144-15:147, 15:149-15:207, 16:35, 16:40, 16:48, 16:110, 17:01:01:02-17:16:02, 17:18-17:51:01:02
<b>22<sup>4</sup></b>	100 bp 135 bp	1070 bp	*12:15, 12:23 *12:99:01-12:99:02, 12:235	*15:02:14 *03:03:40, <b>B*40:02:21<sup>w</sup></b>
<b>23<sup>4</sup></b>	590 bp 105 bp	1070 bp	*12:21, 12:203 *12:205	*05:106:02, 08:01:19 <b>*06:56, 07:620, 15:147, A*02:362, A*26:85, B*13:80, B*18:116, B*35:326, B*44:38</b>
<b>24</b>	140 bp 185 bp 135 bp 185 bp 425 bp	1070 bp	*12:26, 12:63 *12:172, 12:201 *12:99:01, 12:235 *12:43 *12:28, 12:135	*03:465, 07:470, 16:36 *03:477, 03:496, 04:337 *03:03:40 *04:01:05, 06:02:38, 06:76:02, 07:447
<b>25<sup>4</sup></b>	80 bp 155 bp  430 bp	1070 bp	*12:39N *12:02:06, 12:02:08, 12:02:12, 12:21, 12:118, 12:149 *12:167, 12:243	*02:02:44, 04:01:05, 05:106:02, 07:413, 07:422, 08:01:19, 08:02:02, <b>B*27:05:27, B*40:02:21</b> *02:12, 02:49 <sup>w</sup> , 02:55:01 <sup>w</sup> -02:55:02 <sup>w</sup> , 02:115, 04:226 <sup>w</sup> , 15:03 <sup>w</sup> , 15:16 <sup>w</sup> , 15:25
<b>26<sup>5</sup></b>	355 bp	<b>800 bp</b>	*12:44	*03:38:01-03:38:02, 03:69, 03:130, 03:136, 03:163, 03:246, 03:274, 03:297, 03:461, 04:80, 04:100, 04:178, 04:370, 06:14, 06:143, 07:10, 07:43:01, 07:196, 07:367, 07:568, 15:03, 15:16, 15:25



101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

**Lot No.: 7K1**

**Lot-specific information**

<b>27<sup>4</sup></b>	100 bp 150 bp  175 bp 215 bp 295 bp	1070 bp	*12:30 *12:03:19, 12:03:32, 12:155Q *12:94 *12:36, 12:153 *12:101	*07:214, 07:429 *01:04, 14:02:21, 16:02:13  *01:129, 14:48, <b>A*02:605Q</b> *16:103 *14:84, 16:81, 16:143
<b>28<sup>6</sup></b>	165 bp 275 bp  350 bp	1070 bp	*12:232N *12:50  *12:45, 12:166	*01:32:01-01:32:02, 02:56, 03:102, 03:263:01-03:263:02, 04:180:01, 05:217, 06:20, 07:81, 07:168, 07:450, 08:123, 08:139, 14:82, 14:92, 15:126, 16:98, 16:102, 16:110 *02:159, 02:161, 05:81, 06:87, 07:24, 07:218, 14:65, 16:13, 16:61
<b>29<sup>4</sup></b>	125 bp  185 bp  210 bp	1070 bp	*12:38, 12:104N, 12:219N  *12:42Q, 12:172  *12:29, 12:86, 16:119	*07:820N, 14:93N, <b>B*15:181N</b> <b>B*57:50,</b> *03:477, 03:496, 04:337, 07:513Q, <b>B*46:51Q</b>
<b>30<sup>4,7</sup></b>	90 bp  200 bp 230 bp	1070 bp	*12:32, 12:102, 12:144, 12:185 *12:162 *12:34	*02:51, 05:08, 05:52, 05:89, 06:41, 08:29, 08:31, <b>B*15:33, B*15:248</b> *02:64
<b>31</b>	135 bp 180 bp 240 bp	1070 bp	*12:47, 12:84N, 12:123 *12:42Q, 12:80N *12:164	*02:133, <b>A*11:197, A*26:67, A*68:95</b> *07:513Q, <b>B*46:51Q</b> *15:67
<b>32<sup>4</sup></b>	55 bp   115 bp  180 bp	1070 bp	*12:54, 12:188   *12:37  *12:62	*01:59, 01:118, 02:65, 03:130, 03:140, 03:243, 04:114, 04:383, 05:20, 06:82, 06:210, 07:49, 07:210, 07:238, 07:247, 07:403, 14:04, 14:64, 14:77, 15:85, 15:181, 16:57, <b>A*03:267, A*68:46,</b> <b>B*07:253</b> *07:204:01, 07:482, <b>A*02:211:01,</b> <b>A*02:594, A*02:817, A*24:261,</b> <b>A*24:445N, A*68:76:01-68:76:02</b> <b>A*02:335</b>
<b>33</b>	135 bp 210 bp	1070 bp	*12:105N, 12:219N *12:100	*07:820N
<b>34<sup>4</sup></b>	85 bp 235 bp  260 bp	1070 bp	*12:171 *12:109  *12:125	*02:170, 03:171, 03:211:01, 04:144, 05:93, 06:73, 08:20, 08:40
<b>35</b>	195 bp  260 bp	1070 bp	*12:110, 12:143, 12:278  *12:125	*01:173, 03:411, 04:375, 04:380, 06:227, 15:154
<b>36</b>	180 bp	1070 bp	*12:62, 12:111, 12:149	*03:282, 07:413, 07:422, 08:18, <b>A*02:335, B*35:339, B*40:346</b>
<b>37<sup>4</sup></b>	90 bp 225 bp 285 bp	1070 bp	*12:163 *12:108 *12:73	<b>B*14:51, B*53:34, B*58:21</b>
<b>38<sup>4</sup></b>	105 bp 145 bp	1070 bp	*12:15, 12:113, 12:282 *12:156	*15:02:14, <b>B*40:02:21, B*57:01:24</b> *04:01:05, 14:02:08

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

**Lot No.: 7K1**

**Lot-specific information**

<b>39</b>	200 bp 225 bp	1070 bp	*12:136 *12:59, 12:82	*04:280 <sup>w</sup> *07:102, 07:351, <b>A*25:55, B*07:13, B*07:15, B*07:160, B*42:18, B*67:02:01:01-67:02:01:02</b>
<b>40<sup>4,7</sup></b>	65 bp	1070 bp	*12:04:01-12:04:02:02, 12:41, 12:60, 12:72, 12:135, 12:146, 12:154	*02:12, 02:49, 02:55:01-02:55:02, 03:15, 03:163, 03:297, 03:461, 04:01:01:01-04:01:01:29, 04:01:01:31- 04:01:47, 04:01:49-04:01:106:02, 04:01:108-04:01:120, 04:03:01:01- 04:09N, 04:12-04:20, 04:23-04:28, 04:30-04:35:02, 04:37-04:44, 04:46- 04:54:02, 04:56-04:96, 04:98:01- 04:113, 04:115N-04:147, 04:149- 04:152, 04:154-04:168, 04:170N- 04:171, 04:173N-04:209, 04:211- 04:213, 04:216-04:249, 04:251- 04:284, 04:286-04:332, 04:334- 04:345, 04:348-04:382, 04:384- 04:405, 05:42, 05:46, 06:02:01:01- 06:02:01:43, 06:02:03-06:02:23, 06:02:25-06:02:50:02, 06:02:52- 06:04:02, 06:06-06:10, 06:12- 06:43:02, 06:45-06:64, 06:66-06:81, 06:83-06:123, 06:125-06:137, 06:139- 06:146, 06:148-06:197, 06:199- 06:209, 06:211:01:01N-06:216, 06:218-06:247, 06:249-06:251, 06:253-06:292, 07:07, 07:09, 07:76:01-07:76:02, 07:315, 07:406, 07:559, 07:598, 07:656, 14:49, 15:03, 15:16, 16:25, 17:01:01:02-17:01:14, 17:01:16-17:14, 17:16:01-17:21, 17:23-17:39, 17:41-17:51:01:02, 18:01:01:01-18:14
<b>41</b>	215 bp 240 bp	1070 bp	*12:90, 12:148N *12:164	<b>B*56:08</b> *15:67
<b>42<sup>4</sup></b>	125 bp	1070 bp	*12:122	<b>*03:03:40, B*35:01:54, B*35:08:10, B*35:12:03, B*44:03:24, B*51:01:79, B*53:01:06, B*57:01:24, B*58:01:33</b>
<b>43</b>	180 bp	1070 bp	*12:28, 12:132, 12:135, 12:146, 12:287	*02:49, 02:75, 02:115, 04:01:01:01- 04:01:01:29, 04:01:01:31-04:01:09, 04:01:11-04:01:22, 04:01:24-04:01:73, 04:01:74 <sup>w</sup> , 04:01:75-04:01:120, 04:03:01:01-04:07:01, 04:08-04:10, 04:12-04:20, 04:23-04:26, 04:28- 04:32, 04:34-04:51, 04:53-04:54:02, 04:56-04:106, 04:108-04:115N, 04:117-04:129, 04:131-04:166:01, 04:167-04:168, 04:170N-04:171, 04:173N-04:230, 04:232-04:282, 04:284-04:404, 05:25, 05:42, 06:02:72, 06:05, 06:76:02, 07:01:74, 07:02:09,

**101.624-12 – including *Taq* pol., IFU-01**  
**101.624-12u – without *Taq* pol., IFU-02**

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
**“Instructions for Use” (IFU)**

**Lot No.: 7K1**

**Lot-specific information**

				07:583, 08:28, 08:137, 08:168, 14:116, 15:25, 15:62, 15:169, 16:26, 16:46, 16:55, 16:64
<b>44</b>	145 bp	1070 bp	*12:16:01, 12:147, 12:195:02, 12:217	*02:14:01-02:14:02, 02:107, 02:164, 04:42:01-04:42:02, 04:220, 05:43, 06:02:72, 06:05, 07:01:74, 07:02:09, 08:37, 15:23:01-15:23:02, 15:63, 15:138, 15:158, 16:21, 16:80
	290 bp		*12:160	
<b>45</b>	245 bp	1070 bp	*12:181	*07:01:01:01-07:01:02:08, 07:01:04-07:01:88, 07:01:90-07:02:20, 07:02:22-07:02:104:02, 07:02:106:01-07:33N, 07:35-07:40, 07:42-07:165, 07:167-07:294, 07:296-07:347N, 07:349-07:622, 07:624-07:765, 07:767-07:798, 07:800-07:831, 07:833-07:837
<b>46</b>	175 bp 190 bp	1070 bp	*12:165, *12:143, 12:278	*02:148 *01:173, 03:411, 04:375, 04:380, 06:227, 15:154
<b>47</b>	370 bp	1070 bp	*12:02:01-12:02:02:08, 12:02:04-12:02:07, 12:02:09-12:02:11, 12:02:13-12:02:34, 12:03:29, 12:03:32, 12:08, 12:14:01-12:14:02, 12:16:01-12:18:02, 12:22, 12:27, 12:30, 12:36, 12:40-12:41, 12:56, 12:64, 12:67-12:69, 12:72-12:74, 12:80N, 12:83-12:86, 12:103-12:106, 12:112, 12:114, 12:117, 12:123-12:124, 12:126-12:128, 12:130, 12:132, 12:134, 12:136-12:137, 12:142, 12:145-12:146, 12:148N, 12:151, 12:161, 12:164, 12:166, 12:168-12:169, 12:177, 12:179, 12:183, 12:193, 12:196, 12:198, 12:200, 12:204, 12:207, 12:212, 12:214, 12:217, 12:219N, 12:224, 12:226, 12:228, 12:231, 12:234, 12:236N, 12:239-12:241, 12:243, 12:247, 12:250, 12:252, 12:255, 12:261, 12:263, 12:268, 12:270N, 12:275, 12:279-12:281, 12:285, 12:287, 12:294, 12:298, 12:303-12:304	*01:21, 02:12, 02:49, 02:55:01-02:55:02, 02:115, 04:01:55, 04:15:01, 04:15:03, 04:294, 04:367, 06:02:37, 06:127:02, 06:264, 07:01:07, 07:04:03, 07:14, 07:27:02, 07:31:02, 07:43:02, 07:199:02, 07:526:01, 07:681, 17:01:10

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

48 <sup>8</sup>	-	-	Negative Control
-----------------	---	---	------------------

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-C\*12 high resolution SSP typings.

When the primers in a primer mix can give rise to HLA-specific PCR products of more than one length this is indicated if the size difference is more than 20 base pairs. Size differences of 20 base pairs or less are not given. For high resolution SSP kits, the alleles listed are specified according to amplicon length.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>For several HLA Class I alleles 1st and/or 4th exon(s) and beyond, as well as intron nucleotide sequences, are not available. In these instances it is not known whether some of the primers of the SSP sets are completely matched with the target sequences or not. Assumption is made that unknown sequences in these regions are conserved within allelic groups.

<sup>4</sup>HLA-specific PCR products shorter than 125 base pairs have a lower intensity and are less sharp than longer PCR products.

<sup>5</sup>Primer mixes 5 and 26 may give rise to a lower yield of HLA-specific PCR product than the other C\*12 primer mixes

<sup>6</sup>Primer mix 28 has a most pronounced tendency to giving rise to primer oligomer formation.

<sup>7</sup>Primer mixes 30 and 40 may have tendencies of unspecific amplifications.

<sup>8</sup>Primer mix 48 contains a negative control, which will amplify a majority of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 200 base pairs.

#### Abbreviations

w: might be weakly amplified.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information  
**PRIMER SPECIFICATION**

Well No.	1	2	3	4	5	6	7	8	9	10	11	12
Length of spec. PCR product	235	100	220	340	130	75	225	95	95	155	230	135
						150		155				
						415		195				
								245				
Length of int. pos. control <sup>1</sup>	800	1070	800	1070	1070	1070	1070	1070	1070	1070	1070	1070
5'-primer(s) <sup>2</sup>	98	419	357	1 <sup>st</sup> I	201	28	118	98	289	361	361	201
	5'-CTA 3'	5'-gTC 3'	5'-CTT 3'	5'-CgA 3'	5'-CCA 3'	5'-TCA 3'	5'-CCg 3'	5'-CTA 3'	5'-Agg 3'	5'-AgT 3'	5'-AgT 3'	5'-CCA 3'
			361			431		420				213
			5'-AgT 3'			5'-CgT 3'		5'-TTA 3'				5'-CCC 3'
			361			499						
			5'-AgT 3'			5'-TCg 3'						
						504						
						5'-CAT 3'						
3'-primer(s) <sup>3</sup>	289	477	538	302	289	270	302	214	341	474	538	302
	5'-AgC 3'	5'-gCA 3'	5'-CCA 3'	5'-ggT 3'	5'-AgT 3'	5'-TAG 3'	5'-ggT 3'	5'-CCA 3'	5'-Cgg 3'	5'-gCA 3'	5'-CCg 3'	5'-ggC 3'
	289				289	538		251			559	
	5'-AgC 3'				5'-AgT 3'	5'-CCA 3'		5'-CCT 3'			5'-CAT 3'	
	295							301				
	5'-TCC 3'							5'-gCC 3'				
								474				
								5'-gCA 3'				
Well No.	1	2	3	4	5	6	7	8	9	10	11	12

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

Well No.	13	14	15	16	17	18	19	20	21	22	23	24
Length of spec.	105	100	140	185	130	145	100	105	250	100	105	135
PCR product	150	150		225	565	175		175		135	140	185
						245		230		590	185	425
						270						
Length of int.	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070
pos. control <sup>1</sup>												
5'-primer(s) <sup>2</sup>	368	142	201	2 <sup>nd</sup> I	201	98	257	201	2 <sup>nd</sup> I	176	201	98
	5'-gTT 3'	5'-TCT 3'	5'-CCA 3'	5'-CCA 3'	5'-CCg 3'	5'-CTA 3'	5'-CCC 3'	5'-CCA 3'	5'-CCA 3'	5'-gCA 3'	5'-CCA 3'	5'-CTA 3'
	375	364			379	201	477	361		379	368	341
	5'-Tgg 3'	5'-ggT 3'			5'-ACg 3'	5'-CCA 3'	5'-gCT 3'	5'-AgT 3'		5'-ACg 3'	5'-gTA 3'	5'-ggA 3'
	412	368			386	3 <sup>rd</sup> I				379	3 <sup>rd</sup> I	379
	5'-ATA 3'	5'-gTC 3'			5'-gCT 3'	5'-Cgg 3'				5'-ACg 3'	5'-Cgg 3'	5'-ACg 3'
										420		
										5'-TTA 3'		
										3 <sup>rd</sup> I		
										5'-Cgg 3'		
3'-primer(s) <sup>3</sup>	474	201	299	473	474	295	311	332	539	474	288	241
	5'-gCA 3'	5'-CTT 3'	5'-TCT 3'	5'-CAA 3'	5'-gCA 3'	5'-TCC 3'	5'-ggT 3'	5'-TCC 3'	5'-TCA 3'	5'-gCA 3'	5'-gCg 3'	5'-CgT 3'
	477	474		512		308	539	343		658	430	474
	5'-gCA 3'	5'-gCA 3'		5'-CCA 3'		5'-Tcg 3'	5'-TCA 3'	5'-T 3'		5'-gTg 3'	5'-gCT 3'	5'-gCA 3'
						326		426			476	
						5'-TgC 3'		5'-TCC 3'			5'-CgA 3'	
						728		506			728	
						5'-CCg 3'		5'-TgT 3'			5'-CCg 3'	
								538			739	
								5'-gCA 3'			5'-gTT 3'	
								559				
								5'-CAT 3'				
								562				
								5'-gCg 3'				
Well No.	13	14	15	16	17	18	19	20	21	22	23	24

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

Well No.	25	26	27	28	29	30	31	32	33	34	35	36
Length of spec.	80	355	100	165	125	90	135	55	135	85	195	180
PCR product	155		150	275	185	200	180	115	210	235	260	
	430		175	350	210	230	240	180		260		
			215									
			295									
Length of int.	1070	<b>800</b>	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070
pos. control <sup>1</sup>												
5'-primer(s) <sup>2</sup>	28	289	127	385	419	347	98	173	419	625	625	173
	5'-TCC 3'	5'-Agg 3'	5'-gAC 3'	5'-ggT 3'	5'-gTC 3'	5'-gTg 3'	5'-CTA 3'	5'-CgA 3'	5'-gTC 3'	5'-ACg 3'	5'-ACg 3'	5'-CgA 3'
	173		228	463	3 <sup>rd</sup> I	370	419	302		652	694	173
	5'-CgC 3'		5'-ATg 3'	5'-TgA 3'	5'-Cgg 3'	5'-ACT 3'	5'-gTC 3'	5'-gAA 3'		5'-CCA 3'	5'-gCA 3'	5'-CgC 3'
	252		361	569		485		379		802	1046	710
	5'-TgA 3'		5'-AgT 3'	5'-gA 3'		5'-CAA 3'		5'-ACC 3'		5'-AgC 3'	5'-TgT 3'	5'-gAA 3'
	361		375			490						
	5'-AgC 3'		5'-Tgg 3'			5'-CgT 3'						
	361					499						
	5'-AgA 3'					5'-TCg 3'						
3'-primer(s) <sup>3</sup>	289	353	289	3 <sup>rd</sup> I	502	538	175	312	508	846	846	312
	5'-AgC 3'	5'-TgA 3'	5'-AgC 3'	5'-CTC 3'	5'-CTA 3'	5'-CCA 3'	5'-CCA 3'	5'-Agg 3'	5'-CTA 3'	5'-CAC 3'	5'-CAC 3'	5'-Agg 3'
	474		477		505		197	453	514		1087	846
	5'-gCA 3'		5'-gCA 3'		5'-gCC 3'		5'-gAT 3'	5'-TCT 3'	5'-CTA 3'		5'-AgC 3'	5'-CAC 3'
			506		508		201		589			
			5'-TgT 3'		5'-CTA 3'		5'-CTT 3'		5'-CTg 3'			
			545		564		251					
			5'-AgA 3'		5'-ACC 3'		5'-CCT 3'					
			613		587		296					
			5'-gCA 3'		5'-CCC 3'		5'-CTg 3'					
					739		564					
					5'-gTT 3'		5'-ACC 3'					
Well No.	25	26	27	28	29	30	31	32	33	34	35	36

Well No.	37	38	39	40	41	42	43	44	45	46	47
Length of spec.	90	105	200	65	215	125	180	145	245	175	370
PCR product	225	145	225		240			290		190	
	285										
Length of int.	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070	1070
pos. control <sup>1</sup>											
5'-primer(s) <sup>2</sup>	98	368	103	289	98	474	201	97	648	404	289
	5'-CTA 3'	5'-gTT 3'	5'-CCT 3'	5'-Agg 3'	5'-CTA 3'	5'-ACT 3'	5'-CCA 3'	5'-TCg 3'	5'-CAC 3'	5'-CCA 3'	5'-Agg 3'
		409	112					832		1046	
		5'-ggC 3'	5'-CCg 3'					5'-AAC 3'		5'-TgT 3'	
			130								
			5'-AgA 3'								
3'-primer(s) <sup>3</sup>	146	474	289	312	265	559	341	201	853	538	369
	5'-CCC 3'	5'-gCA 3'	5'-AgC 3'	5'-AgT 3'	5'-CTA 3'	5'-CAg 3'	5'-CgT 3'	5'-CTT 3'	5'-CAT 3'	5'-CCA 3'	5'-CCg 3'
	284				277			956		1087	
	5'-gTT 3'				5'-gCC 3'			5'-CAg 3'		5'-AgC 3'	
	340				296						
	5'-ggT 3'				5'-CTg 3'						
Well No.	37	38	39	40	41	42	43	44	45	46	47

**101.624-12 – including *Taq* pol., IFU-01**  
**101.624-12u – without *Taq* pol., IFU-02**

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
**“Instructions for Use” (IFU)**

**Lot No.: 7K1**

**Lot-specific information**

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

<sup>2</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.



101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-C*12 SSP primer set <sup>2</sup>																				
				Well																
				1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
				Prod. No.:	202014601	202014602	202014603	202014604	202014605	202014606	202014607	202014608	202014609	202014610	202014611	202014612	202014613	202014614	202014615	202014616
IHW cell line <sup>1</sup>		C*																		
1	9001	SA	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*07:01	*15:05	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*12:02		+	+	-	-	-	-	-	-	+	-	-	+	-	-	-	-
4	9275	GU373	*03:04	*04:01	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*06:02		-	-	+	+	-	-	+	-	-	-	-	-	-	-	-	-
6	9353	SM	*03:04	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*05:01		-	-	-	+	+	-	+	-	-	-	-	-	-	-	-	-
8	9025	DEU	*04:01		-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*12:03		+	-	+	-	-	-	-	-	+	+	-	+	-	-	-	-
10	9107	LKT3	*01:02		-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
11	9051	PITOUT	*16:01		-	-	-	-	+	-	-	-	-	-	-	+	-	-	-	-
12	9052	DBB	*06:02		-	-	+	+	-	-	+	-	-	-	-	-	-	-	-	-
13	9004	JESTHOM	*01:02		-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
14	9071	OLGA	*01:02	*03:04	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
15	9075	DKB	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*02:02		-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*03:03	*07:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257	32367	*01:02	*07:05	-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
19	9038	BM16	*07:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*12:03		+	-	+	-	-	-	-	-	+	+	-	+	-	-	-	-
23	9124	IHL	*01:02	*15:02	-	-	-	+	+	-	-	-	-	-	+	-	-	-	-	-
24	9035	JBUSH	*12:03		+	-	+	-	-	-	-	-	+	+	-	+	-	-	-	-
25	9049	IBW9	*08:02		-	-	-	-	+	-	-	-	-	-	-	+	-	-	-	-
26	9285	WT49	*07:18		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*07:04	*15:29	-	-	-	+	+	-	-	-	-	-	-	+	-	-	-	-
28	9320	BEL5GB	*05:01	*16:01	-	-	-	+	+	-	+	-	-	-	-	+	-	-	-	-
29	9050	MOU	*16:01		-	-	-	-	+	-	-	-	-	-	-	+	-	-	-	-
30	9021	RSH	*17:01		+	-	-	+	-	-	+	-	+	-	-	-	-	-	-	+
31	9019	DUCAF	*05:01		-	-	-	+	+	-	+	-	-	-	-	-	-	-	-	-
32	9297	HAG	*17:03		+	-	-	+	-	-	+	-	+	-	-	-	-	-	-	+
33	9098	MT14B	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*12:03		+	-	+	-	-	-	-	-	+	+	-	+	-	-	-	-
35	9302	SSTO	*05:01		-	-	-	+	+	-	+	-	-	-	-	-	-	-	-	-
36	9024	KT17	*03:03	*04:01	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315	CML	*02:02	*07:01	-	-	-	+	+	-	-	-	-	-	-	-	-	-	-	-
40	9134	WHONP199	*01:02	*06:02	-	-	+	+	-	-	+	-	-	-	+	-	-	-	-	-
41	9055	H0301	*08:02		-	-	-	-	+	-	-	-	-	-	-	+	-	-	-	-
42	9066	TAB089	*01:02		-	-	-	-	-	-	-	-	-	-	+	-	-	-	-	-
43	9076	T7526	*01:02	*08:01	-	-	-	-	+	-	-	-	-	-	+	+	-	-	-	-
44	9057	TEM	*12:03		+	-	+	-	-	-	-	-	+	+	-	+	-	-	-	-
45	9239	SHJO	*06:02	*17:01	+	-	+	+	-	-	+	-	+	-	-	-	-	-	-	+
46	9013	SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*07:04	*15:02	-	-	-	+	+	-	-	-	-	-	-	+	-	-	-	-
48	9303	TER-ND	*04:01	*16:01	-	-	-	+	+	-	-	-	-	-	-	+	-	-	-	-



101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-C*12 SSP primer set <sup>2</sup>																				
				Well																
				17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	
				Prod. No.:	202014617	202014618	202014619	202014620	202014621	202014622	202014623	202014624	202014625	202014626	202014627	202014628	202014629	202014630	202014631	202014632
IHCW cell line <sup>1</sup>			C*																	
1	9001	SA	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280	LK707	*07:01	*15:05	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
3	9011	E4181324	*12:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275	GU373	*03:04	*04:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009	KAS011	*06:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353	SM	*03:04	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020	QBL	*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025	DEU	*04:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026	YAR	*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107	LKT3	*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051	PITOUT	*16:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052	DBB	*06:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004	JESTHOM	*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071	OLGA	*01:02	*03:04	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075	DKB	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037	SWEIG007	*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282	CTM3953540	*03:03	*07:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257	32367	*01:02	*07:05	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038	BM16	*07:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059	SLE005	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064	AMALA	*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056	KOSE	*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124	IHL	*01:02	*15:02	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
24	9035	JBUSH	*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049	IBW9	*08:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285	WT49	*07:18		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191	CH1007	*07:04	*15:29	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
28	9320	BEL5GB	*05:01	*16:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050	MOU	*16:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021	RSH	*17:01		-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
31	9019	DUCAF	*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297	HAG	*17:03		-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
33	9098	MT14B	*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104	DHIF	*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302	SSTO	*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024	KT17	*03:03	*04:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065	HHKB	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099	LZL	*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315	CML	*02:02	*07:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134	WHONP199	*01:02	*06:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055	H0301	*08:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066	TAB089	*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076	T7526	*01:02	*08:01	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
44	9057	TEM	*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239	SHJO	*06:02	*17:01	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
46	9013	SCHU	*07:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045	TUBO	*07:04	*15:02	-	-	-	-	+	-	-	-	-	-	-	-	-	-	-	-
48	9303	TER-ND	*04:01	*16:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-



101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

CELL LINE VALIDATION SHEET																				
HLA-C*12 SSP primer set <sup>2</sup>																				
				Prod. No.:	Well															
					33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	
					202014633	202014634	202014635	202014636	202014637	202014638	202014639	202014640	202014641	202014642	202014643	202014644	202014645	202014646	202014647	
	IHWC cell line <sup>1</sup>		C*																	
1	9001 SA		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
2	9280 LK707		*07:01	*15:05	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
3	9011 E4181324		*12:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	+
4	9275 GU373		*03:04	*04:01	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	-
5	9009 KAS011		*06:02		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
6	9353 SM		*03:04	*07:02	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
7	9020 QBL		*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
8	9025 DEU		*04:01		-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	-
9	9026 YAR		*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3		*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051 PITOUT		*16:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB		*06:02		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
13	9004 JESTHOM		*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA		*01:02	*03:04	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB		*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007		*02:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540		*03:03	*07:01	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
18	9257 32367		*01:02	*07:05	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
19	9038 BM16		*07:01		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
20	9059 SLE005		*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA		*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056 KOSE		*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL		*01:02	*15:02	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035 JBUSH		*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9		*08:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
26	9285 WT49		*07:18		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
27	9191 CH1007		*07:04	*15:29	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
28	9320 BEL5GB		*05:01	*16:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
29	9050 MOU		*16:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH		*17:01		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
31	9019 DUCAF		*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
32	9297 HAG		*17:03		-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
33	9098 MT14B		*03:04		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF		*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO		*05:01		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
36	9024 KT17		*03:03	*04:01	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	-
37	9065 HHKB		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
38	9099 LZL		*03:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315 CML		*02:02	*07:01	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
40	9134 WHONP199		*01:02	*06:02	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
41	9055 H0301		*08:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
42	9066 TAB089		*01:02		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526		*01:02	*08:01	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
44	9057 TEM		*12:03		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO		*06:02	*17:01	-	-	-	-	-	-	-	+	-	-	-	-	-	-	-	-
46	9013 SCHU		*07:02		-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
47	9045 TUBO		*07:04	*15:02	-	-	-	-	-	-	-	-	-	-	-	-	+	-	-	-
48	9303 TER-ND		*04:01	*16:01	-	-	-	-	-	-	-	+	-	-	+	-	-	-	-	-



**101.624-12 – including *Taq* pol., IFU-01**  
**101.624-12u – without *Taq* pol., IFU-02**

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
**“Instructions for Use” (IFU)**

**Lot No.: 7K1**

**Lot-specific information**

<sup>1</sup>The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

<sup>2</sup>The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 6, 8, 13 to 15, 17 to 20, 22 to 39, 41, 42, 44 and 46 were available. The specificities of the primers in primer solutions 6, 8, 13 to 15, 17, 19, 20, 22, 24 to 28, 30, 32, 38, 39, 42 and 44 were tested by separately adding additional 5'-primers, respectively additional 3'-primers. In primer solutions 18, 23, 29, 31, 33, 37 and 41 it was only possible to test the 5'-primers, the 3'-primers were not possible to test. In primer solutions 34 to 36 and 46 it was only possible to test the 3'-primers, the 5'-primers were not possible to test. In primer solutions 3, 6, 13, 14, 17, 19, 25, 27, 28, 30, 32, 39 and 44 one or more 5'-primers were not possible to test, and in primer solutions 1, 8, 11, 16, 20, 22, 24 and 27 one or more 3'-primers were not possible to test.

101.624-12 – including *Taq* pol., IFU-01  
101.624-12u – without *Taq* pol., IFU-02

Visit [www.labproducts.caredx.com](http://www.labproducts.caredx.com) for  
“Instructions for Use” (IFU)

Lot No.: **7K1**

Lot-specific information

**ADDRESSES:**

**Manufacturer:**

**Caredx AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-508 939 00

**Fax:** +46-8-717 88 18

**E-mail:** [orders-se@caredx.com](mailto:orders-se@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

**Distributed by:**

**CareDx GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [orders-at@caredx.com](mailto:orders-at@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

**CareDx Lab Solutions Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-653-78171

**Fax:** 610-344-7989

**E-mail:** [orders-us@caredx.com](mailto:orders-us@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx GmbH**.