

## Olerup SSP® HLA-Cw\*08

Product number:	101.623-12 – including <i>Taq</i> polymerase
Lot number:	79F
Expiry date:	2011-May-01
Number of tests:	12
Number of wells per test:	16
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

**This Product Description is only valid for Lot No. 79F.**

### CHANGES COMPARED TO THE PREVIOUS OLERUP SSP® HLA-Cw\*08 LOT

The HLA-Cw\*08 specificity and interpretation tables have been updated for the HLA-Cw alleles described since the previous *Olerup SSP®* HLA-Cw\*08 lot was made (**Lot No. 23E**).

The primers of the wells detailed below have been exchanged, added or modified compared to the previous lot.

Well	5'-primer	3'-primer	rationale
6	-	Added	New primer added for the Cw*0823 allele.
9	Exchanged	Exchanged	New primer pair for improved resolution.
10	Added	Added	New primer pair for the Cw*0817 allele.
11	Added	Added	New primer pair for the Cw*0820 allele.
14	Added	Added	New primer pair for the Cw*0818 allele.
15	Added	Added	New primer pair for the Cw*0819 allele.
16	Added	Added	New primer pair for the Cw*0822 allele.

## PRODUCT DESCRIPTION

### HLA-Cw\*08 SSP typing

#### CONTENT

The primer set contains 5'- and 3'-primers for identifying the Cw\*0801 to Cw\*0823 alleles.

#### PLATE LAYOUT

Each HLA-Cw\*08 test consists of 16 PCR reactions in a 16 well cut PCR plate.

1	2	3	4	5	6	7	8
9	10	11	12	13	14	15	16

The 16 well PCR plate is marked with 'Cw\*08' in silver/gray ink.

Well No. 1 is marked with the Lot No. '79F'.

The PCR plates are heat-sealed with a PCR-compatible foil.

**Please note:** When removing each 16 well PCR plate, make sure that the remaining plates stay sealed. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### INTERPRETATION

The interpretation of HLA-Cw\*08 SSP subtypings will be influenced by other HLA-Cw alleles, as primer mixes 1, 3, 5, 7, 8, 11, 13 and 16 amplify non-HLA-Cw\*08 alleles. In addition, primer mix 1 will amplify the B\*5802 allele, primer mix 5 will amplify the B\*1533 allele and primer mix 16 will amplify B\*6702.

#### UNIQUELY IDENTIFIED ALLELES

All the HLA-Cw\*08 alleles, i.e. **Cw\*0801 to Cw\*0823**, recognized by the HLA Nomenclature Committee in May 2009<sup>1</sup> will be amplified by the primers in the HLA-Cw\*08 SSP kit.

The HLA-Cw\*08 subtyping kit cannot distinguish the Cw\*080101 to Cw\*080102 alleles.

<sup>1</sup>HLA-B alleles listed on the IMGT/HLA web page 2009-May-09, release 2.25.2, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

#### RESOLUTION IN HOMO- AND HETEROZYGOTES

The twenty-three HLA-Cw\*08 alleles can be combined in 276 homozygous and heterozygous combinations. 106 of these genotypes do not give rise to unique amplification patterns. The different sizes of the specific PCR products generated by primer mixes 6, 10, 11, 14 to 16 were not considered in these calculations.

Lot No.: **79F**

Lot-specific information

[www.olerup.com](http://www.olerup.com)

+++++--- -++++--- 0802,0806 = 0803,0823 = 0806,0823  
+++++--- -++++--- 0805,0820 = 0812,0821  
+++++--- -++++--- 0805,0808 = 0818,0821  
+++++--- -+++++--- 0805,0816 = 0819,0821  
+++++--- -++++--- 0801,0805 = 0802,0821 = 0805,0821  
+++++--- -+++++--- 0802,0809 = 0809,0817 = 0811,0817  
+++++--- -++++--- 0808,0812 = 0818,0820  
+++++--- -+++++--- 0812,0816 = 0819,0820  
+++++--- -++++--- 0801,0812 = 0802,0820 = 0812,0820  
+++++--- -+++++--- 0801,0818 = 0802,0808 = 0808,0818  
+++++--- -+++++--- 0801,0819 = 0802,0816 = 0816,0819  
+++++--- -+++++--- 0801,0813 = 0804,0816  
++-++++- -++++--- 0806,0815 = 0806,0821  
++-++++- -+++++--- 0814,0815 = 0814,0821  
++-++++- -++++--- 0803,0815 = 0803,0821  
++-++++- -+++++--- 0806,0814 = 0806,0822  
++-++++- -++++--- 0801,0806 = 0803,0806 = 0806,0806  
++-++++- -+++++--- 0801,0814 = 0803,0814 = 0803,0822 = 0814,0814  
= 0814,0822  
++-++++- -++++--- 0801,0803 = 0803,0803  
++-++++- -+++++--- 0809,0815 = 0809,0821  
++-++++- -++++--- 0811,0815 = 0811,0821  
++-++++- -+++++--- 0810,0815 = 0810,0821  
++-++++- -++++--- 0815,0820 = 0820,0821  
++-++++- -+++++--- 0808,0815 = 0808,0821  
++-++++- -+++++--- 0815,0816 = 0816,0821  
++-++++- -+++++--- 0815,0822 = 0821,0822  
++-++++- -++++--- 0801,0815 = 0801,0821 = 0815,0821 = 0821,0821  
++-++++- -+++++--- 0801,0809 = 0809,0809 = 0809,0811  
++-++++- -++++--- 0801,0811 = 0811,0811  
++-++++- -+++++--- 0801,0820 = 0820,0820  
++-++++- -+++++--- 0801,0808 = 0808,0808  
++-++++- -+++++--- 0801,0822 = 0822,0822  
+-+----- -+++++--- 0802,0813 = 0804,0819 = 0813,0819  
--+++-+--- -++++--- 0805,0823 = 0815,0823  
--+++-+--- -++++--- 0805,0807 = 0807,0815  
--+++-+--- -+++++--- 0805,0817 = 0815,0817  
--+++-+--- -++++--- 0805,0812 = 0812,0815  
--+++-+--- -+++++--- 0805,0818 = 0815,0818  
--+++-+--- -+++++--- 0805,0819 = 0815,0819  
--+++-+--- -++++--- 0802,0805 = 0802,0815 = 0805,0805 = 0805,0815  
--+++-+--- -++++--- 0802,0823 = 0823,0823  
--+++-+--- -++++--- 0802,0807 = 0807,0807  
--+++-+--- -+++++--- 0802,0817 = 0817,0817  
--+++-+--- -++++--- 0802,0812 = 0812,0812  
--+++-+--- -+++++--- 0802,0819 = 0819,0819

0801 = 080101-080102



## SPECIFICITY TABLE

### HLA-Cw\*08 SSP subtyping

Specificities and sizes of the PCR products of the 16 primer mixes used for HLA-Cw\*08 SSP subtyping

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	Amplified HLA-Cw*08 alleles	Other amplified HLA Class I alleles <sup>3</sup>
<b>1<sup>5</sup></b>	250 bp	<b>800 bp</b>	080101-080102, 0803, 0804, 0806, 0808-0811, 0813, 0814, 0816, 0820-0822	0122, 0203, 021601-021602, 0218, 040401-040402, 0406, 0413, 0434, 0511, 0517, 0604, 121401-121402, 1218, 1220, 1406, 150201-1507, 1509-1513, 1515-1521, 1701-1705, <b>B*5802</b>
<b>2<sup>4</sup></b>	115 bp	1070 bp	080101-080102, 0803, 0806, 0808-0811, 0814, 0816, 0820-0822	
<b>3<sup>4</sup></b>	115 bp	<b>800 bp</b>	0802, 0804, 0805, 0807, 0812, 0813, 0817-0819, 0823	05010101-050104, 0503-0508, 0510, 0511, 0513-0516, 0518-0525, 0741
<b>4<sup>4</sup></b>	110 bp	1070 bp	0803, 0806, 0814	
<b>5</b>	155 bp	1070 bp	0805, 0815, 0821	0113, 0509, 0517, 1705, <b>B*1533</b>
<b>6<sup>6</sup></b>	270, 310 bp	1070 bp	0806, 0823	
<b>7<sup>4</sup></b>	100 bp	1070 bp	0807	0523
<b>8</b>	225 bp	1070 bp	0809, 0811	0504, 0768
<b>9</b>	495 bp	1070 bp	0810	
<b>10<sup>4,5,7</sup></b>	110, 140 bp	<b>800 bp</b>	0809, 0817	
<b>11<sup>8</sup></b>	280, 465 bp	<b>800 bp</b>	0812, 0820	0516
<b>12</b>	165 bp	<b>800 bp</b>	080101-0809, 0811, 0812, 0814, 0815, 0817, 0819-0823	
<b>13<sup>4</sup></b>	105 bp	<b>800 bp</b>	0802, 0805, 0807, 0812, 0817-0819, 0823	05010101-050104, 0503-0508, 0510, 0512-0516, 0518-0525, 0741
<b>14<sup>9</sup></b>	170, 280 bp	1070 bp	0808, 0818	
<b>15<sup>10</sup></b>	260, 490 bp	1070 bp	0813, 0816, 0819	
<b>16<sup>4,5,11</sup></b>	100, 540 bp	1070 bp	0814, 0822	0121, 0605, 1216, <b>B*6702</b>

<sup>1</sup>Alleles are assigned by the presence of specific PCR product(s). However, the sizes of the specific PCR products may be helpful in the interpretation of HLA-Cw\*08 SSP subtypings.

When the primers in a primer mix can give rise to specific PCR products of more than one length this is indicated if the size difference is 20 base pairs or more. Size differences shorter than 20 base pairs are not given. For high resolution SSP kits the respective lengths of the specific PCR product(s) of the alleles amplified by these primer mixes are given.

Nonspecific amplifications, i.e. a ladder or a smear of bands, may sometimes be seen. GC-rich primers have a higher tendency of giving rise to nonspecific amplifications than other primers.

PCR fragments longer than the control bands may sometimes be observed. Such bands should be disregarded and do not influence the interpretation of the SSP typings.

PCR fragments migrating faster than the control bands, but slower than a 400 bp fragment may be seen in some gel read-outs. Such bands can be disregarded and do not influence the interpretation of the SSP typings.

Some primers may give rise to primer oligomer artifacts. Sometimes this phenomenon is an inherent feature of the primer pair(s) of a primer mix. More often it is due to other factors such as too low amount of DNA in the PCR reactions, taking too long time in setting up the PCR reactions, working at elevated room temperature or using thermal cyclers that are not pre-heated.

<sup>2</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-Cw\*08 SSP subtyping.

In addition, wells number 3 and 10 to 13 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band.

In the presence of a specific amplification the intensity of the control band often decreases.

<sup>3</sup>Due to the sharing of sequence motifs between HLA-Cw alleles some non-HLA-Cw\*08 alleles will be amplified by primer mixes 1, 3, 5, 7, 8, 11, 13 and 16. In addition, primer mix 1 will amplify the B\*5802 allele, primer mix 5 will amplify the B\*1533 allele and primer mix 16 will amplify B\*6702.

<sup>4</sup>Short specific PCR fragments are less intense and not as sharp as longer specific bands.

<sup>5</sup>Primer mixes 1, 11 and 16 have a tendency of giving rise to non-specific amplifications.

<sup>6</sup>Primer mix 6: Specific PCR fragment of 270 bp in the Cw\*0806 allele. Specific PCR fragment of 310 bp in the Cw\*0823 allele.

<sup>7</sup>Primer mix 10: Specific PCR fragment of 110 bp in the Cw\*0809 allele. Specific PCR fragment of 140 bp in the Cw\*0817 allele.

<sup>8</sup>Primer mix 11: Specific PCR fragment of 280 bp in the Cw\*0812 and in the Cw\*0516 allele. Specific PCR fragment of 465 bp in the Cw\*0820 allele.

<sup>9</sup>Primer mix 14: Specific PCR fragment of 170 bp in the Cw\*0818 allele. Specific PCR fragment of 280 bp in the Cw\*0808 allele.

<sup>10</sup>Primer mix 15: Specific PCR fragment of 260 bp in the Cw\*0813 and \*0816 alleles. Specific PCR fragment of 490 bp in the Cw\*0819 allele.

<sup>11</sup>Primer mix 16: Specific PCR fragment of 100 bp in the Cw\*0814 and Cw\*0121, 0605, 1216 and in the B\*6702 allele. Specific PCR fragment of 540 bp in the Cw\*0822 allele.

<b>INTERPRETATION TABLE</b>								
<b>HLA-Cw*08 SSP subtyping</b>								
<b>Amplification patterns of the HLA-Cw*0801 to 0823 alleles</b>								
	<b>Well<sup>4</sup></b>							
	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>
<b>Length of spec.</b>	<b>250</b>	<b>115</b>	<b>115</b>	<b>110</b>	<b>155</b>	<b>270</b>	<b>100</b>	<b>225</b>
<b>PCR product</b>						<b>310</b>		
<b>Length of int.</b>	<b>800</b>	<b>1070</b>	<b>800</b>	<b>1070</b>	<b>1070</b>	<b>1070</b>	<b>1070</b>	<b>1070</b>
<b>pos. control<sup>1</sup></b>								
<b>5'-primer<sup>2</sup></b>	<b>2<sup>nd</sup> I</b>	<b>527</b>	<b>527</b>	<b>527</b>	<b>176</b>	<b>2<sup>nd</sup> I</b>	<b>453</b>	<b>419</b>
	<small>5'-CCA<sup>3'</sup></small>	<small>5'-TAC<sup>3'</sup></small>	<small>5'-TgA<sup>3'</sup></small>	<small>5'-TAC<sup>3'</sup></small>	<small>5'-gCA<sup>3'</sup></small>	<small>5'-CCA<sup>3'</sup></small>	<small>5'-AAT<sup>3'</sup></small>	<small>5'-gTC<sup>3'</sup></small>
					<b>485</b>			
					<small>5'-CAA<sup>3'</sup></small>			
<b>3'-primer<sup>3</sup></b>	<b>539</b>	<b>601</b>	<b>601</b>	<b>595</b>	<b>289</b>	<b>559</b>	<b>512</b>	<b>601</b>
	<small>5'-TCA<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>	<small>5'-CCT<sup>3'</sup></small>	<small>5'-AgC<sup>3'</sup></small>	<small>5'-CgC<sup>3'</sup></small>	<small>5'-CCA<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>
					<b>601</b>	<b>599</b>		
					<small>5'-CTC<sup>3'</sup></small>	<small>5'-TCC<sup>3'</sup></small>		
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>
<b>HLA-Cw allele</b>								
<b>*080101-080102</b>	<b>1</b>	<b>2</b>						
<b>*0802</b>			<b>3</b>					
<b>*0803</b>	<b>1</b>	<b>2</b>		<b>4</b>				
<b>*0804</b>	<b>1</b>		<b>3</b>					
<b>*0805</b>			<b>3</b>		<b>5</b>			
<b>*0806</b>	<b>1</b>	<b>2</b>		<b>4</b>		<b>6</b>		
<b>*0807</b>			<b>3</b>				<b>7</b>	
<b>*0808</b>	<b>1</b>	<b>2</b>						
<b>*0809</b>	<b>1</b>	<b>2</b>						<b>8</b>
<b>*0810</b>	<b>1</b>	<b>2</b>						
<b>*0811</b>	<b>1</b>	<b>2</b>						<b>8</b>
<b>*0812</b>			<b>3</b>					
<b>*0813</b>	<b>1</b>		<b>3</b>					
<b>*0814</b>	<b>1</b>	<b>2</b>		<b>4</b>				
<b>*0815</b>					<b>5</b>			
<b>*0816</b>	<b>1</b>	<b>2</b>						
<b>*0817</b>			<b>3</b>					
<b>*0818</b>			<b>3</b>					
<b>*0819</b>			<b>3</b>					
<b>*0820</b>	<b>1</b>	<b>2</b>						
<b>*0821</b>	<b>1</b>	<b>2</b>			<b>5</b>			
<b>*0822</b>	<b>1</b>	<b>2</b>						
<b>*0823</b>			<b>3</b>			<b>6</b>		
<b>Well No.</b>	<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	<b>8</b>

<b>INTERPRETATION TABLE</b>								
<b>HLA-Cw*08 SSP subtyping</b>								
<b>Amplification patterns of the HLA-Cw*0801 to 0823 alleles</b>								
<b>Well<sup>4</sup></b>								
<b>9</b>	<b>10</b>	<b>11</b>	<b>12</b>	<b>13</b>	<b>14</b>	<b>15</b>	<b>16</b>	
<b>495</b>	<b>110</b>	<b>280</b>	<b>165</b>	<b>105</b>	<b>170</b>	<b>260</b>	<b>100</b>	<b>Length of spec.</b>
	<b>140</b>	<b>465</b>			<b>280</b>	<b>490</b>	<b>540</b>	<b>PCR product</b>
<b>1070</b>	<b>800</b>	<b>800</b>	<b>800</b>	<b>800</b>	<b>1070</b>	<b>1070</b>	<b>1070</b>	<b>Length of int. pos. control<sup>1</sup></b>
<b>312</b>	<b>176</b>	<b>361</b>	<b>176</b>	<b>539</b>	<b>173</b>	<b>1<sup>st</sup>  </b>	<b>142</b>	<b>5'-primer<sup>2</sup></b>
<small>5'-AAA<sup>3'</sup></small>	<small>5'-gCA<sup>3'</sup></small>	<small>5'-AgT<sup>3'</sup></small>	<small>5'-gCA<sup>3'</sup></small>	<small>5'-gCg<sup>3'</sup></small>	<small>5'-CgC<sup>3'</sup></small>	<small>5'-CgA<sup>3'</sup></small>	<small>5'-TCT<sup>3'</sup></small>	
	<b>527</b>	<b>652</b>			<b>363</b>	<b>176</b>	<b>972</b>	
	<small>5'-TAC<sup>3'</sup></small>	<small>5'-CCA<sup>3'</sup></small>			<small>5'-AgC<sup>3'</sup></small>	<small>5'-gCA<sup>3'</sup></small>	<small>5'-CTA<sup>3'</sup></small>	
<b>526</b>	<b>277</b>	<b>601</b>	<b>302</b>	<b>601</b>	<b>302</b>	<b>175</b>	<b>201</b>	<b>3'-primer<sup>3</sup></b>
<small>5'-CgT<sup>3'</sup></small>	<small>5'-gCA<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>	<small>5'-ggC<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>	<small>5'-ggC<sup>3'</sup></small>	<small>5'-CCg<sup>3'</sup></small>	<small>5'-CTT<sup>3'</sup></small>	
	<b>598</b>	<b>956</b>			<b>601</b>	<b>387</b>	<b>1034</b>	
	<small>5'-CTC<sup>3'</sup></small>	<small>5'-CAg<sup>3'</sup></small>			<small>5'-CTT<sup>3'</sup></small>	<small>5'-TCC<sup>3'</sup></small>	<small>5'-AgT<sup>3'</sup></small>	
<b>9</b>	<b>10</b>	<b>11</b>	<b>12</b>	<b>13</b>	<b>14</b>	<b>15</b>	<b>16</b>	<b>Well No.</b>
								<b>HLA-Cw allele</b>
			<b>12</b>					<b>*080101-080102</b>
			<b>12</b>	<b>13</b>				<b>*0802</b>
			<b>12</b>					<b>*0803</b>
			<b>12</b>					<b>*0804</b>
			<b>12</b>	<b>13</b>				<b>*0805</b>
			<b>12</b>					<b>*0806</b>
			<b>12</b>	<b>13</b>				<b>*0807</b>
			<b>12</b>		<b>14</b>			<b>*0808</b>
	<b>10</b>		<b>12</b>					<b>*0809</b>
<b>9</b>								<b>*0810</b>
			<b>12</b>					<b>*0811</b>
		<b>11</b>	<b>12</b>	<b>13</b>				<b>*0812</b>
						<b>15</b>		<b>*0813</b>
			<b>12</b>				<b>16</b>	<b>*0814</b>
			<b>12</b>					<b>*0815</b>
						<b>15</b>		<b>*0816</b>
	<b>10</b>		<b>12</b>	<b>13</b>				<b>*0817</b>
				<b>13</b>	<b>14</b>			<b>*0818</b>
			<b>12</b>	<b>13</b>		<b>15</b>		<b>*0819</b>
		<b>11</b>	<b>12</b>					<b>*0820</b>
			<b>12</b>					<b>*0821</b>
			<b>12</b>				<b>16</b>	<b>*0822</b>
			<b>12</b>	<b>13</b>				<b>*0823</b>
<b>9</b>	<b>10</b>	<b>11</b>	<b>12</b>	<b>13</b>	<b>14</b>	<b>15</b>	<b>16</b>	<b>Well No.</b>

Lot No.: **79F**

Lot-specific information

www.olerup.com

Length of spec.	250	115	115	110	155	270	100	225
PCR product						310		
Well No.	1	2	3	4	5	6	7	8
*0113, 0509					5			
*0122, 0203, 021601, 021602, 0218, 040401, 040402, 0406, 0413, 0434, 0604, 121401-121402, 1218, 1220, 1406, 150201-1507, 1509-1513, 1515-1521, 1701-1704	1							
*0121, 0605, 1216								
*05010101-050104, 0503, 0505-0508, 0510, 0513-0515, 0518-0522, 0524, 0525, 0741			3					
*0504			3					8
*0511	1		3					
*0512								
*0516			3					
*0517, 1705	1				5			
*0523			3			7		
*0768								8
HLA-Cw allele								
Well No.	1	2	3	4	5	6	7	8
B*1533					5			
B*5802	1							
B*6702								
Well No.	1	2	3	4	5	6	7	8

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The two different control primer pairs give rise to either an internal positive control band of 1070 base pairs, for most wells, or a band of 800 base pairs, for some wells.

Well number 1 contains the primer pair giving rise to the shorter, 800 bp, internal positive control band in order to help in the correct orientation of the HLA-Cw\*08 SSP subtyping. In addition, wells number 3 and 10 to 13 contain the primer pair giving rise to the shorter, 800 bp, internal positive control band.

<sup>2</sup>The nucleotide position, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon or the 1<sup>st</sup> or 2<sup>nd</sup> intron, matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The nucleotide position, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. The sequence of the 3 terminal nucleotides of the primer is given.



Lot No.: **79F**

Lot-specific information

www.olerup.com

495	110	280	165	105	170	260	100	Length of spec.
	140	465			280	490	540	PCR product
9	10	11	12	13	14	15	16	Well No.
								*0113, 0509
								*0122, 0203, 021601, 021602, 0218, 040401, 040402, 0406, 0413, 0434, 0604, 121401-121402, 1218, 1220, 1406, 150201-1507, 1509-1513, 1515-1521, 1701-1704
							16	*0121, 0605, 1216
				13				*05010101-050104, 0503, 0505-0508, 0510, 0513-0515, 0518-0522, 0524, 0525, 0741
				13				*0504
								*0511
				13				*0512
		11		13				*0516
								*0517, 1705
				13				*0523
								*0768
								HLA-Cw allele
9	10	11	12	13	14	15	16	Well No.
								B*1533
								B*5802
							16	B*6702
9	10	11	12	13	14	15	16	Well No.

<sup>4</sup>Primer mix 6: Specific PCR fragment of 270 bp in the Cw\*0806 allele. Specific PCR fragment of 310 bp in the Cw\*0823 allele.

Primer mix 10: Specific PCR fragment of 110 bp in the Cw\*0809 allele. Specific PCR fragment of 140 bp in the Cw\*0817 allele.

Primer mix 11: Specific PCR fragment of 280 bp in the Cw\*0812 and in the Cw\*0516 allele. Specific PCR fragment of 465 bp in the Cw\*0820 allele.

Primer mix 14: Specific PCR fragment of 170 bp in the Cw\*0818 allele. Specific PCR fragment of 280 bp in the Cw\*0808 allele.

Primer mix 15: Specific PCR fragment of 260 bp in the Cw\*0813 and \*0816 alleles. Specific PCR fragment of 490 bp in the Cw\*0819 allele.

Primer mix 16: Specific PCR fragment of 100 bp in the Cw\*0814 and Cw\*0121, 0605, 1216 and in the B\*6702 allele. Specific PCR fragment of 540 bp in the Cw\*0822 allele.

CELL LINE VALIDATION SHEET																				
HLA-Cw*08 SSP primer set																				
				Prod. No.:	Well															
					1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16
					200629001	200629002	200958703	200958704	200842805	200958706	200958707	200629008	200958709	200958710	200958711	200629012	200629013	200958714	200958715	200958716
	IHWC cell line		Cw*																	
1	9001 SA		*0702		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
2	9280 LK707		*0701	*1505	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
3	9011 E4181324		*1202		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
4	9275 GU373		*0304	*0401	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
5	9009 KAS011		*0602		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
6	9353 SM		*0304	*0702	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
7	9020 QBL		*0501		-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
8	9025 DEU		*0401		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
9	9026 YAR		*1203		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
10	9107 LKT3		*0102		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	9051 PITOUT		*1601		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
12	9052 DBB		*0602		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
13	9004 JESTHOM		*0102		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
14	9071 OLGA		*0102	*0304	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
15	9075 DKB		*0304		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
16	9037 SWEIG007		*0202		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
17	9282 CTM3953540		*0303	*0701	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
18	9257 32367		*0102	*0705	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
19	9038 BM16		*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
20	9059 SLE005		*0304		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
21	9064 AMALA		*0303		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	9056 KOSE		*1203		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
23	9124 IHL		*0102	*1502	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
24	9035 JBUSH		*1203		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
25	9049 IBW9		*0802		-	-	+	-	-	-	-	-	-	-	-	+	+	-	-	-
26	9285 WT49		*0701		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
27	9191 CH1007		*0704	*1505	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
28	9320 BEL5GB		*0501	*1601	-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
29	9050 MOU		*1601		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
30	9021 RSH		*1701		+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
31	9019 DUCAF		*0501		-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
32	9297 HAG		*1701	*1703	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
33	9098 MT14B		*0304		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
34	9104 DHIF		*1203		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
35	9302 SSTO		*0501		-	-	+	-	-	-	-	-	-	-	-	-	+	-	-	-
36	9024 KT17		*0303	*0401	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
37	9065 HHKB		*0702		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
38	9099 LZL		*0303		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
39	9315 CML		*0202	*0701	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
40	9134 WHONP199		*0602		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
41	9055 H0301		*0802		-	-	+	-	-	-	-	-	-	-	-	+	+	-	-	-
42	9066 TAB089		*0102		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
43	9076 T7526		*0102	*0801	+	+	-	-	-	-	-	-	-	-	-	+	-	-	-	-
44	9057 TEM		*1203		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
45	9239 SHJO		*0602	*1701	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
46	9013 SCHU		*0702		-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
47	9045 TUBO		*0704	*1502	+	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
48	9303 TER-ND		*0401	*1601	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

## CERTIFICATE OF ANALYSIS

### **Olerup SSP® HLA-Cw\*08 SSP**

**Product number:** 101.623-12 – including *Taq* polymerase  
**Lot number:** 79F  
**Expiry date:** 2011-May-01  
**Number of tests:** 12  
**Number of wells per test:** 16

#### **Well specifications:**

Well No.	Production No.	Well No.	Production No.
1	2006-290-01	9	2009-587-09
2	2006-290-02	10	2009-587-10
3	2009-587-03	11	2009-587-11
4	2009-587-04	12	2006-290-12
5	2008-428-05	13	2006-290-13
6	2009-587-06	14	2009-587-14
7	2009-587-07	15	2009-587-15
8	2006-290-08	16	2009-587-16

The specificity of each primer solution of the HLA-Cw\*08 primer set has been tested against 48 well characterized cell line DNAs.

No DNAs carrying the alleles to be amplified by primer solutions 4 to 11 and 14 to 16 were available. The specificity of the primers in primer solutions 5, 7, 8, 9, 11, 14, 15 and 16 were tested by separately adding one additional 5'-primer, respectively one additional 3'-primer. In primer solutions 4, 6 and 10 it was only possible to test the 5'-primer, the 3'-primer was not possible to test. In primer solution 11 and 14, one 5'-primer was not possible to test.

**Results:** No false positive or false negative amplifications were obtained.

**Date of approval:** 2009-June-08

**Approved by:**

**Quality Control, Supervisor**

Lot No.: **79F**

Lot-specific information

[www.olerup.com](http://www.olerup.com)

## Declaration of Conformity

**Product name:** *Olerup* SSP® HLA-Cw\*08  
**Product number:** 101.623-12  
**Lot number:** 79F

**Intended use:** HLA-Cw\*08 high resolution histocompatibility testing

**Manufacturer:** *Olerup* SSP AB  
Hasselstigen 1  
SE-133 33 Saltsjöbaden, Sweden  
**Phone:** +46-8-717 88 27  
**Fax:** +46-8-717 88 18

We, *Olerup* SSP AB, hereby declare that this product, to which this Declaration of Conformity relates is in conformity with the following Standard(s) and other normative document(s) ISO 9001:2000 and ISO 13485:2003, following the provisions of the 98/79/EC Directive on *in vitro* diagnostic medical devices, Annex III, as transposed into the national laws of the Member States of the European Union.

The Technical Documentation File is maintained at *Olerup* SSP AB, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

The Authorized Representative located within the Community is: *Olerup* SSP AB.

Saltsjöbaden, Sweden  
2009-June-08

Olle Olerup  
Managing Director









HLA-Cw\*08  
101.623-12 – including *Taq* polymerase

Product Insert

Page 16 of 16

General “Instructions for Use”

IFU-01 Rev. No. 00 can be downloaded from

Lot No.: **79F**

Lot-specific information

[www.olerup.com](http://www.olerup.com)

**ADDRESSES:**

**Manufacturer:**

**Olerup SSP AB**, Hasselstigen 1, SE-133 33 Saltsjöbaden, Sweden.

**Tel:** +46-8-717 88 27

**Fax:** +46-8-717 88 18

**E-mail:** [info-ssp@olerup.com](mailto:info-ssp@olerup.com)

**Web page:** <http://www.olerup.com>

**Distributed by:**

**Olerup GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [support-at@olerup.com](mailto:support-at@olerup.com)

**Web page:** <http://www.olerup.com>

**Olerup Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-OLERUP1

**Fax:** 610-344-7989

**E-mail:** [info.us@olerup.com](mailto:info.us@olerup.com)

**Web page:** <http://www.olerup.com>

For information on *Olerup* SSP distributors worldwide, contact **Olerup GmbH**.

