

104.201-12 – including *Taq* polymerase104.201-12u – without *Taq* polymeraseLot No.: **0L4**Visit <https://labproducts.caredx.com> for

“Instructions for Use” (IFU)

Lot-specific information

## Olerup SSP<sup>®</sup> KIR HLA Ligand

Product number:	104.201-12 – including <i>Taq</i> polymerase 104.201-12u – without <i>Taq</i> polymerase
Lot number:	0L4
Expiry date:	2024-03-01
Number of tests:	12
Number of wells per test:	7+1
Storage - pre-aliquoted primers:	dark at -20°C
- PCR Master Mix:	-20°C
- Adhesive PCR seals	RT
- Product Insert	RT

### This Product Description is only valid for Lot No. 0L4.

Complete product documentation consists of generic Instructions for Use (IFU), lot specific Product Insert, Worksheet and Certificate.

### CHANGES COMPARED TO THE PREVIOUS OLERUP SSP<sup>®</sup> KIR HLA LIGAND LOT (8H1)

- The product documentation has been updated for new alleles of IMGT 3.38.0

The KIR HLA Ligand specificity and interpretation tables have been updated for the HLA-A, HLA-B and HLA-C alleles described since the previous *Olerup SSP<sup>®</sup>* KIR HLA Ligand lot was made (**Lot No. 8H1**).

The KIR HLA Ligand primer set is unchanged compared to the previous *Olerup SSP<sup>®</sup>* KIR HLA Ligand (**Lot No. 8H1**).

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **OL4**

Lot-specific information

Well **8** contains Negative Control primer pairs, that will amplify the majority of the *Olerup SSP*<sup>®</sup> HLA Class I, DRB, DQB1, DPB1 and DQA1 amplicons as well as all the amplicons generated by the control primer pairs matching the human growth hormone gene.

HLA-specific PCR product sizes range from 75 to 200 base pairs.

The PCR product generated by the positive control primer pair is 200 base pairs.

Length of PCR product	105	200	105	80	75	80	85
<b>5'-primer<sup>1</sup></b>	<b>164</b>	<b>340</b>	<b>440</b>	<b>45</b>	<b>45</b>	<b>43</b>	<b>36</b>
	5'-CAC <sup>3'</sup>	5'-Agg <sup>3'</sup>	5'-TTA <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-Tgg <sup>3'</sup>	5'-TAC <sup>3'</sup>
							<b>36</b>
							5'-TAT <sup>3'</sup>
<b>3'-primer<sup>2</sup></b>	<b>231</b>	<b>2<sup>nd</sup> I</b>	<b>507</b>	<b>59</b>	<b>58</b>	<b>57</b>	<b>47</b>
	5'-TgC <sup>3'</sup>	5'-AAA <sup>3'</sup>	5'-TTg <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ggC <sup>3'</sup>	5'-CTC <sup>3'</sup>	5'-ACA <sup>3'</sup>
							<b>48</b>
							5'-gCA <sup>3'</sup>
							<b>48</b>
							5'-gCC <sup>3'</sup>
							<b>52</b>
							5'-TgT <sup>3'</sup>
<b>A*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>B*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>C*</b>	<b>+</b>	<b>+</b>	<b>+</b>				
<b>DRB1</b>				<b>+</b>	<b>+</b>		
<b>DRB3</b>				<b>+</b>	<b>+</b>		
<b>DRB5</b>				<b>+</b>			
<b>DQB1</b>					<b>+</b>		
<b>DPB1</b>						<b>+</b>	
<b>DQA1</b>							<b>+</b>

<sup>1</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon, matching the specificity-determining 3'-end of the primer is given. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>2</sup>The nucleotide position for HLA class I genes and the codon for HLA class II genes, in the 2<sup>nd</sup> or 3<sup>rd</sup> exon or the 2<sup>nd</sup> intron, matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide and codon numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **0L4**

Lot-specific information

## PRODUCT DESCRIPTION

### KIR HLA Ligand SSP typing

#### CONTENT

The primer set contains 5'- and 3'-primers for determining KIR HLA Ligand nucleotide sequence motifs;

HLA-A<sup>Bw4+</sup> alleles,

HLA-B<sup>Bw4+</sup> alleles encoding Isoleucine or Threonine at position 80,

HLA-B<sup>Bw4+</sup> alleles encoding Aspartic acid at position 77 and Threonine at position 80,

HLA-B<sup>Bw6+</sup> alleles encoding Asparagine at position 80,

HLA-C alleles encoding Asparagine or Lysine at position 80.

#### PLATE LAYOUT

Each test consists of 8 PCR reactions in an 8 well cut PCR plate.

1	2	3	4	5	6	7	NC
---	---	---	---	---	---	---	----

The 8 well cut PCR plate is marked with ‘LIG’ in silver/gray ink.

Well No. 1 is marked with the Lot Number ‘0L4’.

Wells 1 and 2: HLA-C KIR ligand primers.

Wells 3, 4, 6 and 7: HLA-B KIR ligand primers.

Well 5: HLA-A KIR ligand primers.

Well 8 – Negative Control (NC).

A faint row of numbers is seen between wells 1 and 2 or wells 7 and 8 of the PCR trays. These stem from the manufacture of the trays, and should be disregarded. The PCR plates are covered with a PCR-compatible foil.

**Please note:** When removing each 8 well PCR plate, make sure that the remaining plates stay covered. Use a scalpel or a similar instrument to carefully cut the foil between the plates.

#### UNIQUELY IDENTIFIED ALLELES

The HLA-A, HLA-B and HLA-C alleles recognized by the HLA Nomenclature Committee in October 2019<sup>1,2,3</sup> have been considered in the Specificity and Interpretation Tables.

<sup>1</sup>A minor subset of rare B and C alleles with the listed target motifs will not be amplified by the kit due to downstream sequence variations.

<sup>2</sup>HLA-A, HLA-B and HLA-C alleles listed on the IMGT/HLA web page 2019-October-17, release 3.38.0, [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla).

<sup>3</sup>Alleles that have been deleted from or renamed in the official WHO HLA Nomenclature up to and including the last IMGT/HLA database release can be retrieved from web page <http://hla.alleles.org/alleles/deleted.html>.

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **OL4**

Lot-specific information

## PROTOCOL

### DNA EXTRACTION

Extracted, highly pure DNA is needed for SSP typings. DNA samples to be used for PCR-SSP HLA typing should be re-suspended in dH<sub>2</sub>O. The A260/A280 ratio should be 1.6 – 2.0 by UV spectrophotometry for optimal band visualization during electrophoresis.

We recommend automated DNA extraction with the QIAGEN EZ1 DSP DNA Blood System. ACD blood should be used as starting material.

Alternatively, the DNA can be extracted by any preferred method yielding pure DNA. When using alternative methods, the DNA concentration should be adjusted to 30 ng/μl. **Do not use heparinised blood with these methods.**

Recommended DNA concentration using:  
EZ1-extracted DNA, 15 ng/μl.  
DNA extracted by other methods, 30 ng/μl.

Concentrations exceeding 50 ng/μl will increase the risk for nonspecific amplifications and weak extra bands, especially for HLA Class I high resolution SSP typings. If necessary, dilute the extracted DNA in dH<sub>2</sub>O.

***DNA samples should not be re-suspended in solutions containing chelating agents such as EDTA, above 0.5 mM in concentration.***

DNA samples may be used immediately after extraction or stored at +4°C for up to 2 weeks with no adverse effects on results. DNA samples can be stored at -20°C or colder for 9 months. The purity and concentration of extracted DNA samples that have been stored for a prolonged period should be tested for acceptability prior to HLA typing.

DNA samples should be shipped at +4°C or colder to preserve their integrity during transport.

### PCR AMPLIFICATION

#### **104.201-12 – including *Taq* polymerase**

For one KIR HLA Ligand typing, begin by adding to well 8, i.e. the well with the negative control primer pairs:

7 μl dH<sub>2</sub>O

3 μl PCR Master Mix complete with *Taq*,

then add at room temperature in a 0.5 ml tube:

9 x 2 μl = 18 μl DNA (30 ng/μl)

9 x 3 μl = 27 μl PCR Master Mix with *Taq* – mix well before taking your aliquot

9 x 5 μl = 45 μl dH<sub>2</sub>O

**104.201-12 – including *Taq* polymerase**  
**104.201-12u – without *Taq* polymerase**

Visit <https://labproducts.caredx.com> for  
“Instructions for Use” (IFU)

**Lot No.: OL4**

**Lot-specific information**

Mix well, dispense 10 µl of the DNA-PCR Master Mix-H<sub>2</sub>O mixture into each of the 7 wells of a KIR HLA Ligand typing, i.e. wells 1 to 7. Cover the primer tray(s) with the provided adhesive seals. Check that all reaction wells are completely covered to prevent evaporative loss during PCR amplification.

**104.201-12u – without *Taq* polymerase**

For one KIR HLA Ligand typing, begin by adding at room temperature in a 0.5 ml tube:

10 x 3 µl = 30 µl PCR Master Mix without *Taq* – mix well before taking your aliquot

0,8 µl *Taq* polymerase (5 units/µl)

Mix well, dispense 3 µl of the PCR Master Mix-*Taq* mixture from the 0.5 ml tube into well No. 8, i.e. the well with the negative control primer pairs. Then add 7 µl dH<sub>2</sub>O to well 8.

Then add at room temperature to the 0.5 ml tube containing 30 + 0,8 - 3 = 27,8 µl PCR Master Mix-*Taq* mixture:

9 x 2 µl = 18 µl DNA (30 ng/µl)

9 x 5 µl – 0.8 µl = 44.2 µl dH<sub>2</sub>O

Mix well, dispense 10 µl of the DNA-PCR Master Mix-*Taq*-H<sub>2</sub>O mixture into each of the 7 wells of a KIR HLA Ligand typing, i.e. wells 1 to 7. Cover the primer tray(s) with the provided adhesive seals. Check that all reaction wells are completely covered to prevent evaporative loss during PCR amplification.

Use a 96 well thermal cycler with a heated lid. The temperature gradient across the heating block should be < 1°C.

**PCR cycling parameters:**

1. 1 cycle	94°C	2 min	denaturation
2. 10 cycles	94°C	10 sec.	denaturation
	65°C	60 sec.	annealing and extension
3. 20 cycles	94°C	10 sec.	denaturation
	61°C	50 sec.	annealing
	72°C	30 sec.	extension
4. End - hold	RT		if less than 8 hours
	4°C		if longer than 8 hours

Total reaction volume in each well, 10 µl.

The same PCR cycling parameters are used for all the *Olerup SSP*<sup>®</sup> kits.

**AGAROSE GEL ELECTROPHORESIS**

Prepare a 2% (w/v) agarose gel in 0.5 x TBE buffer. Dissolve the agarose by boiling in a microwave oven. Let the gel solution cool to 60°C. Stain the gel prior

104.201-12 – including *Taq* polymerase  
104.201-12u – without *Taq* polymerase

Visit <https://labproducts.caredx.com> for  
“Instructions for Use” (IFU)

Lot No.: **OL4**

**Lot-specific information**

to casting with ethidium bromide (10 mg/ml), 5 µl per 100 ml gel solution. For maximal ease of handling use our ethidium bromide dropper bottle (Product No. 103.301-10), 1 drop of ethidium bromide solution per 50-75 ml of gel, or our GelRed™ dropper bottle (Product No. 103.302-05) 4 drops per 100-120 ml of gel solution. **Note: Ethidium bromide is a powerful carcinogen. Handle with appropriate personal protective equipment.**

Load the PCR products, preferably using an 8-channel pipette. Load a DNA size marker (100 base pair ladder, Product No. 103.202-100 or DNA Size Marker for short gel runs 103.203-100) in one well per row.

Run the gel in 0.5 x TBE buffer, without re-circulation of the buffer, for 15-20 minutes at 8-10 V/cm.

**DOCUMENTATION AND INTERPRETATION**

Put the gel on a UV transilluminator and document by photography.

Record the presence and absence of specific PCR products. The relative lengths of the specific PCR products are helpful in the interpretation of the results.

Record the presence and relative lengths of the internal positive control bands. The differently sized control bands will help in the correct orientation of the typing as well as in kit identification.

Lanes without either control band or specific PCR products should be repeated.

Interpret the typings with the ***lot-specific Interpretation and Specificity Tables***.

**PCR MASTER MIXES**

The PCR Master Mix complete with *Taq* polymerase contains:

<i>Taq</i> polymerase	0.4 unit per 10 µl SSP reaction
nucleotides	final concentration of each dNTP is 200 µM
PCR buffer	final concentrations: 50 mM KCl, 1.5 mM MgCl <sub>2</sub> , 10 mM Tris-HCl pH 8.3, 0.001% w/v gelatin
glycerol	final concentration of glycerol is 5%
cresol red	final concentration of cresol red is 100 µg/ml

**The same PCR Master Mix complete with *Taq* is used for all Olerup SSP® kits.**

The PCR Master Mix without *Taq* polymerase contains:

nucleotides	final concentration of each dNTP is 200 µM
PCR buffer	final concentrations: 50 mM KCl, 1.5 mM MgCl <sub>2</sub> , 10 mM Tris-HCl pH 8.3, 0.001% w/v gelatin
glycerol	final concentration of glycerol is 5%
cresol red	final concentration of cresol red is 100 µg/ml

**The same PCR Master Mix without *Taq* is used for all Olerup SSP® kits.**

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **OL4**

Lot-specific information

## SPECIFICITY TABLE

### KIR HLA Ligand SSP typing

Specificities and sizes of the PCR products of the 7+1 primer mixes used for KIR HLA Ligand SSP.

Primer Mix	Size of spec. PCR product <sup>1</sup>	Size of control band <sup>2</sup>	KIR HLA Ligand nucleotide sequence motif	Amplified HLA alleles <sup>3,4</sup>
1 <sup>7</sup>	345 bp	800 bp	HLA-C <sup>Asn80</sup>	C*01:02:01:01-01:02:50, 01:02:52-01:13, 01:15:01-01:45, 01:47-01:153, 01:155-01:170, 01:172-01:187, 02:27:01-02:27:02, 02:65, 02:87, 02:115, 02:131, 03:02:01-03:06:02, 03:08-03:09, 03:11:01-03:11:02, 03:13:01:01-03:14, 03:16-03:28, 03:30-03:44, 03:46-03:98, 03:100-03:162, 03:164-03:267, 03:269-03:296:02, 03:298-03:394, 03:396N-03:398, 03:400-03:414, 03:416-03:437, 03:439-03:449N, 03:451-03:460, 03:462N-03:500, 03:502-03:504, 04:11, 04:29, 04:36, 04:55, 04:114, 04:172, 04:346, 04:383, 05:20, 06:11, 06:82, 06:147, 06:210, 06:217, 06:248, 06:252, 07:01:01:01-07:01:02:08, 07:01:04-07:02:77, 07:02:79-07:04:02:01, 07:04:03-07:06:03, 07:08, 07:10-07:33N, 07:35-07:75, 07:77-07:78:01, 07:79-07:114, 07:116-07:294, 07:296-07:314:03, 07:316-07:327, 07:329N-07:360, 07:362-07:405, 07:407-07:425, 07:427-07:458, 07:460-07:477, 07:479-07:558:01:02, 07:560-07:575:02, 07:577-07:597, 07:599-07:655, 07:657-07:672N, 07:674-07:715, 07:717-07:837, 08:01:01:01-08:09, 08:11-08:63, 08:65-08:145, 08:147-08:203, 08:205-08:206, 12:02:01-12:02:22, 12:02:24-12:03:61, 12:06-12:08, 12:10:01-12:20, 12:22-12:32, 12:34, 12:36-12:40, 12:42Q-12:59, 12:61-12:71, 12:73-12:134, 12:136-12:145, 12:147-12:153, 12:155Q-12:229, 12:231-12:304, 14:02:01:01-14:11, 14:13-14:48, 14:50-14:107, 14:109-14:116, 15:07:01:01-15:07:01:02, 15:25, 15:43, 15:85, 15:144, 15:181, 16:01:01:01-16:01:36, 16:04:01:01-16:04:01:03,



**104.201-12 – including Taq polymerase**

Visit <https://labproducts.caredx.com> for

**104.201-12u – without Taq polymerase**

**“Instructions for Use” (IFU)**

**Lot No.: OL4**

**Lot-specific information**

				16:04:03-16:04:05, 16:06-16:08, 16:10-16:11, 16:13-16:18, 16:20-16:24, 16:26-16:36, 16:38-16:45, 16:49-16:59, 16:61-16:62, 16:64-16:68, 16:71-16:73, 16:75-16:76, 16:78-16:83, 16:86-16:87, 16:92-16:93, 16:95-16:98, 16:100, 16:105-16:106, 16:109-16:114, 16:116-16:119, 16:122, 16:124-16:128, 16:130-16:131, 16:134-16:135, 16:137-16:139, 16:141-16:142, 16:146-16:152, 16:154, 16:157-16:162, 16:164, 17:22
<b>2<sup>8</sup></b>	<b>350 bp</b>	<b>800 bp</b>	<b>HLA-C<sup>Lys80</sup></b>	C*01:14, 01:154, 02:02:01-02:02:03, 02:02:06-02:26:03, 02:28-02:40:02, 02:42-02:64, 02:66-02:86, 02:88-02:114, 02:116-02:130, 02:132-02:186, 03:07:01:01-03:07:02, 03:10, 03:15, 03:29, 03:45, 03:163, 03:268, 03:297, 03:450, 03:461, 04:01:01:01-04:01:01:29, 04:01:01:31-04:01:120, 04:03:01:01-04:10, 04:12-04:20, 04:23-04:28, 04:30-04:35:02, 04:37-04:54:02, 04:56-04:113, 04:115N-04:171, 04:173N-04:249, 04:251-04:345, 04:347-04:382, 04:384-04:405, 05:01:01:01-05:01:50, 05:03-05:19, 05:21-05:227, 06:02:01:01-06:02:01:43, 06:02:03-06:10, 06:12-06:81, 06:83-06:146, 06:148-06:209, 06:211:01:01N-06:216, 06:218-06:247, 06:249-06:251, 06:253-06:292, 07:07, 07:09, 07:76:01-07:76:02, 07:315, 07:328, 07:406, 07:559, 07:598, 07:656, 08:10, 12:04:01-12:05, 12:09, 12:21, 12:33, 12:41, 12:60, 12:72, 12:135, 12:146, 12:154, 14:12, 14:49, 14:108, 15:02:01:01-15:06:03, 15:08:01-15:13:01:02, 15:15-15:19, 15:21-15:24, 15:26-15:42, 15:44:01-15:84Q, 15:86-15:143, 15:145N-15:180, 15:182-15:207, 16:02:01:01-16:02:17, 16:09, 16:12, 16:19, 16:25, 16:37, 16:46-16:48, 16:60, 16:63, 16:69-16:70, 16:74, 16:77N, 16:84, 16:88-16:91, 16:99, 16:101-16:104, 16:107-16:108, 16:115, 16:120-16:121, 16:123N, 16:132N-16:133, 16:136, 16:140, 16:143-16:145, 16:153, 16:155-16:156, 16:163, 17:01:01:02-17:21, 17:23-17:51:01:02, 18:01:01:01-18:14





**104.201-12 – including Taq polymerase**  
**104.201-12u – without Taq polymerase**

Visit <https://labproducts.caredx.com> for  
**“Instructions for Use” (IFU)**

**Lot No.: OL4**

**Lot-specific information**

<b>3<sup>5</sup></b>	350 bp	<b>800 bp</b>	HLA-B <sup>Bw4+Thr80</sup>	B*07:149, 08:02, 08:117, 13:01:01:01-13:04, 13:06-13:08, 13:10-13:23, 13:25-13:38, 13:40-13:66, 13:68-13:139N, 15:36, 15:89:01-15:89:02, 15:115, 15:256, 15:339, 15:505, 18:09, 27:01, 27:142, 37:01:02, 37:10, 38:02:01:01-38:04, 38:08, 38:15, 38:18, 38:23, 38:29, 38:35, 38:43-38:50, 38:62, 38:64, 38:72, 38:74-38:76, 38:79, 38:82, 38:84-38:85, 38:160, 38:162, 40:47, 40:96, 40:110, 40:157, 40:201, 44:02:01:01-44:02:44, 44:02:46-44:05:02, 44:05:04, 44:07-44:08, 44:10-44:17, 44:19N-44:24, 44:26-44:45, 44:47-44:49, 44:51-44:74, 44:76-44:89, 44:91-44:94, 44:96-44:128:02, 44:130, 44:132-44:151, 44:153-44:169, 44:171N-44:195N, 44:197-44:245, 44:247-44:253, 44:255-44:256, 44:258-44:276, 44:278-44:344, 44:346, 44:437, 44:439-44:453, 44:456-44:480Q, 47:04, 49:02, 51:54, 51:78:01-51:78:02, 52:20, 53:09, 53:11-53:13, 53:31, 53:36, 53:38, 56:07, 59:08
<b>4<sup>5</sup></b>	350 bp	<b>1070 bp</b>	HLA-B <sup>Bw4+Ile80</sup>	B*07:36, 07:38, 07:81, 07:180, 07:219, 08:03, 08:52, 08:78, 15:13:01-15:13:02, 15:16:01:01-15:17:07, 15:23-15:24:02, 15:67, 15:87, 15:95, 15:157, 15:162, 15:168, 15:177, 15:196, 15:208, 15:216, 15:222, 15:230, 15:254, 15:268, 15:273, 15:356, 15:361-15:362, 15:396, 15:403, 15:408, 15:411, 15:418, 15:423-15:424, 15:442, 15:446, 15:462, 15:500, 15:516, 15:523, 15:532, 15:546Q, 15:550, 15:555, 18:67, 18:136, 27:02:01:01-27:02:05, 27:30, 27:53, 27:57, 27:62, 27:65N, 27:75, 27:77, 27:83, 27:95, 27:102, 27:119, 27:126, 27:134, 27:157, 27:163, 27:171-27:172, 27:176N, 27:181, 27:188, 27:197, 27:203, 27:213, 37:34, 38:01:01:01-38:01:16, 38:05-38:07, 38:09-38:14, 38:16, 38:19-38:22, 38:24-38:28, 38:30-38:34N, 38:36-38:42, 38:51-38:61, 38:63, 38:65-38:71, 38:73, 38:77-38:78, 38:80N-38:81, 38:83N, 38:86, 38:161, 40:13, 40:19, 40:109, 40:117, 40:292, 40:340, 40:394, 41:46, 41:60, 44:06, 44:18, 44:25, 44:50:01, 44:50:03, 44:95,

**104.201-12 – including Taq polymerase**

Visit <https://labproducts.caredx.com> for

**104.201-12u – without Taq polymerase**

**“Instructions for Use” (IFU)**

**Lot No.: OL4**

**Lot-specific information**

				48:18, 49:01:01:01-49:01:16, 49:03-49:14, 49:16-49:65, 51:01:01:01-51:24:05, 51:26-51:46, 51:48-51:53, 51:55-51:77, 51:79- 51:307, 52:01:01:01-52:01:28, 52:01:30-52:19, 52:21:01-52:94N, 53:01:01:01-53:02, 53:04-53:08:02, 53:10, 53:14-53:30, 53:32-53:35, 53:37, 53:39-53:54, 53:56-53:58, 54:12, 55:103, 56:21, 56:65, 57:01:01:01-57:11, 57:13-57:127, 58:01:01:01-58:02:02, 58:04- 58:16:02, 58:18-58:29, 58:31N- 58:63, 58:65-58:120, 59:01:01:01- 59:07, 59:09-59:10N
<b>5</b>	370 bp	1070 bp	HLA-A <sup>Bw4+</sup>	A*01:95, 01:289, 01:298, 02:81, 02:87, 02:112, 02:124, 02:129, 02:136, 02:571, 02:829, 03:152, 03:219, 11:178, 11:190, 23:01:01:01-23:68, 23:70-23:89, 23:91N-23:102, 24:02:01:01- 24:03:04, 24:05:01-24:11N, 24:13:01-24:15, 24:17-24:18, 24:20:01:01-24:27, 24:29-24:43, 24:45N-24:64, 24:66-24:88, 24:90:01N-24:99, 24:101-24:108, 24:110-24:128, 24:130-24:210, 24:212-24:240N, 24:242-24:289, 24:291-24:372, 24:374-24:405, 24:407-24:423, 24:425N-24:470, 24:472-24:477, 25:01:01:01-25:58, 25:60-25:64, 29:13, 31:07-31:08, 31:10, 32:01:01:01-32:133N, 68:36
<b>6</b>	350 bp	1070 bp	HLA-B <sup>Bw4+, Thr80</sup>	B*07:27, 07:236, 07:273, 08:126, 15:43, 18:54, 27:03-27:05:36, 27:05:38-27:07:06, 27:09-27:11, 27:13-27:17, 27:19:01:01-27:21, 27:23-27:25, 27:27-27:29, 27:31- 27:32, 27:34-27:39, 27:41, 27:43, 27:45-27:48, 27:50:01-27:52, 27:54-27:56, 27:58-27:61, 27:63- 27:64N, 27:66N-27:74, 27:76, 27:78-27:82, 27:84-27:88, 27:90:01-27:94N, 27:96:01-27:101, 27:103-27:118, 27:120-27:125, 27:127-27:133, 27:135-27:141, 27:143-27:152, 27:154-27:155, 27:158-27:162, 27:164, 27:166- 27:170, 27:173-27:175, 27:177- 27:180, 27:182-27:187, 27:189- 27:196, 27:198-27:202, 27:205- 27:212N, 27:214-27:221, 35:329, 37:01:01:01-37:01:01:09, 37:01:03- 37:04:02, 37:06:01-37:09, 37:12- 37:13, 37:15-37:33N, 37:35-37:36, 37:38-37:66, 37:68-37:91, 38:17, 40:188, 44:257,



**104.201-12 – including Taq polymerase**  
**104.201-12u – without Taq polymerase**

Visit <https://labproducts.caredx.com> for  
**“Instructions for Use” (IFU)**

**Lot No.: OL4**

**Lot-specific information**

				47:01:01:02-47:01:05, 47:05-47:10, 53:03, 53:55, 55:90, 56:46
<b>7</b>	350 bp	1070 bp	HLA-B <sup>Bw6+</sup>	B*07:02:01:01-07:02:77, 07:02:79-07:26, 07:28-07:35, 07:37:01-07:37:02, 07:39-07:80, 07:82-07:148, 07:150-07:179, 07:181N-07:191, 07:193-07:218, 07:220-07:235, 07:237-07:272N, 07:274-07:395, 08:01:01:01-08:01:55, 08:04:01-08:05, 08:07-08:51, 08:53:01-08:77, 08:79-08:116, 08:118-08:125, 08:127-08:130, 08:132-08:135, 08:137-08:138, 08:140-08:152, 08:154-08:156, 08:158-08:249, 13:09, 13:39, 13:67, 14:01:01:01-14:86, 15:01:01:01-15:01:04, 15:01:06-15:12:03, 15:14-15:15:01:02, 15:18:01:01-15:21:01:02, 15:25:01:01-15:35, 15:37-15:40:02, 15:42, 15:44-15:58, 15:60-15:66, 15:68-15:86, 15:88, 15:90-15:94N, 15:96-15:99, 15:101-15:114, 15:116-15:129, 15:131-15:156, 15:158-15:161, 15:163-15:167, 15:169-15:176, 15:178-15:195, 15:197:01-15:207, 15:209N-15:215, 15:217-15:221, 15:223-15:229, 15:231-15:253, 15:255, 15:257-15:267, 15:269-15:272N, 15:274-15:276, 15:278-15:338, 15:340-15:355, 15:357-15:360, 15:363:01-15:376, 15:378-15:395, 15:397-15:402, 15:404-15:407, 15:409-15:410, 15:412-15:417, 15:419-15:422, 15:425-15:429, 15:431-15:441, 15:443-15:445, 15:447-15:461, 15:463N-15:499, 15:501-15:504, 15:506-15:515, 15:517-15:522, 15:524-15:531, 15:533-15:545, 15:547-15:549N, 15:551-15:554, 15:556, 18:01:01:01-18:08, 18:10-18:15, 18:17N-18:53, 18:55-18:66, 18:68-18:135:01:02, 18:137-18:178, 18:180-18:189, 27:08, 27:12:01:01-27:12:01:03, 27:18, 27:26, 27:33, 27:40, 27:42, 27:44, 27:89, 27:153, 27:165 <sup>w</sup> , 27:204, 35:01:01:01-35:01:38, 35:01:40-35:02:08, 35:02:10-35:43:01, 35:43:03-35:72, 35:74-35:221, 35:223-35:247, 35:249-35:292, 35:294-35:328, 35:330-35:423, 35:425-35:466, 37:05, 37:11, 37:14 <sup>w</sup> , 37:37, 37:67, 38:156-38:159, 39:01:01:01-39:01:01:12, 39:01:03:01-39:01:14, 39:01:16-39:20, 39:22-39:103,





**104.201-12 – including *Taq* polymerase**

Visit <https://labproducts.caredx.com> for

**104.201-12u – without *Taq* polymerase**

**“Instructions for Use” (IFU)**

**Lot No.: OL4**

**Lot-specific information**

<sup>6</sup>Primer mix 8 contains a negative control, which will amplify the majority of HLA amplicons as well as the amplicons generated by the control primer pairs matching the human growth hormone gene. HLA-specific PCR product sizes range from 75 to 200 base pairs and the PCR product generated by the HGH positive control primer pair is 200 base pairs.

<sup>7</sup>This lot will not amplify the C\*01:46 allele in primer mix 1.

<sup>8</sup>The alleles C\*05:32 (80<sup>Arg</sup>), C\*15:60 (80<sup>Gln</sup>) and C\*15:71(80<sup>Ile</sup>) will be amplified in primer mix 2.

Abbreviations

w: might be weakly amplified.

Asn: asparagine

Asp: aspartic acid

Ile: isoleucine

Lys: lysine

Thr: threonine

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **0L4**

Lot-specific information

### PRIMER SPECIFICATION

Well No.	1	2	3	4	5	6	7
Length of spec. PCR product	345	350	350	350	370	350	350
Length of int. pos. control <sup>1</sup>	800	800	800	1070	1070	1070	1070
5'-primer(s) <sup>2</sup>	1 <sup>st</sup>   5' -CgA 3'	1 <sup>st</sup>   5' -CgA 3'	1 <sup>st</sup>   5' -CAg 3'	1 <sup>st</sup>   5' -CAg 3'	1 <sup>st</sup>   5' -gCA 3'	1 <sup>st</sup>   5' -CAg 3'	1 <sup>st</sup>   5' -CAg 3'
3'-primer(s) <sup>3</sup>	310 5' -gTT 3'	312 5' -AgT 3'	309 5' -gTg 3'	309 5' -ATC 3'	317 5' -ggA 3'	310 5' -ggT 3'	311 5' -ggT 3'
	311 5' -gAT 3'	312 5' -AgT 3'	312 5' -gCC 3'	312 5' -gCA 3'			
	311 5' -ggT 3'	312 5' -AgT 3'	313 5' -Cgg 3'				
	316 5' -gCT 3'		315 5' -AgT 3'				
Well No.	1	2	3	4	5	6	7

<sup>1</sup>The internal positive control primer pairs amplify segments of the human growth hormone gene. The internal positive control bands are 1070 or 800 base pairs respectively, well distribution as outlined in the table. Well number 1 contains the shorter, 800 bp, internal positive control band. The well distribution of the internal controls can help in orientation of the kit on gel photo, as well as allow for kit identification. In the presence of a specific amplification the intensity of the control band often decreases.

<sup>2</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

<sup>3</sup>The nucleotide position matching the specificity-determining 3'-end of the primer is given in the anti-sense direction. Nucleotide numbering as on the [www.ebi.ac.uk/imgt/hla](http://www.ebi.ac.uk/imgt/hla) web site. The sequence of the 3 terminal nucleotides of the primer is given.

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **OL4**

Lot-specific information

<b>CELL LINE VALIDATION SHEET</b>			<b>KIR HLA Ligand primer set<sup>2</sup></b>							
			<b>Well</b>							
			<b>1</b>	<b>2</b>	<b>3</b>	<b>4</b>	<b>5</b>	<b>6</b>	<b>7</b>	
			<b>Prod. No.:</b>	201896201	201896202	201896203	201896204	201896205	201896206	201896207
<b>IHWC cell line<sup>1</sup></b>										
1	9001	SA	+	-	-	-	+	-	+	
2	9280	LK707	+	+	-	+	-	-	+	
3	9011	E4181324	+	-	-	+	-	-	-	
4	9275	GU373	+	+	-	+	-	-	+	
5	9009	KAS011	-	+	-	-	-	+	-	
6	9353	SM	+	-	-	+	-	-	+	
7	9020	QBL	-	+	-	-	-	-	+	
8	9025	DEU	-	+	-	-	-	-	+	
9	9026	YAR	+	-	-	+	-	-	-	
10	9107	LKT3	+	-	-	-	+	-	+	
11	9051	PITOUT	+	-	+	-	-	-	-	
12	9052	DBB	-	+	-	+	-	-	-	
13	9004	JESTHOM	+	-	-	-	-	+	-	
14	9071	OLGA	+	-	-	-	-	-	+	
15	9075	DKB	+	-	-	-	+	-	+	
16	9037	SWEIG007	-	+	-	-	-	-	+	
17	9282	CTM3953540	+	-	-	-	-	-	+	
18	9257	32367	+	-	-	-	-	-	+	
19	9038	BM16	+	-	-	-	-	-	+	
20	9059	SLE005	+	-	-	-	-	-	+	
21	9064	AMALA	+	-	-	-	-	-	+	
22	9056	KOSE	+	-	-	-	-	-	+	
23	9124	IHL	+	+	-	-	-	-	+	
24	9035	JBUSH	+	-	-	+	+	-	-	
25	9049	IBW9	+	-	-	-	-	-	+	
26	9285	WT49	+	-	-	+	-	-	-	
27	9191	CH1007	+	+	-	+	+	-	+	
28	9320	BEL5GB	+	+	+	-	-	-	-	
29	9050	MOU	+	-	+	-	-	-	-	
30	9021	RSH	-	+	-	-	-	-	+	
31	9019	DUCAF	-	+	-	-	-	-	+	
32	9297	HAG	-	+	-	-	-	-	+	
33	9098	MT14B	+	-	-	-	-	-	+	
34	9104	DHIF	+	-	-	+	-	-	-	
35	9302	SSTO	-	+	+	-	+	-	-	
36	9024	KT17	+	+	-	-	-	-	+	
37	9065	HHKB	+	-	-	-	-	-	+	
38	9099	LZL	+	-	-	-	-	-	+	
39	9315	CML	+	+	-	-	-	+	+	
40	9134	WHONP199	+	+	+	-	-	-	+	
41	9055	H0301	+	-	-	-	-	-	+	
42	9066	TAB089	+	-	-	-	-	-	+	
43	9076	T7526	+	-	-	-	-	-	+	
44	9057	TEM	+	-	-	+	-	-	-	
45	9239	SHJO	-	+	-	-	+	-	+	
46	9013	SCHU	+	-	-	-	-	-	+	
47	9045	TUBO	+	+	-	+	-	-	-	
48	9303	TER-ND	+	+	+	-	-	-	+	

<sup>1</sup>The provided cell line HLA specificities are retrieved from the <http://www.ihwg.org/hla> web site. The specificity of an individual cell line may thus be subject to change.

<sup>2</sup>The specificity of each primer solution in the kit has been tested against 48 well characterized cell line DNAs and where applicable, additional cell line DNAs.

In primer solutions 1 to 4 one, two or three of the 3'-primers were not possible to test.



104.201-12 – including *Taq* polymeraseVisit <https://labproducts.caredx.com> for104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **OL4**

Lot-specific information

**TRADEMARKS USED IN THIS DOCUMENT/PRODUCT***Olerup SSP*<sup>®</sup> is a registered trademark of *CareDx AB*.*Qiagen*<sup>™</sup> is a trademark of QIAGEN.**WARRANTY**

*CareDx AB* warrants its products to the original purchaser against defects in materials and workmanship under normal use and application. *CareDx AB*'s sole obligation under this warranty shall be to replace, at no charge, any product that does not meet the performance standards stated on the product specification sheet.

This warranty applies only to products that have been handled and stored in accordance with *CareDx AB*'s recommendations, and does not apply to products that have been the subject of alternation, misuse, or abuse.

All claims under this warranty must be directed to *CareDx AB* in writing and must be accompanied by a copy of the purchaser's invoice. This warranty is in lieu of all other warranties, expressed or implied, including the warranties of merchantability and fitness for a particular purpose. In no case shall *CareDx AB* be liable for incidental or consequential damages.

This product may not be reformulated, repacked or resold in any form without the written consent of *CareDx AB*, Franzengatan 5, SE-112 51 Stockholm, Sweden.

Handle all samples as if capable of transmitting disease. All work should be performed wearing gloves and appropriate protection.

**GUARANTEE**

*CareDx AB* guarantees that the primers in the *Olerup SSP*<sup>®</sup> typing trays have the specificities given in the lot-specific Specificity and Interpretation Tables of the product insert.

When stored at –20°C, the dried primers are stable for 48 months from the date of manufacture.

When stored at –20°C, the PCR Master Mix including *Taq* polymerase and the PCR Master Mix without *Taq* polymerase are stable for 51 months from the date of manufacture.

104.201-12 – including *Taq* polymerase

Visit <https://labproducts.caredx.com> for

104.201-12u – without *Taq* polymerase

“Instructions for Use” (IFU)

Lot No.: **0L4**

Lot-specific information

**ADDRESSES:**

**Manufacturer:**

**Caredx AB**, Franzengatan 5, SE-112 51 Stockholm, Sweden.

**Tel:** +46-8-508 939 00

**Fax:** +46-8-717 88 18

**E-mail:** [orders-se@caredx.com](mailto:orders-se@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

**Distributed by:**

**CareDx GmbH**, Löwengasse 47 / 6, AT-1030 Vienna, Austria.

**Tel:** +43-1-710 15 00

**Fax:** +43-1-710 15 00 10

**E-mail:** [orders-at@caredx.com](mailto:orders-at@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

**CareDx Lab Solutions Inc.**, 901 S. Bolmar St., Suite R, West Chester, PA 19382

**Tel:** 1-877-653-78171

**Fax:** 610-344-7989

**E-mail:** [orders-us@caredx.com](mailto:orders-us@caredx.com)

**Web page:** <https://labproducts.caredx.com/>

For information on CareDx distributors worldwide, contact **CareDx GmbH**.