

Olerup SSP[®] DQ low resolution

Product number: 101.201-48/12 - including *Taq* pol.
101.201-48u/12u – without *Taq* pol.
Lot number: 41E
Expiry date: 2010-May-01
Number of tests: 48 tests – Product No. 101.201-48
12 tests – Product No. 101.201-12
Number of wells per test: 8

Changes compared to the previous DQ low resolution Lot (X48):

Well	5'-primer	3'-primer	rationale
7	Modified	-	Increased yield of specific PCR product
8	Exchanged	-	Primer exchanged to amplify the DQB1*0403 allele.

THE NUMBER OF WELLS is unchanged.

ALLELE COVERAGE:

DQB1*050101 to 0505, DQB1*060101 to 0634, DQB1*020101 to 0205, DQB1*030101 to 0321 and DQB1*0401 to 0403, i.e. all the currently recognized DQB1 alleles will be grouped into their corresponding serological specificities; www.ebi.ac.uk/imgt/hla, 2008-April-08, release 2.21.0.

MODIFICATIONS MADE DUE TO COMMENTS FROM CUSTOMERS:

No suggestions received.